


Data Path : Z:\HPCHEM1\BNA_M\Data\BM041415\
 Data File : BM000954.D
 Acq On : 13 Apr 2015 17:47
 Operator : TP/IZ
 Sample : DFTPP44
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP44

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM041415.M
 Title : SVOA CALIBRATION
 Last Update : Wed Apr 15 06:41:33 2015



AutoFind: Scans 2510, 2511, 2512; Background Corrected with Scan 2505

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.0	107848	PASS
68	69	0.00	2	1.5	1423	PASS
69	198	0.00	100	42.7	95800	PASS
70	69	0.00	2	0.5	465	PASS
127	198	10	80	49.9	111954	PASS
197	198	0.00	2	0.4	858	PASS
198	198	100	100	100.0	224490	PASS
199	198	5	9	7.0	15683	PASS
275	198	10	60	23.2	52162	PASS
365	198	1	100	2.3	5222	PASS
441	443	0.01	100	76.6	26972	PASS
442	198	50	100	79.6	178613	PASS
443	442	15	24	19.7	35226	PASS

m/z	Abundance
36.00	375.0
38.10	1461.0
39.10	7510.0
40.00	4502.0
44.00	4794.0
45.00	317.0
49.00	1032.0
50.10	21912.0
51.10	94024.0
52.10	4539.0
55.00	660.0
56.00	2689.0
57.00	6413.0
58.10	345.0
61.00	1646.0
62.10	1270.0
63.00	3348.0
64.00	330.0
65.00	1419.0
68.10	1084.0
69.00	79360.0
70.00	439.0
73.10	1392.0
74.00	8001.0
75.00	12339.0
76.00	4063.0
77.10	90176.0
78.00	6627.0
79.00	5631.0
80.00	4436.0
81.00	5954.0
81.90	1395.0
83.00	1695.0
85.10	999.0
86.00	1841.0
87.00	833.0
88.00	313.0
91.00	1265.0
92.10	1671.0
93.00	9397.0
93.90	686.0
94.90	375.0
95.80	453.0
98.00	7932.0
99.00	7092.0
99.90	406.0
101.00	3239.0
103.10	1326.0
103.90	1821.0
105.10	1737.0
105.90	534.0
107.00	26552.0
108.10	3891.0
108.90	762.0
110.00	46440.0
111.00	7635.0
112.00	1071.0
112.90	319.0
116.00	1479.0
117.00	19960.0
118.10	1371.0
120.00	311.0
121.90	1510.0
122.90	2188.0
123.90	1302.0
125.00	972.0
125.80	373.0
127.00	86968.0
128.00	6329.0
129.00	32848.0
130.00	2668.0
131.10	797.0
133.10	586.0
134.00	1173.0
135.00	2581.0
135.90	1038.0
137.00	1844.0
137.80	301.0
138.10	355.0
139.90	569.0
141.00	4991.0
142.00	1326.0
143.00	1165.0
145.90	938.0
147.00	2043.0
147.90	5233.0
148.90	1282.0
149.90	415.0
151.10	611.0
151.90	382.0
153.00	1370.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

154.00	1128.0
155.10	2321.0
156.00	3546.0
156.90	718.0
158.00	907.0
158.90	646.0
160.00	1375.0
161.10	1531.0
161.90	668.0
164.00	325.0
165.00	1508.0
165.90	1594.0
167.00	8395.0
168.00	7240.0
169.10	599.0
170.80	604.0
172.00	670.0
173.00	1007.0
173.90	1443.0
175.10	2794.0
176.10	1056.0
176.90	1650.0
178.10	535.0
179.00	6676.0
180.00	5035.0
181.00	1992.0
182.10	341.0
183.70	321.0
185.00	2942.0
186.00	20896.0
187.00	6086.0
188.20	950.0
189.00	1396.0
190.00	419.0
190.90	1249.0
192.00	1917.0
193.00	2245.0
193.90	861.0
196.00	5657.0
196.80	1004.0
198.00	154176.0
199.00	11882.0
200.00	947.0
201.60	936.0
203.00	1045.0
204.00	5831.0
205.00	9596.0
206.10	39640.0
207.00	8840.0
208.00	1856.0
209.00	815.0
210.30	1213.0
211.10	1424.0
212.10	484.0
214.90	535.0
216.10	796.0
217.00	10050.0
218.00	931.0
220.20	373.0
221.10	8994.0
222.90	1879.0
224.00	20896.0
225.00	5469.0
226.00	1003.0
227.00	8795.0
228.10	1599.0
229.00	1913.0
231.00	1250.0
234.10	681.0
234.80	424.0
236.00	598.0
237.00	1864.0
238.90	306.0
241.10	433.0
242.00	1039.0
243.00	1177.0
244.10	14475.0
245.00	2000.0
246.00	3127.0
247.00	728.0
249.00	464.0
252.80	423.0
254.10	547.0
255.00	79624.0
256.00	10990.0
257.00	867.0
258.00	4601.0
259.00	1384.0
265.00	1941.0
273.00	1434.0
274.10	5909.0
275.00	31256.0
276.00	4832.0
277.00	2446.0
278.00	973.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

281.10	492.0
282.90	470.0
285.00	575.0
293.00	643.0
294.00	418.0
296.00	8756.0
297.00	1342.0
303.00	960.0
313.90	630.0
315.00	1116.0
316.10	541.0
323.00	2832.0
324.10	365.0
326.90	621.0
328.00	388.0
333.90	1218.0
335.00	522.0
341.10	701.0
346.00	594.0
352.00	577.0
353.00	620.0
353.90	904.0
364.90	2957.0
365.80	469.0
370.90	325.0
372.10	1083.0
373.00	408.0
383.00	502.0
402.00	497.0
403.10	589.0
404.00	320.0
420.90	414.0
421.80	759.0
423.00	4749.0
424.10	3006.0
425.00	462.0
441.00	13718.0
442.00	88928.0
443.10	17880.0
444.00	2098.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

m/z	Abundance
37.10	452.0
38.10	1764.0
39.10	10107.0
40.00	4804.0
44.00	5622.0
45.00	593.0
49.00	829.0
50.10	32760.0
51.00	131200.0
52.10	6605.0
54.90	712.0
56.00	4033.0
57.00	8642.0
57.90	664.0
61.00	1548.0
62.00	1648.0
63.00	5505.0
64.10	688.0
65.00	2576.0
67.90	1781.0
69.00	117736.0
70.00	553.0
73.00	1878.0
74.00	12353.0
75.00	18784.0
76.00	6018.0
77.10	131648.0
78.10	8872.0
79.00	7871.0
80.10	6425.0
81.00	9775.0
82.00	2593.0
83.00	1837.0
85.10	1828.0
86.00	2711.0
87.00	1295.0
88.00	432.0
91.00	2228.0
92.00	3088.0
93.00	14656.0
94.00	968.0
95.00	409.0
96.00	1416.0
96.90	339.0
98.00	10853.0
99.00	9929.0
99.90	837.0
101.00	4986.0
102.90	2197.0
104.00	3057.0
105.00	3064.0
106.00	1019.0
107.00	41080.0
108.00	6582.0
109.10	1061.0
110.00	70408.0
111.00	11745.0
112.00	1718.0
112.90	517.0
116.00	1986.0
117.00	31128.0
118.00	2012.0
118.90	360.0
120.00	489.0
120.90	334.0
122.00	2229.0
123.00	4473.0
124.00	1696.0
125.00	2024.0
126.10	383.0
127.00	137600.0
128.00	11770.0
129.00	52680.0
130.00	4202.0
130.90	1126.0
132.00	454.0
132.80	557.0
133.90	1618.0
135.00	4447.0
136.00	1539.0
137.00	2615.0
137.90	485.0
139.00	353.0
140.00	647.0
141.00	6653.0
142.10	1998.0
143.10	1732.0
144.10	392.0
144.90	352.0
145.90	1165.0
147.00	3937.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

148.00	8359.0
149.10	2193.0
150.10	454.0
151.20	1214.0
151.90	357.0
152.90	2095.0
154.00	1256.0
155.10	3400.0
156.00	5345.0
157.00	1248.0
158.00	1429.0
159.00	999.0
160.00	1816.0
161.00	2764.0
161.90	900.0
162.80	304.0
164.10	302.0
165.00	1941.0
165.90	2045.0
167.00	13964.0
168.00	9048.0
169.00	1463.0
170.20	349.0
170.90	809.0
172.00	1429.0
172.90	992.0
174.00	2932.0
175.10	4928.0
176.00	1627.0
176.90	3043.0
178.00	955.0
179.00	10895.0
180.00	7940.0
181.00	3156.0
182.00	328.0
182.90	326.0
183.90	715.0
185.00	5093.0
186.00	37008.0
187.00	10432.0
188.00	1583.0
189.00	2452.0
190.00	423.0
191.00	1463.0
192.00	3306.0
193.00	3973.0
194.00	681.0
195.00	336.0
196.00	8329.0
197.00	1468.0
198.00	277824.0
199.00	18560.0
200.00	1498.0
200.80	337.0
201.40	1122.0
203.00	1865.0
204.00	10274.0
205.10	15337.0
206.10	66048.0
207.10	13312.0
208.10	2915.0
209.10	1377.0
210.10	1249.0
210.80	3034.0
215.00	726.0
216.10	1865.0
217.00	18288.0
218.00	2265.0
219.80	544.0
221.10	11136.0
221.90	3072.0
222.90	3619.0
224.00	37096.0
225.00	9442.0
226.10	1792.0
227.00	14926.0
228.00	1922.0
228.90	3061.0
230.00	449.0
230.90	1300.0
232.00	473.0
234.00	978.0
235.00	827.0
236.00	688.0
237.00	2467.0
238.10	423.0
238.90	506.0
240.00	622.0
241.00	674.0
242.10	1281.0
243.00	2041.0
244.10	28880.0
245.00	3775.0
246.00	5325.0
246.90	1046.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

249.00	1124.0
250.90	362.0
252.00	313.0
253.00	828.0
254.00	1081.0
255.00	144832.0
256.00	21104.0
257.00	1882.0
258.00	7234.0
259.00	1106.0
261.10	351.0
265.00	3519.0
266.10	397.0
270.90	388.0
272.00	450.0
273.00	3873.0
274.00	9683.0
275.00	61280.0
276.00	8120.0
277.00	4412.0
278.00	995.0
281.00	552.0
283.00	868.0
284.00	411.0
285.00	1014.0
293.00	1257.0
294.00	742.0
295.10	338.0
296.00	16768.0
297.00	1815.0
301.00	363.0
303.10	1790.0
303.90	698.0
314.00	805.0
315.00	1547.0
316.00	1145.0
321.00	311.0
322.10	358.0
323.00	5289.0
324.00	888.0
327.00	1078.0
327.90	420.0
331.80	345.0
333.10	476.0
334.00	3459.0
335.00	971.0
341.10	836.0
345.90	1173.0
352.00	1456.0
352.90	1341.0
353.90	1586.0
354.90	415.0
365.00	5760.0
366.00	891.0
371.90	2387.0
373.00	571.0
382.90	888.0
389.90	346.0
402.00	1238.0
403.10	989.0
404.10	892.0
421.10	1123.0
422.00	1435.0
423.00	11224.0
424.00	5228.0
425.00	782.0
441.00	29464.0
442.00	204096.0
443.00	40704.0
444.00	3234.0
445.10	347.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

m/z	Abundance
37.00	436.0
38.10	1301.0
39.10	8009.0
40.00	4644.0
44.00	4677.0
48.90	775.0
50.00	23040.0
51.10	98320.0
52.10	4442.0
55.00	369.0
56.00	2291.0
57.00	6367.0
57.90	419.0
61.00	1562.0
62.00	1293.0
63.00	3940.0
64.10	555.0
65.10	1852.0
68.00	1405.0
69.00	90304.0
70.10	404.0
73.00	1295.0
74.00	8522.0
75.00	14630.0
76.00	4391.0
77.10	104472.0
78.10	7396.0
79.00	6243.0
80.00	4702.0
81.00	7264.0
82.00	1911.0
83.00	1861.0
84.10	511.0
85.00	1510.0
86.00	1725.0
87.00	1270.0
87.90	493.0
89.00	396.0
90.90	2070.0
92.00	2395.0
93.00	11634.0
94.00	928.0
94.90	384.0
96.10	776.0
98.00	8872.0
99.00	7857.0
100.00	552.0
101.00	4528.0
102.90	1514.0
104.00	2292.0
104.90	2098.0
105.90	703.0
107.00	31800.0
108.00	5181.0
109.00	857.0
110.00	60712.0
111.00	8136.0
111.90	981.0
113.10	371.0
116.10	1516.0
117.00	26176.0
118.10	1839.0
118.80	321.0
119.90	498.0
122.10	1821.0
123.10	3094.0
124.00	1918.0
125.00	1797.0
125.90	613.0
127.00	111296.0
128.00	7823.0
129.00	44824.0
130.00	4335.0
131.00	745.0
132.80	673.0
133.90	1337.0
135.00	4041.0
136.10	948.0
137.10	2107.0
137.70	500.0
140.00	702.0
141.00	5471.0
142.00	1441.0
143.00	1416.0
143.90	466.0
144.90	301.0
146.00	1242.0
147.00	3387.0
148.00	6919.0
149.00	1785.0
150.00	622.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

151.00	603.0
151.80	426.0
153.10	1635.0
154.00	1689.0
155.00	3318.0
156.00	4859.0
156.90	827.0
157.90	1040.0
158.90	1047.0
160.00	1571.0
161.00	2405.0
161.90	860.0
164.00	347.0
164.90	1899.0
166.10	2118.0
167.00	11787.0
168.00	7196.0
169.00	1075.0
170.00	523.0
170.80	423.0
172.10	1476.0
173.10	1637.0
174.00	2205.0
175.10	4021.0
176.00	1111.0
176.90	1526.0
178.00	684.0
179.00	8916.0
180.00	7335.0
181.00	2855.0
182.10	461.0
184.00	790.0
185.00	4514.0
186.00	33688.0
187.00	9245.0
188.10	965.0
189.00	2630.0
190.90	1561.0
192.10	2844.0
192.90	3568.0
193.90	802.0
196.00	7911.0
196.80	1570.0
198.00	241472.0
199.00	16608.0
200.10	1697.0
201.50	824.0
202.90	1350.0
204.00	8250.0
205.00	14975.0
206.00	61312.0
207.10	12053.0
208.00	2584.0
209.10	1288.0
210.20	1000.0
211.00	2460.0
213.10	354.0
215.00	930.0
216.00	1392.0
217.00	17984.0
218.00	2587.0
220.20	806.0
221.10	10522.0
223.10	3431.0
224.00	33064.0
225.00	8527.0
226.00	1074.0
227.00	13898.0
228.00	2011.0
229.00	3089.0
229.90	448.0
231.00	1121.0
232.10	315.0
233.00	322.0
234.10	883.0
234.90	1046.0
235.90	687.0
237.10	2004.0
238.80	336.0
239.90	432.0
240.90	596.0
242.00	1616.0
243.00	1548.0
244.10	27056.0
245.10	3022.0
246.00	4698.0
247.00	984.0
248.90	1464.0
252.90	514.0
254.10	883.0
255.00	137344.0
256.00	19704.0
257.10	2156.0
258.00	7128.0
259.10	1324.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

264.90	3439.0
265.90	418.0
273.00	3946.0
274.00	9270.0
275.00	63952.0
276.00	8866.0
277.00	4343.0
277.90	1174.0
281.10	523.0
282.10	450.0
283.10	581.0
283.90	691.0
285.10	819.0
293.00	1019.0
294.00	400.0
294.90	309.0
296.00	16400.0
297.00	2267.0
301.90	367.0
303.00	2194.0
304.00	694.0
308.10	313.0
314.00	1137.0
314.90	1350.0
315.90	723.0
320.80	439.0
321.20	447.0
322.10	371.0
323.00	4869.0
324.10	1030.0
327.10	1125.0
328.10	744.0
331.80	322.0
332.10	338.0
332.80	477.0
333.90	3483.0
335.00	751.0
340.90	547.0
346.00	785.0
352.10	1646.0
353.10	1319.0
354.00	2089.0
365.00	6951.0
366.00	1106.0
371.20	432.0
372.00	3071.0
373.00	861.0
383.00	869.0
383.90	355.0
389.80	384.0
390.90	364.0
402.00	1320.0
403.00	1513.0
403.90	824.0
420.90	994.0
421.90	1155.0
423.00	11547.0
424.00	5134.0
425.00	700.0
441.00	37736.0
442.00	242816.0
443.00	47096.0
444.10	3753.0
445.10	319.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

Scan 2505 (17.421 min): BM000954.D
DFTPP44

m/z	Abundance
36.00	314.0
40.00	4443.0
44.00	4711.0
45.00	387.0
72.90	692.0
190.90	450.0
192.90	482.0
207.00	3609.0
208.10	781.0
209.00	407.0
281.00	561.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44