


Data Path : Z:\HPCHEM1\BNA_M\Data\BM041415\
 Data File : BM000980.D
 Acq On : 14 Apr 2015 10:37
 Operator : TP/IZ
 Sample : DFTPP45
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP45

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM041415.M
 Title : SVOA CALIBRATION
 Last Update : Wed Apr 15 06:41:33 2015



AutoFind: Scans 2510, 2511, 2512; Background Corrected with Scan 2505

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.3	115560	PASS
68	69	0.00	2	1.6	1751	PASS
69	198	0.00	100	41.7	106360	PASS
70	69	0.00	2	0.5	548	PASS
127	198	10	80	49.9	127320	PASS
197	198	0.00	2	0.4	928	PASS
198	198	100	100	100.0	254954	PASS
199	198	5	9	6.6	16916	PASS
275	198	10	60	23.4	59557	PASS
365	198	1	100	2.4	6184	PASS
441	443	0.01	100	75.2	30680	PASS
442	198	50	100	81.4	207533	PASS
443	442	15	24	19.7	40784	PASS

m/z	Abundance
37.00	411.0
38.00	1085.0
39.10	8381.0
40.00	5018.0
41.20	303.0
44.00	4738.0
49.10	1089.0
50.10	27160.0
51.10	112152.0
52.10	6118.0
53.10	300.0
55.00	542.0
56.00	3317.0
57.00	8292.0
58.00	592.0
61.00	1208.0
62.00	1521.0
63.00	4576.0
64.00	766.0
65.00	1864.0
68.00	1666.0
69.00	103792.0
70.00	648.0
73.10	1281.0
74.00	9169.0
75.00	16145.0
76.00	5244.0
77.10	120040.0
78.10	8265.0
79.00	7553.0
80.00	5142.0
81.00	8601.0
82.00	2285.0
83.00	1922.0
83.90	348.0
85.00	1547.0
86.00	2319.0
87.00	1296.0
88.00	439.0
91.00	2045.0
92.00	2672.0
93.00	12032.0
94.00	672.0
95.10	396.0
96.00	610.0
97.00	329.0
98.00	8904.0
99.00	8926.0
100.00	813.0
101.00	5421.0
101.90	354.0
103.00	1642.0
104.00	3318.0
105.00	2185.0
106.00	976.0
107.00	32912.0
108.00	4990.0
109.00	1142.0
110.00	65112.0
111.00	9569.0
112.10	989.0
113.00	484.0
116.00	2227.0
117.00	25536.0
117.90	2242.0
118.90	390.0
120.00	609.0
122.00	2111.0
123.00	3399.0
123.90	1486.0
125.10	1548.0
126.10	417.0
127.00	116528.0
128.00	9976.0
129.00	44864.0
130.00	3260.0
131.10	961.0
132.60	363.0
133.10	437.0
134.00	1525.0
134.90	3332.0
136.00	1345.0
137.10	1974.0
138.00	462.0
138.90	306.0
139.90	426.0
141.00	5721.0
142.00	2366.0
143.10	1137.0
144.90	379.0
145.90	803.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

147.00	2573.0
148.00	6638.0
149.00	1586.0
150.90	664.0
151.80	439.0
152.20	499.0
153.00	1861.0
154.10	1012.0
155.10	2835.0
156.00	4318.0
157.20	1111.0
157.80	1355.0
159.10	642.0
160.00	1980.0
161.00	2912.0
162.00	939.0
163.00	331.0
163.90	386.0
165.00	2134.0
166.10	1115.0
167.00	10262.0
168.00	7783.0
168.80	785.0
169.10	910.0
169.90	411.0
171.10	491.0
171.90	1162.0
173.00	1252.0
174.00	2072.0
175.00	4413.0
176.10	1558.0
177.00	1998.0
178.10	747.0
179.00	7283.0
180.00	6194.0
181.00	2951.0
181.80	353.0
182.20	435.0
184.10	752.0
185.00	3591.0
186.00	27576.0
187.00	8387.0
188.00	1401.0
188.90	1493.0
191.00	1320.0
191.90	3198.0
193.00	3120.0
194.10	581.0
196.00	5744.0
197.00	905.0
198.00	214400.0
199.00	16020.0
200.00	1405.0
201.20	1168.0
201.60	1058.0
202.90	1373.0
204.00	7563.0
205.00	11782.0
206.10	55640.0
207.10	11181.0
208.00	2737.0
209.00	1294.0
210.00	1094.0
211.00	2165.0
212.10	451.0
215.00	512.0
216.00	1219.0
217.00	13073.0
218.00	1833.0
221.00	9226.0
222.00	1912.0
223.10	2397.0
224.10	28760.0
225.00	7550.0
226.10	1187.0
227.00	11428.0
228.00	1960.0
229.00	3098.0
229.80	532.0
231.00	899.0
233.00	365.0
233.90	1087.0
235.00	1054.0
236.10	740.0
236.90	1650.0
237.90	406.0
239.10	481.0
240.00	600.0
241.00	851.0
242.10	1489.0
243.00	1589.0
244.10	21000.0
245.10	2880.0
246.00	3650.0
247.00	802.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

249.00	679.0
252.10	374.0
253.00	567.0
254.00	558.0
255.00	106968.0
256.00	16201.0
257.00	1414.0
258.00	5192.0
259.00	540.0
265.00	2148.0
265.90	475.0
270.90	421.0
273.00	3157.0
274.00	7490.0
275.00	47848.0
276.00	6042.0
277.00	3447.0
277.90	730.0
282.90	446.0
285.00	727.0
292.10	301.0
293.00	766.0
293.90	361.0
296.00	11696.0
296.90	1718.0
303.00	1694.0
303.90	495.0
314.00	494.0
314.90	1030.0
315.80	690.0
321.10	358.0
323.10	3538.0
324.10	614.0
327.00	899.0
327.90	380.0
332.10	352.0
333.00	380.0
334.10	2677.0
335.00	589.0
341.10	614.0
346.00	951.0
352.10	1209.0
353.10	536.0
354.20	952.0
355.10	333.0
364.90	4214.0
366.00	620.0
372.10	1871.0
372.90	397.0
383.00	492.0
402.10	810.0
403.00	1368.0
403.90	507.0
421.10	931.0
422.10	1025.0
423.00	6048.0
424.00	2867.0
424.90	469.0
441.00	19808.0
442.00	125704.0
443.10	28448.0
444.00	2022.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

m/z	Abundance
36.00	343.0
37.00	544.0
38.00	1598.0
39.00	10028.0
40.00	4894.0
41.10	578.0
44.00	4777.0
45.10	387.0
49.00	1083.0
50.10	32560.0
51.10	137664.0
52.00	6485.0
53.00	340.0
55.00	771.0
56.00	3538.0
57.00	9189.0
58.00	531.0
61.00	1832.0
62.00	1780.0
63.10	5307.0
64.00	534.0
65.10	2032.0
68.10	1852.0
69.00	124552.0
69.90	592.0
73.10	1538.0
74.00	12013.0
75.00	19136.0
76.00	6673.0
77.10	142144.0
78.10	9933.0
79.00	9188.0
79.90	6972.0
81.00	11000.0
82.00	2234.0
83.00	1991.0
83.90	458.0
84.90	1771.0
86.00	2860.0
87.00	1331.0
87.90	548.0
91.00	2322.0
92.00	3170.0
93.00	14897.0
94.00	1342.0
95.20	427.0
95.90	1329.0
98.00	12387.0
99.00	10170.0
99.90	816.0
101.00	5739.0
101.80	311.0
103.00	2086.0
104.00	3833.0
105.00	2848.0
106.00	1229.0
107.00	43048.0
108.00	6372.0
109.00	1550.0
110.00	78912.0
111.00	11620.0
112.00	1632.0
112.90	444.0
115.90	1764.0
117.00	32304.0
118.00	2905.0
118.90	401.0
120.00	442.0
122.00	3131.0
123.00	4512.0
123.90	1580.0
124.90	2048.0
126.10	929.0
127.00	150912.0
128.10	10702.0
129.00	57488.0
130.00	4093.0
131.10	915.0
132.10	700.0
133.00	530.0
133.90	2032.0
135.00	4358.0
136.00	1609.0
136.90	2313.0
138.00	636.0
138.90	331.0
139.90	873.0
141.00	6758.0
141.90	2092.0
143.00	2086.0
144.90	419.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

145.80	830.0
146.10	846.0
147.00	4030.0
148.00	8155.0
148.90	2345.0
150.00	379.0
151.30	1118.0
152.00	645.0
152.90	2319.0
154.10	2199.0
155.10	4347.0
156.00	5139.0
157.00	1165.0
158.10	1296.0
159.00	1199.0
160.00	2254.0
161.00	3810.0
162.00	824.0
163.10	427.0
164.00	477.0
165.00	2417.0
165.90	1845.0
167.00	12866.0
168.00	8215.0
169.00	1447.0
169.70	399.0
171.00	706.0
172.10	1322.0
173.00	2085.0
174.10	2924.0
175.00	5253.0
176.00	1773.0
177.00	2714.0
178.00	714.0
179.00	10909.0
180.00	7529.0
181.00	3734.0
182.00	514.0
183.10	300.0
183.90	893.0
185.00	5325.0
186.00	41192.0
187.10	11508.0
188.00	1595.0
189.00	2768.0
189.90	352.0
191.00	1545.0
192.00	3318.0
193.00	4478.0
194.00	980.0
195.10	474.0
196.00	9302.0
197.00	1880.0
198.00	304192.0
199.00	20184.0
200.10	964.0
201.70	1283.0
202.90	1561.0
204.00	8780.0
205.00	18192.0
206.10	77120.0
207.10	12153.0
208.00	3071.0
209.00	1513.0
210.10	979.0
210.90	2277.0
213.00	367.0
214.90	1073.0
216.10	1532.0
217.00	20960.0
218.00	2377.0
221.00	13763.0
222.00	3061.0
223.00	3932.0
224.10	40184.0
225.00	9990.0
226.00	837.0
227.00	16050.0
228.10	2190.0
229.10	3552.0
229.80	466.0
231.00	1236.0
233.00	346.0
234.00	916.0
235.10	1282.0
236.10	1038.0
237.00	1792.0
238.90	567.0
240.00	583.0
240.90	811.0
242.00	1747.0
243.10	2235.0
244.10	31664.0
245.10	3958.0
246.00	5401.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

247.00	912.0
247.90	329.0
249.00	1338.0
252.00	479.0
253.10	592.0
254.00	1163.0
255.00	159744.0
256.00	24256.0
257.00	2619.0
258.00	8829.0
259.10	1662.0
260.10	362.0
261.00	394.0
265.00	3325.0
266.10	601.0
271.20	340.0
272.00	512.0
273.00	5190.0
274.00	13458.0
275.00	67064.0
276.10	9460.0
277.00	6878.0
277.90	1121.0
281.20	554.0
283.00	648.0
285.20	738.0
292.10	309.0
293.10	1178.0
293.90	358.0
295.00	439.0
296.00	18648.0
297.10	2191.0
302.10	421.0
303.10	1718.0
304.10	865.0
308.00	506.0
310.00	368.0
313.90	760.0
314.90	1845.0
316.00	1070.0
321.00	911.0
322.00	376.0
323.00	5268.0
324.00	976.0
326.90	1070.0
328.00	487.0
331.90	429.0
333.00	383.0
334.00	3649.0
335.00	1158.0
340.80	1023.0
346.00	1014.0
352.00	1960.0
352.90	929.0
354.00	1829.0
364.90	7358.0
366.00	1057.0
369.90	312.0
371.00	538.0
372.00	3183.0
373.10	373.0
383.00	1235.0
384.00	434.0
390.00	514.0
402.00	1226.0
402.90	2097.0
403.80	704.0
421.00	1694.0
422.00	1847.0
423.00	13196.0
424.00	3863.0
424.90	886.0
441.00	34992.0
442.00	241984.0
443.00	45456.0
444.00	5046.0
445.00	348.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

m/z	Abundance
38.00	1192.0
39.10	7242.0
40.00	4769.0
41.10	731.0
44.00	4822.0
45.00	368.0
49.00	842.0
50.10	23136.0
51.10	96864.0
52.10	4609.0
55.00	555.0
56.00	2663.0
57.00	6751.0
58.10	360.0
60.90	1209.0
62.00	940.0
63.00	3672.0
64.00	716.0
65.00	1741.0
67.90	1735.0
69.00	90736.0
70.00	406.0
73.00	1146.0
74.00	8999.0
75.00	14672.0
76.10	5283.0
77.10	103608.0
78.00	6119.0
79.00	6362.0
80.00	5857.0
81.00	7115.0
82.00	2105.0
83.00	1871.0
84.00	436.0
85.00	1538.0
86.00	2153.0
86.90	793.0
87.90	490.0
89.10	309.0
91.00	1701.0
92.00	1999.0
93.00	11180.0
93.90	673.0
96.00	810.0
98.00	8562.0
99.00	7488.0
100.00	564.0
101.00	4687.0
102.90	1689.0
104.00	2735.0
105.00	2152.0
106.00	1023.0
107.00	32080.0
108.00	5366.0
109.10	811.0
110.00	58200.0
111.00	8386.0
112.00	1015.0
113.00	310.0
116.00	2238.0
117.00	24952.0
118.00	2201.0
119.00	457.0
120.10	561.0
122.00	2105.0
123.00	2526.0
123.90	1158.0
125.00	1441.0
126.10	559.0
127.00	114520.0
128.00	8334.0
129.00	41976.0
129.90	3495.0
130.90	588.0
132.00	486.0
133.00	720.0
134.00	1274.0
135.00	3187.0
136.00	1095.0
136.90	1678.0
140.00	639.0
141.00	5473.0
142.00	1151.0
143.00	1297.0
144.10	412.0
145.10	357.0
146.00	1341.0
147.00	3114.0
148.00	7244.0
149.00	1790.0
150.10	535.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

150.80	498.0
151.20	537.0
152.00	653.0
153.00	2021.0
154.00	1445.0
155.00	3143.0
156.00	4524.0
157.00	1026.0
158.10	1067.0
159.00	962.0
160.00	1652.0
161.00	2728.0
161.90	957.0
162.90	393.0
164.00	374.0
165.10	2251.0
165.90	1999.0
167.00	12489.0
168.00	6414.0
168.90	695.0
170.00	374.0
171.10	411.0
172.10	1345.0
173.00	1210.0
174.00	2481.0
175.00	4265.0
175.90	999.0
176.20	995.0
177.00	2370.0
178.00	753.0
179.00	9808.0
180.00	6552.0
181.00	2730.0
182.00	341.0
183.90	1079.0
185.00	3874.0
186.00	32032.0
187.00	8493.0
188.00	1226.0
189.00	2200.0
190.00	508.0
191.00	1346.0
191.90	2469.0
193.00	3275.0
194.10	789.0
196.10	6660.0
198.00	246272.0
199.00	14544.0
200.10	1110.0
201.40	1428.0
202.90	1206.0
204.00	7872.0
205.00	15266.0
206.10	63184.0
207.10	12912.0
208.00	2536.0
208.90	1453.0
209.90	863.0
211.10	2697.0
215.00	507.0
216.00	1251.0
217.00	16488.0
217.90	1780.0
221.10	12087.0
221.80	2673.0
223.00	3580.0
224.00	35680.0
225.00	7984.0
226.10	888.0
227.00	14656.0
228.00	1784.0
228.90	2725.0
229.90	369.0
230.90	1349.0
233.00	347.0
234.00	1281.0
235.10	1336.0
235.90	948.0
237.10	1689.0
238.90	423.0
240.10	400.0
241.00	608.0
242.00	1453.0
243.10	1966.0
244.00	26456.0
245.00	3321.0
246.00	4854.0
247.00	1290.0
247.90	408.0
249.10	1124.0
251.00	306.0
252.10	319.0
253.00	479.0
254.10	1163.0
255.00	134272.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

256.00	18840.0
256.90	1721.0
258.00	7950.0
259.00	983.0
265.00	3763.0
266.10	519.0
273.00	4113.0
274.00	10898.0
275.00	63760.0
276.00	7969.0
277.00	4766.0
277.80	805.0
281.10	373.0
283.10	706.0
284.00	318.0
285.10	1020.0
291.90	586.0
293.00	1402.0
293.80	556.0
295.00	520.0
296.00	16672.0
296.90	2345.0
301.90	440.0
303.00	1596.0
303.90	355.0
308.00	336.0
314.10	817.0
315.00	1453.0
316.10	1037.0
321.90	349.0
323.10	5622.0
324.00	847.0
327.00	878.0
328.10	525.0
333.00	500.0
334.10	3134.0
335.10	1434.0
340.80	563.0
346.00	1274.0
346.90	387.0
352.00	1463.0
353.00	1139.0
354.10	2450.0
355.00	390.0
365.00	6981.0
365.80	1327.0
371.10	307.0
372.00	2890.0
373.00	731.0
383.10	878.0
383.90	464.0
390.00	301.0
402.10	1332.0
403.00	2272.0
404.00	704.0
405.00	356.0
421.00	1334.0
422.00	1583.0
423.00	12037.0
424.10	4020.0
425.00	449.0
441.00	37240.0
442.00	254912.0
443.00	48448.0
444.10	3609.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45

Scan 2505 (17.421 min): BM000980.D
DFTPP45

m/z	Abundance
36.00	490.0
40.00	4428.0
43.00	329.0
44.10	4549.0
48.90	331.0
73.00	668.0
96.10	334.0
133.00	354.0
147.00	360.0
190.90	328.0
193.00	302.0
207.00	3833.0
207.90	370.0
209.00	682.0

Instrument :
BNA_M
ClientSampleId :
DFTPP45