


Data Path : Z:\HPCHEM1\BNA M\DATA\BM041416\
 Data File : BM004945.D
 Acq On : 14 Apr 2016 13:21
 Operator : UM/SJ
 Sample : DFTPP51
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP51

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM041416.M
 Title : SVOA CALIBRATION
 Last Update : Fri Apr 15 00:36:20 2016



AutoFind: Scans 2539, 2540, 2541; Background Corrected with Scan 2534

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.3	27098	PASS
68	69	0.00	2	2.0	570	PASS
69	198	0.00	100	32.0	28586	PASS
70	69	0.00	2	0.4	107	PASS
127	198	10	80	43.6	38981	PASS
197	198	0.00	2	0.7	598	PASS
198	198	100	100	100.0	89442	PASS
199	198	5	9	6.7	5987	PASS
275	198	10	60	26.0	23253	PASS
365	198	1	100	3.1	2787	PASS
441	443	0.01	100	80.2	12573	PASS
442	198	50	100	90.6	81018	PASS
443	442	15	24	19.4	15679	PASS

m/z	Abundance
41.00	585.0
42.00	330.0
43.00	1028.0
44.00	989.0
45.00	454.0
47.00	876.0
47.90	418.0
49.00	4817.0
51.00	1547.0
55.00	460.0
57.00	421.0
60.00	406.0
73.00	383.0
83.00	348.0
84.00	4013.0
86.00	2519.0
88.00	470.0
167.00	357.0
207.10	490.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

m/z	Abundance
38.00	343.0
39.00	1865.0
41.10	560.0
42.00	398.0
43.00	941.0
44.00	889.0
45.00	465.0
47.00	805.0
47.90	397.0
49.00	4569.0
50.10	5880.0
51.10	25512.0
52.00	1531.0
55.00	536.0
56.00	876.0
57.00	1876.0
60.00	324.0
61.00	359.0
62.00	423.0
63.00	1006.0
65.00	541.0
68.00	518.0
69.00	25400.0
73.00	475.0
74.00	2624.0
75.00	4327.0
76.00	1483.0
77.10	31584.0
78.00	2279.0
79.00	1997.0
80.00	1498.0
81.00	2261.0
82.00	625.0
83.00	795.0
84.00	3481.0
85.00	634.0
86.00	3176.0
87.00	361.0
87.90	424.0
91.00	652.0
92.00	622.0
93.00	3310.0
98.00	2899.0
99.00	2287.0
101.00	1341.0
103.00	367.0
104.00	758.0
105.00	765.0
106.00	302.0
107.00	10087.0
108.00	1485.0
109.10	352.0
110.00	17144.0
111.00	2720.0
112.00	397.0
116.00	579.0
117.00	7880.0
118.00	687.0
122.00	610.0
123.00	1027.0
124.00	540.0
125.00	418.0
126.20	336.0
127.00	32488.0
128.00	2597.0
129.00	12693.0
130.00	1082.0
134.10	376.0
135.00	1080.0
136.00	416.0
137.00	626.0
141.00	1621.0
142.00	528.0
143.00	343.0
147.00	953.0
148.00	2088.0
149.00	580.0
153.00	543.0
154.00	318.0
155.00	947.0
156.10	1484.0
157.90	317.0
160.00	522.0
161.00	784.0
165.00	619.0
166.00	577.0
167.00	3573.0
168.00	2034.0
169.00	394.0
172.00	303.0
172.90	467.0
174.00	791.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

175.00	1367.0
176.00	478.0
177.00	576.0
179.00	2567.0
180.00	1891.0
181.00	891.0
185.00	1256.0
186.10	9728.0
187.10	2744.0
188.00	355.0
189.00	685.0
191.00	323.0
192.00	696.0
193.00	929.0
196.00	1887.0
197.10	428.0
198.00	70328.0
199.00	4793.0
200.00	373.0
201.50	375.0
203.10	431.0
204.00	2498.0
205.00	4044.0
206.10	18432.0
207.10	2821.0
208.10	726.0
211.00	696.0
216.00	411.0
217.00	4693.0
218.00	614.0
221.10	2810.0
221.90	670.0
223.00	1050.0
224.10	9903.0
225.10	2552.0
227.00	4126.0
228.00	542.0
229.00	921.0
231.00	303.0
235.00	311.0
237.10	441.0
242.00	480.0
243.00	606.0
244.10	7479.0
245.00	1060.0
246.00	1490.0
247.00	394.0
249.00	353.0
254.10	354.0
255.10	36896.0
256.00	5802.0
257.00	502.0
258.00	2140.0
259.00	338.0
265.10	881.0
273.00	1081.0
274.00	3065.0
275.10	16656.0
276.10	2339.0
277.00	1520.0
293.00	341.0
296.00	4486.0
297.00	622.0
303.00	528.0
315.00	562.0
323.00	1368.0
334.10	823.0
346.00	380.0
352.10	464.0
353.00	310.0
354.00	552.0
365.00	1763.0
365.90	308.0
372.10	775.0
403.10	451.0
421.00	357.0
422.00	365.0
423.00	2356.0
424.10	836.0
441.10	7351.0
442.10	48520.0
443.10	9822.0
444.00	998.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

m/z	Abundance
38.10	401.0
39.10	2171.0
40.00	354.0
41.00	559.0
42.00	350.0
43.00	838.0
44.00	788.0
45.00	477.0
47.00	888.0
48.00	387.0
49.00	4105.0
50.10	7566.0
51.10	33080.0
52.00	1645.0
55.00	592.0
56.00	1066.0
57.00	2459.0
60.00	394.0
61.00	541.0
62.00	472.0
63.10	1360.0
65.10	733.0
68.10	659.0
69.00	32960.0
73.00	564.0
74.00	3693.0
75.00	5645.0
76.10	1933.0
77.10	40864.0
78.10	2873.0
79.00	2592.0
80.00	1959.0
81.00	3037.0
82.00	845.0
83.00	815.0
84.00	3309.0
84.90	848.0
86.00	3085.0
87.00	419.0
87.90	546.0
91.00	833.0
92.00	800.0
93.00	4296.0
94.00	369.0
98.00	3743.0
99.00	2813.0
100.00	301.0
101.00	1665.0
103.00	530.0
104.00	1020.0
105.00	919.0
106.00	336.0
107.00	13476.0
108.00	2031.0
109.00	379.0
110.00	24368.0
111.00	3656.0
112.00	473.0
116.00	674.0
117.00	10676.0
118.00	800.0
122.10	738.0
123.00	1408.0
124.00	626.0
125.10	528.0
127.00	45216.0
128.00	3386.0
129.00	16992.0
130.00	1492.0
131.00	369.0
134.00	532.0
135.00	1413.0
136.00	641.0
137.10	900.0
141.00	2397.0
142.00	876.0
143.00	510.0
146.10	412.0
147.00	1262.0
148.00	2695.0
149.00	633.0
153.00	754.0
154.00	599.0
155.00	1424.0
156.10	2060.0
157.10	362.0
157.90	302.0
159.00	418.0
160.00	714.0
161.10	1059.0
162.00	416.0
165.00	822.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

166.00	737.0
167.00	5298.0
168.10	2700.0
169.00	530.0
172.00	515.0
173.00	545.0
174.00	1038.0
175.10	1922.0
176.00	553.0
177.00	930.0
178.00	325.0
179.00	3546.0
180.10	2697.0
181.00	1135.0
182.00	340.0
185.00	1714.0
186.10	14354.0
187.10	4167.0
188.00	488.0
189.00	903.0
191.00	492.0
192.00	1194.0
193.00	1408.0
194.00	336.0
196.10	3086.0
197.10	748.0
198.00	105216.0
199.00	7081.0
200.00	439.0
201.40	362.0
203.00	734.0
204.10	3830.0
205.10	6409.0
206.10	26760.0
207.10	3918.0
208.00	944.0
209.00	328.0
210.10	359.0
211.00	986.0
216.00	613.0
217.00	7295.0
218.10	802.0
221.10	4125.0
223.00	1559.0
224.10	14879.0
225.10	3777.0
226.10	380.0
227.00	6069.0
228.00	945.0
229.00	1330.0
231.00	541.0
234.00	400.0
235.00	424.0
236.00	342.0
237.00	596.0
242.00	778.0
243.10	906.0
244.10	11357.0
245.10	1724.0
246.00	2357.0
247.00	495.0
249.00	451.0
254.20	553.0
255.10	58904.0
256.10	9034.0
257.00	795.0
258.00	3308.0
259.00	547.0
265.00	1512.0
273.00	1910.0
274.00	5006.0
275.10	26744.0
276.10	3624.0
277.00	2384.0
278.00	425.0
285.00	412.0
293.00	471.0
296.00	7219.0
297.00	1034.0
303.10	885.0
304.00	300.0
314.00	363.0
315.00	774.0
316.00	490.0
323.10	2445.0
324.00	472.0
327.00	459.0
334.10	1624.0
335.00	455.0
346.00	601.0
352.00	682.0
353.00	478.0
354.10	869.0
365.00	3336.0
366.00	534.0
372.10	1125.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

383.00	340.0
402.00	617.0
403.00	787.0
421.10	659.0
422.00	550.0
423.10	4881.0
424.10	1191.0
441.10	14012.0
442.10	90056.0
443.10	17392.0
444.10	1638.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

m/z	Abundance
38.00	418.0
39.00	1862.0
41.00	613.0
42.10	380.0
43.00	925.0
44.00	801.0
45.00	442.0
47.00	832.0
48.00	446.0
49.00	4165.0
50.00	6666.0
51.10	27344.0
52.10	1331.0
55.00	546.0
56.00	947.0
57.00	2305.0
60.00	396.0
61.00	416.0
62.00	448.0
63.10	1160.0
65.10	585.0
68.10	533.0
69.00	27400.0
70.00	321.0
73.10	668.0
74.10	2981.0
75.00	4526.0
76.10	1658.0
77.10	33008.0
78.10	2287.0
79.00	2020.0
80.00	1657.0
81.00	2394.0
82.00	686.0
83.00	903.0
84.00	3836.0
85.00	798.0
86.00	2927.0
87.00	396.0
88.00	472.0
91.00	661.0
92.10	606.0
93.00	4109.0
94.00	324.0
98.00	3141.0
99.00	2614.0
101.00	1368.0
103.00	588.0
104.00	884.0
105.00	842.0
106.10	351.0
107.00	11241.0
108.00	1661.0
109.10	357.0
110.00	20248.0
111.00	3112.0
112.00	402.0
116.00	602.0
117.00	9479.0
118.00	686.0
122.00	818.0
123.00	1149.0
124.00	463.0
124.90	562.0
127.00	39240.0
128.00	3026.0
129.00	14802.0
130.00	1250.0
134.00	510.0
135.00	1331.0
136.00	490.0
137.00	768.0
140.90	2027.0
142.00	561.0
143.00	511.0
146.00	402.0
147.00	1137.0
148.00	2628.0
149.00	428.0
152.90	649.0
154.00	479.0
155.10	1069.0
156.00	1769.0
157.00	463.0
157.90	452.0
159.00	353.0
159.90	605.0
161.00	1101.0
162.00	315.0
165.00	814.0
166.00	615.0
167.00	4672.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

168.00	2439.0
169.00	348.0
172.00	397.0
173.00	594.0
174.00	934.0
175.10	1691.0
176.00	441.0
177.00	945.0
178.10	352.0
179.00	3459.0
180.00	2289.0
181.00	865.0
185.10	1640.0
186.10	12745.0
187.00	3824.0
188.00	518.0
189.00	774.0
190.90	450.0
192.00	1099.0
193.00	1408.0
194.00	335.0
196.00	2500.0
197.10	620.0
198.00	92784.0
199.00	6089.0
200.00	506.0
201.60	300.0
203.00	594.0
204.10	3287.0
205.10	5842.0
206.10	25048.0
207.10	3483.0
208.00	839.0
209.00	328.0
210.00	377.0
211.10	952.0
216.00	586.0
217.00	6501.0
218.00	774.0
221.00	3806.0
221.90	822.0
223.00	1313.0
224.10	13989.0
225.10	3709.0
227.00	5928.0
228.00	797.0
229.00	1181.0
231.10	539.0
234.00	420.0
235.00	404.0
237.10	493.0
242.10	716.0
243.10	667.0
244.10	10908.0
245.10	1491.0
246.00	2272.0
247.00	465.0
249.00	391.0
254.20	540.0
255.10	56024.0
256.10	7917.0
257.00	624.0
258.00	3109.0
259.00	478.0
265.00	1412.0
273.00	1788.0
274.10	4868.0
275.10	26360.0
276.00	3240.0
277.00	2214.0
278.00	391.0
285.00	440.0
293.00	520.0
296.00	7526.0
297.00	906.0
303.10	978.0
304.00	300.0
314.00	314.0
315.00	841.0
316.00	556.0
323.10	2311.0
324.00	509.0
327.00	511.0
333.00	304.0
334.00	1579.0
335.10	454.0
346.00	602.0
352.00	772.0
353.00	536.0
354.00	673.0
365.00	3263.0
366.00	478.0
372.10	1302.0
373.00	412.0
383.00	403.0
402.00	620.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51

403.10	888.0
404.00	336.0
421.00	747.0
422.10	782.0
423.10	5315.0
424.10	1365.0
441.10	16358.0
442.10	104480.0
443.10	19824.0
444.10	1723.0

Instrument :
BNA_M
ClientSampleId :
DFTPP51