

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM041421\
 Data File : BM029463.D
 Acq On : 14 Apr 2021 10:41
 Operator : CG/JU
 Sample : PB135558BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS558

Manual Integrations
 APPROVED

mohammad
 4/15/2021 11:45:38 AM

Quant Time: Apr 14 11:27:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM041221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 14 10:06:38 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	81101	20.00	ng/ul	0.00
20) Naphthalene-d8	10.46	136	329250	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.32	164	187653	20.00	ng/ul	0.00
64) Phenanthrene-d10	17.06	188	339792	20.00	ng/ul	0.00
78) Chrysene-d12	21.24	240	311702	20.00	ng/ul	0.00
87) Perylene-d12	23.44	264	312539	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.19	96	14944	6.68	ng/uL	0.00
4) Pyridine-d5	3.59	84	195900	33.05	ng/ul	0.00
7) Phenol-d5	6.85	99	246357	33.18	ng/ul	0.00
9) Bis-(2-Chloroethyl)ether-d	7.02	67	161519	33.33	ng/ul	0.00
11) 2-Chlorophenol-d4	7.22	132	192949	34.69	ng/ul	0.00
15) 4-Methylphenol-d8	8.39	113	188527	33.19	ng/ul	0.00
21) Nitrobenzene-d5	8.83	128	86212	33.82	ng/ul	0.00
24) 2-Nitrophenol-d4	9.55	143	93085	33.54	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.08	165	176828	33.88	ng/ul	0.00
31) 4-Chloroaniline-d4	10.60	131	210837	28.39	ng/ul	0.00
46) Dimethylphthalate-d6	13.73	166	502377	34.73	ng/ul	0.00
49) Acenaphthylene-d8	14.01	160	635846	36.84	ng/ul	0.00
54) 4-Nitrophenol-d4	14.52	143	90657	32.75	ng/ul	0.00
60) Fluorene-d10	15.31	176	408968	34.34	ng/ul	0.00
65) 4,6-Dinitro-2-methylphenol	15.43	200	74906	33.98	ng/ul	0.00
73) Anthracene-d10	17.16	188	601220	36.61	ng/ul	0.00
80) Pyrene-d10	19.45	212	609778	36.91	ng/ul	0.00
91) Benzo(a)pyrene-d12	23.30	264	630417	36.98	ng/ul	0.00

Target Compounds

					Ovalue
2) 1,4-Dioxane	3.22	88	31350	13.551	ng/uL 94
5) Pyridine	3.61	79	208468	32.817	ng/ul 97
6) Benzaldehyde	6.83	77	150934	36.521	ng/ul 99
8) Phenol	6.88	94	265244	33.651	ng/ul 99
10) Bis(2-Chloroethyl)ether	7.11	93	202483	33.812	ng/ul 99
12) 2-Chlorophenol	7.25	128	207788	35.051	ng/ul 97
13) 2-Methylphenol	8.12	108	190748	33.586	ng/ul 99
14) 2,2'-oxybis(1-Chloropropan	8.20	45	320946	34.254	ng/ul 98
16) Acetophenone	8.50	105	312077	33.116	ng/ul 99
17) N-Nitroso-di-n-propylamine	8.49	70	182529	35.435	ng/ul 95
18) 4-Methylphenol	8.45	108	206427	33.738	ng/ul 96
19) Hexachloroethane	8.75	117	94281	35.209	ng/ul 97
22) Nitrobenzene	8.87	77	267334	34.471	ng/ul 99
23) Isophorone	9.40	82	458284	34.773	ng/ul 98
25) 2-Nitrophenol	9.58	139	107603	34.639	ng/ul 99
26) 2,4-Dimethylphenol	9.65	107	234972	34.428	ng/ul 98
27) Bis(2-Chloroethoxy)methane	9.88	93	263928	34.267	ng/ul 98
29) 2,4-Dichlorophenol	10.11	162	181776	34.831	ng/ul 98
30) Naphthalene	10.51	128	627589	34.234	ng/ul 99
32) 4-Chloroaniline	10.62	127	219253	28.599	ng/ul 99
33) Hexachlorobutadiene	10.80	225	106683	34.584	ng/ul 97
34) Caprolactam	11.40	113	53066m	32.712	ng/ul

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM041421\
 Data File : BM029463.D
 Acq On : 14 Apr 2021 10:41
 Operator : CG/JU
 Sample : PB135558BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS558

Manual Integrations
 APPROVED

mohammad
 4/15/2021 11:45:38 AM

Quant Time: Apr 14 11:27:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM041221.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 14 10:06:38 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) 4-Chloro-3-methylphenol	11.75	107	205450	35.435	ng/ul	99
36) 2-Methylnaphthalene	12.13	142	436880	34.732	ng/ul	95
37) 1-Methylnaphthalene	12.35	142	420556	33.470	ng/ul	96
39) 1,2,4,5-Tetrachlorobenzene	12.50	216	186448	33.951	ng/ul	98
40) Hexachlorocyclopentadiene	12.48	237	118166	31.158	ng/ul	98
41) 2,4,6-Trichlorophenol	12.74	196	130746	35.167	ng/ul	96
42) 2,4,5-Trichlorophenol	12.81	196	135490	33.959	ng/ul	97
43) 1,1'-Biphenyl	13.15	154	546100	34.274	ng/ul	99
44) 2-Chloronaphthalene	13.19	162	414055	34.037	ng/ul	99
45) 2-Nitroaniline	13.39	65	153566	36.624	ng/ul	94
47) Dimethylphthalate	13.78	163	527407	35.591	ng/ul	100
48) 2,6-Dinitrotoluene	13.90	165	104275	36.484	ng/ul	99
50) Acenaphthylene	14.04	152	631133	34.720	ng/ul	99
51) 3-Nitroaniline	14.22	138	95105	30.796	ng/ul	98
52) Acenaphthene	14.38	153	457038	34.185	ng/ul	99
53) 2,4-Dinitrophenol	14.43	184	55339	31.291	ng/ul	96
55) 4-Nitrophenol	14.53	109	97280	34.679	ng/ul	97
56) Dibenzofuran	14.72	168	618398	34.871	ng/ul	99
57) 2,4-Dinitrotoluene	14.69	165	148850	36.803	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.94	232	109494	35.225	ng/ul	99
59) Diethylphthalate	15.15	149	560834	36.237	ng/ul	98
61) Fluorene	15.37	166	537627	35.558	ng/ul	98
62) 4-Chlorophenyl-phenylether	15.36	204	240447	35.586	ng/ul	99
63) 4-Nitroaniline	15.39	138	107255	35.743	ng/ul	95
66) 4,6-Dinitro-2-methylphenol	15.44	198	80366	35.180	ng/ul	93
67) N-Nitrosodiphenylamine	15.58	169	442514	37.500	ng/ul	100
68) 4-Bromophenyl-phenylether	16.26	248	130666	38.066	ng/ul	99
69) Hexachlorobenzene	16.37	284	145782	37.700	ng/ul	97
70) Atrazine	16.53	200	147980	35.142	ng/ul	98
71) Pentachlorophenol	16.71	266	87214	35.095	ng/ul	96
72) Phenanthrene	17.10	178	768102	36.652	ng/ul	99
74) Anthracene	17.19	178	798057	37.158	ng/ul	100
75) Pentachlorobenzene	14.64	250	172881	34.317	ng/ul	99
76) Carbazole	17.46	167	704050	35.506	ng/ul	99
77) Di-n-butylphthalate	18.03	149	923217	38.639	ng/ul	100
79) Fluoranthene	19.12	202	815755	37.541	ng/ul	98
81) Pyrene	19.48	202	852692	37.076	ng/ul	99
82) Butylbenzylphthalate	20.39	149	400809	38.905	ng/ul	96
83) 3,3'-Dichlorobenzidine	21.16	252	232572	34.048	ng/ul	99
84) Benzo(a)anthracene	21.22	228	777483	37.532	ng/ul	100
85) Bis(2-ethylhexyl)phthalate	21.16	149	640127	39.741	ng/ul	100
86) Chrysene	21.27	228	750114	37.492	ng/ul	99
88) Di-n-octyl phthalate	22.03	149	1080606	36.555	ng/ul	100
89) Benzo(b)fluoranthene	22.79	252	777836	36.856	ng/ul	99
90) Benzo(k)fluoranthene	22.83	252	790277	37.373	ng/ul	99
92) Benzo(a)pyrene	23.35	252	707878	37.809	ng/ul	99
93) Indeno(1,2,3-cd)pyrene	25.66	276	919826	38.699	ng/ul	99
94) Dibenzo(a,h)anthracene	25.68	278	782630	39.131	ng/ul	99
95) Benzo(a,h,i)perylene	26.34	276	745155	38.747	ng/ul	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM041421\
Data File : BM029463.D
Acq On : 14 Apr 2021 10:41
Operator : CG/JU
Sample : PB135558BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS558


Manual Integrations
APPROVED

mohammad
4/15/2021 11:45:38 AM

Quant Time: Apr 14 11:27:11 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM041221.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 14 10:06:38 2021
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 13.551 ng/uL

RT: 3.22 min Scan# 56

Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M





ClientSampled :

SLCS558

Tot Ion:	88	Resp:	31350
Ion	Ratio	Lower	Upper
88	100		
43	44.0	32.4	48.6
58	77.2	65.9	98.9

Manual Integrations
APPROVED

mohammad
4/15/2021 11:45:38 AM



#5

Pyridine

Concen: 32.817 ng/uL




RT: 3.61 min Scan# 123


Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Tgt Ion:	79	Resp:	208468
Ion	Ratio	Lower	Upper
79	100		
52	85.1	66.6	99.8
51	42.8	32.2	48.2





#6

Benzaldehyde

Concen: 36.521 ng/ul

RT: 6.83 min Scan# 670

Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M





ClientSampleId :

SLCS558

Tot Ion:	77	Resp:	150934
Ion	Ratio	Lower	Upper
77	100		
105	84.0	67.4	101.2
106	80.5	65.3	97.9

Manual Integrations
APPROVED

mohammad
4/15/2021 11:45:38 AM



#8

Phenol

Concen: 33.651 ng/ul




RT: 6.88 min Scan# 679


Delta R.T. 0.01 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Tgt Ion:	94	Resp:	265244
Ion	Ratio	Lower	Upper
94	100		
65	34.2	27.4	41.0
66	44.8	35.3	52.9





#10





Bis(2-Chloroethyl)ether
 Concen: 33.812 ng/ul
 RT: 7.11 min Scan# 718
 Delta R.T. 0.00 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Instrument :
 BNA_M
 ClientSampleId :
 SLCS558

Tot Ion	Ratio	Lower	Upper
93	100		
63	86.4	68.5	102.7
95	32.3	26.5	39.7

Manual Integrations
 APPROVED




mohammad
 4/15/2021 11:45:38 AM




#12

2-Chlorophenol
 Concen: 35.051 ng/ul
 RT: 7.25 min Scan# 741
 Delta R.T. 0.00 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
128	100		
64	55.4	42.3	63.5
130	32.0	24.6	37.0





#13

2-Methylphenol

Concen: 33.586 ng/ul

RT: 8.12 min Scan# 890

Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M

ClientSampled :

SLCS558

Tot Ion:108 Resp: 190748

Ion Ratio Lower Upper

108 100





107 91.3 72.6 109.0

Manual Integrations

APPROVED

mohammad

4/15/2021 11:45:38 AM



#14

2,2'-oxybis(1-Chloropropane)

Concen: 34.254 ng/ul

RT: 8.20 min Scan# 904

Delta R.T. -0.01 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41




Tgt Ion: 45 Resp: 320946


Ion Ratio Lower Upper

45 100

77 14.8 11.4 17.2

79 10.5 9.4 14.0









#16
Acetophenone
Concen: 33.116 ng/ul
RT: 8.50 min Scan# 954
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampled :
SLCS558

Tot Ion:	105	Resp:	312077
Ion Ratio	Lower	Upper	
105	100		
77	89.3	70.8	106.2
51	35.1	26.9	40.3




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#17
N-Nitroso-di-n-propylamine
Concen: 35.435 ng/ul
RT: 8.49 min Scan# 952
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion:	70	Resp:	182529
Ion Ratio	Lower	Upper	
70	100		
42	71.8	53.0	79.4
101	8.9	7.0	10.4
130	17.0	13.8	20.6









#18
4-Methylphenol
Concen: 33.738 ng/ul
RT: 8.45 min Scan# 946
Delta R.T. 0.01 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampled :
SLCS558

Tot Ion	Ratio	Resp	Lower	Upper
108	100	206427		
107	111.2	85.8	128.6	




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#19
Hexachloroethane
Concen: 35.209 ng/ul
RT: 8.75 min Scan# 997
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Resp	Lower	Upper
117	100	94281		
201	80.7	61.6	92.4	
199	49.0	38.0	57.0	









#22
Nitrobenzene
Concen: 34.471 ng/ul
RT: 8.87 min Scan# 1018
Delta R.T. 0.01 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampleId :
SLCS558

Tot Ion	Ratio	Lower	Upper
77	100		
123	36.1	29.2	43.8
65	14.2	11.2	16.8




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#23
Isophorone
Concen: 34.773 ng/ul
RT: 9.40 min Scan# 1107
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
82	100		
95	7.1	5.8	8.8
138	15.3	13.0	19.4









#25
2-Nitrophenol
Concen: 34.639 ng/ul
RT: 9.58 min Scan# 1138
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampleId :
SLCS558

Tot Ion	Ratio	Resp	Lower	Upper
139	100	107603		
65	67.3		55.1	82.7
109	36.5		29.2	43.8




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#26
2,4-Dimethylphenol
Concen: 34.428 ng/ul
RT: 9.65 min Scan# 1149
Delta R.T. 0.01 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Resp	Lower	Upper
107	100	234972		
121	49.5		38.5	57.7
122	81.7		63.8	95.8





#27

Bis(2-Chloroethoxy)methane

Concen: 34.267 ng/ul

RT: 9.88 min Scan# 1189

Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M





ClientSampleId :

SLCS558

Tot Ion	93	Resp	263928
Ion Ratio	Lower	Upper	
93	100		
95	32.2	26.9	40.3
123	10.1	8.2	12.2

Manual Integrations
APPROVED

mohammad
4/15/2021 11:45:38 AM



#29

2,4-Dichlorophenol

Concen: 34.831 ng/ul




RT: 10.11 min Scan# 1228


Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Tgt Ion	162	Resp	181776
Ion Ratio	Lower	Upper	
162	100		
164	64.7	49.9	74.9
98	40.1	33.0	49.4









#30
Naphthalene
Concen: 34.234 ng/ul
RT: 10.51 min Scan# 1296
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampleId :
SLCS558

Tot Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.9	13.3
127	13.4	10.4	15.6




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#32
4-Chloroaniline
Concen: 28.599 ng/ul
RT: 10.62 min Scan# 1315
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
127	100		
129	32.0	25.0	37.4









#33
Hexachlorobutadiene
Concen: 34.584 ng/ul
RT: 10.80 min Scan# 1345
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampled :
SLCS558

Tot Ion	Ratio	Lower	Upper
225	100		
223	60.6	51.3	76.9
227	65.1	53.4	80.0




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#34
Caprolactam
Concen: 32.712 ng/ul m
RT: 11.40 min Scan# 1447
Delta R.T. 0.01 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
113	100		
55	194.2	144.8	217.2
56	142.8	115.4	173.0









#35
4-Chloro-3-methylphenol
Concen: 35.435 ng/ul
RT: 11.75 min Scan# 1507
Delta R.T. 0.01 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampleId :
SLCS558

Tot Ion	Ratio	Lower	Upper
107	100		
144	25.3	20.3	30.5
142	76.2	61.6	92.4




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#36
2-Methylnaphthalene
Concen: 34.732 ng/ul
RT: 12.13 min Scan# 1571
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
142	100		
141	88.3	74.6	112.0









#37
1-Methylnaphthalene
Concen: 33.470 ng/ul
RT: 12.35 min Scan# 1608
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampleId :
SLCS558

Tot Ion	Ratio	Lower	Upper
142	100		
141	95.0	73.2	109.8
116	4.3	3.5	5.3




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#39
1,2,4,5-Tetrachlorobenzene
Concen: 33.951 ng/ul
RT: 12.50 min Scan# 1634
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
216	100		
214	77.1	63.0	94.4
179	25.2	19.3	28.9
108	24.7	18.6	28.0





#40

Hexachlorocyclopentadiene

Concen: 31.158 ng/ul

RT: 12.48 min Scan# 1631

Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M





ClientSampled :

SLCS558

Tot Ion:	237	Resp:	118166
Ion Ratio	Lower	Upper	
237	100		
235	61.9	50.8	76.2
272	11.9	10.6	15.8

Manual Integrations
APPROVED

mohammad
4/15/2021 11:45:38 AM



#41

2,4,6-Trichlorophenol

Concen: 35.167 ng/ul




RT: 12.74 min Scan# 1675


Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Tgt Ion:	196	Resp:	130746
Ion Ratio	Lower	Upper	
196	100		
198	93.7	78.5	117.7
200	30.3	24.1	36.1









#42
 2,4,5-Trichlorophenol
 Concen: 33.959 ng/ul
 RT: 12.81 min Scan# 1687
 Delta R.T. 0.00 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Instrument :
 BNA_M
 ClientSampleId :
 SLCS558

Tot Ion	Ratio	Resp	Lower	Upper
196	100	135490		
198	93.7	77.2	115.8	
200	30.4	25.0	37.6	




Manual Integrations
 APPROVED


mohammad
 4/15/2021 11:45:38 AM



#43
 1,1'-Biphenyl
 Concen: 34.274 ng/ul
 RT: 13.15 min Scan# 1745
 Delta R.T. 0.01 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	546100		
153	41.0	33.5	50.3	
76	14.2	11.8	17.6	









#44
 2-Chloronaphthalene
 Concen: 34.037 ng/ul
 RT: 13.19 min Scan# 1751
 Delta R.T. 0.00 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Instrument :
 BNA_M
 ClientSampleId :
 SLCS558

Tot Ion	Ratio	Lower	Upper
162	100		
164	32.7	26.0	39.0
127	42.2	32.8	49.2




Manual Integrations
 APPROVED


mohammad
 4/15/2021 11:45:38 AM



#45
 2-Nitroaniline
 Concen: 36.624 ng/ul
 RT: 13.39 min Scan# 1786
 Delta R.T. 0.00 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
65	100		
92	59.7	51.5	77.3
138	78.4	67.0	100.6









#47
Dimethylphthalate
Concen: 35.591 ng/ul
RT: 13.78 min Scan# 1852
Delta R.T. 0.01 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampled :
SLCS558

Tot Ion:	163	Resp:	527407
Ion Ratio	Lower	Upper	
163	100		
194	3.7	2.7	4.1
164	10.2	8.1	12.1




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#48
2,6-Dinitrotoluene
Concen: 36.484 ng/ul
RT: 13.90 min Scan# 1872
Delta R.T. 0.01 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion:	165	Resp:	104275
Ion Ratio	Lower	Upper	
165	100		
89	76.6	62.1	93.1
121	24.4	19.6	29.4





#50

Acenaphthylene

Concen: 34.720 ng/ul

RT: 14.04 min Scan# 1896

Delta R.T. 0.01 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M





ClientSampled :

SLCS558

Tot Ion:	152	Resp:	631133
Ion Ratio	Lower	Upper	
152	100		
151	19.9	16.4	24.6
153	12.6	10.2	15.4

Manual Integrations
APPROVED

mohammad
4/15/2021 11:45:38 AM



#51

3-Nitroaniline

Concen: 30.796 ng/ul




RT: 14.22 min Scan# 1927


Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Tgt Ion:	138	Resp:	95105
Ion Ratio	Lower	Upper	
138	100		
108	11.0	9.3	13.9
92	139.0	109.1	163.7





#52

Acenaphthene

Concen: 34.185 ng/ul

RT: 14.38 min Scan# 1954

Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M

ClientSampled :

SLCS558

Tot Ion:153 Resp: 457038

Ion Ratio Lower Upper


153 100

152 49.2 38.3 57.5

154 89.3 71.3 106.9

Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM





Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E



Time--> 14.30 14.35 14.40 14.45



#53

2,4-Dinitrophenol

Concen: 31.291 ng/ul

RT: 14.43 min Scan# 1963

Delta R.T. 0.01 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41


Tgt Ion:184 Resp: 55339

Ion Ratio Lower Upper

184 100

63 94.5 80.6 120.8


154 73.2 60.0 90.0




Abundance Ion 184.00 (183.70 to 184.70): E


Ion 63.00 (62.70 to 63.70): BM0

Ion 154.00 (153.70 to 154.70): E



Time--> 14.40 14.45









#55
4-Nitrophenol
Concen: 34.679 ng/ul
RT: 14.53 min Scan# 1980
Delta R.T. 0.01 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampleId :
SLCS558

Tot Ion	Ion	Ratio	Resp	Lower	Upper
109	100				
139	103.3	86.1	129.1		
65	138.4	112.6	168.8		




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#56
Dibenzofuran
Concen: 34.871 ng/ul
RT: 14.72 min Scan# 2011
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
168	100				
139	42.0	33.1	49.7		









#57
 2,4-Dinitrotoluene
 Concen: 36.803 ng/ul
 RT: 14.69 min Scan# 2006
 Delta R.T. 0.01 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Instrument :
 BNA_M
 ClientSampleId :
 SLCS558

Tot Ion	Ratio	Resp	Lower	Upper
165	100	148850		
63	61.2	48.6	72.8	
182	2.7	2.4	3.6	




Manual Integrations
 APPROVED


mohammad
 4/15/2021 11:45:38 AM



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 35.225 ng/ul
 RT: 14.94 min Scan# 2050
 Delta R.T. 0.00 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Resp	Lower	Upper
232	100	109494		
131	63.1	49.3	73.9	
130	3.4	2.8	4.2	
166	38.6	30.7	46.1	









#59
Diethylphthalate
Concen: 36.237 ng/ul
RT: 15.15 min Scan# 2085
Delta R.T. 0.01 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampled :
SLCS558

Tot Ion	Ratio	Lower	Upper
149	100		
177	19.6	16.3	24.5
150	11.9	10.0	15.0




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#61
Fluorene
Concen: 35.558 ng/ul
RT: 15.37 min Scan# 2122
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
166	100		
165	97.5	79.8	119.6
167	13.0	10.7	16.1









#62
4-Chlorophenyl-phenylether
Concen: 35.586 ng/ul
RT: 15.36 min Scan# 2121
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampled :
SLCS558

Tot Ion	Ratio	Resp	Lower	Upper
204	100	240447		
206	31.3	26.1	39.1	
141	70.7	55.8	83.8	




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#63
4-Nitroaniline
Concen: 35.743 ng/ul
RT: 15.39 min Scan# 2125
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	107255		
92	62.3	46.4	69.6	
108	107.9	83.0	124.4	









#66
 4,6-Dinitro-2-methylphenol
 Concen: 35.180 ng/ul
 RT: 15.44 min Scan# 2135
 Delta R.T. 0.00 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Instrument :
 BNA_M
 ClientSampled :
 SLCS558

Tot Ion	Ratio	Resp	Lower	Upper
198	100	80366		
182	4.6		3.5	5.3
77	41.8		29.8	44.6




Manual Integrations
 APPROVED


mohammad
 4/15/2021 11:45:38 AM



#67
 N-Nitrosodiphenylamine
 Concen: 37.500 ng/ul
 RT: 15.58 min Scan# 2158
 Delta R.T. 0.01 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Resp	Lower	Upper
169	100	442514		
168	68.5		55.1	82.7
167	38.0		30.5	45.7









#68
4-Bromophenyl-phenylether
Concen: 38.066 ng/ul
RT: 16.26 min Scan# 2273
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampled :
SLCS558

Tot Ion	Ratio	Lower	Upper
248	100		
250	99.3	78.6	117.8
141	95.9	78.2	117.4




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#69
Hexachlorobenzene
Concen: 37.700 ng/ul
RT: 16.37 min Scan# 2292
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
284	100		
142	50.6	42.6	64.0
249	32.2	26.4	39.6









#70
Atrazine
Concen: 35.142 ng/ul
RT: 16.53 min Scan# 2320
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampleId :
SLCS558

Tot Ion	Ratio	Lower	Upper
200	100		
173	25.8	20.6	31.0
215	43.1	36.4	54.6




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#71
Pentachlorophenol
Concen: 35.095 ng/ul
RT: 16.71 min Scan# 2350
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
266	100		
264	62.0	49.8	74.8
268	62.7	55.1	82.7









#72
Phenanthrene
Concen: 36.652 ng/ul
RT: 17.10 min Scan# 2417
Delta R.T. 0.01 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampleId :
SLCS558

Tot Ion	Ratio	Lower	Upper
178	100		
179	15.3	11.8	17.8
176	19.9	15.5	23.3




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#74
Anthracene
Concen: 37.158 ng/ul
RT: 17.19 min Scan# 2432
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
178	100		
179	15.3	12.2	18.2
176	19.0	15.2	22.8





#75

Pentachlorobenzene

Concen: 34.317 ng/uL

RT: 14.64 min Scan# 1998

Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M


ClientSampleId :

SLCS558




Tot Ion:	250	Resp:	172881
Ion Ratio	Lower	Upper	
250	100		
248	64.8	51.4	77.2
252	63.9	50.6	76.0

Manual Integrations
APPROVED

mohammad
4/15/2021 11:45:38 AM



Abundance Ion 250.00 (249.70 to 250.70): E
Ion 248.00 (247.70 to 248.70): E
Ion 252.00 (251.70 to 252.70): E



#76

Carbazole

Concen: 35.506 ng/uL


RT: 17.46 min Scan# 2478

Delta R.T. 0.01 min



Lab File: BM029463.D


Acq: 14 Apr 2021 10:41

Tgt Ion:	167	Resp:	704050
Ion Ratio	Lower	Upper	
167	100		
166	22.5	17.7	26.5
139	13.7	11.4	17.0



Abundance Ion 167.00 (166.70 to 167.70): E
Ion 166.00 (165.70 to 166.70): E
Ion 139.00 (138.70 to 139.70): E





#77

Di-n-butylphthalate

Concen: 38.639 ng/ul

RT: 18.03 min Scan# 2575

Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M





ClientSampled :

SLCS558

Tot Ion:	149	Resp:	923217
Ion Ratio	Lower	Upper	
149	100		
150	9.2	7.4	11.0
104	6.4	5.3	7.9

Manual Integrations
APPROVED

mohammad
4/15/2021 11:45:38 AM



#79

Fluoranthene

Concen: 37.541 ng/ul




RT: 19.12 min Scan# 2759


Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Tgt Ion:	202	Resp:	815755
Ion Ratio	Lower	Upper	
202	100		
101	12.4	10.5	15.7
100	9.9	8.6	12.8









#81
Pvrene
Concen: 37.076 ng/ul
RT: 19.48 min Scan# 2821
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampleId :
SLCS558

Tot Ion	Ratio	Resp Lower	Resp Upper
202	100		
101	14.5	12.0	18.0
100	12.1	9.8	14.8




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#82
Butylbenzylphthalate
Concen: 38.905 ng/ul
RT: 20.39 min Scan# 2975
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
91	84.0	70.1	105.1
206	17.5	13.4	20.2









#83
 3,3'-Dichlorobenzidine
 Concen: 34.048 ng/ul
 RT: 21.16 min Scan# 3106
 Delta R.T. 0.00 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Instrument :
 BNA_M
 ClientSampled :
 SLCS558

Tot Ion	Ratio	Lower	Upper
252	100		
254	64.6	52.3	78.5
126	14.4	11.2	16.8




Manual Integrations
 APPROVED


mohammad
 4/15/2021 11:45:38 AM



#84
 Benzo(a)anthracene
 Concen: 37.532 ng/ul
 RT: 21.22 min Scan# 3117
 Delta R.T. 0.00 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
228	100		
229	19.9	15.7	23.5
226	27.8	22.3	33.5





#85

Bis(2-ethylhexyl)phthalate

Concen: 39.741 ng/ul

RT: 21.16 min Scan# 3107

Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M


ClientSampleId :

SLCS558

Tot Ion:	149	Resp:	640127
Ion Ratio		Lower	Upper
149	100		
167	25.2	20.0	30.0
279	3.2	2.6	3.8

Manual Integrations
APPROVED

mohammad
4/15/2021 11:45:38 AM






Abundance

Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#86

Chrysene

Concen: 37.492 ng/ul


RT: 21.27 min Scan# 3126

Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Tgt Ion:	228	Resp:	750114
Ion Ratio		Lower	Upper
228	100		
226	31.2	24.2	36.4
229	19.3	15.4	23.2






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#88

Di-n-octyl phthalate

Concen: 36.555 ng/ul

RT: 22.03 min Scan# 3255

Delta R.T. 0.00 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M


ClientSampled :

SLCS558




Tgt Ion:149 Resp: 1080606

Manual Integrations
APPROVED

mohammad
4/15/2021 11:45:38 AM



Abundance Ion 149.00 (148.70 to 149.70): E



#89

Benzo(b)fluoranthene

Concen: 36.856 ng/ul

RT: 22.79 min Scan# 3383


Delta R.T. 0.01 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Tgt Ion:252 Resp: 777836



Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.7	26.5
125	11.4	9.4	14.2




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#90
Benzo(k)fluoranthene
Concen: 37.373 ng/ul
RT: 22.83 min Scan# 3390
Delta R.T. 0.00 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Instrument :
BNA_M
ClientSampleId :
SLCS558

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.3	17.3	25.9
125	12.3	9.9	14.9




Manual Integrations
APPROVED


mohammad
4/15/2021 11:45:38 AM



#92
Benzo(a)pyrene
Concen: 37.809 ng/ul
RT: 23.35 min Scan# 3479
Delta R.T. 0.01 min
Lab File: BM029463.D
Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.5	26.3
125	13.7	11.1	16.7









#93
 Indeno(1,2,3-cd)pyrene
 Concen: 38.699 ng/ul
 RT: 25.66 min Scan# 3872
 Delta R.T. 0.00 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Instrument :
 BNA_M
 ClientSampleId :
 SLCS558

Tot Ion	Ratio	Lower	Upper
276	100		
138	29.4	24.2	36.4
277	25.6	20.6	30.8




Manual Integrations
 APPROVED


mohammad
 4/15/2021 11:45:38 AM



#94
 Dibenzo(a,h)anthracene
 Concen: 39.131 ng/ul
 RT: 25.68 min Scan# 3875
 Delta R.T. 0.01 min
 Lab File: BM029463.D
 Acq: 14 Apr 2021 10:41

Tgt Ion	Ratio	Lower	Upper
278	100		
139	18.9	15.2	22.8
279	23.2	19.1	28.7





#95

Benzo(a,h,i)perylene

Concen: 38.747 ng/ul

RT: 26.34 min Scan# 3988

Delta R.T. 0.01 min

Lab File: BM029463.D

Acq: 14 Apr 2021 10:41

Instrument :

BNA_M

ClientSampleId :

SLCS558

Tot Ion:	276	Resp:	745155
Ion Ratio	Lower	Upper	
276	100		
138	25.0	21.0	31.6
277	23.5	18.7	28.1

Manual Integrations
APPROVED

mohammad
4/15/2021 11:45:38 AM

