


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM041421\
 Data File : BM029460.D
 Acq On : 14 Apr 2021 08:49
 Operator : CG/JU
 Sample : DFTPP31
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP031

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM041221.M
 Title : SVOA CALIBRATION
 Last Update : Wed Apr 14 10:06:38 2021



AutoFind: Scans 2485, 2486, 2487; Background Corrected with Scan 2480

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	60.1	136877	PASS
68	69	0.00	2	1.8	2536	PASS
69	198	0.00	100	60.5	137922	PASS
70	69	0.00	2	0.6	790	PASS
127	198	10	80	58.2	132589	PASS
197	198	0.00	2	0.6	1338	PASS
198	198	100	100	100.0	227818	PASS
199	198	5	9	6.8	15552	PASS
275	198	10	60	25.5	58008	PASS
365	198	1	100	3.9	8808	PASS
441	198	0.01	100	9.5	21651	PASS
442	198	50	100	60.3	137269	PASS
443	442	15	24	19.5	26788	PASS

Scan 2480 (17.474 min): BM029460.D
DFTPP31

m/z	Abundance
40.00	250.0
44.00	733.0
49.10	157.0
73.10	994.0
146.90	329.0
206.90	1138.0
207.90	152.0
221.00	247.0
280.90	288.0
282.00	154.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031

m/z	Abundance
35.90	197.0
37.00	950.0
38.00	1617.0
39.10	8900.0
40.10	799.0
41.00	341.0
44.00	1071.0
45.00	287.0
49.00	1323.0
50.10	31472.0
51.10	111368.0
52.10	6219.0
53.00	188.0
54.90	622.0
56.00	3542.0
57.00	6672.0
57.80	285.0
61.00	1501.0
62.00	1494.0
63.00	4198.0
64.00	521.0
65.10	2428.0
66.00	182.0
66.90	287.0
68.10	1947.0
69.00	109960.0
70.00	694.0
73.00	1769.0
74.00	12282.0
75.00	17968.0
76.10	5828.0
77.10	125800.0
78.10	9013.0
79.00	8149.0
80.00	6663.0
81.00	8895.0
82.00	2156.0
83.00	1749.0
84.00	424.0
85.00	1752.0
86.00	2981.0
87.00	1011.0
88.00	487.0
88.90	161.0
91.00	1989.0
92.00	2205.0
93.00	15582.0
94.00	1071.0
94.90	293.0
96.00	577.0
97.10	367.0
98.00	12906.0
99.00	9308.0
99.90	758.0
101.00	4520.0
101.90	332.0
103.00	1443.0
104.00	2372.0
105.00	2486.0
106.00	956.0
107.00	32760.0
108.00	4639.0
109.10	1073.0
110.00	49856.0
111.00	8908.0
112.00	1334.0
112.90	446.0
116.00	2047.0
117.00	36928.0
118.00	2885.0
118.90	337.0
120.10	327.0
121.00	169.0
122.00	2145.0
123.00	3732.0
124.00	1802.0
125.00	1312.0
127.00	93832.0
128.00	7680.0
129.00	39872.0
130.00	3433.0
130.90	732.0
132.00	367.0
132.90	221.0
134.00	1158.0
135.00	3247.0
135.90	1274.0
137.00	2076.0
137.90	360.0
138.90	266.0
139.90	614.0
141.00	5731.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031

142.10	1664.0
143.00	985.0
143.80	304.0
144.90	291.0
145.90	1092.0
147.00	2586.0
148.00	8833.0
149.00	2257.0
149.90	447.0
150.90	690.0
152.00	343.0
153.00	1867.0
154.10	1698.0
155.00	2584.0
156.10	4296.0
157.10	902.0
157.90	925.0
159.00	727.0
160.00	1217.0
161.00	2061.0
162.00	745.0
163.00	415.0
164.00	190.0
165.00	2224.0
166.00	1664.0
167.00	12868.0
168.00	8958.0
169.00	1276.0
170.00	490.0
171.00	423.0
171.90	940.0
173.00	1238.0
174.00	2238.0
175.00	4004.0
176.00	1023.0
177.00	2063.0
178.00	741.0
179.00	8546.0
180.00	5633.0
181.00	2151.0
182.00	555.0
183.00	151.0
184.00	621.0
185.00	3276.0
186.00	25568.0
187.10	7532.0
188.00	1036.0
189.00	1980.0
189.90	546.0
191.00	971.0
192.00	2188.0
193.10	2952.0
194.00	554.0
194.90	192.0
196.00	4330.0
196.70	833.0
198.00	148672.0
199.00	10930.0
199.90	943.0
201.20	397.0
203.10	1390.0
204.00	7374.0
205.00	12120.0
206.10	45880.0
207.00	7403.0
208.00	1635.0
208.90	886.0
210.10	861.0
211.10	1723.0
212.90	164.0
214.90	601.0
216.10	1159.0
217.00	13738.0
218.10	1816.0
219.00	199.0
220.00	180.0
221.00	4303.0
221.90	1670.0
223.00	2741.0
224.10	24328.0
225.00	6480.0
226.10	823.0
227.00	12685.0
228.00	2113.0
229.00	2274.0
230.00	282.0
231.00	689.0
232.10	173.0
233.00	171.0
234.10	832.0
235.00	880.0
236.00	522.0
237.00	1295.0
239.00	496.0
239.90	440.0
241.00	716.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031

242.00	1255.0
243.10	1554.0
244.10	17664.0
245.00	2113.0
246.00	4324.0
247.00	878.0
248.00	151.0
249.00	585.0
250.90	252.0
252.00	171.0
253.00	489.0
254.10	721.0
255.00	87232.0
256.00	13416.0
257.00	1212.0
258.00	6900.0
259.00	1311.0
260.10	152.0
261.00	210.0
263.80	209.0
265.00	2975.0
265.80	452.0
267.00	225.0
269.90	161.0
270.80	189.0
272.10	244.0
273.00	3071.0
274.10	6982.0
275.00	34424.0
276.00	4819.0
277.00	4122.0
278.00	978.0
278.90	158.0
281.20	351.0
283.10	380.0
283.80	255.0
285.00	613.0
289.00	203.0
292.20	184.0
292.90	851.0
293.90	261.0
295.00	264.0
296.00	13585.0
297.00	1834.0
301.00	205.0
303.00	1170.0
303.90	317.0
309.70	150.0
314.00	560.0
314.90	1444.0
316.00	789.0
321.10	242.0
322.00	225.0
323.00	2387.0
323.90	635.0
326.90	669.0
328.00	442.0
332.00	211.0
333.00	245.0
334.00	1780.0
334.90	601.0
340.90	285.0
345.90	966.0
346.90	162.0
352.10	944.0
352.90	580.0
354.00	904.0
355.10	323.0
365.00	5184.0
366.00	717.0
371.00	196.0
372.00	1343.0
372.90	355.0
382.90	346.0
389.90	208.0
401.90	561.0
403.00	755.0
403.80	159.0
404.90	201.0
421.10	459.0
421.90	485.0
423.00	3662.0
423.90	1428.0
424.80	194.0
441.00	10403.0
442.00	67552.0
443.10	13390.0
444.00	1738.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031

m/z	Abundance
37.00	981.0
38.10	2745.0
39.10	13457.0
40.00	1137.0
41.00	474.0
43.10	185.0
44.00	804.0
45.00	260.0
49.00	1586.0
50.10	46960.0
51.10	161728.0
52.10	9173.0
53.00	383.0
55.00	983.0
56.00	5029.0
57.10	10067.0
58.00	400.0
59.00	190.0
61.00	1848.0
62.00	2692.0
63.00	6182.0
64.10	717.0
65.00	3233.0
66.00	230.0
66.90	392.0
68.10	3113.0
69.00	164608.0
70.00	755.0
73.10	2184.0
74.10	19536.0
75.00	28088.0
76.10	10157.0
77.10	187648.0
78.10	13367.0
79.00	13206.0
80.00	9608.0
81.00	13780.0
82.00	3478.0
83.00	2776.0
83.90	382.0
85.00	2483.0
86.00	4022.0
87.00	1966.0
88.00	660.0
89.00	331.0
90.10	154.0
91.00	2648.0
92.00	3445.0
93.00	25000.0
94.00	1685.0
95.00	398.0
96.10	941.0
97.00	498.0
98.00	18760.0
99.00	13227.0
100.00	1245.0
101.00	6337.0
102.00	342.0
103.00	2235.0
104.00	4791.0
105.00	4679.0
106.10	1419.0
107.00	52104.0
108.00	7734.0
109.10	1840.0
110.00	80584.0
111.00	13920.0
112.00	1628.0
113.00	587.0
115.10	160.0
116.00	3027.0
117.00	59552.0
118.00	4228.0
118.90	513.0
120.00	763.0
120.90	206.0
122.00	4005.0
123.00	5956.0
123.90	2664.0
125.00	2041.0
127.00	159936.0
128.00	11845.0
129.00	66096.0
130.00	5811.0
131.00	985.0
132.00	498.0
133.00	502.0
134.00	1985.0
135.00	5532.0
136.00	2530.0
137.10	3336.0
138.00	555.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031

139.00	518.0
140.10	930.0
141.00	9519.0
142.00	2801.0
143.00	2145.0
144.00	766.0
144.90	430.0
145.90	1886.0
147.00	4459.0
148.00	13722.0
149.00	2672.0
150.00	537.0
151.10	1486.0
151.70	522.0
153.00	2959.0
154.00	2059.0
155.00	4838.0
156.00	7708.0
157.00	1290.0
158.00	2028.0
158.90	1169.0
160.00	3575.0
161.00	4017.0
162.00	1205.0
163.00	351.0
164.10	560.0
165.00	3251.0
166.00	2905.0
167.00	22104.0
168.00	12696.0
169.00	2103.0
170.00	632.0
170.90	841.0
172.00	1888.0
173.00	2482.0
174.00	3854.0
175.10	6904.0
176.10	1328.0
177.00	3524.0
178.00	1277.0
179.00	15178.0
180.00	10642.0
181.00	4192.0
182.00	627.0
183.10	284.0
184.00	1039.0
185.00	6483.0
186.10	45648.0
187.00	13414.0
188.00	1772.0
189.00	3365.0
190.00	536.0
191.00	1566.0
192.00	4267.0
193.00	5333.0
194.00	1047.0
195.00	624.0
196.00	8699.0
197.10	1949.0
198.00	273792.0
199.00	18544.0
200.00	1712.0
201.50	966.0
203.00	2471.0
204.00	13249.0
205.00	21496.0
206.10	80800.0
207.10	11770.0
208.00	3444.0
208.90	1152.0
210.10	1525.0
211.10	3339.0
212.10	468.0
213.00	452.0
215.00	1040.0
216.00	2387.0
217.00	26312.0
218.00	3569.0
219.00	414.0
220.10	542.0
221.10	9034.0
222.00	2673.0
223.00	5574.0
224.10	46048.0
225.10	11855.0
226.00	1449.0
227.00	23408.0
228.00	3933.0
229.00	4445.0
230.00	620.0
231.00	1902.0
232.00	208.0
233.00	348.0
234.00	1506.0
235.00	1781.0
235.90	1187.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031

237.10	2125.0
238.00	407.0
239.00	1053.0
239.90	675.0
241.00	1818.0
242.00	2668.0
243.10	2760.0
244.10	32576.0
245.10	4437.0
246.00	8414.0
247.00	1554.0
247.80	396.0
249.00	1306.0
250.00	246.0
250.80	359.0
251.90	358.0
253.00	849.0
254.10	1616.0
255.00	167552.0
256.00	24560.0
257.10	2211.0
258.00	12321.0
259.00	2054.0
259.90	423.0
261.00	285.0
263.00	159.0
264.10	381.0
265.00	5296.0
266.10	973.0
267.00	163.0
270.00	261.0
270.90	414.0
272.10	547.0
273.00	5459.0
274.00	12952.0
275.00	66984.0
276.00	9250.0
277.00	7050.0
278.00	1233.0
279.00	267.0
281.10	306.0
281.90	260.0
282.90	796.0
284.10	408.0
285.00	1099.0
286.10	217.0
288.80	254.0
289.90	245.0
292.00	322.0
293.00	1458.0
294.00	464.0
295.10	416.0
296.00	25248.0
297.00	3711.0
297.90	183.0
301.10	330.0
301.90	463.0
303.00	2445.0
304.10	704.0
309.00	217.0
309.90	228.0
312.90	339.0
314.00	1086.0
314.90	3085.0
316.00	1627.0
317.00	283.0
321.00	798.0
322.10	277.0
323.10	6554.0
324.00	1260.0
324.90	155.0
326.00	174.0
326.90	1422.0
328.00	770.0
332.10	357.0
333.00	567.0
334.00	4537.0
335.00	937.0
340.00	156.0
341.00	695.0
346.00	1498.0
347.00	350.0
352.00	1881.0
353.00	1345.0
354.00	1790.0
355.00	240.0
359.00	160.0
365.00	9847.0
366.00	1334.0
370.10	195.0
370.80	449.0
372.00	2600.0
373.00	815.0
383.00	708.0
383.90	442.0
389.90	393.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031

391.00	323.0
392.00	177.0
402.10	1099.0
403.00	1418.0
404.00	368.0
404.90	151.0
421.00	1275.0
422.00	1422.0
423.00	7877.0
424.00	2398.0
425.10	284.0
440.10	159.0
441.00	22936.0
442.00	150656.0
443.00	29408.0
444.10	3332.0
444.90	163.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031

m/z	Abundance
36.00	225.0
37.00	821.0
38.00	1977.0
39.10	11147.0
40.00	989.0
41.10	387.0
43.00	224.0
43.90	687.0
45.00	384.0
46.80	217.0
48.00	185.0
49.10	1445.0
50.10	39152.0
51.10	137536.0
52.10	6265.0
53.10	299.0
55.00	889.0
56.00	3692.0
57.00	8501.0
58.00	345.0
61.00	1686.0
62.10	2239.0
63.10	5316.0
64.10	567.0
65.00	2243.0
66.10	178.0
66.80	171.0
68.00	2549.0
69.00	139200.0
70.00	923.0
73.00	1949.0
74.10	15075.0
75.00	23416.0
76.10	8591.0
77.10	160320.0
78.10	10622.0
79.00	10846.0
80.00	7986.0
81.00	12189.0
82.00	2511.0
83.00	2591.0
84.00	287.0
85.00	2316.0
86.00	3884.0
87.00	1849.0
88.00	470.0
89.00	334.0
91.00	2284.0
92.00	2911.0
93.00	20752.0
94.00	1419.0
94.80	513.0
96.00	778.0
97.00	527.0
98.00	16800.0
99.00	11260.0
99.90	932.0
101.00	5792.0
102.00	268.0
103.00	2144.0
104.00	3916.0
105.00	3895.0
106.00	1466.0
107.00	45976.0
108.00	7271.0
109.00	1371.0
110.00	70960.0
111.00	11694.0
112.00	1310.0
113.00	525.0
113.90	221.0
114.80	270.0
116.00	2740.0
117.00	54496.0
118.00	3641.0
119.00	299.0
120.00	699.0
121.00	263.0
122.10	2836.0
123.00	5239.0
124.00	2447.0
125.00	2102.0
127.00	144000.0
128.00	11434.0
129.00	62720.0
130.00	5426.0
131.00	1212.0
131.90	418.0
132.90	298.0
134.00	1811.0
135.00	4919.0
136.00	2252.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031

137.00	2655.0
138.00	739.0
138.90	348.0
139.90	739.0
141.00	8671.0
142.00	2461.0
143.00	1963.0
144.00	493.0
145.00	414.0
146.00	1692.0
147.00	4894.0
148.00	13058.0
149.00	2019.0
149.90	356.0
151.10	809.0
152.00	697.0
153.00	2740.0
154.00	2139.0
155.00	4527.0
156.00	7473.0
156.90	1096.0
158.00	1849.0
159.00	1227.0
160.00	2983.0
161.00	4002.0
161.90	984.0
163.00	460.0
164.00	387.0
165.00	3269.0
166.00	2886.0
167.00	21056.0
168.00	10788.0
169.00	1562.0
170.00	530.0
170.90	605.0
172.00	1645.0
173.00	2214.0
174.00	3747.0
175.00	7473.0
176.00	1499.0
177.00	3360.0
178.00	1148.0
179.00	12858.0
180.00	9272.0
181.00	3958.0
182.00	587.0
183.00	471.0
184.00	1152.0
185.00	6131.0
186.00	42504.0
187.00	12480.0
188.00	1362.0
188.90	3333.0
190.00	717.0
191.00	1612.0
192.00	4891.0
193.00	4696.0
194.00	1155.0
194.90	401.0
196.00	7668.0
197.10	2065.0
198.00	260992.0
199.00	17184.0
200.00	1590.0
201.20	927.0
201.70	763.0
203.00	2611.0
204.00	12159.0
205.00	20496.0
206.10	79344.0
207.10	11180.0
208.00	3443.0
208.90	1316.0
210.00	1453.0
211.10	3451.0
211.90	534.0
213.00	285.0
215.00	1019.0
216.00	2012.0
217.00	25528.0
218.00	3417.0
218.90	275.0
220.10	505.0
221.10	8217.0
222.00	3324.0
223.00	5327.0
224.10	44880.0
225.00	12506.0
226.10	1238.0
227.00	23544.0
228.00	3249.0
229.00	4283.0
230.00	759.0
231.00	1653.0
232.00	277.0
233.00	369.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031

234.00	1837.0
235.00	1755.0
236.00	1390.0
237.00	2024.0
238.10	292.0
239.00	1034.0
240.00	646.0
241.00	1350.0
242.00	2730.0
243.10	2939.0
244.10	33264.0
245.00	4197.0
246.00	7670.0
247.00	1534.0
247.90	319.0
249.00	1072.0
250.00	191.0
250.90	407.0
251.80	275.0
253.00	736.0
254.10	1993.0
255.00	175040.0
256.00	25520.0
257.00	2197.0
258.00	13000.0
258.90	2032.0
259.90	332.0
261.00	248.0
264.00	275.0
265.00	5306.0
266.00	678.0
269.90	240.0
270.90	405.0
272.00	519.0
273.00	5241.0
274.00	14126.0
275.00	72616.0
276.00	9394.0
277.00	7265.0
278.00	1465.0
278.90	360.0
281.10	330.0
281.90	328.0
283.10	798.0
284.00	566.0
285.00	1363.0
285.90	276.0
289.00	214.0
290.10	310.0
291.90	368.0
293.00	1488.0
294.00	433.0
295.10	505.0
296.00	27200.0
297.00	3764.0
298.10	170.0
300.80	332.0
302.00	382.0
303.00	3081.0
304.00	664.0
305.00	177.0
308.00	399.0
309.00	239.0
310.00	397.0
313.00	245.0
314.00	1309.0
315.00	3452.0
316.10	1439.0
316.90	213.0
319.80	191.0
321.00	675.0
321.90	490.0
323.00	6774.0
324.00	1084.0
326.90	1349.0
328.00	703.0
329.00	238.0
332.00	536.0
333.00	674.0
334.00	4422.0
335.10	1007.0
338.80	181.0
340.90	1149.0
342.10	205.0
345.90	1869.0
346.90	378.0
351.00	208.0
352.00	1928.0
353.00	1335.0
354.10	1791.0
355.00	283.0
358.90	164.0
365.00	11394.0
365.90	1993.0
370.10	193.0
371.00	555.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031

372.00	3618.0
372.90	684.0
383.00	912.0
384.00	235.0
390.00	437.0
391.00	266.0
392.00	231.0
401.00	178.0
402.00	1503.0
403.00	1715.0
403.90	696.0
420.90	1736.0
422.00	1580.0
423.00	10581.0
424.10	3076.0
424.90	414.0
441.10	31616.0
442.00	193600.0
443.10	37568.0
444.00	3372.0
445.10	159.0

Instrument :
BNA_M
ClientSampleId :
DFTPP031