

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041422\
 Data File : BM034672.D
 Acq On : 14 Apr 2022 17:21
 Operator : CG/JU
 Sample : N2266-12
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 YAZ04

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/15/2022
 Supervised By :Yogesh Patel 04/23/2022

Quant Time: Apr 15 00:47:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 14 23:42:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.866	152	114608	20.000	ng/u1	# 0.00
20) Naphthalene-d8	10.678	136	471002	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.518	164	273387	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.254	188	531710	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.436	240	488913	20.000	ng/u1	# 0.00
88) Perylene-d12	23.812	264	556971	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.249	96	17854	5.555	ng/uL	0.00
4) Pyridine-d5	3.678	84	57420	7.403	ng/u1	0.00
7) Phenol-d5	7.031	99	63706	6.678	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.195	67	211444	33.265	ng/u1	0.00
11) 2-Chlorophenol-d4	7.395	132	189030	25.635	ng/u1	0.00
15) 4-Methylphenol-d8	8.584	113	126099	16.923	ng/u1	0.00
21) Nitrobenzene-d5	9.031	128	111802	32.298	ng/u1	0.00
24) 2-Nitrophenol-d4	9.760	143	110487	30.657	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.301	165	177028	26.086	ng/u1	0.00
31) 4-Chloroaniline-d4	10.813	131	233871m	24.980	ng/u1	0.00
46) Dimethylphthalate-d6	13.924	166	681308	34.489	ng/u1	0.00
49) Acenaphthylene-d8	14.207	160	801193	31.555	ng/u1	0.00
54) 4-Nitrophenol-d4	14.754	143	6565m	2.180	ng/u1	0.04
60) Fluorene-d10	15.507	176	606851	35.549	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.624	200	92011	28.346	ng/u1	0.00
73) Anthracene-d10	17.354	188	924928	36.486	ng/u1	0.00
81) Pyrene-d10	19.648	212	1134782	42.526	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.659	264	1068969	38.154	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041422\
Data File : BM034672.D
Acq On : 14 Apr 2022 17:21
Operator : CG/JU
Sample : N2266-12
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
YAZ04

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/23/2022

Quant Time: Apr 15 00:47:01 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Apr 14 23:42:00 2022
Response via : Initial Calibration

