

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041422\
 Data File : BM034686.D
 Acq On : 15 Apr 2022 02:27
 Operator : CG/JU
 Sample : N2316-09
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ETNQ8

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/15/2022
 Supervised By :Yogesh Patel 04/23/2022

Quant Time: Apr 15 03:51:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 14 23:42:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.866	152	139930	20.000	ng/u1	0.00
20) Naphthalene-d8	10.672	136	568637	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.513	164	335836	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.254	188	652261	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.430	240	608912	20.000	ng/u1	# 0.00
88) Perylene-d12	23.801	264	669799	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.243	96	10509	2.678	ng/uL	0.00
4) Pyridine-d5	3.678	84	54625	5.768	ng/u1	0.00
7) Phenol-d5	7.025	99	147790	12.690	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.190	67	116395	14.998	ng/u1	0.00
11) 2-Chlorophenol-d4	7.390	132	130173	14.458	ng/u1	0.00
15) 4-Methylphenol-d8	8.578	113	83832	9.215	ng/u1	0.00
21) Nitrobenzene-d5	9.025	128	62228m	14.890	ng/u1	0.00
24) 2-Nitrophenol-d4	9.754	143	63246m	14.536	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.296	165	111946m	13.664	ng/u1	0.00
31) 4-Chloroaniline-d4	10.813	131	115588	10.226	ng/u1	0.00
46) Dimethylphthalate-d6	13.919	166	434290	17.897	ng/u1	0.00
49) Acenaphthylene-d8	14.201	160	489371	15.690	ng/u1	0.00
54) 4-Nitrophenol-d4	14.725	143	25144m	6.795	ng/u1	0.01
60) Fluorene-d10	15.501	176	409911	19.547	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.625	200	47919	12.034	ng/u1	0.00
73) Anthracene-d10	17.348	188	632727	20.346	ng/u1	0.00
81) Pyrene-d10	19.642	212	819555	24.660	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.648	264	741308	22.002	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041422\
Data File : BM034686.D
Acq On : 15 Apr 2022 02:27
Operator : CG/JU
Sample : N2316-09
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ETNQ8

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/23/2022

Quant Time: Apr 15 03:51:41 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040922.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Apr 14 23:42:00 2022
Response via : Initial Calibration

