


Data Path : Z:\HPCHEM1\BNA_M\Data\BM041515\SOM01.2\
 Data File : BM001007.D
 Acq On : 15 Apr 2015 12:50
 Operator : TP/IZ
 Sample : DFTPP46
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP46

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM01.2-EPA-BM041515.M
 Title : SVOA CALIBRATION
 Last Update : Thu Apr 16 05:47:45 2015



AutoFind: Scans 2508, 2509, 2510; Background Corrected with Scan 2502

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.9	155567	PASS
68	69	0.00	2	1.6	2237	PASS
69	198	0.00	100	41.6	143970	PASS
70	69	0.00	2	0.5	717	PASS
127	198	10	80	48.5	167914	PASS
197	198	0.00	2	0.5	1811	PASS
198	198	100	100	100.0	346453	PASS
199	198	5	9	6.8	23642	PASS
275	198	10	60	22.7	78698	PASS
365	198	1	100	2.4	8211	PASS
441	443	0.01	100	77.3	42349	PASS
442	198	50	100	80.8	279850	PASS
443	442	15	24	19.6	54786	PASS

m/z	Abundance
37.00	770.0
38.10	1266.0
39.10	10482.0
40.00	6145.0
41.10	1330.0
42.20	366.0
43.10	800.0
44.00	7302.0
44.90	622.0
49.10	1239.0
50.10	34736.0
51.00	140096.0
52.10	6726.0
53.00	514.0
54.90	994.0
56.00	3788.0
57.00	11052.0
58.00	627.0
61.00	1719.0
62.10	1605.0
63.00	5084.0
64.20	676.0
65.00	2888.0
66.90	357.0
68.00	1713.0
69.00	131264.0
70.00	416.0
73.10	1079.0
74.00	12552.0
75.00	19832.0
76.10	6775.0
77.10	143488.0
78.00	9663.0
79.00	10699.0
79.90	7671.0
81.00	9917.0
82.00	3361.0
83.00	2522.0
84.00	521.0
84.90	1855.0
86.00	2674.0
87.10	1360.0
88.00	449.0
88.90	313.0
91.00	1671.0
92.00	3236.0
93.00	14941.0
94.00	1312.0
95.00	522.0
96.00	792.0
98.00	11272.0
99.00	10726.0
100.10	1108.0
101.00	6019.0
102.00	477.0
103.00	2071.0
104.00	3053.0
105.10	4328.0
105.90	1010.0
107.00	40144.0
108.00	6045.0
109.10	979.0
110.00	71944.0
111.00	12828.0
112.10	1445.0
113.00	705.0
115.10	315.0
116.10	2577.0
117.00	32280.0
118.00	2505.0
119.00	429.0
119.80	406.0
121.00	385.0
122.00	2754.0
123.00	6530.0
124.00	2358.0
125.00	1981.0
126.10	862.0
127.00	138880.0
128.00	10313.0
129.00	56400.0
130.00	4835.0
131.10	843.0
131.90	422.0
134.00	1410.0
135.00	5055.0
136.00	1735.0
137.00	1823.0
140.00	642.0
141.00	6804.0
142.10	1813.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

142.90	1438.0
144.10	536.0
144.80	343.0
145.90	1247.0
147.00	3286.0
148.00	7863.0
149.10	1598.0
150.00	593.0
151.10	850.0
151.90	413.0
153.00	2051.0
154.00	1503.0
155.10	4343.0
156.00	5734.0
157.00	910.0
158.00	1670.0
159.00	1051.0
160.00	2300.0
161.00	3380.0
162.00	1194.0
162.90	314.0
163.10	307.0
164.90	2637.0
166.00	1573.0
167.00	11622.0
168.00	8233.0
169.00	1612.0
170.20	385.0
170.90	417.0
171.90	1268.0
172.90	1232.0
174.00	3093.0
175.10	5096.0
176.00	1305.0
177.00	2489.0
177.90	559.0
179.00	10790.0
180.00	7097.0
181.00	3473.0
182.00	614.0
184.00	486.0
185.00	5045.0
186.10	38024.0
187.00	10200.0
187.90	1191.0
189.00	2206.0
190.00	422.0
191.10	1373.0
192.00	3447.0
193.00	3313.0
193.80	558.0
195.00	472.0
196.00	7507.0
197.00	1779.0
198.00	265280.0
199.00	19144.0
200.00	1354.0
200.90	826.0
201.50	870.0
203.00	1361.0
204.00	8368.0
205.00	17056.0
206.10	65568.0
207.00	11270.0
208.00	2267.0
209.00	911.0
210.10	1267.0
211.10	2025.0
211.80	462.0
215.00	848.0
216.00	1452.0
217.00	18088.0
218.00	2473.0
220.00	374.0
221.00	11040.0
221.70	3391.0
223.00	4131.0
224.00	35912.0
225.10	9127.0
226.10	1188.0
227.00	13217.0
227.90	2115.0
229.00	2449.0
230.10	431.0
231.00	1369.0
233.90	1143.0
235.20	912.0
235.80	512.0
236.10	499.0
237.00	1990.0
237.80	305.0
238.90	455.0
239.90	757.0
241.00	949.0
242.00	2004.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

243.10	1301.0
244.10	26368.0
245.00	3189.0
246.00	6129.0
247.00	1094.0
248.90	827.0
253.10	580.0
253.90	705.0
255.00	131200.0
256.00	18472.0
257.00	1990.0
258.00	6732.0
258.90	975.0
264.00	374.0
265.00	2843.0
265.90	657.0
270.90	383.0
271.90	353.0
273.00	4633.0
274.00	9571.0
275.10	56632.0
276.00	7760.0
277.00	4357.0
277.90	1042.0
283.20	386.0
285.00	520.0
293.10	1314.0
295.00	636.0
296.00	13659.0
297.00	2201.0
303.00	1826.0
303.90	347.0
313.90	605.0
314.90	1487.0
316.10	1053.0
321.10	333.0
323.10	3779.0
324.00	949.0
326.90	968.0
328.10	438.0
333.00	337.0
334.00	3257.0
335.10	1249.0
341.00	494.0
346.00	723.0
351.90	989.0
353.10	1008.0
354.00	1057.0
355.00	340.0
365.00	4997.0
366.00	580.0
369.90	308.0
370.90	431.0
372.10	2754.0
372.80	551.0
383.00	639.0
383.70	312.0
402.00	1034.0
403.00	1100.0
403.90	417.0
420.90	949.0
422.00	929.0
423.00	7206.0
424.00	3693.0
425.10	521.0
441.00	25072.0
442.00	156992.0
443.00	34824.0
444.10	2958.0

Instrument :

BNA_M

ClientSampleId :

DFTPP46

m/z	Abundance
35.90	367.0
36.90	779.0
38.10	2082.0
39.10	14106.0
40.00	5997.0
41.10	1585.0
42.00	306.0
43.10	568.0
44.00	6987.0
45.00	847.0
49.00	1022.0
50.00	47944.0
51.10	193344.0
52.10	9608.0
53.00	655.0
55.00	1289.0
56.00	5070.0
57.00	15535.0
57.90	838.0
61.10	2234.0
62.00	2779.0
63.00	7228.0
64.00	847.0
65.10	3430.0
66.00	329.0
68.00	2519.0
69.00	174720.0
70.00	819.0
73.10	1940.0
74.00	18312.0
75.00	24968.0
76.10	10396.0
77.10	198848.0
78.00	13958.0
79.00	11988.0
80.00	9910.0
81.00	14199.0
82.10	3245.0
83.00	3288.0
84.00	677.0
85.00	2488.0
86.00	4152.0
87.00	1272.0
88.10	658.0
91.00	3305.0
92.10	3859.0
93.00	22856.0
94.00	1402.0
95.10	316.0
96.10	1293.0
96.90	445.0
98.00	15248.0
99.00	13331.0
100.10	1014.0
101.00	7076.0
101.90	375.0
103.10	2488.0
104.10	4214.0
105.00	5156.0
106.00	1783.0
107.00	57184.0
108.10	8359.0
109.00	1549.0
110.00	106624.0
111.10	15540.0
112.10	2179.0
113.10	808.0
116.00	2836.0
117.00	44760.0
118.00	2987.0
119.00	487.0
119.90	687.0
122.00	4108.0
123.00	8645.0
123.90	2808.0
125.00	2369.0
125.90	556.0
127.00	204288.0
128.00	14849.0
129.00	80104.0
130.00	6556.0
130.90	1766.0
131.80	826.0
133.30	408.0
133.90	2375.0
135.10	6064.0
136.00	2983.0
137.00	3791.0
137.90	1080.0
139.00	693.0
140.00	940.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

141.00	10773.0
142.00	3127.0
143.00	2253.0
144.00	711.0
145.00	712.0
146.00	2351.0
147.10	4722.0
148.00	11004.0
149.00	2891.0
149.90	508.0
151.00	1390.0
151.80	828.0
153.00	2715.0
154.00	2557.0
155.00	6095.0
156.10	8599.0
157.00	1323.0
158.10	2351.0
159.10	1306.0
160.00	2832.0
161.00	3834.0
162.00	1633.0
163.10	524.0
164.10	606.0
164.90	3798.0
166.00	3210.0
167.00	18480.0
168.00	11171.0
169.10	2312.0
169.90	755.0
170.90	950.0
172.00	2019.0
173.00	2149.0
174.00	4802.0
175.00	7465.0
176.00	1946.0
177.00	3067.0
177.90	1350.0
179.00	15032.0
180.00	11064.0
181.10	4501.0
182.00	678.0
182.90	820.0
183.90	1209.0
185.00	7738.0
186.10	53424.0
187.00	16624.0
188.10	1867.0
189.00	3363.0
190.10	633.0
190.90	1576.0
192.00	5028.0
193.00	5289.0
193.90	898.0
194.90	680.0
196.00	11037.0
196.80	1897.0
198.00	412032.0
199.00	27408.0
200.00	1868.0
201.60	2071.0
203.00	2229.0
204.00	13419.0
205.00	24592.0
206.10	110736.0
207.10	14650.0
208.00	4049.0
209.10	880.0
210.40	2342.0
211.00	4684.0
212.00	773.0
215.10	1091.0
216.00	2008.0
217.00	27840.0
218.10	3664.0
218.90	339.0
221.00	18760.0
221.90	3694.0
223.00	5914.0
224.10	56928.0
225.00	13797.0
226.00	2285.0
227.00	21432.0
228.10	3453.0
229.10	4622.0
230.00	686.0
231.00	1787.0
232.90	587.0
234.00	2083.0
234.90	1518.0
236.00	1162.0
237.00	2647.0
238.00	390.0
239.00	830.0
240.00	1084.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

240.90	1790.0
242.10	2374.0
243.00	2573.0
244.10	43248.0
245.00	5057.0
246.00	7559.0
247.00	2026.0
248.00	493.0
249.00	2096.0
250.90	392.0
252.20	372.0
253.00	1250.0
254.00	1638.0
255.00	221568.0
256.10	31792.0
257.10	2872.0
258.00	12082.0
259.00	1601.0
261.20	358.0
265.00	5765.0
266.00	710.0
270.00	350.0
271.00	344.0
271.90	340.0
273.00	6570.0
274.00	15250.0
275.00	91688.0
276.00	12318.0
277.00	6913.0
278.00	1528.0
278.90	305.0
280.90	470.0
282.10	300.0
282.90	989.0
283.90	752.0
285.00	1228.0
285.90	459.0
292.00	311.0
293.00	1761.0
294.00	515.0
295.00	507.0
296.00	24264.0
297.00	3284.0
302.00	304.0
303.00	2937.0
304.00	973.0
307.90	312.0
309.00	361.0
312.90	386.0
314.10	1464.0
315.00	2900.0
316.10	1615.0
317.10	474.0
320.90	916.0
322.00	410.0
323.10	8988.0
324.10	2037.0
326.90	1471.0
328.10	427.0
332.10	370.0
333.00	605.0
334.00	4559.0
335.00	1044.0
335.90	333.0
341.10	1102.0
342.00	569.0
345.90	2321.0
346.90	607.0
352.10	2307.0
352.90	1648.0
354.00	2003.0
354.90	579.0
359.00	389.0
365.00	9594.0
366.00	1608.0
369.90	485.0
371.00	770.0
372.10	3929.0
373.00	884.0
377.00	336.0
383.10	1430.0
384.00	324.0
391.80	328.0
402.00	2243.0
402.90	2297.0
404.10	979.0
421.00	1859.0
422.00	2520.0
423.00	17488.0
424.10	5391.0
424.80	583.0
425.10	474.0
441.00	47992.0
442.00	322816.0
443.00	62792.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

444.10

5375.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

m/z	Abundance
36.00	354.0
37.00	711.0
38.00	1584.0
39.10	10400.0
40.00	6377.0
41.00	1097.0
42.90	511.0
44.00	7619.0
45.10	1031.0
49.10	1162.0
50.10	32888.0
51.10	134272.0
52.10	7068.0
54.90	1053.0
56.00	3662.0
57.00	11199.0
58.00	561.0
61.00	1840.0
62.10	1774.0
63.10	5363.0
64.10	926.0
65.00	2836.0
67.10	418.0
68.00	2480.0
69.00	125928.0
70.00	917.0
73.10	1303.0
74.00	12659.0
75.00	20696.0
76.00	6844.0
77.10	145536.0
78.10	9895.0
79.00	9286.0
80.10	6630.0
81.00	9298.0
82.00	2275.0
83.00	2404.0
84.00	573.0
85.00	2228.0
86.00	3232.0
87.00	1155.0
88.00	532.0
88.90	360.0
91.10	2981.0
92.00	3201.0
93.00	16270.0
94.00	1114.0
96.10	641.0
97.00	317.0
98.00	11675.0
99.00	9881.0
100.00	737.0
101.00	6056.0
101.80	367.0
103.00	1647.0
104.00	3094.0
105.00	3935.0
106.10	1152.0
107.00	43808.0
108.00	7393.0
109.00	1122.0
110.00	79792.0
111.00	11267.0
111.90	1680.0
113.00	552.0
115.90	3131.0
117.00	35664.0
117.90	2492.0
119.90	740.0
121.10	380.0
122.00	3413.0
123.00	4904.0
123.90	2593.0
125.00	2504.0
125.80	632.0
127.00	160576.0
128.00	11320.0
129.00	61704.0
130.00	5691.0
131.00	820.0
131.90	581.0
132.90	339.0
134.00	1582.0
135.00	4833.0
136.00	2041.0
137.10	2244.0
138.00	779.0
139.00	405.0
140.00	611.0
141.00	8686.0
141.90	3025.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

143.00	1626.0
143.90	500.0
144.90	368.0
146.00	1573.0
147.00	4390.0
148.00	8164.0
149.00	1814.0
150.00	854.0
151.00	1093.0
151.60	778.0
152.20	435.0
153.00	2484.0
153.90	1695.0
155.00	5232.0
156.00	7197.0
157.10	1295.0
157.90	1545.0
159.00	1236.0
160.00	2214.0
161.00	3927.0
161.90	1218.0
163.20	304.0
163.90	301.0
165.00	2945.0
166.00	2523.0
167.00	16552.0
168.00	10600.0
169.10	1830.0
170.00	653.0
170.80	688.0
172.00	1478.0
173.00	1744.0
174.00	3477.0
175.10	6838.0
176.00	2322.0
176.90	2539.0
178.00	925.0
179.00	12640.0
180.00	8854.0
181.00	4074.0
182.00	825.0
182.70	435.0
184.00	1201.0
185.00	6085.0
186.00	46256.0
187.00	13925.0
187.90	1708.0
189.00	2955.0
189.90	625.0
191.00	1743.0
192.00	4365.0
193.00	4456.0
193.90	813.0
195.10	602.0
196.00	11638.0
197.00	1759.0
198.00	362048.0
199.00	24376.0
200.00	1670.0
200.90	823.0
201.70	1645.0
203.00	2752.0
204.00	12741.0
205.00	19632.0
206.10	91744.0
207.10	12075.0
208.00	3190.0
209.10	1254.0
210.30	1833.0
211.00	3113.0
214.90	706.0
216.00	1700.0
217.00	21792.0
218.10	3356.0
219.20	418.0
220.20	655.0
221.00	16752.0
223.10	4476.0
224.10	49736.0
225.10	12245.0
226.00	1673.0
227.00	19864.0
228.00	3564.0
229.00	4697.0
230.00	679.0
231.10	2012.0
233.10	623.0
234.00	1551.0
235.00	1760.0
236.10	861.0
236.90	2462.0
238.80	480.0
240.10	709.0
240.90	851.0
242.00	2779.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

243.00	2514.0
244.10	40600.0
245.10	5134.0
246.00	7946.0
246.90	1008.0
247.90	571.0
248.90	1161.0
250.00	379.0
251.00	347.0
252.90	1005.0
253.90	1171.0
255.00	194432.0
256.00	28488.0
257.10	2042.0
258.00	10337.0
259.10	1155.0
265.00	4996.0
266.00	764.0
271.00	643.0
272.10	482.0
273.00	6488.0
274.00	14089.0
275.00	87776.0
276.00	10186.0
277.00	5699.0
277.90	1121.0
278.80	480.0
281.20	322.0
283.00	834.0
284.00	344.0
285.10	1175.0
286.00	317.0
293.10	1895.0
294.00	616.0
294.90	616.0
296.00	21760.0
297.00	3785.0
302.00	377.0
303.10	2251.0
304.00	1054.0
310.10	430.0
313.90	1079.0
315.00	2340.0
316.00	1429.0
317.00	337.0
321.00	735.0
321.90	412.0
323.10	8690.0
324.10	1639.0
327.10	1090.0
328.00	791.0
328.90	313.0
331.90	591.0
333.00	823.0
334.00	4944.0
335.10	1126.0
341.00	776.0
346.00	1559.0
352.00	2823.0
353.00	1523.0
354.00	3186.0
355.00	588.0
365.00	10043.0
366.00	1486.0
371.10	773.0
372.00	3479.0
373.10	1133.0
383.00	1213.0
384.00	338.0
390.10	645.0
390.90	385.0
391.90	359.0
401.90	1442.0
403.10	2899.0
404.00	902.0
421.10	3729.0
422.00	2577.0
423.00	17480.0
424.00	5164.0
425.10	609.0
441.00	53984.0
442.00	359744.0
443.00	66744.0
444.00	5861.0
445.00	565.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46

m/z	Abundance
40.00	6226.0
41.10	1289.0
41.90	466.0
43.00	513.0
44.00	7090.0
45.10	575.0
49.00	332.0
51.00	337.0
55.10	438.0
56.00	371.0
57.20	715.0
73.00	572.0
77.00	349.0
79.10	471.0
104.90	427.0
123.00	1430.0
167.00	643.0
179.00	390.0
192.90	302.0
207.00	1241.0
208.10	309.0
208.90	325.0

Instrument :
BNA_M
ClientSampleId :
DFTPP46