

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM041715\
 Data File : BM001039.D
 Acq On : 17 Apr 2015 19:45
 Operator : TP/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.439

Quant Time: Apr 18 00:14:50 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM041715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 17 15:43:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
3	Naphthalene	1.029	1.015	1.4	104	0.00
4 SURR	2-Methylnaphthalene-d10	0.524	0.507	3.2	103	0.00
5	2-Methylnaphthalene	0.693	0.670	3.3	103	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
7	Acenaphthylene	1.770	1.692	4.4	98	0.00
8 C	Acenaphthene	1.383	1.365	1.3	101	0.00
9	Fluorene	1.612	1.581	1.9	100	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
11	Pentachlorophenol	0.094	0.066	29.8#	71	0.00
12	Phenanthrene	1.160	1.134	2.2	100	0.00
13	Anthracene	1.047	1.028	1.8	99	0.00
14 SURR	Fluoranthene-d10	0.880	0.853	3.1	98	0.00
15 C	Fluoranthene	1.286	1.248	3.0	100	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
17	Pyrene	1.525	1.511	0.9	99	0.00
18	Benzo(a)anthracene	1.262	1.198	5.1	95	0.00
19	Chrysene	1.384	1.392	-0.6	103	0.00
20 I	Perylene-d12	1.000	1.000	0.0	102	0.00
21	Benzo(b)fluoranthene	1.476	1.497	-1.4	108	0.01
22	Benzo(k)fluoranthene	1.476	1.410	4.5	93	0.01
23 C	Benzo(a)pyrene	1.382	1.340	3.0	99	0.01
24	Indeno(1,2,3-cd)pyrene	1.540	1.527	0.8	101	0.01
25	Dibenzo(a,h)anthracene	1.248	1.243	0.4	102	0.00
26	Benzo(g,h,i)perylene	1.360	1.356	0.3	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0