

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041717\
 Data File : BM009653.D
 Acq On : 17 Apr 2017 18:23
 Operator : SJ/MA
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV07

Quant Time: Apr 17 19:01:30 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM041717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 17 18:22:30 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.638	0.583	8.6	91	0.00
3 S	1,4-Dioxane-d8	0.543	0.493	9.2	92	0.00
4	Benzaldehyde	1.127	1.321	-17.2	93	0.00
5 S	Phenol-d5	1.823	1.798	1.4	96	0.00
6	Phenol	1.968	1.936	1.6	95	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.282	1.270	0.9	94	0.00
8	Bis(2-Chloroethyl)ether	1.508	1.477	2.1	96	0.00
9 S	2-Chlorophenol-d4	1.267	1.240	2.1	94	0.00
10	2-Chlorophenol	1.324	1.357	-2.5	100	0.00
11	2-Methylphenol	1.377	1.370	0.5	97	0.00
12	2,2'-oxybis(1-Chloropropane	2.465	2.455	0.4	97	0.00
13 S	4-Methylphenol-d8	1.347	1.310	2.7	95	0.00
14	Acetophenone	2.305	2.279	1.1	95	0.00
15 P	N-Nitroso-di-n-propylamine	1.361	1.381	-1.5	95	0.00
16	4-Methylphenol	1.505	1.437	4.5	91	0.00
17	Hexachloroethane	0.641	0.641	0.0	96	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
19 S	Nitrobenzene-d5	0.154	0.156	-1.3	95	0.00
20	Nitrobenzene	0.509	0.505	0.8	96	0.00
21	Isophorone	0.831	0.827	0.5	93	0.00
22 S	2-Nitrophenol-d4	0.163	0.155	4.9	89	0.00
23 C	2-Nitrophenol	0.182	0.182	0.0	95	0.00
24	2,4-Dimethylphenol	0.428	0.430	-0.5	96	0.00
25	Bis(2-Chloroethoxy)methane	0.500	0.513	-2.6	97	0.00
26 S	2,4-Dichlorophenol-d3	0.323	0.312	3.4	93	0.00
27 C	2,4-Dichlorophenol	0.324	0.316	2.5	91	0.00
28	Naphthalene	1.044	1.029	1.4	93	0.00
29 S	4-Chloroaniline-d4	0.325	0.389	-19.7	93	0.00
30	4-Chloroaniline	0.339	0.391	-15.3	95	0.00
31 C	Hexachlorobutadiene	0.242	0.235	2.9	91	0.00
32	Caprolactam	0.106	0.098	7.5	90	0.00
33 C	4-Chloro-3-methylphenol	0.365	0.361	1.1	93	0.00
34	2-Methylnaphthalene	0.738	0.742	-0.5	95	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
36	1,2,4,5-Tetrachlorobenzene	0.736	0.720	2.2	95	0.00
37	Hexachlorocyclopentadiene	0.401	0.365	9.0	96	0.00
38 C	2,4,6-Trichlorophenol	0.471	0.451	4.2	93	0.00
39	2,4,5-Trichlorophenol	0.476	0.474	0.4	100	0.00
40	1,1'-Biphenyl	1.677	1.657	1.2	95	0.00
41	2-Chloronaphthalene	1.324	1.286	2.9	96	0.00
42	2-Nitroaniline	0.506	0.505	0.2	95	0.00
43 S	Dimethylphthalate-d6	1.613	1.591	1.4	95	0.00
44	Dimethylphthalate	1.622	1.584	2.3	94	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041717\
 Data File : BM009653.D
 Acq On : 17 Apr 2017 18:23
 Operator : SJ/MA
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV07

Quant Time: Apr 17 19:01:30 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM041717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 17 18:22:30 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.324	0.318	1.9	91	0.00
46 S	Acenaphthylene-d8	2.042	1.999	2.1	93	0.00
47	Acenaphthylene	2.049	1.981	3.3	93	0.00
48	3-Nitroaniline	0.320	0.323	-0.9	96	0.00
49 C	Acenaphthene	1.378	1.346	2.3	95	0.00
50	2,4-Dinitrophenol	0.201	0.166	17.4	88	0.00
51 S	4-Nitrophenol-d4	0.290	0.259	10.7	91	0.00
52	4-Nitrophenol	0.338	0.318	5.9	91	0.00
53	Dibenzofuran	1.984	1.939	2.3	94	0.00
54	2,4-Dinitrotoluene	0.480	0.462	3.7	93	0.00
55	2,3,4,6-Tetrachlorophenol	0.437	0.418	4.3	91	0.00
56	Diethylphthalate	1.658	1.622	2.2	95	0.00
57 S	Fluorene-d10	1.386	1.339	3.4	94	0.00
58	Fluorene	1.646	1.589	3.5	94	0.00
59	4-Chlorophenyl-phenylether	0.855	0.835	2.3	94	0.00
60	4-Nitroaniline	0.350	0.319	8.9	89	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.127	0.119	6.3	97	0.00
63	4,6-Dinitro-2-methylphenol	0.134	0.129	3.7	97	0.00
64	N-Nitrosodiphenylamine	0.611	0.611	0.0	96	0.00
65	4-Bromophenyl-phenylether	0.230	0.228	0.9	95	0.00
66	Hexachlorobenzene	0.252	0.248	1.6	94	0.00
67	Atrazine	0.237	0.227	4.2	89	0.00
68 C	Pentachlorophenol	0.156	0.140	10.3	92	0.00
69	Phenanthrene	1.122	1.105	1.5	94	0.00
70 S	Anthracene-d10	0.969	0.941	2.9	93	0.00
71	Anthracene	1.168	1.163	0.4	95	0.00
72	Carbazole	1.026	0.995	3.0	94	0.00
73	Di-n-butylphthalate	1.242	1.205	3.0	92	0.00
74 C	Fluoranthene	1.392	1.304	6.3	92	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
76 S	Pyrene-d10	0.861	0.836	2.9	94	0.00
77	Pyrene	1.136	1.105	2.7	94	0.00
78	Butylbenzylphthalate	0.469	0.450	4.1	92	0.00
79	3,3'-Dichlorobenzidine	0.406	0.399	1.7	92	0.00
80	Benzo(a)anthracene	1.171	1.135	3.1	94	0.00
81	Bis(2-ethylhexyl)phthalate	0.694	0.673	3.0	95	0.00
82	Chrysene	1.055	1.036	1.8	94	0.00
83 I	Perylene-d12	1.000	1.000	0.0	101	0.00
84	Di-n-octyl phthalate	1.331	1.219	8.4	94	0.00
85	Benzo(b)fluoranthene	1.231	1.240	-0.7	99	0.00
86	Benzo(k)fluoranthene	1.172	1.130	3.6	93	0.00
87 S	Benzo(a)pyrene-d12	0.905	0.888	1.9	96	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041717\
Data File : BM009653.D
Acq On : 17 Apr 2017 18:23
Operator : SJ/MA
Sample : SSTDICV020
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SICV07

Quant Time: Apr 17 19:01:30 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM041717.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Apr 17 18:22:30 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.186	1.162	2.0	95	0.00
89	Indeno(1,2,3-cd)pyrene	1.386	1.355	2.2	96	0.00
90	Dibenzo(a,h)anthracene	1.172	1.155	1.5	96	0.00
91	Benzo(a,h,i)perylene	1.133	1.124	0.8	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0