


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041724\
Data File : BM045310.D
Acq On : 17 Apr 2024 10:40
Operator : MA/JU
Sample : DFTPP021
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP021

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM040524.MA.M
Title : SVOA CALIBRATION
Last Update : Wed Apr 17 12:42:01 2024



AutoFind: Scans 2438, 2439, 2440; Background Corrected with Scan 2433

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	60.1	27896	PASS
68	69	0.00	2	2.0	544	PASS
69	198	0.00	100	60.1	27867	PASS
70	69	0.00	2	0.6	159	PASS
127	198	10	80	64.4	29888	PASS
197	198	0.00	2	0.8	377	PASS
198	198	100	100	100.0	46400	PASS
199	198	5	9	7.5	3501	PASS
275	198	10	60	26.7	12398	PASS
365	198	1	100	5.5	2567	PASS
441	198	0.01	100	12.3	5713	PASS
442	198	50	100	78.9	36603	PASS
443	442	15	24	18.9	6914	PASS

m/z	Abundance
40.00	293.0
43.90	595.0
72.90	413.0
192.80	152.0
206.90	775.0
208.00	172.0
208.90	215.0
280.90	181.0

Instrument :

BNA_M

ClientSampleId :

DFTPP021

m/z	Abundance
36.90	162.0
38.00	451.0
39.00	1876.0
39.90	645.0
43.90	549.0
49.00	203.0
50.00	7022.0
51.00	24688.0
52.00	1127.0
56.00	657.0
57.00	1463.0
60.90	297.0
61.80	300.0
63.00	979.0
64.00	238.0
64.90	396.0
68.00	550.0
69.00	24096.0
69.90	219.0
73.00	598.0
74.00	3289.0
75.00	4360.0
76.00	2244.0
77.00	33344.0
78.00	2459.0
79.00	1793.0
79.90	1635.0
80.90	2022.0
81.90	557.0
83.00	444.0
83.80	160.0
84.90	342.0
85.90	678.0
87.00	282.0
91.00	421.0
91.90	454.0
93.00	4155.0
93.90	226.0
95.90	182.0
98.00	3262.0
99.00	2873.0
100.90	1080.0
102.90	303.0
103.80	596.0
104.90	538.0
106.20	188.0
106.90	8499.0
107.90	1159.0
108.80	160.0
109.90	11531.0
110.90	2090.0
111.90	244.0
112.90	175.0
115.90	465.0
116.90	10327.0
117.90	648.0
122.00	379.0
122.90	877.0
123.90	366.0
124.90	344.0
127.00	23824.0
127.90	1712.0
128.90	9878.0
129.90	888.0
131.00	186.0
133.80	290.0
134.90	863.0
135.90	252.0
136.90	483.0
137.80	158.0
140.90	1357.0
141.90	347.0
142.80	295.0
145.70	258.0
146.80	902.0
147.90	2177.0
148.90	510.0
149.80	163.0
151.10	192.0
153.00	390.0
153.90	238.0
155.00	756.0
155.90	984.0
157.90	191.0
158.90	275.0
159.70	355.0
160.90	648.0
161.80	243.0
164.90	483.0
166.00	399.0
166.90	3944.0
167.90	2546.0

Instrument :
BNA_M
ClientSampleId :
DFTPP021

169.00	380.0
171.80	217.0
173.00	455.0
173.80	556.0
175.00	842.0
175.90	266.0
176.90	394.0
178.90	1766.0
180.00	1392.0
180.90	763.0
184.00	254.0
185.00	912.0
186.00	6088.0
187.00	1567.0
188.00	283.0
189.00	389.0
190.70	279.0
192.00	964.0
193.00	775.0
193.80	281.0
195.90	926.0
196.90	380.0
197.90	34200.0
198.90	2643.0
199.80	187.0
201.60	236.0
202.90	510.0
204.00	2029.0
204.90	2510.0
206.00	11138.0
207.00	2003.0
207.80	688.0
208.90	233.0
210.10	260.0
210.80	212.0
214.90	166.0
215.90	477.0
216.90	2971.0
217.90	408.0
221.00	1525.0
222.10	437.0
223.00	625.0
224.00	6196.0
224.90	1395.0
226.00	185.0
226.90	2888.0
227.80	411.0
228.90	546.0
231.00	178.0
233.80	194.0
235.90	389.0
236.80	259.0
241.90	404.0
243.00	380.0
243.90	4160.0
245.00	544.0
245.90	953.0
248.90	305.0
252.90	153.0
254.90	22240.0
255.90	3181.0
257.00	420.0
257.90	1437.0
258.90	293.0
264.90	689.0
272.90	807.0
273.90	1908.0
274.90	7766.0
276.00	1206.0
276.90	1035.0
280.90	239.0
282.90	204.0
292.90	251.0
295.90	3110.0
296.80	608.0
302.90	304.0
314.90	363.0
315.90	214.0
322.90	695.0
323.80	175.0
326.80	211.0
333.90	388.0
334.80	158.0
345.90	338.0
352.00	407.0
352.90	179.0
354.00	231.0
354.90	160.0
365.00	1241.0
365.90	225.0
371.80	327.0
382.80	196.0
401.80	159.0
402.90	365.0
420.90	201.0
421.90	236.0

Instrument :
BNA_M
ClientSampleId :
DFTPP021

423.00	1111.0
424.00	304.0
441.00	2694.0
442.00	18256.0
443.00	3564.0
443.90	223.0

Instrument :
BNA_M
ClientSampleId :
DFTPP021

m/z	Abundance
36.90	259.0
37.90	424.0
39.00	2494.0
40.00	512.0
43.90	641.0
48.90	292.0
50.00	10699.0
51.00	32624.0
52.00	1706.0
54.90	298.0
56.00	874.0
56.90	2205.0
60.90	406.0
62.00	544.0
63.00	1099.0
65.00	677.0
67.90	696.0
69.00	33360.0
70.00	258.0
73.10	793.0
74.00	4835.0
75.00	6578.0
76.10	2741.0
77.00	47264.0
78.00	3404.0
79.00	3085.0
80.00	1887.0
81.00	3059.0
81.90	708.0
82.90	499.0
83.90	226.0
84.70	226.0
85.90	805.0
86.90	353.0
88.00	163.0
91.00	427.0
91.90	476.0
93.00	5223.0
93.90	410.0
96.00	352.0
98.00	4008.0
98.90	3178.0
100.00	213.0
100.90	1589.0
102.90	418.0
103.90	868.0
105.00	947.0
105.80	315.0
107.00	12707.0
107.90	1785.0
109.00	347.0
109.90	16600.0
111.00	3250.0
112.00	458.0
112.90	273.0
115.90	529.0
116.90	15000.0
117.90	1052.0
119.00	169.0
122.00	715.0
122.90	1246.0
123.90	651.0
124.90	621.0
126.10	208.0
127.00	36672.0
128.00	2917.0
128.90	16472.0
130.00	1254.0
130.90	408.0
133.00	165.0
133.80	382.0
134.90	1358.0
135.80	520.0
136.90	801.0
137.80	224.0
139.80	225.0
140.90	1939.0
141.90	520.0
142.90	533.0
145.90	318.0
146.90	1131.0
147.90	3611.0
148.90	594.0
150.00	217.0
151.00	208.0
152.90	685.0
153.90	441.0
154.90	1384.0
155.90	1504.0
156.80	288.0
157.80	216.0
159.00	316.0

Instrument :
BNA_M
ClientSampleId :
DFTPP021

159.90	747.0
160.90	1106.0
162.00	320.0
164.80	610.0
165.90	1035.0
166.90	5899.0
167.90	3708.0
168.90	558.0
170.00	175.0
171.00	253.0
171.70	220.0
172.80	340.0
174.00	780.0
175.00	1600.0
175.80	344.0
176.90	952.0
177.80	164.0
178.10	164.0
178.90	3228.0
179.90	2030.0
180.90	1096.0
183.80	261.0
184.90	1490.0
186.00	9843.0
187.00	2755.0
187.90	512.0
188.80	567.0
190.70	412.0
191.90	1177.0
192.90	971.0
194.00	230.0
195.90	1566.0
197.00	343.0
197.90	54992.0
198.90	4292.0
200.00	353.0
201.20	181.0
201.70	283.0
202.90	617.0
204.00	2991.0
204.90	4221.0
206.00	17776.0
207.00	3695.0
207.90	812.0
208.80	495.0
209.90	288.0
210.90	842.0
214.80	357.0
216.00	535.0
216.90	5403.0
218.00	805.0
221.00	2393.0
221.80	599.0
222.90	1356.0
224.00	9589.0
225.00	2778.0
225.90	377.0
226.90	4487.0
227.90	645.0
229.00	1073.0
229.90	201.0
230.90	365.0
233.80	222.0
234.90	525.0
235.80	217.0
236.80	573.0
239.90	163.0
240.90	307.0
241.90	693.0
243.00	752.0
243.90	7699.0
244.90	1271.0
245.90	1783.0
246.80	385.0
248.80	308.0
252.90	188.0
254.00	220.0
254.90	38784.0
255.90	5746.0
256.90	333.0
257.90	2518.0
258.80	622.0
264.90	1396.0
265.90	164.0
272.90	1355.0
273.90	2845.0
275.00	15030.0
276.00	2162.0
276.90	1520.0
277.90	334.0
280.90	289.0
282.90	270.0
284.90	177.0
292.90	258.0
295.90	6198.0
296.90	882.0

Instrument :
BNA_M
ClientSampleId :
DFTPP021

302.00	152.0
303.00	919.0
303.90	199.0
313.90	301.0
314.90	1031.0
315.80	388.0
322.90	1528.0
324.00	192.0
326.90	448.0
333.10	166.0
333.90	1071.0
334.90	219.0
340.90	259.0
345.90	364.0
352.00	570.0
352.90	454.0
353.90	381.0
364.90	3243.0
365.80	444.0
371.90	769.0
373.00	192.0
382.90	342.0
402.00	291.0
403.00	628.0
420.90	310.0
422.00	476.0
423.00	2385.0
424.00	890.0
441.00	6941.0
442.00	41560.0
443.00	8782.0
443.90	739.0

Instrument :
BNA_M
ClientSampleId :
DFTPP021

m/z	Abundance
38.00	571.0
39.00	1756.0
40.00	395.0
43.90	570.0
49.00	241.0
50.00	7828.0
51.00	26376.0
52.00	1474.0
55.90	853.0
56.90	1249.0
60.80	264.0
61.90	402.0
63.00	875.0
64.90	349.0
68.10	387.0
69.00	26144.0
73.00	709.0
74.00	3545.0
75.00	4571.0
76.00	2120.0
77.00	36912.0
78.00	2513.0
79.00	2684.0
80.00	1757.0
81.00	1953.0
82.00	724.0
82.90	382.0
84.90	318.0
86.00	560.0
86.80	216.0
90.90	425.0
91.90	517.0
93.00	3910.0
93.90	298.0
95.90	423.0
97.90	3156.0
98.90	2627.0
99.90	155.0
100.90	1361.0
102.80	323.0
104.00	722.0
104.90	1029.0
106.00	307.0
107.00	9955.0
108.10	1383.0
108.90	267.0
109.90	13795.0
110.90	2706.0
112.00	230.0
116.00	563.0
116.90	11314.0
117.80	555.0
118.90	179.0
121.90	513.0
122.90	872.0
123.80	517.0
124.70	369.0
125.10	232.0
125.90	212.0
127.00	29168.0
127.90	2226.0
128.90	12669.0
129.90	1241.0
131.00	274.0
133.90	418.0
134.90	1299.0
135.90	518.0
137.00	707.0
139.80	177.0
140.90	1630.0
141.90	502.0
142.90	381.0
146.00	361.0
146.90	1129.0
147.90	2604.0
148.90	666.0
150.90	178.0
152.90	521.0
154.00	585.0
155.00	1079.0
156.00	1218.0
156.90	271.0
157.80	246.0
158.90	364.0
159.90	661.0
160.90	776.0
161.80	392.0
162.70	150.0
164.90	816.0
165.90	426.0
166.90	4832.0
168.00	3201.0

Instrument :
BNA_M
ClientSampleId :
DFTPP021

168.90	472.0
171.80	346.0
172.90	449.0
173.90	979.0
175.00	1664.0
175.90	247.0
176.80	529.0
177.80	208.0
178.90	2585.0
179.90	1966.0
180.90	954.0
183.70	155.0
184.00	154.0
184.90	1277.0
185.90	8592.0
186.90	2384.0
187.90	198.0
188.90	584.0
190.80	369.0
191.90	876.0
193.00	1046.0
193.80	180.0
196.00	1695.0
196.80	407.0
197.90	50008.0
198.90	3567.0
199.90	431.0
201.50	307.0
202.90	574.0
203.90	2532.0
205.00	4193.0
206.00	15208.0
207.00	3010.0
208.00	841.0
209.10	597.0
210.20	332.0
210.90	738.0
214.80	272.0
215.90	446.0
216.90	4470.0
217.90	641.0
218.80	206.0
221.00	2332.0
221.90	726.0
222.90	949.0
224.00	8928.0
225.00	2276.0
225.90	222.0
226.90	4683.0
227.90	726.0
228.90	764.0
230.80	175.0
233.80	323.0
234.90	354.0
235.80	304.0
236.90	504.0
238.90	222.0
240.90	227.0
241.80	623.0
243.00	479.0
243.90	7849.0
245.00	1155.0
245.90	1551.0
246.70	194.0
247.10	256.0
249.00	152.0
252.90	203.0
253.80	283.0
254.90	35984.0
255.90	5314.0
256.90	402.0
257.90	2363.0
258.90	528.0
264.90	1115.0
266.10	163.0
270.90	169.0
272.90	1338.0
273.90	3107.0
275.00	14398.0
275.90	1863.0
277.00	1557.0
277.90	270.0
282.90	170.0
284.90	222.0
293.00	344.0
296.00	5404.0
296.90	736.0
301.90	154.0
302.90	588.0
303.80	154.0
313.90	217.0
315.00	875.0
316.10	277.0
323.00	1646.0
323.90	362.0
326.70	261.0

Instrument :
BNA_M
ClientSampleId :
DFTPP021

328.00	252.0
334.00	1227.0
335.00	183.0
345.80	469.0
352.10	475.0
353.10	257.0
354.00	410.0
354.80	170.0
365.00	3217.0
365.90	374.0
372.00	705.0
373.00	194.0
382.90	195.0
401.90	455.0
403.00	619.0
403.80	210.0
421.00	478.0
421.90	445.0
423.00	2385.0
423.90	819.0
441.00	7504.0
442.00	49992.0
443.00	8395.0
443.90	844.0

Instrument :
BNA_M
ClientSampleId :
DFTPP021