

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041815\
 Data File : BM001043.D
 Acq On : 17 Apr 2015 23:20
 Operator : TP/IZ
 Sample : SSTDICC0.5
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC0.5

Quant Time: Apr 18 05:59:13 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 18 05:45:49 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	16076	5.00	ng	0.00
4) Naphthalene-d8	10.36	136	57186	5.00	ng	0.00
8) Acenaphthene-d10	14.24	164	29792	5.00	ng	0.00
13) Phenanthrene-d10	16.99	188	64022	5.00	ng	0.00
19) Chrysene-d12	21.19	240	55201	5.00	ng	0.00
25) Perylene-d12	23.36	264	50133	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.74	82	1443	0.50	ng	0.00
9) 2-Fluorobiphenyl	12.86	172	4920	0.53	ng	0.00
21) Terphenyl-d14	19.63	244	3617	0.54	ng	0.00
Target Compounds						Ovalue
2) 1,4-Dioxane	3.08	88	1105	0.81	ng	# 100
3) n-Nitrosodimethylamine	3.42	42	856	0.57	ng	# 95
6) Naphthalene	10.41	128	6818	0.51	ng	99
7) 2-Methylnaphthalene	12.04	142	4373	0.52	ng	98
10) Acenaphthylene	13.95	152	6276	0.48	ng	99
11) Acenaphthene	14.29	154	4502	0.49	ng	99
12) Fluorene	15.30	166	5187	0.51	ng	100
14) Hexachlorobenzene	16.31	284	1747	0.47	ng	# 100
15) Pentachlorophenol	16.64	266	352	0.38	ng	91
16) Phenanthrene	17.03	178	8528	0.49	ng	100
17) Anthracene	17.11	178	7578	0.51	ng	99
18) Fluoranthene	19.05	202	8607	0.49	ng	100
20) Pyrene	19.42	202	8985	0.50	ng	100
22) Benzo(a)anthracene	21.17	228	7520	0.48	ng	99
23) Chrysene	21.22	228	8573	0.52	ng	99
24) Indeno(1,2,3-cd)pyrene	25.52	276	8448	0.49	ng	100
26) Benzo(b)fluoranthene	22.70	252	7560	0.48	ng	98
27) Benzo(k)fluoranthene	22.74	252	7787	0.51	ng	98
28) Benzo(a)pyrene	23.25	252	6630	0.49	ng	97
29) Dibenzo(a,h)anthracene	25.54	278	6644	0.50	ng	97
30) Benzo(g,h,i)perylene	26.18	276	7249	0.50	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041815\
Data File : BM001043.D
Acq On : 17 Apr 2015 23:20
Operator : TP/IZ
Sample : SSTDICC0.5
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC0.5

Quant Time: Apr 18 05:59:13 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 18 05:45:49 2015
Response via : Initial Calibration

