

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041815\
 Data File : BM001048.D
 Acq On : 18 Apr 2015 06:50
 Operator : TP/IZ
 Sample : SSTDICC010
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC010

Quant Time: Apr 18 07:41:29 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 18 05:45:49 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	17193	5.00	ng	0.00
4) Naphthalene-d8	10.36	136	62176	5.00	ng	0.00
8) Acenaphthene-d10	14.24	164	33559	5.00	ng	0.00
13) Phenanthrene-d10	16.99	188	143167	5.00	ng	0.00
19) Chrysene-d12	21.19	240	131433	5.00	ng	0.00
25) Perylene-d12	23.36	264	119218	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.74	82	43702	14.06	ng	0.00
9) 2-Fluorobiphenyl	12.86	172	113599	10.87	ng	0.00
21) Terphenyl-d14	19.63	244	100198	6.28	ng	0.00
Target Compounds						Ovalue
2) 1,4-Dioxane	3.07	88	16507	11.33	ng	# 100
3) n-Nitrosodimethylamine	3.38	42	21678	13.46	ng	98
6) Naphthalene	10.41	128	154868	10.73	ng	99
7) 2-Methylnaphthalene	12.04	142	115038	12.55	ng	96
10) Acenaphthylene	13.95	152	203403	13.68	ng	100
11) Acenaphthene	14.29	154	116523	11.35	ng	100
12) Fluorene	15.29	166	142888	12.35	ng	100
14) Hexachlorobenzene	16.31	284	46564	5.63	ng	# 100
15) Pentachlorophenol	16.64	266	20462	9.79	ng	96
16) Phenanthrene	17.03	178	221747	5.70	ng	99
17) Anthracene	17.11	178	227085	6.86	ng	100
18) Fluoranthene	19.05	202	249637	6.42	ng	99
20) Pyrene	19.42	202	261012	6.09	ng	100
22) Benzo(a)anthracene	21.17	228	231780	6.16	ng	96
23) Chrysene	21.22	228	218279	5.56	ng	97
24) Indeno(1,2,3-cd)pyrene	25.53	276	241896	5.94	ng	99
26) Benzo(b)fluoranthene	22.70	252	217159	5.78	ng	97
27) Benzo(k)fluoranthene	22.74	252	221759	6.16	ng	96
28) Benzo(a)pyrene	23.25	252	200772	6.22	ng	97
29) Dibenzo(a,h)anthracene	25.54	278	189303	5.95	ng	98
30) Benzo(g,h,i)perylene	26.18	276	197679	5.76	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041815\
Data File : BM001048.D
Acq On : 18 Apr 2015 06:50
Operator : TP/IZ
Sample : SSTDIC010
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDIC010

Quant Time: Apr 18 07:41:29 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 18 05:45:49 2015
Response via : Initial Calibration

