

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041815\
 Data File : BM001055.D
 Acq On : 18 Apr 2015 11:00
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001EC

Manual Integrations
 APPROVED

apatel
 4/20/2015 4:34:21 PM

Quant Time: Apr 20 08:25:04 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 08:05:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	16721	5.00	ng	0.00
4) Naphthalene-d8	10.36	136	60092	5.00	ng	0.00
8) Acenaphthene-d10	14.24	164	31772	5.00	ng	0.00
13) Phenanthrene-d10	16.99	188	69451	5.00	ng	0.00
19) Chrysene-d12	21.19	240	59295	5.00	ng	0.00
25) Perylene-d12	23.36	264	53525	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.74	82	3190	1.02	ng	0.00
9) 2-Fluorobiphenyl	12.86	172	10313	0.98	ng	0.00
21) Terphenyl-d14	19.63	244	7860	1.00	ng	0.00
Target Compounds						Ovalue
2) 1,4-Dioxane	3.08	88	1543m	0.82	ng	
3) n-Nitrosodimethylamine	3.41	42	1670	0.95	ng	# 99
6) Naphthalene	10.41	128	14131	0.99	ng	100
7) 2-Methylnaphthalene	12.04	142	9137	0.99	ng	99
10) Acenaphthylene	13.95	152	13914	0.99	ng	100
11) Acenaphthene	14.29	154	9311	0.95	ng	99
12) Fluorene	15.29	166	11146	0.98	ng	99
14) Hexachlorobenzene	16.31	284	3719	0.97	ng	# 100
15) Pentachlorophenol	16.64	266	660	0.75	ng	95
16) Phenanthrene	17.03	178	17830	0.97	ng	99
17) Anthracene	17.11	178	16671	0.98	ng	100
18) Fluoranthene	19.05	202	18560	0.97	ng	100
20) Pyrene	19.42	202	19521	0.99	ng	100
22) Benzo(a)anthracene	21.17	228	16859	0.98	ng	98
23) Chrysene	21.22	228	18204	1.01	ng	97
24) Indeno(1,2,3-cd)pyrene	25.52	276	18387	1.00	ng	100
26) Benzo(b)fluoranthene	22.70	252	16571	0.96	ng	99
27) Benzo(k)fluoranthene	22.74	252	17003	1.03	ng	98
28) Benzo(a)pyrene	23.25	252	14583	0.99	ng	99
29) Dibenzo(a,h)anthracene	25.53	278	14489	1.00	ng	100
30) Benzo(g,h,i)perylene	26.18	276	15530	0.99	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041815\
Data File : BM001055.D
Acq On : 18 Apr 2015 11:00
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001EC

Manual Integrations
APPROVED

apatel
4/20/2015 4:34:21 PM

Quant Time: Apr 20 08:25:04 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 20 08:05:22 2015
Response via : Initial Calibration

