

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041815\
 Data File : BM001057.D
 Acq On : 18 Apr 2015 13:24
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001

Manual Integrations
 APPROVED

apatel
 4/20/2015 4:34:25 PM

Quant Time: Apr 20 08:26:36 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 08:05:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	15527	5.00	ng	0.00
4) Naphthalene-d8	10.36	136	55714	5.00	ng	0.00
8) Acenaphthene-d10	14.24	164	30057	5.00	ng	0.00
13) Phenanthrene-d10	16.99	188	67323	5.00	ng	0.00
19) Chrysene-d12	21.19	240	57954	5.00	ng	0.00
25) Perylene-d12	23.36	264	52016	5.00	ng	0.00
System Monitoring Compounds						
5) Nitrobenzene-d5	8.74	82	2900	1.00	ng	0.00
9) 2-Fluorobiphenyl	12.86	172	9694	0.98	ng	0.00
21) Terphenyl-d14	19.63	244	7732	1.01	ng	0.00
Target Compounds						Ovalue
2) 1,4-Dioxane	3.08	88	1408m	0.80	ng	
3) n-Nitrosodimethylamine	3.41	42	1579	0.96	ng	# 99
6) Naphthalene	10.41	128	13163	0.99	ng	100
7) 2-Methylnaphthalene	12.04	142	8624	1.01	ng	100
10) Acenaphthylene	13.95	152	13087	0.99	ng	99
11) Acenaphthene	14.29	154	8850	0.95	ng	98
12) Fluorene	15.29	166	10871	1.02	ng	100
14) Hexachlorobenzene	16.31	284	3644	0.98	ng	# 100
15) Pentachlorophenol	16.64	266	603	0.72	ng	94
16) Phenanthrene	17.03	178	17257	0.97	ng	100
17) Anthracene	17.11	178	16037	0.98	ng	100
18) Fluoranthene	19.05	202	18194	0.98	ng	100
20) Pyrene	19.42	202	19067	0.99	ng	100
22) Benzo(a)anthracene	21.17	228	15995	0.95	ng	99
23) Chrysene	21.22	228	17623	1.00	ng	100
24) Indeno(1,2,3-cd)pyrene	25.52	276	17787	0.99	ng	100
26) Benzo(b)fluoranthene	22.70	252	16278	0.97	ng	99
27) Benzo(k)fluoranthene	22.74	252	15970	1.00	ng	99
28) Benzo(a)pyrene	23.25	252	14073	0.98	ng	100
29) Dibenzo(a,h)anthracene	25.53	278	14113	1.00	ng	100
30) Benzo(g,h,i)perylene	26.17	276	15077	0.98	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041815\
Data File : BM001057.D
Acq On : 18 Apr 2015 13:24
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001

Manual Integrations
APPROVED

apatel
4/20/2015 4:34:25 PM

Quant Time: Apr 20 08:26:36 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 20 08:05:22 2015
Response via : Initial Calibration

