

Data Path : Z:\HPCHEM1\BNA M\Data\BM041815\
 Data File : BM001049.D
 Acq On : 18 Apr 2015 07:26
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM041815

Quant Time: Apr 20 15:57:34 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 08:05:22 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.564	0.495	12.2	94	0.00
3	n-Nitrosodimethylamine	0.528	0.505	4.4	105	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
5 S	Nitrobenzene-d5	0.261	0.267	-2.3	110	0.00
6	Naphthalene	1.194	1.173	1.8	108	0.00
7	2-Methylnaphthalene	0.770	0.761	1.2	107	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	108	0.00
9 S	2-Fluorobiphenyl	1.651	1.618	2.0	107	0.00
10	Acenaphthylene	2.209	2.185	1.1	106	0.00
11	Acenaphthene	1.546	1.446	6.5	105	0.00
12	Fluorene	1.781	1.773	0.4	107	-0.02
13 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
14	Hexachlorobenzene	0.276	0.268	2.9	109	0.00
15	Pentachlorophenol	0.068	0.053	22.1	85	0.00
16	Phenanthrene	1.327	1.275	3.9	107	0.00
17	Anthracene	1.220	1.200	1.6	107	0.00
18	Fluoranthene	1.384	1.356	2.0	107	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
20	Pvrene	1.657	1.651	0.4	108	0.00
21 S	Terphenyl-d14	0.663	0.651	1.8	105	0.00
22	Benzo(a)anthracene	1.449	1.408	2.8	107	0.00
23	Chrysene	1.526	1.496	2.0	106	0.00
24	Indeno(1,2,3-cd)pyrene	1.557	1.541	1.0	107	0.00
25 I	Perylene-d12	1.000	1.000	0.0	106	0.00
26	Benzo(b)fluoranthene	1.607	1.600	0.4	106	0.00
27	Benzo(k)fluoranthene	1.542	1.567	-1.6	108	0.00
28 C	Benzo(a)pyrene	1.374	1.375	-0.1	107	0.00
29	Dibenzo(a,h)anthracene	1.359	1.365	-0.4	107	0.00
30	Benzo(g,h,i)perylene	1.472	1.454	1.2	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0