

Data Path : Z:\HPCHEM1\BNA M\Data\BM041815\
 Data File : BM001055.D
 Acq On : 18 Apr 2015 11:00
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 20 08:25:04 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 08:05:22 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.564	0.461	18.3	83	0.00
3	n-Nitrosodimethylamine	0.528	0.499	5.5	99	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
5 S	Nitrobenzene-d5	0.261	0.265	-1.5	105	0.00
6	Naphthalene	1.194	1.176	1.5	104	0.00
7	2-Methylnaphthalene	0.770	0.760	1.3	103	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
9 S	2-Fluorobiphenyl	1.651	1.623	1.7	102	0.00
10	Acenaphthylene	2.209	2.190	0.9	102	0.00
11	Acenaphthene	1.546	1.465	5.2	102	0.00
12	Fluorene	1.781	1.754	1.5	101	-0.02
13 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
14	Hexachlorobenzene	0.276	0.268	2.9	103	0.00
15	Pentachlorophenol	0.068	0.048	29.4#	73	0.00
16	Phenanthrene	1.327	1.284	3.2	102	0.00
17	Anthracene	1.220	1.200	1.6	101	0.00
18	Fluoranthene	1.384	1.336	3.5	100	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
20	Pvrene	1.657	1.646	0.7	102	0.00
21 S	Terphenyl-d14	0.663	0.663	0.0	101	0.00
22	Benzo(a)anthracene	1.449	1.422	1.9	102	0.00
23	Chrysene	1.526	1.535	-0.6	103	0.00
24	Indeno(1,2,3-cd)pyrene	1.557	1.550	0.4	102	0.00
25 I	Perylene-d12	1.000	1.000	0.0	102	0.00
26	Benzo(b)fluoranthene	1.607	1.548	3.7	99	0.00
27	Benzo(k)fluoranthene	1.542	1.588	-3.0	105	0.00
28 C	Benzo(a)pyrene	1.374	1.362	0.9	101	0.00
29	Dibenzo(a,h)anthracene	1.359	1.353	0.4	102	0.00
30	Benzo(g,h,i)perylene	1.472	1.451	1.4	102	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0