

Data Path : Z:\HPCHEM1\BNA M\Data\BM041815\
 Data File : BM001057.D
 Acq On : 18 Apr 2015 13:24
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 20 08:26:36 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 08:05:22 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.564	0.453	19.7	76	0.00
3	n-Nitrosodimethylamine	0.528	0.508	3.8	94	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
5 S	Nitrobenzene-d5	0.261	0.260	0.4	96	0.00
6	Naphthalene	1.194	1.181	1.1	97	0.00
7	2-Methylnaphthalene	0.770	0.774	-0.5	97	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
9 S	2-Fluorobiphenyl	1.651	1.613	2.3	96	0.00
10	Acenaphthylene	2.209	2.177	1.4	96	0.00
11	Acenaphthene	1.546	1.472	4.8	97	0.00
12	Fluorene	1.781	1.808	-1.5	99	-0.02
13 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
14	Hexachlorobenzene	0.276	0.271	1.8	101	0.00
15	Pentachlorophenol	0.068	0.045	33.8#	66	0.00
16	Phenanthrene	1.327	1.282	3.4	99	0.00
17	Anthracene	1.220	1.191	2.4	97	0.00
18	Fluoranthene	1.384	1.351	2.4	98	0.00
19 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
20	Pvrene	1.657	1.645	0.7	99	0.00
21 S	Terphenyl-d14	0.663	0.667	-0.6	99	0.00
22	Benzo(a)anthracene	1.449	1.380	4.8	97	0.00
23	Chrysene	1.526	1.520	0.4	100	0.00
24	Indeno(1,2,3-cd)pyrene	1.557	1.535	1.4	98	0.00
25 I	Perylene-d12	1.000	1.000	0.0	99	0.00
26	Benzo(b)fluoranthene	1.607	1.565	2.6	97	0.00
27	Benzo(k)fluoranthene	1.542	1.535	0.5	99	0.00
28 C	Benzo(a)pyrene	1.374	1.353	1.5	98	0.00
29	Dibenzo(a,h)anthracene	1.359	1.357	0.1	99	0.00
30	Benzo(g,h,i)perylene	1.472	1.449	1.6	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0