

Data Path : Z:\HPCHEM1\BNA M\Data\BM041815\
 Data File : BM001057.D
 Acq On : 18 Apr 2015 13:24
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 20 08:26:36 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 08:05:22 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	96	0.00
2	1,4-Dioxane	1.000	0.804	19.6	76	0.00
3	n-Nitrosodimethylamine	1.000	0.963	3.7	94	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	97	0.00
5 S	Nitrobenzene-d5	1.000	0.997	0.3	96	0.00
6	Naphthalene	1.000	0.990	1.0	97	0.00
7	2-Methylnaphthalene	1.000	1.006	-0.6	97	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	98	0.00
9 S	2-Fluorobiphenyl	1.000	0.977	2.3	96	0.00
10	Acenaphthylene	1.000	0.986	1.4	96	0.00
11	Acenaphthene	1.000	0.952	4.8	97	0.00
12	Fluorene	1.000	1.015	-1.5	99	-0.02
13 I	Phenanthrene-d10	5.000	5.000	0.0	101	0.00
14	Hexachlorobenzene	1.000	0.980	2.0	101	0.00
15	Pentachlorophenol	1.000	0.722	27.8#	66	0.00
16	Phenanthrene	1.000	0.966	3.4	99	0.00
17	Anthracene	1.000	0.976	2.4	97	0.00
18	Fluoranthene	1.000	0.976	2.4	98	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	99	0.00
20	Pvrene	1.000	0.993	0.7	99	0.00
21 S	Terphenyl-d14	1.000	1.006	-0.6	99	0.00
22	Benzo(a)anthracene	1.000	0.953	4.7	97	0.00
23	Chrysene	1.000	0.997	0.3	100	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	0.985	1.5	98	0.00
25 I	Perylene-d12	5.000	5.000	0.0	99	0.00
26	Benzo(b)fluoranthene	1.000	0.974	2.6	97	0.00
27	Benzo(k)fluoranthene	1.000	0.996	0.4	99	0.00
28 C	Benzo(a)pyrene	1.000	0.985	1.5	98	0.00
29	Dibenzo(a,h)anthracene	1.000	0.998	0.2	99	0.00
30	Benzo(g,h,i)perylene	1.000	0.984	1.6	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0