

Data Path : Z:\HPCHEM1\BNA M\Data\BM041815\
 Data File : BM001065.D
 Acq On : 18 Apr 2015 18:11
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Apr 20 08:21:30 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 08:05:22 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	121	0.00
2	1,4-Dioxane	1.000	0.907	9.3	108	0.00
3	n-Nitrosodimethylamine	1.000	0.937	6.3	114	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	122	0.00
5 S	Nitrobenzene-d5	1.000	1.127	-12.7	137	0.00
6	Naphthalene	1.000	0.988	1.2	123	0.00
7	2-Methylnaphthalene	1.000	0.984	1.6	121	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	117	0.00
9 S	2-Fluorobiphenyl	1.000	0.977	2.3	115	0.00
10	Acenaphthylene	1.000	1.067	-6.7	123	0.00
11	Acenaphthene	1.000	0.971	2.9	117	0.00
12	Fluorene	1.000	1.002	-0.2	116	0.00
13 I	Phenanthrene-d10	5.000	5.000	0.0	117	0.00
14	Hexachlorobenzene	1.000	0.987	1.3	118	0.00
15	Pentachlorophenol	1.000	0.914	8.6	111	0.00
16	Phenanthrene	1.000	0.987	1.3	118	0.00
17	Anthracene	1.000	1.020	-2.0	118	0.00
18	Fluoranthene	1.000	1.039	-3.9	121	0.00
19 I	Chrysene-d12	5.000	5.000	0.0	114	0.00
20	Pvrene	1.000	1.069	-6.9	123	0.00
21 S	Terphenyl-d14	1.000	1.027	-2.7	116	0.00
22	Benzo(a)anthracene	1.000	1.106	-10.6	129	0.00
23	Chrysene	1.000	0.939	6.1	108	0.00
24	Indeno(1,2,3-cd)pyrene	1.000	1.056	-5.6	121	0.01
25 I	Perylene-d12	5.000	5.000	0.0	118	0.00
26	Benzo(b)fluoranthene	1.000	0.972	2.8	115	0.00
27	Benzo(k)fluoranthene	1.000	1.052	-5.2	124	0.01
28 C	Benzo(a)pyrene	1.000	1.066	-6.6	125	0.01
29	Dibenzo(a,h)anthracene	1.000	1.012	-1.2	119	0.01
30	Benzo(g,h,i)perylene	1.000	0.992	0.8	118	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0