

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041815\  
 Data File : BM001044.D  
 Acq On : 17 Apr 2015 23:55  
 Operator : TP/IZ  
 Sample : SSTDICC0.8  
 Misc :  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC0.8

Quant Time: Apr 18 05:59:24 2015  
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Sat Apr 18 05:45:49 2015  
 Response via : Initial Calibration

| Internal Standards        | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|---------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4 | 7.58  | 152  | 15932    | 5.00 | ng    | 0.00     |
| 4) Naphthalene-d8         | 10.36 | 136  | 56894    | 5.00 | ng    | 0.00     |
| 8) Acenaphthene-d10       | 14.24 | 164  | 29889    | 5.00 | ng    | 0.00     |
| 13) Phenanthrene-d10      | 16.99 | 188  | 64388    | 5.00 | ng    | 0.00     |
| 19) Chrysene-d12          | 21.18 | 240  | 55030    | 5.00 | ng    | -0.01    |
| 25) Perylene-d12          | 23.35 | 264  | 49425    | 5.00 | ng    | -0.01    |

|                             |       |     |      |      |    |      |
|-----------------------------|-------|-----|------|------|----|------|
| System Monitoring Compounds |       |     |      |      |    |      |
| 5) Nitrobenzene-d5          | 8.74  | 82  | 2476 | 0.87 | ng | 0.00 |
| 9) 2-Fluorobiphenyl         | 12.86 | 172 | 8195 | 0.88 | ng | 0.00 |
| 21) Terphenyl-d14           | 19.63 | 244 | 6076 | 0.91 | ng | 0.00 |


|                            |       |     |       |        |    |       |
|----------------------------|-------|-----|-------|--------|----|-------|
| Target Compounds           |       |     |       | Ovalue |    |       |
| 2) 1,4-Dioxane             | 3.08  | 88  | 1433  | 1.06   | ng | # 100 |
| 3) n-Nitrosodimethylamine  | 3.41  | 42  | 1447  | 0.97   | ng | 96    |
| 6) Naphthalene             | 10.41 | 128 | 11253 | 0.85   | ng | 100   |
| 7) 2-Methylnaphthalene     | 12.04 | 142 | 7290  | 0.87   | ng | 98    |
| 10) Acenaphthylene         | 13.95 | 152 | 10730 | 0.81   | ng | 99    |
| 11) Acenaphthene           | 14.29 | 154 | 7434  | 0.81   | ng | 100   |
| 12) Fluorene               | 15.30 | 166 | 8791  | 0.85   | ng | 100   |
| 14) Hexachlorobenzene      | 16.31 | 284 | 2989  | 0.80   | ng | # 100 |
| 15) Pentachlorophenol      | 16.64 | 266 | 638   | 0.68   | ng | 94    |
| 16) Phenanthrene           | 17.03 | 178 | 14182 | 0.81   | ng | 100   |
| 17) Anthracene             | 17.11 | 178 | 12880 | 0.86   | ng | 99    |
| 18) Fluoranthene           | 19.05 | 202 | 14524 | 0.83   | ng | 99    |
| 20) Pyrene                 | 19.42 | 202 | 14993 | 0.84   | ng | 100   |
| 22) Benzo(a)anthracene     | 21.17 | 228 | 12816 | 0.81   | ng | 99    |
| 23) Chrysene               | 21.22 | 228 | 14083 | 0.86   | ng | 100   |
| 24) Indeno(1,2,3-cd)pyrene | 25.52 | 276 | 14235 | 0.83   | ng | 100   |
| 26) Benzo(b)fluoranthene   | 22.70 | 252 | 13261 | 0.85   | ng | 99    |
| 27) Benzo(k)fluoranthene   | 22.74 | 252 | 12676 | 0.85   | ng | 100   |
| 28) Benzo(a)pyrene         | 23.25 | 252 | 11212 | 0.84   | ng | 99    |
| 29) Dibenzo(a,h)anthracene | 25.53 | 278 | 11260 | 0.85   | ng | 99    |
| 30) Benzo(g,h,i)perylene   | 26.17 | 276 | 12102 | 0.85   | ng | 99    |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM041815\  
Data File : BM001044.D  
Acq On : 17 Apr 2015 23:55  
Operator : TP/IZ  
Sample : SSTDIC0.8  
Misc :  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC0.8

Quant Time: Apr 18 05:59:24 2015  
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat Apr 18 05:45:49 2015  
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.58 min Scan# 631

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55

Instrument :

BNA\_M

ClientSampleId :

SSTDIC0.8


Tot Ion:152 Resp: 15932

Ion Ratio Lower Upper

152 100

150 153.7 123.3 184.9

115 55.2 45.0 67.6




Abundance Ion 152.00 (151.70 to 152.70): E


Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E

7.58



Time-->



#2

1,4-Dioxane

Concen: 1.06 ng

RT: 3.08 min Scan# 39

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55


Tgt Ion: 88 Resp: 1433

Ion Ratio Lower Upper

88 100

58 72.9 0.0 0.0#

43 30.9 0.0 0.0#




Abundance Ion 88.00 (87.70 to 88.70): BM0


Ion 58.00 (57.70 to 58.70): BM0

Ion 43.00 (42.70 to 43.70): BM0

3.08



Time-->



#3

n-Nitrosodimethylamine

Concen: 0.97 ng

RT: 3.41 min Scan# 70

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55

Instrument :

BNA\_M

ClientSampleId :

SSTDIC0.8


Tot Ion: 42 Resp: 1447

Ion Ratio Lower Upper

42 100

74 110.9 92.1 138.1




44 7.0 5.8 8.6



Abundance Ion 42.00 (41.70 to 42.70): BM001044.D (-66) (-)

Ion 74.00 (73.70 to 74.70): BM001044.D (-66) (-)

Ion 44.00 (43.70 to 44.70): BM001044.D (-66) (-)



#4

Naphthalene-d8

Concen: 5.00 ng

RT: 10.36 min Scan# 1074

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55


Tgt Ion: 136 Resp: 56894

Ion Ratio Lower Upper

136 100

137 10.9 8.6 13.0



54 8.1 6.5 9.7




Abundance Ion 136.00 (135.70 to 136.70): BM001044.D (-1066) (-)

Ion 137.00 (136.70 to 137.70): BM001044.D (-1066) (-)

Ion 54.00 (53.70 to 54.70): BM001044.D (-1066) (-)





#5

Nitrobenzene-d5

Concen: 0.87 ng

RT: 8.74 min Scan# 847

Delta R.T. 0.01 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55

Instrument :

BNA\_M

ClientSampleId :

SSTDIC0.8


Tot Ion: 82 Resp: 2476

Ion Ratio Lower Upper

82 100

128 42.7 31.9 47.9




54 60.9 51.3 76.9



Abundance Ion 82.00 (81.70 to 82.70): BM001044.D (-837) (-)

Ion 128.00 (127.70 to 128.70): BM001044.D (-837) (-)

Ion 54.00 (53.70 to 54.70): BM001044.D (-837) (-)



#6

Naphthalene

Concen: 0.85 ng

RT: 10.41 min Scan# 1078

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55


Tgt Ion: 128 Resp: 11253

Ion Ratio Lower Upper

128 100

129 11.9 9.3 13.9



127 12.6 10.0 15.0



Abundance Ion 128.00 (127.70 to 128.70): BM001044.D (-1071) (-)

Ion 129.00 (128.70 to 129.70): BM001044.D (-1071) (-)

Ion 127.00 (126.70 to 127.70): BM001044.D (-1071) (-)





#7

2-Methylnaphthalene

Concen: 0.87 ng

RT: 12.04 min Scan# 1224

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55

Instrument :

BNA\_M

ClientSampleId :

SSTDIC0.8


Tot Ion:142 Resp: 7290

Ion Ratio Lower Upper

142 100

141 84.9 66.7 100.1




115 32.9 25.5 38.3



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#8

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.24 min Scan# 1408

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55


Tgt Ion:164 Resp: 29889

Ion Ratio Lower Upper

164 100

162 97.3 78.5 117.7



160 40.4 32.9 49.3



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E





#9

2-Fluorobiphenyl

Concen: 0.88 ng

RT: 12.86 min Scan# 1303

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55

Instrument :

BNA\_M

ClientSampleId :

SSTDIC0.8


Tot Ion:172 Resp: 8195

Ion Ratio Lower Upper

172 100

171 32.6 25.6 38.4




170 20.4 15.8 23.6



Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#10

Acenaphthylene

Concen: 0.81 ng

RT: 13.95 min Scan# 1391

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55


Tgt Ion:152 Resp: 10730

Ion Ratio Lower Upper

152 100

151 18.6 14.7 22.1



153 13.0 10.2 15.2




Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E





#11

Acenaphthene

Concen: 0.81 ng

RT: 14.29 min Scan# 1411

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55

Instrument :

BNA\_M

ClientSampleId :

SSTDIC0.8


Tot Ion:154 Resp: 7434

Ion Ratio Lower Upper




154 100

153 110.1 88.1 132.1

152 53.6 42.3 63.5



Abundance Ion 154.00 (153.70 to 154.70): E  
Ion 153.00 (152.70 to 153.70): E  
Ion 152.00 (151.70 to 152.70): E



#12

Fluorene

Concen: 0.85 ng

RT: 15.30 min Scan# 1470

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55


Tgt Ion:166 Resp: 8791

Ion Ratio Lower Upper



166 100


165 96.6 77.6 116.4

167 13.4 10.6 16.0



Abundance Ion 166.00 (165.70 to 166.70): E  
Ion 165.00 (164.70 to 165.70): E  
Ion 167.00 (166.70 to 167.70): E









#13  
Phenanthrene-d10  
Concen: 5.00 ng  
RT: 16.99 min Scan# 1558  
Delta R.T. 0.00 min  
Lab File: BM001044.D  
Acq: 17 Apr 2015 23:55




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC0.8

Tot Ion:188 Resp: 64388  
Ion Ratio Lower Upper  
188 100  
94 4.3 3.6 5.4  
80 3.8 3.0 4.6



#14  
Hexachlorobenzene  
Concen: 0.80 ng  
RT: 16.31 min Scan# 1525  
Delta R.T. 0.00 min  
Lab File: BM001044.D  
Acq: 17 Apr 2015 23:55

Tgt Ion:284 Resp: 2989  
Ion Ratio Lower Upper  
284 100  
142 0.0 0.0 0.0  
249 0.0 0.0 0.0







#15  
 Pentachlorophenol  
 Concen: 0.68 ng  
 RT: 16.64 min Scan# 1541  
 Delta R.T. 0.00 min  
 Lab File: BM001044.D  
 Acq: 17 Apr 2015 23:55

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SSTDICC0.8



Tot Ion:266 Resp: 638  
 Ion Ratio Lower Upper  
 266 100  
 268 62.6 46.3 69.5  
 264 79.0 59.8 89.6



Abundance Ion 265.70 (265.40 to 266.40): E  
 Ion 268.00 (267.70 to 268.70): E  
 Ion 264.00 (263.70 to 264.70): E




Time-->




#16  
 Phenanthrene  
 Concen: 0.81 ng  
 RT: 17.03 min Scan# 1560  
 Delta R.T. 0.00 min  
 Lab File: BM001044.D  
 Acq: 17 Apr 2015 23:55


Tgt Ion:178 Resp: 14182  
 Ion Ratio Lower Upper  
 178 100  
 176 18.2 14.6 22.0  
 179 16.1 12.6 19.0




Abundance Ion 178.00 (177.70 to 178.70): E  
 Ion 176.00 (175.70 to 176.70): E  
 Ion 179.00 (178.70 to 179.70): E



Time-->





#17

Anthracene

Concen: 0.86 ng

RT: 17.11 min Scan# 1564

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55

Instrument :

BNA\_M

ClientSampleId :

SSTDIC0.8


Tot Ion:178 Resp: 12880

Ion Ratio Lower Upper

178 100

176 18.1 14.3 21.5




179 15.2 11.8 17.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#18

Fluoranthene

Concen: 0.83 ng

RT: 19.05 min Scan# 1689

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55


Tgt Ion:202 Resp: 14524

Ion Ratio Lower Upper

202 100

101 12.6 10.1 15.1



203 16.9 13.8 20.8




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#19

Chrysene-d12

Concen: 5.00 ng

RT: 21.18 min Scan# 1873

Delta R.T. -0.01 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55

Instrument :

BNA\_M

ClientSampleId :

SSTDIC0.8





Tot Ion:240 Resp: 55030

Ion Ratio Lower Upper

240 100

120 19.4 4.6 6.8#

236 31.0 19.1 28.7#



#20

Pyrene

Concen: 0.84 ng

RT: 19.42 min Scan# 1719

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55




Tgt Ion:202 Resp: 14993


Ion Ratio Lower Upper

202 100

200 20.4 16.3 24.5

203 17.5 13.8 20.6





#21

Terphenyl-d14

Concen: 0.91 ng

RT: 19.63 min Scan# 1737

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55

Instrument :

BNA\_M

ClientSampleId :

SSTDIC0.8


Tot Ion:244 Resp: 6076

Ion Ratio Lower Upper

244 100

212 9.3 7.1 10.7




122 8.6 5.9 8.9



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#22

Benzo(a)anthracene

Concen: 0.81 ng

RT: 21.17 min Scan# 1872

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55


Tgt Ion:228 Resp: 12816

Ion Ratio Lower Upper

228 100

226 24.0 19.0 28.6



229 21.3 16.8 25.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#23

Chrysene

Concen: 0.86 ng

RT: 21.22 min Scan# 1877

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55

Instrument :

BNA\_M

ClientSampleId :

SSTDIC0.8





Tot Ion:228 Resp: 14083

Ion Ratio Lower Upper

228 100

226 25.5 20.6 31.0

229 21.5 17.2 25.8



#24

Indeno(1,2,3-cd)pyrene

Concen: 0.83 ng

RT: 25.52 min Scan# 2289

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55




Tgt Ion:276 Resp: 14235


Ion Ratio Lower Upper

276 100

138 32.7 26.3 39.5

277 24.6 19.5 29.3









#25  
Pervlene-d12  
Concen: 5.00 ng  
RT: 23.35 min Scan# 2081  
Delta R.T. -0.01 min  
Lab File: BM001044.D  
Acq: 17 Apr 2015 23:55




Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC0.8

Tot Ion:264 Resp: 49425  
Ion Ratio Lower Upper  
264 100  
260 26.4 19.0 28.6  
265 20.8 17.9 26.9



#26  
Benzo(b)fluoranthene  
Concen: 0.85 ng  
RT: 22.70 min Scan# 2019  
Delta R.T. 0.00 min  
Lab File: BM001044.D  
Acq: 17 Apr 2015 23:55

Tgt Ion:252 Resp: 13261  
Ion Ratio Lower Upper  
252 100  
253 23.3 18.2 27.2  
125 13.3 10.9 16.3





#27

Benzo(k)fluoranthene

Concen: 0.85 ng

RT: 22.74 min Scan# 2023

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55

Instrument :

BNA\_M

ClientSampleId :

SSTDIC0.8


Tot Ion:252 Resp: 12676

Ion Ratio Lower Upper

252 100

253 23.0 18.3 27.5




125 13.4 10.6 15.8



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#28

Benzo(a)pyrene

Concen: 0.84 ng

RT: 23.25 min Scan# 2072

Delta R.T. 0.00 min

Lab File: BM001044.D

Acq: 17 Apr 2015 23:55


Tgt Ion:252 Resp: 11212

Ion Ratio Lower Upper

252 100

253 23.8 18.6 27.8



125 15.2 11.9 17.9




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#29  
Dibenzo(a,h)anthracene  
Concen: 0.85 ng  
RT: 25.53 min Scan# 2290  
Delta R.T. 0.00 min  
Lab File: BM001044.D  
Acq: 17 Apr 2015 23:55

Instrument :  
BNA\_M  
ClientSampleId :  
SSTDIC0.8

Tot Ion:278 Resp: 11260  
Ion Ratio Lower Upper  
278 100  
139 23.1 18.2 27.2  
279 24.5 19.1 28.7



#30  
Benzo(g,h,i)perylene  
Concen: 0.85 ng  
RT: 26.17 min Scan# 2352  
Delta R.T. 0.00 min  
Lab File: BM001044.D  
Acq: 17 Apr 2015 23:55

Tgt Ion:276 Resp: 12102  
Ion Ratio Lower Upper  
276 100  
277 24.0 19.0 28.4  
138 29.7 23.4 35.2

