

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041823\
 Data File : BM039452.D
 Acq On : 18 Apr 2023 11:22
 Operator : CG/JU
 Sample : SST01046
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST010007

Quant Time: Apr 18 23:16:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 18 18:09:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	223198	20.000	ng/u1	0.00
20) Naphthalene-d8	10.645	136	1027865	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.492	164	679558	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.245	188	1450746	20.000	ng/u1	0.00
79) Chrysene-d12	21.439	240	1238929	20.000	ng/u1	0.00
88) Perylene-d12	23.809	264	1152655	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.216	96	22042	3.743	ng/uL	0.00
4) Pyridine-d5	3.652	84	144266	8.830	ng/u1	0.00
7) Phenol-d5	7.028	99	177880	8.708	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.169	67	125440	9.290	ng/u1	0.00
11) 2-Chlorophenol-d4	7.369	132	149176	9.309	ng/u1	0.00
15) 4-Methylphenol-d8	8.569	113	142962	8.688	ng/u1	0.00
21) Nitrobenzene-d5	9.004	128	75622	9.145	ng/u1	0.00
24) 2-Nitrophenol-d4	9.728	143	85292	9.185	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.281	165	152454	9.542	ng/u1	0.00
31) 4-Chloroaniline-d4	10.798	131	218399	9.152	ng/u1	0.00
46) Dimethylphthalate-d6	13.904	166	539062	9.836	ng/u1	0.00
49) Acenaphthylene-d8	14.186	160	592359	9.621	ng/u1	0.00
54) 4-Nitrophenol-d4	14.739	143	54585	6.581	ng/u1	0.00
60) Fluorene-d10	15.486	176	444795	9.906	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.621	200	78845	8.178	ng/u1	0.00
73) Anthracene-d10	17.345	188	682154	9.777	ng/u1	0.00
81) Pyrene-d10	19.639	212	782115	9.500	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.650	264	599808	9.638	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.257	88	24254	3.812	ng/uL	93
5) Pyridine	3.669	79	157482	9.141	ng/u1	97
6) Benzaldehyde	6.975	77	94611	10.840	ng/u1	96
8) Phenol	7.051	94	185844	8.738	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.263	93	165824	9.351	ng/u1	100
12) 2-Chlorophenol	7.404	128	153987	9.550	ng/u1	98
13) 2-Methylphenol	8.298	108	144855	8.768	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.375	45	244392	9.921	ng/u1	99
16) Acetophenone	8.669	105	260631	9.643	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.645	70	145948	9.756	ng/u1	96
18) 4-Methylphenol	8.634	108	162831	9.130	ng/u1	99
19) Hexachloroethane	8.910	117	71628	9.780	ng/u1	96
22) Nitrobenzene	9.051	77	193095	9.457	ng/u1	99
23) Isophorone	9.575	82	420106	9.649	ng/u1	99
25) 2-Nitrophenol	9.763	139	89344	9.222	ng/u1	96
26) 2,4-Dimethylphenol	9.833	107	189754	9.584	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.063	93	251581	9.826	ng/u1	100
29) 2,4-Dichlorophenol	10.304	162	148368	9.427	ng/u1	98
30) Naphthalene	10.692	128	556126	9.779	ng/u1	100
32) 4-Chloroaniline	10.822	127	214289	9.114	ng/u1	100
33) Hexachlorobutadiene	10.969	225	101878	9.824	ng/u1	99
34) Caprolactam	11.610	113	52211	8.985	ng/u1	87
35) 4-Chloro-3-methylphenol	11.963	107	170452	9.276	ng/u1	99
36) 2-Methylnaphthalene	12.310	142	370794	9.726	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041823\
 Data File : BM039452.D
 Acq On : 18 Apr 2023 11:22
 Operator : CG/JU
 Sample : SSTD01046
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD010007


Quant Time: Apr 18 23:16:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 18 18:09:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.527	142	376295	9.822	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.680	216	192683	9.446	ng/ul	98
40) Hexachlorocyclopentadiene	12.651	237	84956	6.991	ng/ul	97
41) 2,4,6-Trichlorophenol	12.933	196	125763	9.229	ng/ul	98
42) 2,4,5-Trichlorophenol	13.016	196	129427	9.138	ng/ul	98
43) 1,1'-Biphenyl	13.321	154	504149	9.600	ng/ul	100
44) 2-Chloronaphthalene	13.369	162	399784	9.704	ng/ul	99
45) 2-Nitroaniline	13.586	65	108207	8.957	ng/ul	96
47) Dimethylphthalate	13.951	163	532703	9.817	ng/ul	99
48) 2,6-Dinitrotoluene	14.080	165	102517	9.341	ng/ul	97
50) Acenaphthylene	14.216	152	648404	9.845	ng/ul	99
51) 3-Nitroaniline	14.421	138	82818	8.051	ng/ul	99
52) Acenaphthene	14.557	153	443825	9.788	ng/ul	99
53) 2,4-Dinitrophenol	14.639	184	31718	5.430	ng/ul	97
55) 4-Nitrophenol	14.751	109	39093	6.205	ng/ul	91
56) Dibenzofuran	14.892	168	614748	9.974	ng/ul	99
57) 2,4-Dinitrotoluene	14.874	165	143481	9.409	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.127	232	119960	9.463	ng/ul	99
59) Diethylphthalate	15.315	149	558154	9.923	ng/ul	99
61) Fluorene	15.539	166	507504	9.954	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.533	204	250178	9.900	ng/ul	99
63) 4-Nitroaniline	15.586	138	64712	8.364	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.633	198	77411	8.352	ng/ul#	89
67) N-Nitrosodiphenylamine	15.757	169	425423	9.842	ng/ul	99
68) 4-Bromophenyl-phenylether	16.433	248	149992	9.573	ng/ul	98
69) Hexachlorobenzene	16.545	284	174699	9.705	ng/ul	97
70) Atrazine	16.709	200	156873	9.781	ng/ul	98
71) Pentachlorophenol	16.904	266	87165	8.283	ng/ul	99
72) Phenanthrene	17.286	178	790826	9.819	ng/ul	99
74) Anthracene	17.380	178	808868	10.011	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.286	216	203468	9.417	ng/uL	99
76) Pentachlorobenzene	14.810	250	208796	9.818	ng/uL	99
77) Carbazole	17.657	167	680176	9.473	ng/ul	100
78) Di-n-butylphthalate	18.215	149	984234	9.692	ng/ul	100
80) Fluoranthene	19.309	202	911242	9.443	ng/ul	96
82) Pyrene	19.668	202	962785	9.638	ng/ul	98
83) Butylbenzylphthalate	20.568	149	437776	9.284	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.362	252	267794	9.747	ng/ul	97
85) Benzo(a)anthracene	21.421	228	867591	9.682	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.344	149	661820	9.438	ng/ul	99
87) Chrysene	21.474	228	835024	9.769	ng/ul	100
89) Di-n-octyl phthalate	22.262	149	1097977	9.481	ng/ul	100
90) Benzo(b)fluoranthene	23.086	252	776309	9.572	ng/ul	99
91) Benzo(k)fluoranthene	23.133	252	768646	9.650	ng/ul	99
93) Benzo(a)pyrene	23.703	252	663263	9.702	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.262	276	701807	9.333	ng/ul	100
95) Dibenzo(a,h)anthracene	26.279	278	582648	9.345	ng/ul	99
96) Benzo(g,h,i)perylene	27.021	276	566947	9.465	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 18 23:16:45 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 18 18:09:04 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.812 ng/uL
RT: 3.257 min Scan# 11
Delta R.T. 0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion: 88 Resp: 24254
Ion Ratio Lower Upper
88 100
43 33.9 24.6 36.8
58 76.9 56.7 85.1



#5
Pyridine
Concen: 9.141 ng/uL
RT: 3.669 min Scan# 82
Delta R.T. 0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion: 79 Resp: 157482
Ion Ratio Lower Upper
79 100
52 74.9 61.9 92.9
51 38.8 29.7 44.5









#6
Benzaldehyde
Concen: 10.840 ng/ul
RT: 6.975 min Scan# 644
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion: 77 Resp: 94611
Ion Ratio Lower Upper
77 100
105 90.7 76.3 114.5
106 90.9 75.4 113.0



#8
Phenol
Concen: 8.738 ng/ul
RT: 7.051 min Scan# 657
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion: 94 Resp: 185844
Ion Ratio Lower Upper
94 100
65 27.5 22.2 33.4
66 36.2 29.6 44.4





#10

Bis(2-Chloroethyl)ether

Concen: 9.351 ng/ul

RT: 7.263 min Scan# 693

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion: 93 Resp: 165824

Ion Ratio Lower Upper

93 100

63 77.7 61.9 92.9

95 32.2 25.8 38.8



#12

2-Chlorophenol

Concen: 9.550 ng/ul

RT: 7.404 min Scan# 717

Delta R.T. 0.006 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22




Tgt Ion:128 Resp: 153987


Ion Ratio Lower Upper

128 100

64 48.8 38.8 58.2

130 33.5 25.4 38.0





#13

2-Methylphenol

Concen: 8.768 ng/ul

RT: 8.298 min Scan# 869

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :





SSTD010007

Tgt Ion:108 Resp: 144855

Ion Ratio Lower Upper

108 100

107 86.1 71.7 107.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 9.921 ng/ul

RT: 8.375 min Scan# 882

Delta R.T. 0.006 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22




Tgt Ion: 45 Resp: 244392


Ion Ratio Lower Upper

45 100

77 15.9 12.3 18.5

79 12.6 9.8 14.6









#16
Acetophenone
Concen: 9.643 ng/ul
RT: 8.669 min Scan# 91
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:105 Resp: 260631
Ion Ratio Lower Upper
105 100
77 79.6 64.1 96.1
51 29.1 22.2 33.4



#17
N-Nitroso-di-n-propylamine
Concen: 9.756 ng/ul
RT: 8.645 min Scan# 928
Delta R.T. -0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion: 70 Resp: 145948
Ion Ratio Lower Upper
70 100
42 57.8 43.0 64.4
101 9.3 7.4 11.2
130 20.1 16.2 24.2









#18
4-Methylphenol
Concen: 9.130 ng/ul
RT: 8.634 min Scan# 912
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:108 Resp: 162831
Ion Ratio Lower Upper
108 100
107 110.2 88.6 133.0



#19
Hexachloroethane
Concen: 9.780 ng/ul
RT: 8.910 min Scan# 973
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:117 Resp: 71628
Ion Ratio Lower Upper
117 100
201 92.4 78.3 117.5
199 58.7 47.4 71.0









#22
Nitrobenzene
Concen: 9.457 ng/ul
RT: 9.051 min Scan# 997
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion: 77 Resp: 193095
Ion Ratio Lower Upper
77 100
123 44.2 36.1 54.1
65 12.6 10.2 15.2



#23
Isophorone
Concen: 9.649 ng/ul
RT: 9.575 min Scan# 1086
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion: 82 Resp: 420106
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.2
138 16.8 13.2 19.8









#25
2-Nitrophenol
Concen: 9.222 ng/ul
RT: 9.763 min Scan# 1118
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:139 Resp: 89344
Ion Ratio Lower Upper
139 100
65 55.4 40.7 61.1
109 30.1 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 9.584 ng/ul
RT: 9.833 min Scan# 1130
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:107 Resp: 189754
Ion Ratio Lower Upper
107 100
121 48.3 39.1 58.7
122 85.1 68.1 102.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 9.826 ng/ul
 RT: 10.063 min Scan# 1169
 Delta R.T. -0.000 min
 Lab File: BM039452.D
 Acq: 18 Apr 2023 11:22




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010007


Tgt Ion: 93 Resp: 251581
 Ion Ratio Lower Upper
 93 100
 95 32.3 25.8 38.6
 123 10.7 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 9.427 ng/ul
 RT: 10.304 min Scan# 1210
 Delta R.T. -0.000 min
 Lab File: BM039452.D
 Acq: 18 Apr 2023 11:22

Tgt Ion: 162 Resp: 148368
 Ion Ratio Lower Upper
 162 100
 164 63.5 52.4 78.6
 98 37.7 29.4 44.0









#30
Naphthalene
Concen: 9.779 ng/ul
RT: 10.692 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:128 Resp: 556126
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 12.9 10.5 15.7



#32
4-Chloroaniline
Concen: 9.114 ng/ul
RT: 10.822 min Scan# 1298
Delta R.T. 0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:127 Resp: 214289
Ion Ratio Lower Upper
127 100
129 32.5 25.8 38.8









#33
Hexachlorobutadiene
Concen: 9.824 ng/ul
RT: 10.969 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:225 Resp: 101878
Ion Ratio Lower Upper
225 100
223 63.2 50.1 75.1
227 62.3 50.4 75.6



#34
Caprolactam
Concen: 8.985 ng/ul
RT: 11.610 min Scan# 1432
Delta R.T. 0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:113 Resp: 52211
Ion Ratio Lower Upper
113 100
55 158.3 145.5 218.3
56 120.2 103.9 155.9









#35
4-Chloro-3-methylphenol
Concen: 9.276 ng/ul
RT: 11.963 min Scan# 1492
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:107 Resp: 170452
Ion Ratio Lower Upper
107 100
144 25.4 21.0 31.4
142 77.4 62.9 94.3



#36
2-Methylnaphthalene
Concen: 9.726 ng/ul
RT: 12.310 min Scan# 1551
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:142 Resp: 370794
Ion Ratio Lower Upper
142 100
141 87.7 71.1 106.7





#37

1-Methylnaphthalene

Concen: 9.822 ng/ul

RT: 12.527 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion:142 Resp: 376295

Ion Ratio Lower Upper

142 100

141 91.4 72.0 108.0

116 3.9 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 9.446 ng/ul

RT: 12.680 min Scan# 1614

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Tgt Ion:216 Resp: 192683




Ion Ratio Lower Upper


216 100

214 79.7 62.2 93.2

179 20.3 16.5 24.7

108 18.8 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 6.991 ng/ul
RT: 12.651 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:237 Resp: 84956
Ion Ratio Lower Upper
237 100
235 61.7 50.9 76.3
272 14.2 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 9.229 ng/ul
RT: 12.933 min Scan# 1657
Delta R.T. 0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:196 Resp: 125763
Ion Ratio Lower Upper
196 100
198 98.5 76.7 115.1
200 29.0 23.8 35.6





#42

2,4,5-Trichlorophenol

Concen: 9.138 ng/ul

RT: 13.016 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM039452.D


Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007






Tgt Ion:196 Resp: 129427

Ion Ratio Lower Upper

196 100

198 97.0 75.7 113.5

200 30.5 24.6 36.8



#43

1,1'-Biphenyl

Concen: 9.600 ng/ul

RT: 13.321 min Scan# 1723

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22




Tgt Ion:154 Resp: 504149


Ion Ratio Lower Upper

154 100

153 40.0 31.9 47.9

76 14.0 11.5 17.3









#44
2-Chloronaphthalene
Concen: 9.704 ng/ul
RT: 13.369 min Scan# 1730
Delta R.T. 0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:162 Resp: 399784
Ion Ratio Lower Upper
162 100
164 32.5 26.1 39.1
127 37.9 31.4 47.2



#45
2-Nitroaniline
Concen: 8.957 ng/ul
RT: 13.586 min Scan# 1768
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion: 65 Resp: 108207
Ion Ratio Lower Upper
65 100
92 66.3 53.4 80.2
138 97.2 83.2 124.8









#47
Dimethylphthalate
Concen: 9.817 ng/ul
RT: 13.951 min Scan# 1830
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:163 Resp: 532703
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 9.8 8.2 12.2



#48
2,6-Dinitrotoluene
Concen: 9.341 ng/ul
RT: 14.080 min Scan# 1852
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:165 Resp: 102517
Ion Ratio Lower Upper
165 100
89 64.9 50.6 76.0
121 23.6 17.3 25.9









#50
Acenaphthylene
Concen: 9.845 ng/ul
RT: 14.216 min Scan# 1875
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:152 Resp: 648404
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.8
153 12.9 10.4 15.6



#51
3-Nitroaniline
Concen: 8.051 ng/ul
RT: 14.421 min Scan# 1910
Delta R.T. 0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:138 Resp: 82818
Ion Ratio Lower Upper
138 100
108 10.8 8.8 13.2
92 120.1 95.5 143.3





#52

Acenaphthene

Concen: 9.788 ng/ul

RT: 14.557 min Scan# 1933

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion:153 Resp: 443825

Ion Ratio Lower Upper

153 100

152 47.2 38.2 57.4

154 89.4 71.0 106.4



#53

2,4-Dinitrophenol

Concen: 5.430 ng/ul

RT: 14.639 min Scan# 1947

Delta R.T. 0.012 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22




Tgt Ion:184 Resp: 31718


Ion Ratio Lower Upper

184 100

63 77.1 60.0 90.0

154 67.2 51.8 77.6









#55
4-Nitrophenol
Concen: 6.205 ng/ul
RT: 14.751 min Scan# 1966
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:109 Resp: 39093
Ion Ratio Lower Upper
109 100
139 137.8 106.4 159.6
65 146.9 104.8 157.2



#56
Dibenzofuran
Concen: 9.974 ng/ul
RT: 14.892 min Scan# 1990
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:168 Resp: 614748
Ion Ratio Lower Upper
168 100
139 37.9 30.7 46.1





#57

2,4-Dinitrotoluene

Concen: 9.409 ng/ul

RT: 14.874 min Scan# 1986

Delta R.T. 0.006 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion:165 Resp: 143481

Ion Ratio Lower Upper

165 100

63 51.1 39.4 59.2

182 2.6 2.0 3.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 9.463 ng/ul

RT: 15.127 min Scan# 2030

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Tgt Ion:232 Resp: 119960




Ion Ratio Lower Upper


232 100

131 48.0 38.0 57.0

130 2.5 2.0 3.0

166 32.2 24.8 37.2





#59

Diethylphthalate

Concen: 9.923 ng/ul

RT: 15.315 min Scan# 2062

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion:149 Resp: 558154

Ion Ratio Lower Upper

149 100

177 21.2 17.3 25.9

150 11.9 9.8 14.6



#61

Fluorene

Concen: 9.954 ng/ul

RT: 15.539 min Scan# 2100

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22




Tgt Ion:166 Resp: 507504


Ion Ratio Lower Upper

166 100

165 97.9 78.2 117.2

167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 9.900 ng/ul
RT: 15.533 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:204 Resp: 250178
Ion Ratio Lower Upper
204 100
206 32.7 26.8 40.2
141 61.5 48.5 72.7



#63
4-Nitroaniline
Concen: 8.364 ng/ul
RT: 15.586 min Scan# 2108
Delta R.T. 0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:138 Resp: 64712
Ion Ratio Lower Upper
138 100
92 53.8 44.1 66.1
108 79.3 64.6 97.0









#66
4,6-Dinitro-2-methylphenol
Concen: 8.352 ng/ul
RT: 15.633 min Scan# 2116
Delta R.T. -0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:198 Resp: 77411
Ion Ratio Lower Upper
198 100
182 5.3 3.1 4.7#
77 32.0 20.6 31.0#



#67
N-Nitrosodiphenylamine
Concen: 9.842 ng/ul
RT: 15.757 min Scan# 2137
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:169 Resp: 425423
Ion Ratio Lower Upper
169 100
168 66.9 53.2 79.8
167 36.9 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 9.573 ng/ul
RT: 16.433 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:248 Resp: 149992
Ion Ratio Lower Upper
248 100
250 100.8 77.7 116.5
141 76.9 61.0 91.4



#69
Hexachlorobenzene
Concen: 9.705 ng/ul
RT: 16.545 min Scan# 2271
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:284 Resp: 174699
Ion Ratio Lower Upper
284 100
142 37.9 28.9 43.3
249 31.2 23.9 35.9









#70
Atrazine
Concen: 9.781 ng/ul
RT: 16.709 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:200 Resp: 156873
Ion Ratio Lower Upper
200 100
173 26.3 22.2 33.2
215 47.9 38.9 58.3



#71
Pentachlorophenol
Concen: 8.283 ng/ul
RT: 16.904 min Scan# 2332
Delta R.T. 0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:266 Resp: 87165
Ion Ratio Lower Upper
266 100
264 62.7 50.2 75.4
268 66.5 51.9 77.9





#72

Phenanthrene

Concen: 9.819 ng/ul

RT: 17.286 min Scan# 2397

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion:178 Resp: 790826

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.4

176 19.2 15.8 23.8



#74

Anthracene

Concen: 10.011 ng/ul

RT: 17.380 min Scan# 2413

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22




Tgt Ion:178 Resp: 808868


Ion Ratio Lower Upper

178 100

179 15.2 12.1 18.1

176 18.3 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 9.417 ng/uL

RT: 13.286 min Scan# 1717

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion:216 Resp: 203468

Ion Ratio Lower Upper

216 100

214 79.4 63.0 94.4

218 48.5 38.1 57.1



#76

Pentachlorobenzene

Concen: 9.818 ng/uL

RT: 14.810 min Scan# 1976

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22




Tgt Ion:250 Resp: 208796


Ion Ratio Lower Upper

250 100

248 62.3 50.4 75.6

252 62.6 50.2 75.2





#77

Carbazole

Concen: 9.473 ng/ul

RT: 17.657 min Scan# 2460

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion:167 Resp: 680176

Ion Ratio Lower Upper

167 100

166 21.2 17.0 25.4

139 13.5 10.6 15.8



#78

Di-n-butylphthalate

Concen: 9.692 ng/ul

RT: 18.215 min Scan# 2555

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22




Tgt Ion:149 Resp: 984234


Ion Ratio Lower Upper

149 100

150 8.9 7.0 10.6

104 4.9 4.1 6.1









#80
Fluoranthene
Concen: 9.443 ng/ul
RT: 19.309 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:202 Resp: 911242
Ion Ratio Lower Upper
202 100
101 11.4 10.4 15.6
100 8.4 7.7 11.5



#82
Pyrene
Concen: 9.638 ng/ul
RT: 19.668 min Scan# 2802
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:202 Resp: 962785
Ion Ratio Lower Upper
202 100
101 13.8 12.1 18.1
100 11.5 9.6 14.4









#83
 Butylbenzylphthalate
 Concen: 9.284 ng/ul
 RT: 20.568 min Scan# 2955
 Delta R.T. -0.000 min
 Lab File: BM039452.D
 Acq: 18 Apr 2023 11:22




Instrument :
 BNA_M
 ClientSampleId :
 SSTD010007


Tgt Ion:149 Resp: 437776
 Ion Ratio Lower Upper
 149 100
 91 76.0 59.4 89.0
 206 19.4 15.8 23.8



#84
 3,3'-Dichlorobenzidine
 Concen: 9.747 ng/ul
 RT: 21.362 min Scan# 3090
 Delta R.T. 0.006 min
 Lab File: BM039452.D
 Acq: 18 Apr 2023 11:22

Tgt Ion:252 Resp: 267794
 Ion Ratio Lower Upper
 252 100
 254 65.3 50.9 76.3
 126 11.3 10.4 15.6





#85

Benzo(a)anthracene

Concen: 9.682 ng/ul

RT: 21.421 min Scan# 3100

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion:228 Resp: 867591

Ion Ratio Lower Upper

228 100

229 19.2 15.4 23.2

226 26.6 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 9.438 ng/ul

RT: 21.344 min Scan# 3087

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22




Tgt Ion:149 Resp: 661820


Ion Ratio Lower Upper

149 100

167 27.9 21.8 32.6

279 4.0 3.1 4.7





#87

Chrysene

Concen: 9.769 ng/ul

RT: 21.474 min Scan# 3109

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion:228 Resp: 835024

Ion Ratio Lower Upper

228 100

226 29.8 23.9 35.9

229 19.9 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 9.481 ng/ul




RT: 22.262 min Scan# 3243


Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Tgt Ion:149 Resp: 1097977









#90
Benzo(b)fluoranthene
Concen: 9.572 ng/ul
RT: 23.086 min Scan# 3383
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22




Instrument :
BNA_M
ClientSampleId :
SSTD010007


Tgt Ion:252 Resp: 776309
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.2
125 11.1 8.7 13.1



#91
Benzo(k)fluoranthene
Concen: 9.650 ng/ul
RT: 23.133 min Scan# 3391
Delta R.T. -0.000 min
Lab File: BM039452.D
Acq: 18 Apr 2023 11:22

Tgt Ion:252 Resp: 768646
Ion Ratio Lower Upper
252 100
253 21.6 17.6 26.4
125 11.0 8.6 12.8





#93

Benzo(a)pyrene

Concen: 9.702 ng/ul

RT: 23.703 min Scan# 34

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion:252 Resp: 663263

Ion Ratio Lower Upper

252 100

253 22.1 17.4 26.0

125 12.2 9.4 14.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.333 ng/ul

RT: 26.262 min Scan# 3923

Delta R.T. -0.006 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22




Tgt Ion:276 Resp: 701807


Ion Ratio Lower Upper

276 100

138 26.3 21.2 31.8

277 25.3 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 9.345 ng/ul

RT: 26.279 min Scan# 3926

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Instrument :

BNA_M

ClientSampleId :

SSTD010007





Tgt Ion:278 Resp: 582648

Ion Ratio Lower Upper

278 100

139 17.2 13.5 20.3

279 24.2 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 9.465 ng/ul

RT: 27.021 min Scan# 4052

Delta R.T. -0.000 min

Lab File: BM039452.D

Acq: 18 Apr 2023 11:22

Tgt Ion:276 Resp: 566947

Ion Ratio Lower Upper

276 100

138 22.8 18.7 28.1

277 23.4 18.7 28.1

