

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041823\
 Data File : BM039496.D
 Acq On : 19 Apr 2023 15:29
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020112

Quant Time: Apr 19 22:54:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 19 08:18:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	242718	20.000	ng/u1	0.00
20) Naphthalene-d8	10.636	136	1148758	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.483	164	736449	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.236	188	1573289	20.000	ng/u1	0.00
79) Chrysene-d12	21.430	240	1375938	20.000	ng/u1	0.00
88) Perylene-d12	23.794	264	1323437	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.213	96	53492	8.459	ng/uL	0.00
4) Pyridine-d5	3.643	84	367153	20.664	ng/u1	0.00
7) Phenol-d5	7.013	99	456425	20.546	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.160	67	317147	21.599	ng/u1	0.00
11) 2-Chlorophenol-d4	7.360	132	371766	21.334	ng/u1	0.00
15) 4-Methylphenol-d8	8.560	113	376533	21.042	ng/u1	0.00
21) Nitrobenzene-d5	9.001	128	193534	20.941	ng/u1	0.00
24) 2-Nitrophenol-d4	9.725	143	219803	21.180	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.266	165	385460	21.586	ng/u1	0.00
31) 4-Chloroaniline-d4	10.789	131	552273	20.708	ng/u1	0.00
46) Dimethylphthalate-d6	13.901	166	1251266	21.067	ng/u1	0.00
49) Acenaphthylene-d8	14.177	160	1413228	21.180	ng/u1	0.00
54) 4-Nitrophenol-d4	14.718	143	174867	19.454	ng/u1	0.00
60) Fluorene-d10	15.477	176	1026933	21.104	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.613	200	209030	19.993	ng/u1	0.00
73) Anthracene-d10	17.336	188	1588799	20.999	ng/u1	0.00
81) Pyrene-d10	19.636	212	1832797	20.046	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.641	264	1491923	20.880	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.249	88	58416	8.442	ng/uL	97
5) Pyridine	3.660	79	380820	20.327	ng/u1	98
6) Benzaldehyde	6.972	77	248544	26.187	ng/u1	99
8) Phenol	7.037	94	484349	20.941	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.254	93	411212	21.324	ng/u1	99
12) 2-Chlorophenol	7.395	128	371923	21.210	ng/u1	99
13) 2-Methylphenol	8.289	108	374187	20.829	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.366	45	596820	22.283	ng/u1	99
16) Acetophenone	8.660	105	651196	22.155	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.642	70	361107	22.197	ng/u1	96
18) 4-Methylphenol	8.625	108	420903	21.701	ng/u1	98
19) Hexachloroethane	8.901	117	166615	20.921	ng/u1	97
22) Nitrobenzene	9.042	77	491369	21.533	ng/u1	96
23) Isophorone	9.566	82	1037960	21.331	ng/u1	98
25) 2-Nitrophenol	9.754	139	226971	20.961	ng/u1	97
26) 2,4-Dimethylphenol	9.825	107	474613	21.449	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.054	93	615180	21.499	ng/u1	99
29) 2,4-Dichlorophenol	10.295	162	381774	21.704	ng/u1	99
30) Naphthalene	10.683	128	1334901	21.002	ng/u1	100
32) 4-Chloroaniline	10.813	127	545777	20.769	ng/u1	99
33) Hexachlorobutadiene	10.960	225	240845	20.780	ng/u1	99
34) Caprolactam	11.595	113	79419	12.124	ng/u1	99
35) 4-Chloro-3-methylphenol	11.948	107	440443	21.446	ng/u1	98
36) 2-Methylnaphthalene	12.301	142	902714	21.186	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041823\
 Data File : BM039496.D
 Acq On : 19 Apr 2023 15:29
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020112


Quant Time: Apr 19 22:54:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 19 08:18:41 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.519	142	908537	21.218	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.672	216	460731	20.842	ng/ul	98
40) Hexachlorocyclopentadiene	12.642	237	234195	17.782	ng/ul	99
41) 2,4,6-Trichlorophenol	12.924	196	312426	21.156	ng/ul	99
42) 2,4,5-Trichlorophenol	13.001	196	325233	21.190	ng/ul	99
43) 1,1'-Biphenyl	13.319	154	1225189	21.527	ng/ul	100
44) 2-Chloronaphthalene	13.360	162	947150	21.215	ng/ul	100
45) 2-Nitroaniline	13.577	65	290023	22.511	ng/ul	98
47) Dimethylphthalate	13.948	163	1237530	21.045	ng/ul	100
48) 2,6-Dinitrotoluene	14.077	165	252346	21.218	ng/ul	98
50) Acenaphthylene	14.207	152	1531549	21.458	ng/ul	100
51) 3-Nitroaniline	14.407	138	230284	20.658	ng/ul	98
52) Acenaphthene	14.548	153	1046375	21.293	ng/ul	99
53) 2,4-Dinitrophenol	14.624	184	106359	16.800	ng/ul	98
55) 4-Nitrophenol	14.736	109	138285	20.254	ng/ul	97
56) Dibenzofuran	14.883	168	1441329	21.579	ng/ul	100
57) 2,4-Dinitrotoluene	14.866	165	357312	21.622	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.118	232	292870	21.317	ng/ul	99
59) Diethylphthalate	15.313	149	1291637	21.189	ng/ul	99
61) Fluorene	15.536	166	1194485	21.619	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.530	204	589728	21.534	ng/ul	97
63) 4-Nitroaniline	15.577	138	180478	21.113	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.630	198	199368	19.834	ng/ul	94
67) N-Nitrosodiphenylamine	15.748	169	984853	21.010	ng/ul	100
68) 4-Bromophenyl-phenylether	16.424	248	349185	20.549	ng/ul	98
69) Hexachlorobenzene	16.536	284	404522	20.722	ng/ul	97
70) Atrazine	16.707	200	367970	21.156	ng/ul	99
71) Pentachlorophenol	16.895	266	220531	19.325	ng/ul	98
72) Phenanthrene	17.277	178	1828545	20.935	ng/ul	100
74) Anthracene	17.371	178	1854857	21.168	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.283	216	484737	20.688	ng/ul	99
76) Pentachlorobenzene	14.801	250	487537	21.139	ng/ul	100
77) Carbazole	17.648	167	1647660	21.161	ng/ul	100
78) Di-n-butylphthalate	18.206	149	2279425	20.697	ng/ul	99
80) Fluoranthene	19.301	202	2125459	19.833	ng/ul	98
82) Pyrene	19.665	202	2248964	20.272	ng/ul	97
83) Butylbenzylphthalate	20.565	149	1039433	19.849	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.353	252	670540	21.976	ng/ul	98
85) Benzo(a)anthracene	21.412	228	2076425	20.864	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.336	149	1586848	20.376	ng/ul	100
87) Chrysene	21.465	228	1986644	20.928	ng/ul	100
89) Di-n-octyl phthalate	22.253	149	2658974	19.996	ng/ul	100
90) Benzo(b)fluoranthene	23.077	252	1935905	20.789	ng/ul	100
91) Benzo(k)fluoranthene	23.124	252	1924252	21.041	ng/ul	100
93) Benzo(a)pyrene	23.694	252	1646355	20.975	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.247	276	1821865	21.102	ng/ul	99
95) Dibenzo(a,h)anthracene	26.265	278	1523013	21.276	ng/ul	99
96) Benzo(g,h,i)perylene	27.000	276	1421137	20.663	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 19 22:54:26 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 19 08:18:41 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.442 ng/uL
RT: 3.249 min Scan# 30
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion: 88 Resp: 58416
Ion Ratio Lower Upper
88 100
43 31.1 24.6 36.8
58 74.1 56.7 85.1



#5
Pyridine
Concen: 20.327 ng/uL
RT: 3.660 min Scan# 106
Delta R.T. 0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion: 79 Resp: 380820
Ion Ratio Lower Upper
79 100
52 75.7 61.9 92.9
51 36.0 29.7 44.5









#6
Benzaldehyde
Concen: 26.187 ng/ul
RT: 6.972 min Scan# 643
Delta R.T. 0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion: 77 Resp: 248544
Ion Ratio Lower Upper
77 100
105 94.4 76.3 114.5
106 93.8 75.4 113.0



#8
Phenol
Concen: 20.941 ng/ul
RT: 7.037 min Scan# 680
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion: 94 Resp: 484349
Ion Ratio Lower Upper
94 100
65 29.0 22.2 33.4
66 37.1 29.6 44.4









#10
Bis(2-Chloroethyl)ether
Concen: 21.324 ng/ul
RT: 7.254 min Scan# 717
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion: 93 Resp: 411212
Ion Ratio Lower Upper
93 100
63 78.7 61.9 92.9
95 32.2 25.8 38.8



#12
2-Chlorophenol
Concen: 21.210 ng/ul
RT: 7.395 min Scan# 741
Delta R.T. 0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:128 Resp: 371923
Ion Ratio Lower Upper
128 100
64 49.6 38.8 58.2
130 32.1 25.4 38.0









#13
2-Methylphenol
Concen: 20.829 ng/ul
RT: 8.289 min Scan# 893
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:108 Resp: 374187
Ion Ratio Lower Upper
108 100
107 88.4 71.7 107.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 22.283 ng/ul
RT: 8.366 min Scan# 906
Delta R.T. 0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion: 45 Resp: 596820
Ion Ratio Lower Upper
45 100
77 15.2 12.3 18.5
79 11.6 9.8 14.6









#16
Acetophenone
Concen: 22.155 ng/ul
RT: 8.660 min Scan# 911
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:105 Resp: 651196
Ion Ratio Lower Upper
105 100
77 81.7 64.1 96.1
51 28.8 22.2 33.4



#17
N-Nitroso-di-n-propylamine
Concen: 22.197 ng/ul
RT: 8.642 min Scan# 953
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion: 70 Resp: 361107
Ion Ratio Lower Upper
70 100
42 57.8 43.0 64.4
101 9.1 7.4 11.2
130 19.3 16.2 24.2









#18
4-Methylphenol
Concen: 21.701 ng/ul
RT: 8.625 min Scan# 911
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:108 Resp: 420903
Ion Ratio Lower Upper
108 100
107 109.0 88.6 133.0



#19
Hexachloroethane
Concen: 20.921 ng/ul
RT: 8.901 min Scan# 997
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:117 Resp: 166615
Ion Ratio Lower Upper
117 100
201 93.5 78.3 117.5
199 59.3 47.4 71.0









#22
Nitrobenzene
Concen: 21.533 ng/ul
RT: 9.042 min Scan# 1110
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion: 77 Resp: 491369
Ion Ratio Lower Upper
77 100
123 42.2 36.1 54.1
65 13.5 10.2 15.2



#23
Isophorone
Concen: 21.331 ng/ul
RT: 9.566 min Scan# 1110
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion: 82 Resp: 1037960
Ion Ratio Lower Upper
82 100
95 6.5 5.4 8.2
138 15.4 13.2 19.8









#25
2-Nitrophenol
Concen: 20.961 ng/ul
RT: 9.754 min Scan# 1112
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:139 Resp: 226971
Ion Ratio Lower Upper
139 100
65 53.4 40.7 61.1
109 31.5 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 21.449 ng/ul
RT: 9.825 min Scan# 1154
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:107 Resp: 474613
Ion Ratio Lower Upper
107 100
121 48.6 39.1 58.7
122 83.9 68.1 102.1





#27

Bis(2-Chloroethoxy)methane

Concen: 21.499 ng/ul

RT: 10.054 min Scan# 1168

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion: 93 Resp: 615180

Ion Ratio Lower Upper

93 100

95 31.4 25.8 38.6

123 11.0 8.8 13.2



#29

2,4-Dichlorophenol

Concen: 21.704 ng/ul

RT: 10.295 min Scan# 1234

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29




Tgt Ion:162 Resp: 381774


Ion Ratio Lower Upper

162 100

164 65.5 52.4 78.6

98 37.8 29.4 44.0





#30

Naphthalene

Concen: 21.002 ng/ul

RT: 10.683 min Scan# 11

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:128 Resp: 1334901

Ion Ratio Lower Upper

128 100

129 11.0 8.9 13.3

127 13.3 10.5 15.7



#32

4-Chloroaniline

Concen: 20.769 ng/ul

RT: 10.813 min Scan# 1322

Delta R.T. 0.003 min

Lab File: BM039496.D




Acq: 19 Apr 2023 15:29


Tgt Ion:127 Resp: 545777

Ion Ratio Lower Upper

127 100

129 31.8 25.8 38.8





#33

Hexachlorobutadiene

Concen: 20.780 ng/ul

RT: 10.960 min Scan# 1

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:225 Resp: 240845

Ion Ratio Lower Upper

225 100

223 63.3 50.1 75.1

227 64.6 50.4 75.6



#34

Caprolactam

Concen: 12.124 ng/ul

RT: 11.595 min Scan# 1455

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29




Tgt Ion:113 Resp: 79419


Ion Ratio Lower Upper

113 100

55 181.3 145.5 218.3

56 128.3 103.9 155.9









#35
4-Chloro-3-methylphenol
Concen: 21.446 ng/ul
RT: 11.948 min Scan# 11
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:107 Resp: 440443
Ion Ratio Lower Upper
107 100
144 24.9 21.0 31.4
142 80.2 62.9 94.3



#36
2-Methylnaphthalene
Concen: 21.186 ng/ul
RT: 12.301 min Scan# 1575
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:142 Resp: 902714
Ion Ratio Lower Upper
142 100
141 86.7 71.1 106.7





#37

1-Methylnaphthalene

Concen: 21.218 ng/ul

RT: 12.519 min Scan# 1

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:142 Resp: 908537

Ion Ratio Lower Upper

142 100

141 92.1 72.0 108.0

116 4.0 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.842 ng/ul

RT: 12.672 min Scan# 1638

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Tgt Ion:216 Resp: 460731




Ion Ratio Lower Upper


216 100

214 79.4 62.2 93.2

179 20.5 16.5 24.7

108 18.7 14.0 21.0





#40

Hexachlorocyclopentadiene

Concen: 17.782 ng/ul

RT: 12.642 min Scan# 1

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:237 Resp: 234195

Ion Ratio Lower Upper

237 100

235 64.5 50.9 76.3

272 13.6 10.3 15.5



#41

2,4,6-Trichlorophenol

Concen: 21.156 ng/ul

RT: 12.924 min Scan# 1681

Delta R.T. 0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29




Tgt Ion:196 Resp: 312426


Ion Ratio Lower Upper

196 100

198 94.9 76.7 115.1

200 30.4 23.8 35.6





#42

2,4,5-Trichlorophenol

Concen: 21.190 ng/ul

RT: 13.001 min Scan# 1

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:196 Resp: 325233

Ion Ratio Lower Upper

196 100

198 96.0 75.7 113.5

200 30.9 24.6 36.8



#43

1,1'-Biphenyl

Concen: 21.527 ng/ul

RT: 13.319 min Scan# 1748

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29




Tgt Ion:154 Resp: 1225189


Ion Ratio Lower Upper

154 100

153 40.0 31.9 47.9

76 14.1 11.5 17.3





#44

2-Chloronaphthalene

Concen: 21.215 ng/ul

RT: 13.360 min Scan# 1

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:162 Resp: 947150

Ion Ratio Lower Upper

162 100

164 32.7 26.1 39.1

127 39.0 31.4 47.2



#45

2-Nitroaniline

Concen: 22.511 ng/ul

RT: 13.577 min Scan# 1792

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29




Tgt Ion: 65 Resp: 290023


Ion Ratio Lower Upper

65 100

92 66.8 53.4 80.2

138 100.6 83.2 124.8





#47

Dimethylphthalate

Concen: 21.045 ng/ul

RT: 13.948 min Scan# 1855

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:163 Resp: 1237530

Ion Ratio Lower Upper

163 100

194 3.9 3.1 4.7

164 10.1 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 21.218 ng/ul

RT: 14.077 min Scan# 1877

Delta R.T. 0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29




Tgt Ion:165 Resp: 252346


Ion Ratio Lower Upper

165 100

89 61.6 50.6 76.0

121 21.2 17.3 25.9









#50
Acenaphthylene
Concen: 21.458 ng/ul
RT: 14.207 min Scan# 1874
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:152 Resp: 1531549
Ion Ratio Lower Upper
152 100
151 19.9 15.8 23.8
153 13.1 10.4 15.6



#51
3-Nitroaniline
Concen: 20.658 ng/ul
RT: 14.407 min Scan# 1933
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:138 Resp: 230284
Ion Ratio Lower Upper
138 100
108 9.8 8.8 13.2
92 121.7 95.5 143.3





#52

Acenaphthene

Concen: 21.293 ng/ul

RT: 14.548 min Scan# 1932

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:153 Resp: 1046375

Ion Ratio Lower Upper

153 100

152 47.2 38.2 57.4

154 88.2 71.0 106.4



#53

2,4-Dinitrophenol

Concen: 16.800 ng/ul

RT: 14.624 min Scan# 1970

Delta R.T. 0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29




Tgt Ion:184 Resp: 106359


Ion Ratio Lower Upper

184 100

63 76.5 60.0 90.0

154 67.1 51.8 77.6









#55
4-Nitrophenol
Concen: 20.254 ng/ul
RT: 14.736 min Scan# 1964
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:109 Resp: 138285
Ion Ratio Lower Upper
109 100
139 129.9 106.4 159.6
65 134.7 104.8 157.2



#56
Dibenzofuran
Concen: 21.579 ng/ul
RT: 14.883 min Scan# 2014
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:168 Resp: 1441329
Ion Ratio Lower Upper
168 100
139 38.3 30.7 46.1









#57
2,4-Dinitrotoluene
Concen: 21.622 ng/ul
RT: 14.866 min Scan# 2054
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:165 Resp: 357312
Ion Ratio Lower Upper
165 100
63 51.8 39.4 59.2
182 2.4 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.317 ng/ul
RT: 15.118 min Scan# 2054
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:232 Resp: 292870
Ion Ratio Lower Upper
232 100
131 48.2 38.0 57.0
130 2.3 2.0 3.0
166 30.9 24.8 37.2





#59

Diethylphthalate

Concen: 21.189 ng/ul

RT: 15.313 min Scan# 2061

Delta R.T. 0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:149 Resp: 1291637

Ion Ratio Lower Upper

149 100

177 22.0 17.3 25.9

150 12.2 9.8 14.6



#61

Fluorene

Concen: 21.619 ng/ul

RT: 15.536 min Scan# 2125

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29




Tgt Ion:166 Resp: 1194485


Ion Ratio Lower Upper

166 100

165 96.6 78.2 117.2

167 13.5 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 21.534 ng/ul
RT: 15.530 min Scan# 2124
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:204 Resp: 589728
Ion Ratio Lower Upper
204 100
206 32.5 26.8 40.2
141 57.6 48.5 72.7



#63
4-Nitroaniline
Concen: 21.113 ng/ul
RT: 15.577 min Scan# 2132
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:138 Resp: 180478
Ion Ratio Lower Upper
138 100
92 53.5 44.1 66.1
108 80.8 64.6 97.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.834 ng/ul
 RT: 15.630 min Scan# 2116
 Delta R.T. -0.003 min
 Lab File: BM039496.D
 Acq: 19 Apr 2023 15:29




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020112


Tgt Ion:198 Resp: 199368
 Ion Ratio Lower Upper
 198 100
 182 4.1 3.1 4.7
 77 29.4 20.6 31.0



#67
 N-Nitrosodiphenylamine
 Concen: 21.010 ng/ul
 RT: 15.748 min Scan# 2161
 Delta R.T. -0.003 min
 Lab File: BM039496.D
 Acq: 19 Apr 2023 15:29

Tgt Ion:169 Resp: 984853
 Ion Ratio Lower Upper
 169 100
 168 66.4 53.2 79.8
 167 36.5 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 20.549 ng/ul
RT: 16.424 min Scan# 21
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:248 Resp: 349185
Ion Ratio Lower Upper
248 100
250 99.5 77.7 116.5
141 78.3 61.0 91.4



#69
Hexachlorobenzene
Concen: 20.722 ng/ul
RT: 16.536 min Scan# 2295
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:284 Resp: 404522
Ion Ratio Lower Upper
284 100
142 38.5 28.9 43.3
249 31.1 23.9 35.9









#70
Atrazine
Concen: 21.156 ng/ul
RT: 16.707 min Scan# 21
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:200 Resp: 367970
Ion Ratio Lower Upper
200 100
173 26.6 22.2 33.2
215 48.2 38.9 58.3



#71
Pentachlorophenol
Concen: 19.325 ng/ul
RT: 16.895 min Scan# 2356
Delta R.T. 0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:266 Resp: 220531
Ion Ratio Lower Upper
266 100
264 62.6 50.2 75.4
268 62.7 51.9 77.9









#72
Phenanthrene
Concen: 20.935 ng/ul
RT: 17.277 min Scan# 2421
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:178 Resp: 1828545
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.4
176 19.4 15.8 23.8



#74
Anthracene
Concen: 21.168 ng/ul
RT: 17.371 min Scan# 2437
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:178 Resp: 1854857
Ion Ratio Lower Upper
178 100
179 15.3 12.1 18.1
176 19.0 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.688 ng/uL
RT: 13.283 min Scan# 11
Delta R.T. 0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:216 Resp: 484737
Ion Ratio Lower Upper
216 100
214 79.3 63.0 94.4
218 49.0 38.1 57.1



#76
Pentachlorobenzene
Concen: 21.139 ng/uL
RT: 14.801 min Scan# 2000
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29


Tgt Ion:250 Resp: 487537
Ion Ratio Lower Upper
250 100
248 63.0 50.4 75.6
252 62.6 50.2 75.2








#77
 Carbazole
 Concen: 21.161 ng/ul
 RT: 17.648 min Scan# 2459
 Delta R.T. -0.003 min
 Lab File: BM039496.D
 Acq: 19 Apr 2023 15:29

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020112



Tgt Ion:167 Resp: 1647660




Ion	Ratio	Lower	Upper
167	100		
166	21.3	17.0	25.4
139	13.4	10.6	15.8




#78
 Di-n-butylphthalate
 Concen: 20.697 ng/ul
 RT: 18.206 min Scan# 2579
 Delta R.T. -0.003 min
 Lab File: BM039496.D
 Acq: 19 Apr 2023 15:29

Tgt Ion:149 Resp: 2279425

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.0	10.6
104	5.2	4.1	6.1









#80
Fluoranthene
Concen: 19.833 ng/ul
RT: 19.301 min Scan# 21
Delta R.T. 0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:202 Resp: 2125459
Ion Ratio Lower Upper
202 100
101 11.7 10.4 15.6
100 9.1 7.7 11.5



#82
Pyrene
Concen: 20.272 ng/ul
RT: 19.665 min Scan# 2827
Delta R.T. 0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:202 Resp: 2248964
Ion Ratio Lower Upper
202 100
101 13.8 12.1 18.1
100 11.0 9.6 14.4









#83
 Butylbenzylphthalate
 Concen: 19.849 ng/ul
 RT: 20.565 min Scan# 2954
 Delta R.T. 0.003 min
 Lab File: BM039496.D
 Acq: 19 Apr 2023 15:29




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020112


Tgt Ion:149 Resp: 1039433
 Ion Ratio Lower Upper
 149 100
 91 73.3 59.4 89.0
 206 20.4 15.8 23.8



#84
 3,3'-Dichlorobenzidine
 Concen: 21.976 ng/ul
 RT: 21.353 min Scan# 3114
 Delta R.T. 0.003 min
 Lab File: BM039496.D
 Acq: 19 Apr 2023 15:29

Tgt Ion:252 Resp: 670540
 Ion Ratio Lower Upper
 252 100
 254 65.4 50.9 76.3
 126 11.9 10.4 15.6





#85

Benzo(a)anthracene

Concen: 20.864 ng/ul

RT: 21.412 min Scan# 3112

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:228 Resp: 2076425

Ion Ratio Lower Upper

228 100

229 19.8 15.4 23.2

226 27.3 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.376 ng/ul

RT: 21.336 min Scan# 3111

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29




Tgt Ion:149 Resp: 1586848


Ion Ratio Lower Upper

149 100

167 27.3 21.8 32.6

279 3.7 3.1 4.7





#87

Chrysene

Concen: 20.928 ng/ul

RT: 21.465 min Scan# 31

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:228 Resp: 1986644

Ion Ratio Lower Upper

228 100

226 29.8 23.9 35.9

229 20.0 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 19.996 ng/ul




RT: 22.253 min Scan# 3267


Delta R.T. 0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Tgt Ion:149 Resp: 2658974









#90
Benzo(b)fluoranthene
Concen: 20.789 ng/ul
RT: 23.077 min Scan# 3407
Delta R.T. 0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29




Instrument :
BNA_M
ClientSampleId :
SSTD020112


Tgt Ion:252 Resp: 1935905
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.2
125 10.9 8.7 13.1



#91
Benzo(k)fluoranthene
Concen: 21.041 ng/ul
RT: 23.124 min Scan# 3415
Delta R.T. -0.003 min
Lab File: BM039496.D
Acq: 19 Apr 2023 15:29

Tgt Ion:252 Resp: 1924252
Ion Ratio Lower Upper
252 100
253 21.8 17.6 26.4
125 10.8 8.6 12.8





#93

Benzo(a)pyrene

Concen: 20.975 ng/ul

RT: 23.694 min Scan# 3

Delta R.T. 0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:252 Resp: 1646355

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 11.8 9.4 14.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.102 ng/ul

RT: 26.247 min Scan# 3946

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29




Tgt Ion:276 Resp: 1821865


Ion Ratio Lower Upper

276 100

138 25.9 21.2 31.8

277 25.2 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 21.276 ng/ul

RT: 26.265 min Scan# 3924

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Instrument :

BNA_M

ClientSampleId :

SSTD020112





Tgt Ion:278 Resp: 1523013

Ion Ratio Lower Upper

278 100

139 17.1 13.5 20.3

279 23.1 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 20.663 ng/ul

RT: 27.000 min Scan# 4074

Delta R.T. -0.003 min

Lab File: BM039496.D

Acq: 19 Apr 2023 15:29

Tgt Ion:276 Resp: 1421137

Ion Ratio Lower Upper

276 100

138 23.7 18.7 28.1

277 23.8 18.7 28.1

