

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041823\
 Data File : BM039473.D
 Acq On : 19 Apr 2023 01:04
 Operator : CG/JU
 Sample : 02335-17
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 COMJ2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/19/2023
 Supervised By :Jagrut Upadhyay 04/19/2023

Quant Time: Apr 19 02:25:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 18 23:56:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	219313	20.000	ng/u1	0.00
20) Naphthalene-d8	10.639	136	994757	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.486	164	642011	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.239	188	1329510	20.000	ng/u1	0.00
79) Chrysene-d12	21.427	240	1010682	20.000	ng/u1	-0.01
88) Perylene-d12	23.797	264	972218	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.216	96	28124m	4.922	ng/uL	0.00
4) Pyridine-d5	3.646	84	128714	8.017	ng/u1	0.00
7) Phenol-d5	7.022	99	129292	6.441	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.163	67	436798	32.923	ng/u1	0.00
11) 2-Chlorophenol-d4	7.363	132	403195	25.607	ng/u1	0.00
15) 4-Methylphenol-d8	8.563	113	259111	16.025	ng/u1	0.00
21) Nitrobenzene-d5	9.004	128	264548	33.056	ng/u1	0.00
24) 2-Nitrophenol-d4	9.728	143	292627	32.563	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.275	165	429697	27.789	ng/u1	0.00
31) 4-Chloroaniline-d4	10.792	131	569323	24.652	ng/u1	0.00
46) Dimethylphthalate-d6	13.904	166	1790516	34.581	ng/u1	0.00
49) Acenaphthylene-d8	14.180	160	2022880	34.777	ng/u1	0.00
54) 4-Nitrophenol-d4	14.769	143	26004m	3.318	ng/u1	0.04
60) Fluorene-d10	15.480	176	1536823	36.228	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.616	200	263611	29.837	ng/u1	0.00
73) Anthracene-d10	17.339	188	2450003	38.318	ng/u1	0.00
81) Pyrene-d10	19.639	212	2666342	39.702	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.645	264	1971311	37.557	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM041823\
Data File : BM039473.D
Acq On : 19 Apr 2023 01:04
Operator : CG/JU
Sample : 02335-17
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
COMJ2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/19/2023
Supervised By :Jagrut Upadhyay 04/19/2023

Quant Time: Apr 19 02:25:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 18 23:56:45 2023
Response via : Initial Calibration

