

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042015\
 Data File : BM001094.D
 Acq On : 21 Apr 2015 08:46
 Operator : TP/IZ
 Sample : G1859-12
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Quant Time: Apr 21 19:12:32 2015
 Quant Method : Z:\HPCHEM1\BNA M\Methods\SOM01.2-EPA-BM041515.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 16 05:47:45 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	215806	20.00	ng/ul	0.00
18) Naphthalene-d8	10.36	136	866356	20.00	ng/ul	-0.01
35) Acenaphthene-d10	14.24	164	462208	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.99	188	960151	20.00	ng/ul	0.00
75) Chrysene-d12	21.19	240	987478	20.00	ng/ul	0.00
83) Perylene-d12	23.36	264	1002913	20.00	ng/ul	-0.01


System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	2.92	96	273	0.06	ng/uL	-0.13
5) Phenol-d5	6.76	99	436356	25.92	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.91	67	280817	26.77	ng/ul	-0.01
9) 2-Chlorophenol-d4	7.11	132	355770	25.99	ng/ul	0.00
13) 4-Methylphenol-d8	8.29	113	324064	24.39	ng/ul	0.00
19) Nitrobenzene-d5	8.73	128	160409	27.46	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.45	143	180134	27.62	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.99	165	290120	22.82	ng/ul	0.00
29) 4-Chloroaniline-d4	10.50	131	316088	25.51	ng/ul	0.00
43) Dimethylphthalate-d6	13.65	166	824539	26.77	ng/ul	0.00
46) Acenaphthylene-d8	13.93	160	800991	18.17	ng/ul	0.00
51) 4-Nitrophenol-d4	14.46	143	119500	19.96	ng/ul	0.00
57) Fluorene-d10	15.23	176	446876	14.60	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.36	200	57873	11.10	ng/ul	0.00
70) Anthracene-d10	17.08	188	564570	12.99	ng/ul	-0.01
76) Pyrene-d10	19.39	212	478217	10.43	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.21	264	296634	6.61	ng/ul	-0.01


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
44) Dimethylphthalate	13.70	163	137876	4.00	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042015\
Data File : BM001094.D
Acq On : 21 Apr 2015 08:46
Operator : TP/IZ
Sample : G1859-12
Misc :
ALS Vial : 28 Sample Multiplier: 1

Quant Time: Apr 21 19:12:32 2015
Quant Method : Z:\HPCHEM1\BNA M\Methods\SOM01.2-EPA-BM041515.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Apr 16 05:47:45 2015
Response via : Initial Calibration





#44
 Dimethylphthalate
 Concen: 4.00 ng/ul
 RT: 13.70 min Scan# 1872
 Delta R.T. -0.01 min
 Lab File: BM001094.D
 Acq: 21 Apr 2015 08:46

Tot Ion:163	Resp: 137876
Ion Ratio	Lower Upper
163 100	
194 4.1	3.0 4.4
164 10.2	8.2 12.4

