

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042022\
 Data File : BM034728.D
 Acq On : 20 Apr 2022 15:20
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV024

Quant Time: Apr 20 19:08:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM042022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 20 14:02:01 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.951	152	278656	20.000	ng/u1	0.00
20) Naphthalene-d8	10.757	136	1146016	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.580	164	653510	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.321	188	1307501	20.000	ng/u1	0.00
79) Chrysene-d12	21.497	240	1231899	20.000	ng/u1	0.00
88) Perylene-d12	23.903	264	1238835	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	52673	8.132	ng/uL	0.00
4) Pyridine-d5	3.799	84	347846	22.886	ng/u1	0.00
7) Phenol-d5	7.104	99	465439	22.811	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	276899	24.132	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	398226	22.571	ng/u1	0.00
15) 4-Methylphenol-d8	8.657	113	381216	23.035	ng/u1	0.00
21) Nitrobenzene-d5	9.104	128	180854	23.202	ng/u1	0.00
24) 2-Nitrophenol-d4	9.834	143	192085	24.165	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	401279	21.216	ng/u1	0.00
31) 4-Chloroaniline-d4	10.881	131	534661	21.626	ng/u1	0.00
46) Dimethylphthalate-d6	13.992	166	1070568	21.858	ng/u1	0.00
49) Acenaphthylene-d8	14.274	160	1351498	21.621	ng/u1	0.00
54) 4-Nitrophenol-d4	14.763	143	160252	21.202	ng/u1	0.00
60) Fluorene-d10	15.574	176	912090	21.465	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.680	200	143529	21.435	ng/u1	0.00
73) Anthracene-d10	17.421	188	1394941	21.828	ng/u1	0.00
81) Pyrene-d10	19.709	212	1519633	23.099	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.744	264	1380115	21.684	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.422	88	53266	8.089	ng/uL	88
5) Pyridine	3.816	79	361116	22.930	ng/u1#	80
6) Benzaldehyde	7.081	77	251193	23.516	ng/u1	97
8) Phenol	7.134	94	462845	22.987	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.369	93	366009	23.120	ng/u1	94
12) 2-Chlorophenol	7.510	128	397917	22.300	ng/u1	94
13) 2-Methylphenol	8.387	108	356964	23.036	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.481	45	532414	25.674	ng/u1#	84
16) Acetophenone	8.769	105	589391	23.607	ng/u1#	95
17) N-Nitroso-di-n-propyla...	8.763	70	292195	24.745	ng/u1#	85
18) 4-Methylphenol	8.716	108	388759	23.481	ng/u1	99
19) Hexachloroethane	9.039	117	165110	22.306	ng/u1	93
22) Nitrobenzene	9.145	77	413457	23.436	ng/u1	98
23) Isophorone	9.681	82	783672	22.874	ng/u1	97
25) 2-Nitrophenol	9.863	139	214714	24.253	ng/u1	96
26) 2,4-Dimethylphenol	9.928	107	433517	21.858	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.163	93	492117	22.832	ng/u1	99
29) 2,4-Dichlorophenol	10.398	162	388491	21.443	ng/u1	98
30) Naphthalene	10.804	128	1267944	21.515	ng/u1	99
32) 4-Chloroaniline	10.904	127	533834	21.883	ng/u1	99
33) Hexachlorobutadiene	11.104	225	260305	19.115	ng/u1	98
34) Caprolactam	11.675	113	53425	12.322	ng/u1#	70
35) 4-Chloro-3-methylphenol	12.028	107	378523	22.383	ng/u1	95
36) 2-Methylnaphthalene	12.416	142	895073	21.816	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042022\
 Data File : BM034728.D
 Acq On : 20 Apr 2022 15:20
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV024

Quant Time: Apr 20 19:08:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM042022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 20 14:02:01 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.633	142	890298	21.504	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.780	216	471976	20.149	ng/ul	98
40) Hexachlorocyclopentadiene	12.769	237	309600	19.225	ng/ul	97
41) 2,4,6-Trichlorophenol	13.016	196	301247	21.845	ng/ul	95
42) 2,4,5-Trichlorophenol	13.086	196	319279	21.955	ng/ul	99
43) 1,1'-Biphenyl	13.422	154	1216093	22.358	ng/ul	98
44) 2-Chloronaphthalene	13.457	162	928190	21.940	ng/ul	99
45) 2-Nitroaniline	13.651	65	212186	26.873	ng/ul	94
47) Dimethylphthalate	14.039	163	1032373	21.903	ng/ul	98
48) 2,6-Dinitrotoluene	14.151	165	194281	24.906	ng/ul	90
50) Acenaphthylene	14.304	152	1335739	21.695	ng/ul	99
51) 3-Nitroaniline	14.474	138	123003	18.387	ng/ul	98
52) Acenaphthene	14.645	153	867399	21.516	ng/ul	100
53) 2,4-Dinitrophenol	14.680	184	85234	19.536	ng/ul	88
55) 4-Nitrophenol	14.774	109	115460	20.050	ng/ul#	79
56) Dibenzofuran	14.980	168	1269159	21.760	ng/ul	95
57) 2,4-Dinitrotoluene	14.933	165	272356	24.825	ng/ul#	91
58) 2,3,4,6-Tetrachlorophenol	15.204	232	280298	23.937	ng/ul#	92
59) Diethylphthalate	15.404	149	978547	21.355	ng/ul	98
61) Fluorene	15.627	166	1024728	21.724	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.621	204	570704	22.580	ng/ul	98
63) 4-Nitroaniline	15.633	138	98566	16.097	ng/ul#	85
66) 4,6-Dinitro-2-methylph...	15.698	198	152079	22.285	ng/ul#	91
67) N-Nitrosodiphenylamine	15.833	169	860220	21.950	ng/ul	99
68) 4-Bromophenyl-phenylether	16.515	248	353119	23.016	ng/ul	97
69) Hexachlorobenzene	16.633	284	415847	23.021	ng/ul	95
70) Atrazine	16.786	200	327888	21.719	ng/ul	99
71) Pentachlorophenol	16.968	266	238380	21.618	ng/ul	99
72) Phenanthrene	17.362	178	1552076	21.635	ng/ul	99
74) Anthracene	17.457	178	1600236	21.987	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.386	216	487277	20.813	ng/ul	99
76) Pentachlorobenzene	14.904	250	504903	23.411	ng/ul	99
77) Carbazole	17.721	167	1313882	20.650	ng/ul	98
78) Di-n-butylphthalate	18.304	149	1467030	20.327	ng/ul	99
80) Fluoranthene	19.374	202	1743671	22.984	ng/ul	100
82) Pyrene	19.739	202	1777934	22.823	ng/ul	100
83) Butylbenzylphthalate	20.645	149	576146	22.603	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.415	252	481812	18.882	ng/ul	100
85) Benzo(a)anthracene	21.486	228	1669859	21.877	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.427	149	891619	21.884	ng/ul	99
87) Chrysene	21.539	228	1641290	22.114	ng/ul	100
89) Di-n-octyl phthalate	22.362	149	1407503	20.871	ng/ul	100
90) Benzo(b)fluoranthene	23.168	252	1716320	22.250	ng/ul	99
91) Benzo(k)fluoranthene	23.221	252	1556197	21.504	ng/ul	98
93) Benzo(a)pyrene	23.797	252	1632712	21.865	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.391	276	1943509	21.653	ng/ul	99
95) Dibenzo(a,h)anthracene	26.415	278	1685387	21.799	ng/ul	99
96) Benzo(g,h,i)perylene	27.156	276	1648319	21.478	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042022\
 Data File : BM034728.D
 Acq On : 20 Apr 2022 15:20
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV024

Quant Time: Apr 20 19:08:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM042022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 20 14:02:01 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.089 ng/uL
RT: 3.422 min Scan# 51
Delta R.T. 0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion: 88 Resp: 53266
Ion Ratio Lower Upper
88 100
43 30.4 21.0 31.6
58 70.4 47.8 71.8



#5
Pyridine
Concen: 22.930 ng/uL
RT: 3.816 min Scan# 124
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion: 79 Resp: 361116
Ion Ratio Lower Upper
79 100
52 77.6 47.7 71.5#
51 38.9 25.6 38.4#









#6
Benzaldehyde
Concen: 23.516 ng/ul
RT: 7.081 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion: 77 Resp: 251193
Ion Ratio Lower Upper
77 100
105 97.2 80.8 121.2
106 94.0 76.6 115.0



#8
Phenol
Concen: 22.987 ng/ul
RT: 7.134 min Scan# 688
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion: 94 Resp: 462845
Ion Ratio Lower Upper
94 100
65 29.7 22.9 34.3
66 39.4 33.7 50.5









#10
 Bis(2-Chloroethyl)ether
 Concen: 23.120 ng/ul
 RT: 7.369 min Scan# 728
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20

Instrument :
 BNA_M
 ClientSampleId :
 SICV024

Tgt Ion: 93 Resp: 366009




Ion	Ratio	Lower	Upper
93	100		
63	74.5	54.7	82.1
95	31.5	26.6	39.8




#12
 2-Chlorophenol
 Concen: 22.300 ng/ul
 RT: 7.510 min Scan# 752
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20

Tgt Ion: 128 Resp: 397917

Ion	Ratio	Lower	Upper
128	100		
64	44.6	30.6	46.0
130	32.2	25.8	38.6









#13
2-Methylphenol
Concen: 23.036 ng/ul
RT: 8.387 min Scan# 902
Delta R.T. -0.006 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:108 Resp: 356964
Ion Ratio Lower Upper
108 100
107 88.1 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 25.674 ng/ul
RT: 8.481 min Scan# 917
Delta R.T. -0.006 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion: 45 Resp: 532414
Ion Ratio Lower Upper
45 100
77 13.5 17.1 25.7#
79 9.9 12.8 19.2#









#16
Acetophenone
Concen: 23.607 ng/ul
RT: 8.769 min Scan# 966
Delta R.T. -0.006 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:105 Resp: 589391
Ion Ratio Lower Upper
105 100
77 77.0 59.4 89.2
51 29.1 19.3 28.9#



#17
N-Nitroso-di-n-propylamine
Concen: 24.745 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion: 70 Resp: 292195
Ion Ratio Lower Upper
70 100
42 66.6 40.6 60.8#
101 10.3 8.2 12.2
130 22.1 19.9 29.9









#18
4-Methylphenol
Concen: 23.481 ng/ul
RT: 8.716 min Scan# 957
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:108 Resp: 388759
Ion Ratio Lower Upper
108 100
107 109.9 89.0 133.4



#19
Hexachloroethane
Concen: 22.306 ng/ul
RT: 9.039 min Scan# 1012
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:117 Resp: 165110
Ion Ratio Lower Upper
117 100
201 103.5 89.9 134.9
199 66.8 57.2 85.8









#22
Nitrobenzene
Concen: 23.436 ng/ul
RT: 9.145 min Scan# 1030
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion: 77 Resp: 413457
Ion Ratio Lower Upper
77 100
123 46.1 38.2 57.2
65 13.6 10.6 15.8



#23
Isophorone
Concen: 22.874 ng/ul
RT: 9.681 min Scan# 1121
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion: 82 Resp: 783672
Ion Ratio Lower Upper
82 100
95 7.8 6.8 10.2
138 18.7 16.5 24.7









#25
2-Nitrophenol
Concen: 24.253 ng/ul
RT: 9.863 min Scan# 1152
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:139 Resp: 214714
Ion Ratio Lower Upper
139 100
65 49.1 37.8 56.6
109 29.5 26.5 39.7



#26
2,4-Dimethylphenol
Concen: 21.858 ng/ul
RT: 9.928 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:107 Resp: 433517
Ion Ratio Lower Upper
107 100
121 51.2 40.8 61.2
122 86.6 69.0 103.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 22.832 ng/ul
 RT: 10.163 min Scan# 1203
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20




Instrument :
 BNA_M
 ClientSampleId :
 SICV024


Tgt Ion: 93 Resp: 492117
 Ion Ratio Lower Upper
 93 100
 95 32.2 25.4 38.2
 123 11.8 9.5 14.3



#29
 2,4-Dichlorophenol
 Concen: 21.443 ng/ul
 RT: 10.398 min Scan# 1243
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20

Tgt Ion:162 Resp: 388491
 Ion Ratio Lower Upper
 162 100
 164 65.0 52.8 79.2
 98 33.7 26.0 39.0









#30
Naphthalene
Concen: 21.515 ng/ul
RT: 10.804 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:128 Resp: 1267944
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 12.8 9.9 14.9



#32
4-Chloroaniline
Concen: 21.883 ng/ul
RT: 10.904 min Scan# 1329
Delta R.T. -0.006 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:127 Resp: 533834
Ion Ratio Lower Upper
127 100
129 31.9 25.1 37.7









#33
Hexachlorobutadiene
Concen: 19.115 ng/ul
RT: 11.104 min Scan# 11
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:225 Resp: 260305
Ion Ratio Lower Upper
225 100
223 63.5 50.6 75.8
227 65.3 50.3 75.5



#34
Caprolactam
Concen: 12.322 ng/ul
RT: 11.675 min Scan# 1460
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:113 Resp: 53425
Ion Ratio Lower Upper
113 100
55 164.3 100.9 151.3#
56 127.3 81.0 121.4#









#35
4-Chloro-3-methylphenol
Concen: 22.383 ng/ul
RT: 12.028 min Scan# 1520
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:107 Resp: 378523
Ion Ratio Lower Upper
107 100
144 26.4 23.0 34.4
142 82.3 69.8 104.8



#36
2-Methylnaphthalene
Concen: 21.816 ng/ul
RT: 12.416 min Scan# 1586
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:142 Resp: 895073
Ion Ratio Lower Upper
142 100
141 88.7 71.4 107.0





#37

1-Methylnaphthalene

Concen: 21.504 ng/ul

RT: 12.633 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20

Instrument :

BNA_M

ClientSampleId :

SICV024





Tgt Ion:142 Resp: 890298

Ion Ratio Lower Upper

142 100

141 93.4 74.5 111.7

116 3.5 3.0 4.6



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.149 ng/ul

RT: 12.780 min Scan# 1648

Delta R.T. -0.000 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20

Tgt Ion:216 Resp: 471976




Ion Ratio Lower Upper


216 100

214 78.0 62.2 93.2

179 21.1 15.4 23.2

108 17.6 11.8 17.6









#40
Hexachlorocyclopentadiene
Concen: 19.225 ng/ul
RT: 12.769 min Scan# 1
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:237 Resp: 309600
Ion Ratio Lower Upper
237 100
235 64.1 52.9 79.3
272 12.8 11.8 17.8



#41
2,4,6-Trichlorophenol
Concen: 21.845 ng/ul
RT: 13.016 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:196 Resp: 301247
Ion Ratio Lower Upper
196 100
198 95.9 80.8 121.2
200 31.5 26.5 39.7





#42

2,4,5-Trichlorophenol

Concen: 21.955 ng/ul

RT: 13.086 min Scan# 1

Delta R.T. 0.006 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20

Instrument :

BNA_M

ClientSampleId :

SICV024





Tgt Ion:196 Resp: 319279

Ion Ratio Lower Upper

196 100

198 95.8 78.0 117.0

200 30.5 24.5 36.7



#43

1,1'-Biphenyl

Concen: 22.358 ng/ul

RT: 13.422 min Scan# 1757

Delta R.T. -0.000 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20




Tgt Ion:154 Resp: 1216093


Ion Ratio Lower Upper

154 100

153 39.9 32.6 49.0

76 11.3 7.7 11.5









#44
2-Chloronaphthalene
Concen: 21.940 ng/ul
RT: 13.457 min Scan# 1763
Delta R.T. -0.006 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:162 Resp: 928190
Ion Ratio Lower Upper
162 100
164 33.1 26.6 40.0
127 37.6 29.6 44.4



#45
2-Nitroaniline
Concen: 26.873 ng/ul
RT: 13.651 min Scan# 1796
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion: 65 Resp: 212186
Ion Ratio Lower Upper
65 100
92 69.3 57.4 86.2
138 109.4 94.8 142.2









#47
Dimethylphthalate
Concen: 21.903 ng/ul
RT: 14.039 min Scan# 1862
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:163 Resp: 1032373
Ion Ratio Lower Upper
163 100
194 4.5 3.4 5.2
164 10.6 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 24.906 ng/ul
RT: 14.151 min Scan# 1881
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:165 Resp: 194281
Ion Ratio Lower Upper
165 100
89 45.8 44.2 66.4
121 19.9 16.7 25.1









#50
Acenaphthylene
Concen: 21.695 ng/ul
RT: 14.304 min Scan# 1907
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:152 Resp: 1335739
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.8
153 13.2 10.2 15.2



#51
3-Nitroaniline
Concen: 18.387 ng/ul
RT: 14.474 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:138 Resp: 123003
Ion Ratio Lower Upper
138 100
108 11.8 8.3 12.5
92 106.5 87.0 130.4









#52
Acenaphthene
Concen: 21.516 ng/ul
RT: 14.645 min Scan# 1965
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:153 Resp: 867399
Ion Ratio Lower Upper
153 100
152 46.3 37.4 56.0
154 88.8 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 19.536 ng/ul
RT: 14.680 min Scan# 1971
Delta R.T. 0.006 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:184 Resp: 85234
Ion Ratio Lower Upper
184 100
63 45.0 43.8 65.6
154 54.0 49.9 74.9









#55
4-Nitrophenol
Concen: 20.050 ng/ul
RT: 14.774 min Scan# 1987
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:109 Resp: 115460
Ion Ratio Lower Upper
109 100
139 146.8 94.6 142.0#
65 119.7 83.5 125.3



#56
Dibenzofuran
Concen: 21.760 ng/ul
RT: 14.980 min Scan# 2022
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:168 Resp: 1269159
Ion Ratio Lower Upper
168 100
139 34.2 29.8 44.6









#57
2,4-Dinitrotoluene
Concen: 24.825 ng/ul
RT: 14.933 min Scan# 2014
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:165 Resp: 272356
Ion Ratio Lower Upper
165 100
63 32.3 30.2 45.4
182 3.4 2.2 3.2#



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.937 ng/ul
RT: 15.204 min Scan# 2060
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:232 Resp: 280298
Ion Ratio Lower Upper
232 100
131 32.7 32.2 48.4
130 1.6 1.9 2.9#
166 27.1 23.0 34.4









#59
Diethylphthalate
Concen: 21.355 ng/ul
RT: 15.404 min Scan# 2094
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:149 Resp: 978547
Ion Ratio Lower Upper
149 100
177 23.8 17.9 26.9
150 12.9 10.2 15.2



#61
Fluorene
Concen: 21.724 ng/ul
RT: 15.627 min Scan# 2132
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:166 Resp: 1024728
Ion Ratio Lower Upper
166 100
165 97.3 76.9 115.3
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 22.580 ng/ul
RT: 15.621 min Scan# 2131
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:204 Resp: 570704
Ion Ratio Lower Upper
204 100
206 32.4 25.5 38.3
141 53.3 44.2 66.2



#63
4-Nitroaniline
Concen: 16.097 ng/ul
RT: 15.633 min Scan# 2133
Delta R.T. 0.006 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:138 Resp: 98566
Ion Ratio Lower Upper
138 100
92 44.5 36.3 54.5
108 65.4 68.9 103.3#









#66
4,6-Dinitro-2-methylphenol
Concen: 22.285 ng/ul
RT: 15.698 min Scan# 2143
Delta R.T. 0.006 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:198 Resp: 152079
Ion Ratio Lower Upper
198 100
182 4.2 3.8 5.6
77 17.9 18.5 27.7#



#67
N-Nitrosodiphenylamine
Concen: 21.950 ng/ul
RT: 15.833 min Scan# 2167
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:169 Resp: 860220
Ion Ratio Lower Upper
169 100
168 66.9 52.8 79.2
167 35.7 28.1 42.1









#68
4-Bromophenyl-phenylether
Concen: 23.016 ng/ul
RT: 16.515 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:248 Resp: 353119
Ion Ratio Lower Upper
248 100
250 99.4 78.4 117.6
141 59.4 51.4 77.2



#69
Hexachlorobenzene
Concen: 23.021 ng/ul
RT: 16.633 min Scan# 2303
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:284 Resp: 415847
Ion Ratio Lower Upper
284 100
142 25.6 24.6 36.8
249 30.2 24.1 36.1





#70

Atrazine

Concen: 21.719 ng/ul

RT: 16.786 min Scan# 21

Delta R.T. -0.000 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20

Instrument :

BNA_M

ClientSampleId :

SICV024





Tgt Ion:200 Resp: 327888

Ion Ratio Lower Upper

200 100

173 25.4 20.2 30.4

215 47.2 38.4 57.6



#71

Pentachlorophenol

Concen: 21.618 ng/ul

RT: 16.968 min Scan# 2360

Delta R.T. -0.000 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20




Tgt Ion:266 Resp: 238380


Ion Ratio Lower Upper

266 100

264 62.0 49.3 73.9

268 63.6 50.3 75.5









#72
 Phenanthrene
 Concen: 21.635 ng/ul
 RT: 17.362 min Scan# 2427
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20




Instrument :
 BNA_M
 ClientSampleId :
 SICV024


Tgt Ion:178 Resp: 1552076
 Ion Ratio Lower Upper
 178 100
 179 15.7 11.7 17.5
 176 18.6 15.0 22.4



#74
 Anthracene
 Concen: 21.987 ng/ul
 RT: 17.457 min Scan# 2443
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20

Tgt Ion:178 Resp: 1600236
 Ion Ratio Lower Upper
 178 100
 179 15.2 11.9 17.9
 176 18.1 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.813 ng/uL

RT: 13.386 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20

Instrument :

BNA_M

ClientSampleId :

SICV024





Tgt Ion:216 Resp: 487277

Ion Ratio Lower Upper

216 100

214 78.7 62.0 93.0

218 47.7 37.8 56.8



#76

Pentachlorobenzene

Concen: 23.411 ng/uL

RT: 14.904 min Scan# 2009

Delta R.T. -0.000 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20




Tgt Ion:250 Resp: 504903


Ion Ratio Lower Upper

250 100

248 62.3 50.6 75.8

252 63.4 50.2 75.4









#77
 Carbazole
 Concen: 20.650 ng/ul
 RT: 17.721 min Scan# 2488
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20




Instrument :
 BNA_M
 ClientSampleId :
 SICV024


Tgt Ion:167 Resp: 1313882
 Ion Ratio Lower Upper
 167 100
 166 21.3 16.6 25.0
 139 11.5 9.9 14.9



#78
 Di-n-butylphthalate
 Concen: 20.327 ng/ul
 RT: 18.304 min Scan# 2587
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20

Tgt Ion:149 Resp: 1467030
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.4 11.0
 104 4.5 4.2 6.2









#80
 Fluoranthene
 Concen: 22.984 ng/u1
 RT: 19.374 min Scan# 2769
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20




Instrument :
 BNA_M
 ClientSampleId :
 SICV024


Tgt Ion:202 Resp: 1743671
 Ion Ratio Lower Upper
 202 100
 101 7.6 6.2 9.2
 100 6.1 4.7 7.1



#82
 Pyrene
 Concen: 22.823 ng/u1
 RT: 19.739 min Scan# 2831
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20

Tgt Ion:202 Resp: 1777934
 Ion Ratio Lower Upper
 202 100
 101 8.7 7.0 10.4
 100 7.5 5.8 8.6









#83
 Butylbenzylphthalate
 Concen: 22.603 ng/ul
 RT: 20.645 min Scan# 2985
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20




Instrument :
 BNA_M
 ClientSampleId :
 SICV024


Tgt Ion:	149	Resp:	576146
Ion Ratio	Lower	Upper	
149	100		
91	57.9	50.9	76.3
206	25.5	19.3	28.9



#84
 3,3'-Dichlorobenzidine
 Concen: 18.882 ng/ul
 RT: 21.415 min Scan# 3116
 Delta R.T. 0.006 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20

Tgt Ion:	252	Resp:	481812
Ion Ratio	Lower	Upper	
252	100		
254	63.1	50.7	76.1
126	8.0	6.6	9.8





#85

Benzo(a)anthracene

Concen: 21.877 ng/ul

RT: 21.486 min Scan# 3127

Delta R.T. 0.006 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20

Instrument :

BNA_M

ClientSampleId :

SICV024





Tgt Ion:228 Resp: 1669859

Ion Ratio Lower Upper

228 100

229 19.7 15.4 23.2

226 26.3 21.0 31.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.884 ng/ul

RT: 21.427 min Scan# 3118

Delta R.T. -0.000 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20




Tgt Ion:149 Resp: 891619


Ion Ratio Lower Upper

149 100

167 28.1 22.1 33.1

279 5.4 4.2 6.2





#87

Chrysene

Concen: 22.114 ng/ul

RT: 21.539 min Scan# 31

Delta R.T. 0.006 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20

Instrument :

BNA_M

ClientSampleId :

SICV024





Tgt Ion:228 Resp: 1641290

Ion Ratio Lower Upper

228 100

226 28.9 23.4 35.2

229 19.1 15.4 23.0



#89

Di-n-octyl phthalate

Concen: 20.871 ng/ul




RT: 22.362 min Scan# 3277


Delta R.T. 0.006 min

Lab File: BM034728.D

Acq: 20 Apr 2022 15:20

Tgt Ion:149 Resp: 1407503









#90
Benzo(b)fluoranthene
Concen: 22.250 ng/ul
RT: 23.168 min Scan# 3414
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:252 Resp: 1716320
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.0
125 7.8 5.8 8.6



#91
Benzo(k)fluoranthene
Concen: 21.504 ng/ul
RT: 23.221 min Scan# 3423
Delta R.T. 0.006 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:252 Resp: 1556197
Ion Ratio Lower Upper
252 100
253 22.4 17.5 26.3
125 7.6 5.1 7.7









#93
Benzo(a)pyrene
Concen: 21.865 ng/ul
RT: 23.797 min Scan# 3521
Delta R.T. 0.006 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20




Instrument :
BNA_M
ClientSampleId :
SICV024


Tgt Ion:252 Resp: 1632712
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 8.2 6.2 9.4



#94
Indeno(1,2,3-cd)pyrene
Concen: 21.653 ng/ul
RT: 26.391 min Scan# 3962
Delta R.T. -0.000 min
Lab File: BM034728.D
Acq: 20 Apr 2022 15:20

Tgt Ion:276 Resp: 1943509
Ion Ratio Lower Upper
276 100
138 17.9 13.6 20.4
277 25.2 19.9 29.9









#95
 Dibenzo(a,h)anthracene
 Concen: 21.799 ng/ul
 RT: 26.415 min Scan# 3966
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20

Instrument :
 BNA_M
 ClientSampleId :
 SICV024

Tgt Ion:278 Resp: 1685387
 Ion Ratio Lower Upper
 278 100
 139 12.1 9.6 14.4
 279 23.4 19.3 28.9



#96
 Benzo(g,h,i)perylene
 Concen: 21.478 ng/ul
 RT: 27.156 min Scan# 4092
 Delta R.T. -0.000 min
 Lab File: BM034728.D
 Acq: 20 Apr 2022 15:20

Tgt Ion:276 Resp: 1648319
 Ion Ratio Lower Upper
 276 100
 138 16.2 11.8 17.6
 277 23.5 18.8 28.2

