


Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM042022\  
 Data File : BM034721.D  
 Acq On : 20 Apr 2022 09:01  
 Operator : CG/JU  
 Sample : DFTPP048  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP048

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\SFAM-EPA-BM042022.M  
 Title : SVOA CALIBRATION  
 Last Update : Wed Apr 20 14:02:01 2022



AutoFind: Scans 2497, 2498, 2499; Background Corrected with Scan 2491

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 37.3         | 166371     | PASS                |
| 68             | 69              | 0.00            | 2               | 1.6          | 2649       | PASS                |
| 69             | 198             | 0.00            | 100             | 38.0         | 169545     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.5          | 904        | PASS                |
| 127            | 198             | 10              | 80              | 46.9         | 208981     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.5          | 2272       | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 445589     | PASS                |
| 199            | 198             | 5               | 9               | 6.8          | 30203      | PASS                |
| 275            | 198             | 10              | 60              | 28.2         | 125496     | PASS                |
| 365            | 198             | 1               | 100             | 4.1          | 18063      | PASS                |
| 441            | 198             | 0.01            | 100             | 14.7         | 65453      | PASS                |
| 442            | 198             | 50              | 100             | 95.3         | 424683     | PASS                |
| 443            | 442             | 15              | 24              | 19.3         | 81861      | PASS                |

| m/z    | Abundance |
|--------|-----------|
| 36.10  | 401.0     |
| 37.90  | 219.0     |
| 39.00  | 434.0     |
| 40.00  | 8693.0    |
| 41.00  | 717.0     |
| 41.90  | 339.0     |
| 43.00  | 627.0     |
| 44.00  | 6870.0    |
| 45.00  | 450.0     |
| 48.90  | 515.0     |
| 49.90  | 164.0     |
| 51.00  | 285.0     |
| 52.90  | 170.0     |
| 55.00  | 699.0     |
| 56.10  | 253.0     |
| 57.10  | 318.0     |
| 58.90  | 180.0     |
| 63.90  | 181.0     |
| 65.10  | 184.0     |
| 67.10  | 176.0     |
| 68.00  | 151.0     |
| 68.90  | 247.0     |
| 70.00  | 150.0     |
| 73.00  | 3266.0    |
| 74.00  | 365.0     |
| 75.10  | 359.0     |
| 77.10  | 320.0     |
| 78.10  | 289.0     |
| 78.90  | 276.0     |
| 80.90  | 219.0     |
| 83.00  | 201.0     |
| 84.00  | 355.0     |
| 86.10  | 213.0     |
| 87.00  | 232.0     |
| 89.00  | 229.0     |
| 91.00  | 337.0     |
| 92.00  | 167.0     |
| 95.00  | 234.0     |
| 95.90  | 594.0     |
| 96.80  | 245.0     |
| 102.00 | 182.0     |
| 103.00 | 361.0     |
| 105.10 | 281.0     |
| 107.10 | 171.0     |
| 115.00 | 396.0     |
| 117.00 | 245.0     |
| 119.00 | 240.0     |
| 127.90 | 166.0     |
| 129.00 | 189.0     |
| 131.00 | 164.0     |
| 133.00 | 1001.0    |
| 135.00 | 337.0     |
| 136.00 | 219.0     |
| 147.00 | 1264.0    |
| 148.00 | 291.0     |
| 149.00 | 335.0     |
| 151.00 | 175.0     |
| 160.90 | 162.0     |
| 162.90 | 312.0     |
| 164.90 | 162.0     |
| 176.90 | 413.0     |
| 178.00 | 178.0     |
| 178.90 | 207.0     |
| 190.90 | 1153.0    |
| 192.00 | 303.0     |
| 193.10 | 1037.0    |
| 193.90 | 200.0     |
| 206.10 | 177.0     |
| 207.00 | 8821.0    |
| 208.00 | 1911.0    |
| 209.00 | 924.0     |
| 210.00 | 210.0     |
| 221.00 | 2062.0    |
| 222.00 | 503.0     |
| 222.90 | 295.0     |
| 248.80 | 236.0     |
| 253.00 | 212.0     |
| 265.00 | 235.0     |
| 266.90 | 545.0     |
| 267.80 | 246.0     |
| 281.10 | 2420.0    |
| 282.00 | 880.0     |
| 283.10 | 499.0     |
| 283.90 | 200.0     |
| 295.10 | 585.0     |
| 296.10 | 255.0     |
| 326.90 | 195.0     |
| 341.00 | 544.0     |
| 355.10 | 797.0     |
| 356.00 | 357.0     |
| 357.10 | 284.0     |
| 369.10 | 261.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP048

399.10

429.10

163.0

173.0

Instrument :

BNA\_M

ClientSampleId :

DFTPP048

| m/z    | Abundance |
|--------|-----------|
| 35.90  | 438.0     |
| 37.00  | 729.0     |
| 38.00  | 1729.0    |
| 39.00  | 11866.0   |
| 40.00  | 8638.0    |
| 41.10  | 979.0     |
| 41.90  | 291.0     |
| 43.00  | 763.0     |
| 44.00  | 6765.0    |
| 45.00  | 636.0     |
| 46.90  | 223.0     |
| 48.10  | 155.0     |
| 49.00  | 1485.0    |
| 50.00  | 39328.0   |
| 51.00  | 140672.0  |
| 52.00  | 6902.0    |
| 53.00  | 375.0     |
| 55.00  | 1420.0    |
| 56.00  | 4480.0    |
| 57.00  | 9533.0    |
| 58.00  | 559.0     |
| 58.90  | 377.0     |
| 60.00  | 281.0     |
| 61.00  | 1996.0    |
| 62.00  | 1861.0    |
| 63.00  | 6065.0    |
| 63.90  | 676.0     |
| 65.10  | 2774.0    |
| 66.90  | 387.0     |
| 68.00  | 2419.0    |
| 69.00  | 139584.0  |
| 70.00  | 980.0     |
| 73.10  | 3815.0    |
| 74.00  | 14676.0   |
| 75.00  | 22968.0   |
| 76.10  | 7117.0    |
| 77.00  | 154880.0  |
| 78.10  | 11736.0   |
| 79.00  | 11530.0   |
| 80.00  | 7779.0    |
| 81.00  | 11707.0   |
| 82.00  | 3019.0    |
| 83.00  | 2683.0    |
| 83.90  | 431.0     |
| 85.00  | 2193.0    |
| 86.00  | 3824.0    |
| 87.00  | 1700.0    |
| 88.00  | 535.0     |
| 88.90  | 379.0     |
| 90.90  | 2518.0    |
| 92.00  | 2911.0    |
| 93.00  | 18432.0   |
| 94.00  | 1422.0    |
| 94.90  | 356.0     |
| 96.10  | 1442.0    |
| 97.00  | 547.0     |
| 98.00  | 15640.0   |
| 99.00  | 10891.0   |
| 100.00 | 816.0     |
| 101.00 | 6361.0    |
| 101.80 | 340.0     |
| 102.10 | 329.0     |
| 103.00 | 2328.0    |
| 104.00 | 4012.0    |
| 105.00 | 4069.0    |
| 106.00 | 1224.0    |
| 107.00 | 49192.0   |
| 108.00 | 7482.0    |
| 109.00 | 1681.0    |
| 110.00 | 84312.0   |
| 111.00 | 13561.0   |
| 112.00 | 2018.0    |
| 112.90 | 710.0     |
| 115.10 | 445.0     |
| 116.00 | 2613.0    |
| 117.00 | 44488.0   |
| 118.00 | 3623.0    |
| 119.00 | 642.0     |
| 120.00 | 461.0     |
| 121.00 | 454.0     |
| 122.00 | 3208.0    |
| 123.00 | 4937.0    |
| 124.00 | 2125.0    |
| 125.10 | 2369.0    |
| 126.20 | 670.0     |
| 127.00 | 165632.0  |
| 128.00 | 13137.0   |
| 129.00 | 66128.0   |
| 130.00 | 5252.0    |
| 130.90 | 1507.0    |
| 132.00 | 887.0     |
| 133.00 | 1293.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP048

|        |          |
|--------|----------|
| 133.90 | 2447.0   |
| 135.00 | 5916.0   |
| 135.90 | 1910.0   |
| 137.00 | 2774.0   |
| 137.90 | 695.0    |
| 138.80 | 439.0    |
| 140.00 | 854.0    |
| 141.00 | 9016.0   |
| 142.00 | 2861.0   |
| 143.00 | 1847.0   |
| 144.00 | 597.0    |
| 145.00 | 489.0    |
| 146.00 | 2134.0   |
| 147.00 | 5960.0   |
| 148.00 | 11832.0  |
| 149.00 | 3092.0   |
| 150.00 | 718.0    |
| 151.10 | 1122.0   |
| 151.80 | 775.0    |
| 153.00 | 2744.0   |
| 154.00 | 2154.0   |
| 155.00 | 4828.0   |
| 156.00 | 7735.0   |
| 157.00 | 1476.0   |
| 158.00 | 1672.0   |
| 159.00 | 1468.0   |
| 160.00 | 3041.0   |
| 161.00 | 4884.0   |
| 162.00 | 1426.0   |
| 162.90 | 635.0    |
| 163.80 | 446.0    |
| 165.00 | 3676.0   |
| 166.00 | 2675.0   |
| 167.00 | 18744.0  |
| 168.00 | 10029.0  |
| 169.00 | 1577.0   |
| 170.00 | 612.0    |
| 171.00 | 874.0    |
| 171.90 | 1627.0   |
| 173.00 | 2501.0   |
| 174.00 | 4180.0   |
| 175.00 | 6977.0   |
| 176.00 | 1952.0   |
| 177.00 | 3923.0   |
| 178.00 | 1104.0   |
| 179.00 | 14315.0  |
| 180.00 | 9210.0   |
| 181.00 | 4378.0   |
| 181.90 | 864.0    |
| 183.00 | 353.0    |
| 184.00 | 1135.0   |
| 185.00 | 6470.0   |
| 186.00 | 47368.0  |
| 187.00 | 14062.0  |
| 187.90 | 1677.0   |
| 189.00 | 3490.0   |
| 190.00 | 775.0    |
| 191.00 | 2982.0   |
| 192.00 | 4339.0   |
| 193.00 | 6300.0   |
| 194.00 | 1556.0   |
| 195.10 | 986.0    |
| 196.00 | 9547.0   |
| 197.10 | 1407.0   |
| 198.00 | 325504.0 |
| 199.00 | 23472.0  |
| 200.00 | 1701.0   |
| 201.50 | 1185.0   |
| 203.00 | 2646.0   |
| 204.00 | 12652.0  |
| 205.00 | 20896.0  |
| 206.00 | 87760.0  |
| 207.00 | 22008.0  |
| 208.00 | 5140.0   |
| 209.00 | 2294.0   |
| 210.00 | 1586.0   |
| 211.10 | 3262.0   |
| 211.90 | 779.0    |
| 213.00 | 315.0    |
| 215.00 | 1263.0   |
| 216.00 | 2217.0   |
| 217.00 | 24720.0  |
| 218.00 | 3490.0   |
| 219.00 | 353.0    |
| 219.90 | 364.0    |
| 221.00 | 14000.0  |
| 222.00 | 3352.0   |
| 223.00 | 5382.0   |
| 224.00 | 48032.0  |
| 225.00 | 13056.0  |
| 226.00 | 1369.0   |
| 227.00 | 21424.0  |
| 228.00 | 2902.0   |
| 229.00 | 4035.0   |
| 229.90 | 405.0    |
| 231.00 | 1636.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP048

|        |          |
|--------|----------|
| 231.90 | 200.0    |
| 233.10 | 279.0    |
| 234.00 | 1382.0   |
| 235.00 | 1580.0   |
| 235.90 | 1237.0   |
| 237.00 | 1884.0   |
| 238.00 | 296.0    |
| 238.80 | 801.0    |
| 240.00 | 705.0    |
| 241.00 | 1130.0   |
| 242.00 | 2707.0   |
| 243.00 | 2947.0   |
| 244.00 | 37080.0  |
| 245.00 | 4957.0   |
| 246.00 | 7965.0   |
| 247.00 | 1909.0   |
| 247.90 | 520.0    |
| 249.00 | 1590.0   |
| 250.00 | 293.0    |
| 251.10 | 347.0    |
| 251.90 | 313.0    |
| 252.90 | 1168.0   |
| 254.10 | 1415.0   |
| 255.00 | 192256.0 |
| 256.00 | 29456.0  |
| 257.00 | 2311.0   |
| 258.00 | 11478.0  |
| 259.00 | 1845.0   |
| 259.90 | 441.0    |
| 260.90 | 285.0    |
| 262.90 | 171.0    |
| 263.90 | 246.0    |
| 265.00 | 5494.0   |
| 265.90 | 954.0    |
| 266.90 | 615.0    |
| 269.00 | 186.0    |
| 270.00 | 250.0    |
| 270.90 | 609.0    |
| 272.00 | 456.0    |
| 273.00 | 5982.0   |
| 274.00 | 14595.0  |
| 275.00 | 81384.0  |
| 276.00 | 11479.0  |
| 277.00 | 7247.0   |
| 277.90 | 1467.0   |
| 278.90 | 336.0    |
| 281.00 | 2998.0   |
| 282.00 | 880.0    |
| 283.00 | 1426.0   |
| 284.00 | 623.0    |
| 285.00 | 1138.0   |
| 286.00 | 225.0    |
| 288.90 | 245.0    |
| 289.90 | 302.0    |
| 292.10 | 264.0    |
| 293.00 | 1599.0   |
| 294.00 | 522.0    |
| 295.10 | 1167.0   |
| 296.00 | 24248.0  |
| 297.00 | 3711.0   |
| 297.90 | 362.0    |
| 300.90 | 355.0    |
| 302.00 | 600.0    |
| 303.00 | 2964.0   |
| 304.10 | 818.0    |
| 308.00 | 387.0    |
| 308.90 | 176.0    |
| 310.00 | 281.0    |
| 313.00 | 284.0    |
| 314.00 | 1179.0   |
| 315.00 | 2721.0   |
| 316.00 | 1219.0   |
| 317.00 | 338.0    |
| 321.00 | 807.0    |
| 322.00 | 277.0    |
| 323.10 | 7076.0   |
| 324.00 | 1596.0   |
| 324.90 | 322.0    |
| 326.00 | 171.0    |
| 327.00 | 1638.0   |
| 327.90 | 550.0    |
| 328.90 | 231.0    |
| 332.00 | 625.0    |
| 332.90 | 656.0    |
| 334.00 | 5300.0   |
| 335.00 | 1011.0   |
| 335.90 | 182.0    |
| 338.90 | 221.0    |
| 341.00 | 1578.0   |
| 342.10 | 360.0    |
| 343.10 | 185.0    |
| 346.10 | 1539.0   |
| 347.10 | 341.0    |
| 352.00 | 2391.0   |
| 353.00 | 1335.0   |
| 354.10 | 2802.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP048

|        |          |
|--------|----------|
| 355.10 | 1509.0   |
| 356.10 | 463.0    |
| 357.00 | 216.0    |
| 359.00 | 158.0    |
| 365.00 | 10836.0  |
| 365.90 | 1688.0   |
| 369.20 | 234.0    |
| 369.90 | 403.0    |
| 371.00 | 635.0    |
| 372.10 | 4179.0   |
| 373.00 | 809.0    |
| 383.00 | 1226.0   |
| 384.00 | 346.0    |
| 390.00 | 464.0    |
| 391.00 | 293.0    |
| 391.90 | 171.0    |
| 401.00 | 311.0    |
| 402.00 | 1577.0   |
| 403.00 | 2109.0   |
| 404.00 | 730.0    |
| 405.10 | 166.0    |
| 415.10 | 261.0    |
| 416.00 | 157.0    |
| 421.00 | 1663.0   |
| 422.00 | 1486.0   |
| 423.00 | 12301.0  |
| 424.00 | 3065.0   |
| 424.90 | 181.0    |
| 429.10 | 424.0    |
| 441.10 | 35072.0  |
| 442.10 | 229376.0 |
| 443.10 | 44736.0  |
| 444.00 | 4819.0   |
| 445.10 | 427.0    |
| 503.10 | 163.0    |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP048

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 534.0     |
| 37.00  | 973.0     |
| 38.00  | 2782.0    |
| 39.10  | 16195.0   |
| 40.00  | 7840.0    |
| 41.10  | 1169.0    |
| 42.00  | 393.0     |
| 43.00  | 543.0     |
| 44.00  | 6560.0    |
| 45.00  | 849.0     |
| 46.90  | 160.0     |
| 48.10  | 155.0     |
| 49.00  | 1642.0    |
| 50.00  | 54256.0   |
| 51.00  | 201920.0  |
| 52.10  | 10048.0   |
| 52.90  | 502.0     |
| 55.00  | 1438.0    |
| 56.00  | 6060.0    |
| 57.00  | 13373.0   |
| 57.90  | 610.0     |
| 59.00  | 313.0     |
| 60.10  | 286.0     |
| 61.00  | 2913.0    |
| 62.00  | 3044.0    |
| 63.00  | 8003.0    |
| 64.00  | 1446.0    |
| 65.10  | 3747.0    |
| 66.00  | 448.0     |
| 67.00  | 326.0     |
| 68.00  | 3305.0    |
| 69.00  | 207104.0  |
| 70.00  | 985.0     |
| 70.90  | 187.0     |
| 73.10  | 4461.0    |
| 74.00  | 20872.0   |
| 75.00  | 33056.0   |
| 76.00  | 11040.0   |
| 77.10  | 223936.0  |
| 78.10  | 15986.0   |
| 79.00  | 15624.0   |
| 80.00  | 11894.0   |
| 81.00  | 17504.0   |
| 82.00  | 4678.0    |
| 83.00  | 3779.0    |
| 84.00  | 737.0     |
| 85.00  | 2658.0    |
| 86.00  | 5181.0    |
| 87.00  | 2077.0    |
| 88.00  | 990.0     |
| 89.00  | 785.0     |
| 89.80  | 192.0     |
| 91.00  | 4307.0    |
| 92.00  | 4349.0    |
| 93.00  | 28864.0   |
| 93.90  | 2294.0    |
| 95.00  | 506.0     |
| 96.00  | 2054.0    |
| 97.00  | 757.0     |
| 98.00  | 23000.0   |
| 99.00  | 16680.0   |
| 100.00 | 1622.0    |
| 101.00 | 9917.0    |
| 102.00 | 674.0     |
| 103.00 | 3610.0    |
| 104.00 | 6386.0    |
| 105.00 | 6208.0    |
| 106.10 | 2115.0    |
| 107.00 | 70928.0   |
| 108.00 | 11010.0   |
| 109.10 | 2058.0    |
| 110.00 | 126040.0  |
| 111.00 | 20064.0   |
| 112.00 | 2405.0    |
| 112.90 | 769.0     |
| 114.00 | 180.0     |
| 114.90 | 550.0     |
| 116.00 | 3867.0    |
| 117.00 | 66576.0   |
| 118.00 | 4728.0    |
| 118.90 | 701.0     |
| 120.00 | 916.0     |
| 120.90 | 500.0     |
| 122.00 | 5841.0    |
| 123.00 | 8199.0    |
| 123.90 | 3597.0    |
| 125.00 | 3240.0    |
| 126.20 | 1208.0    |
| 127.00 | 248896.0  |
| 128.00 | 19848.0   |
| 129.00 | 102056.0  |
| 130.00 | 8392.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP048

|        |          |
|--------|----------|
| 131.00 | 1643.0   |
| 132.00 | 729.0    |
| 133.00 | 1158.0   |
| 133.90 | 2996.0   |
| 135.00 | 9368.0   |
| 136.00 | 3512.0   |
| 137.00 | 4510.0   |
| 138.00 | 911.0    |
| 138.90 | 587.0    |
| 140.00 | 1401.0   |
| 141.00 | 13704.0  |
| 142.00 | 4215.0   |
| 143.00 | 3083.0   |
| 144.00 | 774.0    |
| 144.90 | 848.0    |
| 146.00 | 2861.0   |
| 147.00 | 9090.0   |
| 148.00 | 17584.0  |
| 149.00 | 3834.0   |
| 150.00 | 861.0    |
| 151.00 | 1854.0   |
| 151.90 | 1109.0   |
| 153.00 | 4545.0   |
| 154.00 | 3419.0   |
| 155.00 | 7618.0   |
| 156.00 | 11798.0  |
| 157.00 | 2410.0   |
| 158.00 | 2539.0   |
| 158.90 | 1841.0   |
| 160.00 | 4721.0   |
| 161.00 | 7072.0   |
| 162.00 | 1863.0   |
| 163.00 | 787.0    |
| 163.90 | 869.0    |
| 165.00 | 5616.0   |
| 166.00 | 4212.0   |
| 167.00 | 28168.0  |
| 168.00 | 14469.0  |
| 169.00 | 2431.0   |
| 170.00 | 1111.0   |
| 170.90 | 1394.0   |
| 171.90 | 2223.0   |
| 173.00 | 3607.0   |
| 174.00 | 5865.0   |
| 175.00 | 11936.0  |
| 176.00 | 2998.0   |
| 177.00 | 5262.0   |
| 177.90 | 1861.0   |
| 179.00 | 21576.0  |
| 180.00 | 15361.0  |
| 181.00 | 6732.0   |
| 181.90 | 1279.0   |
| 182.90 | 480.0    |
| 183.90 | 1992.0   |
| 185.00 | 10658.0  |
| 186.00 | 79816.0  |
| 187.00 | 22560.0  |
| 188.00 | 2613.0   |
| 189.00 | 5179.0   |
| 189.90 | 970.0    |
| 191.00 | 3411.0   |
| 192.00 | 7066.0   |
| 193.00 | 9275.0   |
| 194.00 | 2062.0   |
| 195.00 | 799.0    |
| 196.00 | 13378.0  |
| 197.00 | 2586.0   |
| 198.00 | 530496.0 |
| 199.00 | 35600.0  |
| 200.00 | 3263.0   |
| 201.40 | 2367.0   |
| 203.00 | 4317.0   |
| 204.00 | 22544.0  |
| 205.00 | 35160.0  |
| 206.00 | 149248.0 |
| 207.00 | 29744.0  |
| 208.00 | 7228.0   |
| 209.00 | 2917.0   |
| 210.00 | 2144.0   |
| 211.00 | 6103.0   |
| 213.10 | 286.0    |
| 215.00 | 1937.0   |
| 216.00 | 3508.0   |
| 217.00 | 40400.0  |
| 218.00 | 4856.0   |
| 218.90 | 539.0    |
| 220.20 | 1114.0   |
| 221.10 | 22112.0  |
| 222.00 | 5423.0   |
| 223.00 | 9514.0   |
| 224.00 | 81560.0  |
| 225.00 | 21568.0  |
| 226.00 | 2031.0   |
| 227.00 | 38344.0  |
| 228.00 | 4877.0   |
| 229.00 | 8251.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP048

|        |          |
|--------|----------|
| 230.00 | 869.0    |
| 231.00 | 2675.0   |
| 232.00 | 484.0    |
| 233.00 | 640.0    |
| 234.00 | 2524.0   |
| 235.00 | 2575.0   |
| 236.00 | 1525.0   |
| 237.00 | 2784.0   |
| 237.80 | 430.0    |
| 239.00 | 1505.0   |
| 240.00 | 1029.0   |
| 241.00 | 2229.0   |
| 242.00 | 4833.0   |
| 243.10 | 4531.0   |
| 244.00 | 61944.0  |
| 245.00 | 8465.0   |
| 246.00 | 14381.0  |
| 247.00 | 2613.0   |
| 248.10 | 645.0    |
| 249.00 | 2747.0   |
| 249.90 | 541.0    |
| 250.90 | 948.0    |
| 251.90 | 552.0    |
| 253.00 | 1572.0   |
| 254.10 | 2109.0   |
| 255.00 | 332608.0 |
| 256.00 | 47032.0  |
| 257.00 | 3955.0   |
| 258.00 | 20248.0  |
| 259.00 | 3318.0   |
| 259.90 | 365.0    |
| 261.00 | 387.0    |
| 263.00 | 266.0    |
| 264.00 | 429.0    |
| 265.00 | 8629.0   |
| 266.00 | 1135.0   |
| 266.90 | 854.0    |
| 268.00 | 256.0    |
| 268.90 | 250.0    |
| 269.90 | 497.0    |
| 271.10 | 692.0    |
| 272.00 | 961.0    |
| 273.00 | 10061.0  |
| 274.00 | 24376.0  |
| 275.00 | 151808.0 |
| 276.00 | 20104.0  |
| 277.00 | 13337.0  |
| 278.00 | 2330.0   |
| 279.00 | 508.0    |
| 281.10 | 2917.0   |
| 282.00 | 1154.0   |
| 283.00 | 1951.0   |
| 284.00 | 1344.0   |
| 285.00 | 2575.0   |
| 286.00 | 499.0    |
| 287.90 | 184.0    |
| 289.00 | 694.0    |
| 289.90 | 381.0    |
| 291.00 | 345.0    |
| 291.80 | 640.0    |
| 293.00 | 3080.0   |
| 294.00 | 685.0    |
| 295.00 | 1488.0   |
| 296.00 | 46312.0  |
| 297.00 | 6329.0   |
| 298.00 | 530.0    |
| 300.80 | 432.0    |
| 302.00 | 823.0    |
| 303.00 | 5093.0   |
| 304.00 | 1405.0   |
| 304.90 | 325.0    |
| 308.00 | 496.0    |
| 308.90 | 278.0    |
| 310.00 | 481.0    |
| 311.00 | 184.0    |
| 313.00 | 385.0    |
| 314.00 | 1837.0   |
| 315.00 | 5663.0   |
| 316.00 | 2540.0   |
| 317.00 | 702.0    |
| 320.00 | 220.0    |
| 321.00 | 1364.0   |
| 322.00 | 734.0    |
| 323.00 | 13721.0  |
| 324.00 | 2465.0   |
| 324.80 | 182.0    |
| 326.00 | 302.0    |
| 327.00 | 3403.0   |
| 327.90 | 1333.0   |
| 328.90 | 265.0    |
| 332.00 | 1004.0   |
| 333.00 | 1241.0   |
| 334.00 | 8662.0   |
| 335.00 | 2373.0   |
| 335.90 | 332.0    |
| 339.00 | 312.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP048

|        |          |
|--------|----------|
| 340.00 | 200.0    |
| 341.00 | 1827.0   |
| 342.00 | 607.0    |
| 343.00 | 243.0    |
| 346.00 | 3168.0   |
| 346.90 | 728.0    |
| 350.90 | 258.0    |
| 352.00 | 4170.0   |
| 353.00 | 2774.0   |
| 354.00 | 4457.0   |
| 355.10 | 1505.0   |
| 356.00 | 320.0    |
| 357.20 | 197.0    |
| 358.00 | 169.0    |
| 358.90 | 247.0    |
| 365.00 | 21304.0  |
| 366.00 | 2518.0   |
| 366.80 | 177.0    |
| 369.10 | 291.0    |
| 370.00 | 420.0    |
| 371.10 | 1085.0   |
| 372.10 | 7434.0   |
| 373.10 | 1756.0   |
| 373.90 | 344.0    |
| 383.00 | 2219.0   |
| 384.00 | 507.0    |
| 385.00 | 212.0    |
| 390.00 | 938.0    |
| 391.00 | 535.0    |
| 391.90 | 633.0    |
| 401.00 | 554.0    |
| 402.00 | 2642.0   |
| 403.10 | 4393.0   |
| 404.00 | 1379.0   |
| 405.00 | 200.0    |
| 415.10 | 277.0    |
| 421.00 | 3120.0   |
| 422.00 | 3711.0   |
| 423.10 | 25680.0  |
| 424.10 | 5025.0   |
| 425.00 | 541.0    |
| 429.20 | 361.0    |
| 441.10 | 73672.0  |
| 442.10 | 477184.0 |
| 443.10 | 92040.0  |
| 444.10 | 9403.0   |
| 445.00 | 543.0    |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP048

| m/z    | Abundance |
|--------|-----------|
| 36.00  | 500.0     |
| 37.00  | 740.0     |
| 38.00  | 2499.0    |
| 39.10  | 13669.0   |
| 40.00  | 8601.0    |
| 41.10  | 898.0     |
| 41.90  | 316.0     |
| 43.00  | 744.0     |
| 44.00  | 6917.0    |
| 45.00  | 619.0     |
| 49.00  | 1294.0    |
| 50.10  | 42024.0   |
| 51.10  | 157376.0  |
| 52.00  | 8242.0    |
| 53.00  | 524.0     |
| 54.10  | 178.0     |
| 55.00  | 1149.0    |
| 56.00  | 4633.0    |
| 57.00  | 10824.0   |
| 58.10  | 403.0     |
| 59.00  | 407.0     |
| 60.10  | 176.0     |
| 61.00  | 2190.0    |
| 62.00  | 2280.0    |
| 63.00  | 6077.0    |
| 64.00  | 847.0     |
| 65.10  | 3164.0    |
| 65.90  | 314.0     |
| 67.10  | 422.0     |
| 68.00  | 2676.0    |
| 69.00  | 162688.0  |
| 70.00  | 1196.0    |
| 71.10  | 272.0     |
| 72.00  | 192.0     |
| 73.10  | 3669.0    |
| 74.00  | 17552.0   |
| 75.00  | 25920.0   |
| 76.10  | 8778.0    |
| 77.10  | 181376.0  |
| 78.00  | 12331.0   |
| 79.00  | 13031.0   |
| 80.00  | 9983.0    |
| 81.00  | 13378.0   |
| 82.00  | 3508.0    |
| 83.00  | 2761.0    |
| 83.90  | 691.0     |
| 85.00  | 2279.0    |
| 86.00  | 4084.0    |
| 87.00  | 1749.0    |
| 87.90  | 762.0     |
| 88.90  | 521.0     |
| 91.00  | 3495.0    |
| 92.00  | 3416.0    |
| 93.00  | 22568.0   |
| 94.00  | 1695.0    |
| 94.90  | 394.0     |
| 96.00  | 1805.0    |
| 97.00  | 607.0     |
| 98.00  | 17504.0   |
| 99.00  | 13207.0   |
| 100.00 | 1068.0    |
| 101.00 | 7552.0    |
| 102.00 | 527.0     |
| 103.00 | 2529.0    |
| 104.00 | 4760.0    |
| 105.00 | 4931.0    |
| 105.90 | 1520.0    |
| 107.00 | 59064.0   |
| 108.00 | 9016.0    |
| 109.10 | 1989.0    |
| 110.00 | 103832.0  |
| 111.00 | 15406.0   |
| 112.00 | 2086.0    |
| 113.00 | 727.0     |
| 114.90 | 484.0     |
| 116.00 | 3297.0    |
| 117.00 | 55408.0   |
| 118.00 | 3734.0    |
| 119.00 | 818.0     |
| 120.00 | 784.0     |
| 121.00 | 339.0     |
| 122.00 | 4258.0    |
| 123.00 | 6524.0    |
| 124.00 | 3485.0    |
| 125.00 | 2465.0    |
| 126.10 | 668.0     |
| 127.00 | 212416.0  |
| 128.00 | 16187.0   |
| 129.00 | 84264.0   |
| 130.00 | 6964.0    |
| 131.00 | 1748.0    |
| 132.00 | 834.0     |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP048

|        |          |
|--------|----------|
| 133.00 | 1156.0   |
| 134.00 | 2747.0   |
| 135.00 | 7700.0   |
| 136.00 | 3605.0   |
| 137.00 | 3506.0   |
| 138.00 | 877.0    |
| 138.90 | 479.0    |
| 140.00 | 1037.0   |
| 141.00 | 11471.0  |
| 142.00 | 4114.0   |
| 142.90 | 2861.0   |
| 144.00 | 811.0    |
| 144.90 | 428.0    |
| 146.00 | 1968.0   |
| 147.00 | 7232.0   |
| 148.00 | 15522.0  |
| 149.00 | 3601.0   |
| 150.00 | 1032.0   |
| 150.90 | 1553.0   |
| 151.90 | 1074.0   |
| 153.00 | 4120.0   |
| 154.00 | 2781.0   |
| 155.00 | 6153.0   |
| 156.00 | 10070.0  |
| 157.00 | 1894.0   |
| 158.00 | 2259.0   |
| 159.00 | 2012.0   |
| 160.00 | 3733.0   |
| 161.00 | 5618.0   |
| 162.00 | 1478.0   |
| 163.00 | 774.0    |
| 164.00 | 665.0    |
| 165.00 | 4663.0   |
| 166.00 | 3658.0   |
| 167.00 | 24320.0  |
| 168.00 | 12491.0  |
| 169.00 | 2398.0   |
| 170.00 | 889.0    |
| 171.00 | 1113.0   |
| 171.90 | 2129.0   |
| 173.00 | 3179.0   |
| 174.00 | 5847.0   |
| 175.00 | 10226.0  |
| 176.00 | 3157.0   |
| 177.00 | 4109.0   |
| 177.90 | 1862.0   |
| 179.00 | 19056.0  |
| 180.00 | 12370.0  |
| 181.00 | 5826.0   |
| 182.00 | 1072.0   |
| 183.00 | 584.0    |
| 183.90 | 1543.0   |
| 185.00 | 9000.0   |
| 186.00 | 69552.0  |
| 187.00 | 20368.0  |
| 188.00 | 2212.0   |
| 189.00 | 4533.0   |
| 190.00 | 1011.0   |
| 191.00 | 3395.0   |
| 192.00 | 6062.0   |
| 193.00 | 7878.0   |
| 194.00 | 2038.0   |
| 195.00 | 847.0    |
| 196.00 | 14223.0  |
| 197.10 | 2823.0   |
| 198.00 | 480768.0 |
| 199.00 | 31536.0  |
| 200.00 | 2903.0   |
| 201.40 | 2525.0   |
| 203.00 | 4101.0   |
| 204.00 | 19088.0  |
| 205.00 | 33344.0  |
| 206.00 | 132352.0 |
| 207.00 | 27408.0  |
| 208.00 | 6924.0   |
| 209.00 | 2573.0   |
| 210.00 | 2213.0   |
| 211.10 | 5569.0   |
| 211.90 | 664.0    |
| 213.10 | 546.0    |
| 213.90 | 229.0    |
| 214.90 | 1415.0   |
| 216.00 | 2956.0   |
| 217.00 | 37656.0  |
| 218.00 | 4734.0   |
| 219.00 | 571.0    |
| 221.00 | 19000.0  |
| 222.00 | 5097.0   |
| 223.00 | 9084.0   |
| 224.00 | 76384.0  |
| 225.00 | 19736.0  |
| 226.00 | 2022.0   |
| 227.00 | 33680.0  |
| 228.00 | 4871.0   |
| 229.00 | 7040.0   |
| 230.00 | 1064.0   |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP048

|        |          |
|--------|----------|
| 231.00 | 2951.0   |
| 232.10 | 376.0    |
| 233.00 | 826.0    |
| 234.00 | 2660.0   |
| 235.00 | 2105.0   |
| 236.00 | 1891.0   |
| 236.90 | 2334.0   |
| 237.90 | 502.0    |
| 239.00 | 1600.0   |
| 239.90 | 1109.0   |
| 241.00 | 1812.0   |
| 242.00 | 3774.0   |
| 243.10 | 3762.0   |
| 244.00 | 60640.0  |
| 245.00 | 7810.0   |
| 246.00 | 13191.0  |
| 247.00 | 2332.0   |
| 248.00 | 643.0    |
| 249.00 | 2682.0   |
| 249.90 | 611.0    |
| 251.00 | 700.0    |
| 252.00 | 602.0    |
| 253.00 | 1405.0   |
| 254.10 | 2146.0   |
| 255.00 | 320448.0 |
| 256.00 | 44312.0  |
| 257.00 | 3389.0   |
| 258.00 | 19608.0  |
| 259.00 | 3136.0   |
| 260.00 | 567.0    |
| 260.80 | 443.0    |
| 263.10 | 236.0    |
| 263.90 | 438.0    |
| 265.00 | 8101.0   |
| 266.00 | 1455.0   |
| 267.10 | 902.0    |
| 268.10 | 252.0    |
| 269.00 | 252.0    |
| 269.90 | 277.0    |
| 270.90 | 858.0    |
| 272.00 | 749.0    |
| 273.00 | 10268.0  |
| 274.00 | 24968.0  |
| 275.00 | 143296.0 |
| 276.00 | 18984.0  |
| 277.00 | 13162.0  |
| 278.00 | 1922.0   |
| 279.00 | 549.0    |
| 281.10 | 2694.0   |
| 282.00 | 1022.0   |
| 283.10 | 1648.0   |
| 284.00 | 975.0    |
| 285.00 | 2446.0   |
| 286.00 | 365.0    |
| 287.90 | 171.0    |
| 289.00 | 720.0    |
| 290.00 | 537.0    |
| 291.00 | 244.0    |
| 292.00 | 632.0    |
| 293.00 | 3022.0   |
| 294.00 | 725.0    |
| 295.00 | 1431.0   |
| 296.00 | 45928.0  |
| 297.00 | 6996.0   |
| 297.90 | 356.0    |
| 300.90 | 625.0    |
| 301.90 | 1001.0   |
| 303.00 | 4857.0   |
| 303.90 | 1423.0   |
| 305.00 | 166.0    |
| 305.80 | 153.0    |
| 308.00 | 604.0    |
| 308.90 | 455.0    |
| 310.00 | 589.0    |
| 312.00 | 173.0    |
| 312.90 | 580.0    |
| 314.00 | 2133.0   |
| 315.00 | 5549.0   |
| 316.00 | 2895.0   |
| 317.10 | 439.0    |
| 321.00 | 1579.0   |
| 322.10 | 1004.0   |
| 323.10 | 13929.0  |
| 324.10 | 2370.0   |
| 325.00 | 372.0    |
| 325.90 | 337.0    |
| 327.00 | 3000.0   |
| 328.00 | 1277.0   |
| 329.00 | 252.0    |
| 330.00 | 201.0    |
| 332.00 | 1207.0   |
| 333.10 | 1274.0   |
| 334.00 | 9106.0   |
| 335.00 | 2507.0   |
| 336.00 | 339.0    |
| 339.10 | 289.0    |

Instrument :  
BNA\_M  
ClientSampleId :  
DFTPP048

|        |          |
|--------|----------|
| 340.00 | 234.0    |
| 341.00 | 2266.0   |
| 342.00 | 690.0    |
| 342.90 | 316.0    |
| 346.00 | 3420.0   |
| 346.90 | 647.0    |
| 350.90 | 195.0    |
| 352.00 | 4774.0   |
| 353.00 | 2867.0   |
| 354.10 | 4854.0   |
| 355.10 | 1707.0   |
| 355.90 | 295.0    |
| 356.90 | 338.0    |
| 359.00 | 283.0    |
| 365.00 | 22048.0  |
| 365.90 | 3016.0   |
| 366.90 | 300.0    |
| 369.30 | 238.0    |
| 369.90 | 666.0    |
| 371.00 | 1294.0   |
| 372.10 | 7768.0   |
| 373.00 | 1923.0   |
| 374.00 | 228.0    |
| 376.90 | 196.0    |
| 383.00 | 2067.0   |
| 383.80 | 499.0    |
| 384.90 | 153.0    |
| 390.00 | 1017.0   |
| 391.10 | 685.0    |
| 391.80 | 347.0    |
| 401.00 | 529.0    |
| 402.00 | 2988.0   |
| 403.00 | 4608.0   |
| 404.00 | 1763.0   |
| 404.90 | 276.0    |
| 415.00 | 500.0    |
| 421.00 | 3980.0   |
| 422.00 | 3681.0   |
| 423.10 | 28032.0  |
| 424.10 | 6033.0   |
| 424.90 | 792.0    |
| 429.20 | 243.0    |
| 440.20 | 291.0    |
| 441.10 | 87616.0  |
| 442.10 | 567488.0 |
| 443.10 | 108808.0 |
| 444.00 | 9836.0   |
| 445.10 | 595.0    |
| 489.10 | 164.0    |
| 503.10 | 159.0    |

**Instrument :**  
BNA\_M  
**ClientSampleId :**  
DFTPP048