

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042024\
 Data File : BM045427.D
 Acq On : 20 Apr 2024 16:46
 Operator : MA/JU
 Sample : SSTD3.226
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2034

Quant Time: Apr 21 05:06:07 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM042024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 20 17:19:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.576	152	867	0.400	ng/ul	0.00
4) Naphthalene-d8	10.336	136	2698	0.400	ng/ul	#-0.04
9) Acenaphthene-d10	14.214	164	1510	0.400	ng/ul	-0.01
13) Phenanthrene-d10	16.980	188	2977	0.400	ng/ul	#-0.03
17) Chrysene-d12	21.196	240	1883	0.400	ng/ul	-0.02
23) Perylene-d12	23.411	264	2129	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.171	96	3209	2.761	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.201	88	3292	2.945	ng/ul#	80
14) Pentachlorophenol	16.642	266	2093	3.542	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042024\
Data File : BM045427.D
Acq On : 20 Apr 2024 16:46
Operator : MA/JU
Sample : SST3.226
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD3.2034

Quant Time: Apr 21 05:06:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM042024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 20 17:19:38 2024
Response via : Initial Calibration

