



2-Nitrophenol



Response = $2.023 \times 10^{-1} \times \text{Amt} - 3.229 \times 10^{-2}$
 Coef of Det (r^2) = 0.995497 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM042121.M
 Calibration Table Last Updated: Thu Apr 22 15:55:11 2021

Benzoic acid

Response Ratio



$$\text{Response} = 2.702\text{e-}001 * \text{Amt} - 9.746\text{e-}002$$


Coef of Det (r^2) = 0.993682 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM042121.M

Calibration Table Last Updated: Thu Apr 22 15:55:11 2021

2-Nitroaniline

Response Ratio



$$\text{Response} = 3.476\text{e-}001 * \text{Amt} - 4.970\text{e-}002$$


Coef of Det (r^2) = 0.995234 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM042121.M

Calibration Table Last Updated: Thu Apr 22 15:55:11 2021

3-Nitroaniline

Response Ratio




$$\text{Response} = 3.294\text{e-}001 * \text{Amt} - 4.910\text{e-}002$$

Coef of Det (r^2) = 0.995482 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM042121.M

Calibration Table Last Updated: Thu Apr 22 15:55:11 2021


2,4-Dinitrophenol



Response = 2.243e-001 * Amt - 4.897e-002
 Coef of Det (r^2) = 0.997952 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM042121.M
 Calibration Table Last Updated: Thu Apr 22 15:55:11 2021

4-Nitroaniline

Response Ratio



$$\text{Response} = 3.574\text{e-}001 * \text{Amt} - 6.175\text{e-}002$$

Coef of Det (r^2) = 0.996119 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA M\Methods\8270-BM042121.M

Calibration Table Last Updated: Thu Apr 22 15:55:11 2021