

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042215\
 Data File : BM001107.D
 Acq On : 22 Apr 2015 14:19
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDCCC001

Quant Time: Apr 23 11:10:30 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 08:05:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	22545	5.00	ng	-0.01
4) Naphthalene-d8	10.35	136	79418	5.00	ng	-0.01
8) Acenaphthene-d10	14.22	164	38608	5.00	ng	-0.02
13) Phenanthrene-d10	16.97	188	76777	5.00	ng	-0.02
19) Chrysene-d12	21.18	240	65800	5.00	ng	-0.01
25) Perylene-d12	23.35	264	62803	5.00	ng	-0.01
System Monitoring Compounds						
5) Nitrobenzene-d5	8.73	82	5047	1.22	ng	0.00
9) 2-Fluorobiphenyl	12.85	172	12748	1.00	ng	-0.01
21) Terphenyl-d14	19.62	244	8790	1.01	ng	-0.01
Target Compounds						Ovalue
2) 1,4-Dioxane	3.07	88	2353	0.93	ng	# 100
3) n-Nitrosodimethylamine	3.39	42	2888	1.21	ng	92
6) Naphthalene	10.40	128	18489	0.98	ng	99
7) 2-Methylnaphthalene	12.03	142	11824	0.97	ng	96
10) Acenaphthylene	13.93	152	18643	1.09	ng	99
11) Acenaphthene	14.29	154	11398	0.95	ng	99
12) Fluorene	15.29	166	13598	0.99	ng	98
14) Hexachlorobenzene	16.29	284	4372	1.03	ng	# 100
15) Pentachlorophenol	16.64	266	1655	1.34	ng	93
16) Phenanthrene	17.01	178	20810	1.02	ng	98
17) Anthracene	17.11	178	19300	1.03	ng	99
18) Fluoranthene	19.04	202	21528	1.01	ng	100
20) Pyrene	19.42	202	22595	1.04	ng	100
22) Benzo(a)anthracene	21.17	228	20111	1.05	ng	98
23) Chrysene	21.21	228	19248m	0.96	ng	
24) Indeno(1,2,3-cd)pyrene	25.52	276	23441	1.14	ng	98
26) Benzo(b)fluoranthene	22.70	252	20317	1.01	ng	99
27) Benzo(k)fluoranthene	22.74	252	19008	0.98	ng	100
28) Benzo(a)pyrene	23.25	252	18405	1.07	ng	100
29) Dibenzo(a,h)anthracene	25.53	278	18115	1.06	ng	99
30) Benzo(g,h,i)perylene	26.17	276	19564	1.06	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042215\
Data File : BM001107.D
Acq On : 22 Apr 2015 14:19
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDCCC001

Quant Time: Apr 23 11:10:30 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 20 08:05:22 2015
Response via : Initial Calibration

