

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042215\
 Data File : BM001109.D
 Acq On : 22 Apr 2015 16:22
 Operator : TP/IZ
 Sample : PB82802BSD
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB82802BSD

Quant Time: Apr 23 03:45:09 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 08:05:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	17304	5.00	ng	0.00
4) Naphthalene-d8	10.35	136	58993	5.00	ng	-0.01
8) Acenaphthene-d10	14.22	164	29000	5.00	ng	-0.02
13) Phenanthrene-d10	16.97	188	55830	5.00	ng	-0.02
19) Chrysene-d12	21.18	240	54628	5.00	ng	-0.01
25) Perylene-d12	23.35	264	52342	5.00	ng	-0.01
System Monitoring Compounds						
5) Nitrobenzene-d5	8.73	82	28017	9.09	ng	0.00
9) 2-Fluorobiphenyl	12.85	172	66103	6.90	ng	-0.01
21) Terphenyl-d14	19.62	244	48969	6.76	ng	-0.01
Target Compounds						Ovalue
2) 1,4-Dioxane	3.07	88	10330	5.29	ng	# 100
3) n-Nitrosodimethylamine	3.38	42	15049	8.23	ng	87
6) Naphthalene	10.40	128	87273	6.20	ng	98
7) 2-Methylnaphthalene	12.03	142	55081	6.07	ng	96
10) Acenaphthylene	13.93	152	92019	7.18	ng	99
11) Acenaphthene	14.29	154	56632	6.31	ng	95
12) Fluorene	15.29	166	66922	6.48	ng	99
14) Hexachlorobenzene	16.29	284	21934	7.11	ng	# 100
15) Pentachlorophenol	16.64	266	24805	22.07	ng	91
16) Phenanthrene	17.01	178	99326	6.70	ng	98
17) Anthracene	17.11	178	101685	7.46	ng	99
18) Fluoranthene	19.04	202	112845	7.30	ng	100
20) Pyrene	19.40	202	120912	6.68	ng	99
22) Benzo(a)anthracene	21.16	228	108966	6.88	ng	# 91
23) Chrysene	21.21	228	104513m	6.27	ng	
24) Indeno(1,2,3-cd)pyrene	25.51	276	118622	6.97	ng	99
26) Benzo(b)fluoranthene	22.70	252	114029	6.78	ng	98
27) Benzo(k)fluoranthene	22.74	252	96644	5.99	ng	98
28) Benzo(a)pyrene	23.24	252	102751	7.15	ng	96
29) Dibenzo(a,h)anthracene	25.53	278	95333	6.70	ng	96
30) Benzo(g,h,i)perylene	26.17	276	100435	6.52	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042215\
Data File : BM001109.D
Acq On : 22 Apr 2015 16:22
Operator : TP/IZ
Sample : PB82802BSD
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB82802BSD

Quant Time: Apr 23 03:45:09 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 20 08:05:22 2015
Response via : Initial Calibration

