

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042215\
 Data File : BM001108.D
 Acq On : 22 Apr 2015 15:47
 Operator : TP/IZ
 Sample : PB82802BS
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB82802BS

Quant Time: Apr 22 16:56:33 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 20 08:05:22 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	16594	5.00	ng	-0.01
4) Naphthalene-d8	10.35	136	56192	5.00	ng	-0.01
8) Acenaphthene-d10	14.22	164	26962	5.00	ng	-0.02
13) Phenanthrene-d10	16.97	188	55237	5.00	ng	-0.02
19) Chrysene-d12	21.18	240	54550	5.00	ng	-0.01
25) Perylene-d12	23.35	264	55695	5.00	ng	-0.01
System Monitoring Compounds						
5) Nitrobenzene-d5	8.73	82	29522	10.06	ng	0.00
9) 2-Fluorobiphenyl	12.85	172	69511	7.81	ng	-0.01
21) Terphenyl-d14	19.62	244	52790	7.30	ng	-0.01
Target Compounds						Ovalue
2) 1,4-Dioxane	3.07	88	10765	5.75	ng	# 100
3) n-Nitrosodimethylamine	3.38	42	15451	8.82	ng	# 87
6) Naphthalene	10.40	128	88557	6.60	ng	98
7) 2-Methylnaphthalene	12.03	142	56064	6.48	ng	97
10) Acenaphthylene	13.93	152	93407	7.84	ng	99
11) Acenaphthene	14.29	154	57804	6.93	ng	95
12) Fluorene	15.29	166	68022	7.08	ng	100
14) Hexachlorobenzene	16.29	284	22425	7.35	ng	# 100
15) Pentachlorophenol	16.64	266	26368	23.70	ng	93
16) Phenanthrene	17.01	178	98855	6.74	ng	98
17) Anthracene	17.11	178	103883	7.71	ng	99
18) Fluoranthene	19.05	202	117987	7.71	ng	100
20) Pyrene	19.42	202	126293	6.99	ng	99
22) Benzo(a)anthracene	21.17	228	116625	7.38	ng	99
23) Chrysene	21.22	228	112394	6.75	ng	99
24) Indeno(1,2,3-cd)pyrene	25.52	276	130388	7.67	ng	99
26) Benzo(b)fluoranthene	22.70	252	120547	6.73	ng	97
27) Benzo(k)fluoranthene	22.74	252	109792	6.39	ng	98
28) Benzo(a)pyrene	23.25	252	112685	7.36	ng	97
29) Dibenzo(a,h)anthracene	25.53	278	104582	6.91	ng	97
30) Benzo(g,h,i)perylene	26.17	276	110378	6.73	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042215\
Data File : BM001108.D
Acq On : 22 Apr 2015 15:47
Operator : TP/IZ
Sample : PB82802BS
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB82802BS

Quant Time: Apr 22 16:56:33 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIMPAH-BM041815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 20 08:05:22 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.57 min Scan# 629

Delta R.T. -0.01 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion:152 Resp: 16594

Ion Ratio Lower Upper

152 100

150 161.2 123.3 184.9




115 61.6 45.0 67.6



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 5.75 ng

RT: 3.07 min Scan# 38

Delta R.T. -0.01 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion: 88 Resp: 10765

Ion Ratio Lower Upper

88 100

58 87.3 0.0 0.0#



43 39.9 0.0 0.0#




Abundance Ion 88.00 (87.70 to 88.70): BM0

Ion 58.00 (57.70 to 58.70): BM0

Ion 43.00 (42.70 to 43.70): BM0





#3

n-Nitrosodimethylamine

Concen: 8.82 ng

RT: 3.38 min Scan# 68

Delta R.T. -0.02 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion: 42 Resp: 15451

Ion Ratio Lower Upper

42 100

74 100.7 92.1 138.1




44 5.5 5.8 8.6#



Abundance Ion 42.00 (41.70 to 42.70): BM001108.D (-22) (-)

Ion 74.00 (73.70 to 74.70): BM001108.D (-22) (-)

Ion 44.00 (43.70 to 44.70): BM001108.D (-22) (-)



#4

Naphthalene-d8

Concen: 5.00 ng

RT: 10.35 min Scan# 1073

Delta R.T. -0.01 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion: 136 Resp: 56192

Ion Ratio Lower Upper

136 100

137 10.5 8.6 13.0



54 11.2 6.5 9.7#




Abundance Ion 136.00 (135.70 to 136.70): BM001108.D (-1050) (-)

Ion 137.00 (136.70 to 137.70): BM001108.D (-1050) (-)

Ion 54.00 (53.70 to 54.70): BM001108.D (-1050) (-)





#5

Nitrobenzene-d5

Concen: 10.06 ng

RT: 8.73 min Scan# 845

Delta R.T. -0.01 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion: 82 Resp: 29522

Ion Ratio Lower Upper

82 100

128 41.6 31.9 47.9




54 62.4 51.3 76.9



Abundance Ion 82.00 (81.70 to 82.70): BM001108.D (-792) (-)

Ion 128.00 (127.70 to 128.70): BM001108.D (-792) (-)

Ion 54.00 (53.70 to 54.70): BM001108.D (-792) (-)



#6

Naphthalene

Concen: 6.60 ng

RT: 10.40 min Scan# 1077

Delta R.T. -0.01 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion: 128 Resp: 88557

Ion Ratio Lower Upper

128 100

129 10.6 9.3 13.9



127 12.9 10.0 15.0




Abundance Ion 128.00 (127.70 to 128.70): BM001108.D (-1053) (-)

Ion 129.00 (128.70 to 129.70): BM001108.D (-1053) (-)

Ion 127.00 (126.70 to 127.70): BM001108.D (-1053) (-)





#7

2-Methylnaphthalene

Concen: 6.48 ng

RT: 12.03 min Scan# 1223

Delta R.T. -0.01 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion:142 Resp: 56064

Ion Ratio Lower Upper

142 100

141 86.0 66.7 100.1




115 34.3 25.5 38.3



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#8

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.22 min Scan# 1407

Delta R.T. -0.02 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion:164 Resp: 26962

Ion Ratio Lower Upper

164 100

162 118.8 78.5 117.7#



160 59.8 32.9 49.3#




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E





#9

2-Fluorobiphenyl

Concen: 7.81 ng

RT: 12.85 min Scan# 1302

Delta R.T. -0.01 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion:172 Resp: 69511

Ion Ratio Lower Upper

172 100

171 33.0 25.6 38.4




170 20.7 15.8 23.6



Abundance Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#10

Acenaphthylene

Concen: 7.84 ng

RT: 13.93 min Scan# 1390

Delta R.T. -0.02 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion:152 Resp: 93407

Ion Ratio Lower Upper

152 100

151 19.0 14.7 22.1



153 13.0 10.2 15.2




Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E





#11

Acenaphthene

Concen: 6.93 ng

RT: 14.29 min Scan# 1411

Delta R.T. 0.00 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion:154 Resp: 57804

Ion Ratio Lower Upper




154 100

153 104.8 88.1 132.1

152 50.1 42.3 63.5



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E



#12

Fluorene

Concen: 7.08 ng

RT: 15.29 min Scan# 1469

Delta R.T. -0.02 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion:166 Resp: 68022

Ion Ratio Lower Upper



166 100


165 97.2 77.6 116.4

167 13.2 10.6 16.0



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E





#13

Phenanthrene-d10

Concen: 5.00 ng

RT: 16.97 min Scan# 1557

Delta R.T. -0.02 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion:188 Resp: 55237

Ion Ratio Lower Upper

188 100

94 21.9 3.6 5.4#




80 25.5 3.0 4.6#



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BM0

Ion 80.00 (79.70 to 80.70): BM0



#14

Hexachlorobenzene

Concen: 7.35 ng

RT: 16.29 min Scan# 1524

Delta R.T. -0.02 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion:284 Resp: 22425

Ion Ratio Lower Upper

284 100

142 40.4 0.0 0.0#



249 35.1 0.0 0.0#




Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E









#15
 Pentachlorophenol
 Concen: 23.70 ng
 RT: 16.64 min Scan# 1541
 Delta R.T. 0.00 min
 Lab File: BM001108.D
 Acq: 22 Apr 2015 15:47

Instrument :
 BNA_M
 ClientSampleId :
 PB82802BS

Tot Ion:266 Resp: 26368
 Ion Ratio Lower Upper
 266 100
 268 58.3 46.3 69.5
 264 64.7 59.8 89.6




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#16
 Phenanthrene
 Concen: 6.74 ng
 RT: 17.01 min Scan# 1559
 Delta R.T. -0.02 min
 Lab File: BM001108.D
 Acq: 22 Apr 2015 15:47

Tgt Ion:178 Resp: 98855
 Ion Ratio Lower Upper
 178 100
 176 19.6 14.6 22.0
 179 15.1 12.6 19.0



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E





#17

Anthracene

Concen: 7.71 ng

RT: 17.11 min Scan# 1564

Delta R.T. 0.00 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion:178 Resp: 103883

Ion Ratio Lower Upper

178 100

176 17.7 14.3 21.5




179 15.8 11.8 17.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#18

Fluoranthene

Concen: 7.71 ng

RT: 19.05 min Scan# 1689

Delta R.T. 0.00 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion:202 Resp: 117987

Ion Ratio Lower Upper

202 100

101 12.6 10.1 15.1



203 17.4 13.8 20.8




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#19

Chrysene-d12

Concen: 5.00 ng

RT: 21.18 min Scan# 1873

Delta R.T. -0.01 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion:240 Resp: 54550

Ion Ratio Lower Upper

240 100

120 17.6 4.6 6.8#




236 31.1 19.1 28.7#



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#20

Pyrene

Concen: 6.99 ng

RT: 19.42 min Scan# 1719

Delta R.T. 0.00 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion:202 Resp: 126293

Ion Ratio Lower Upper

202 100

200 20.6 16.3 24.5



203 17.8 13.8 20.6




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#21

Terphenyl-d14

Concen: 7.30 ng

RT: 19.62 min Scan# 1736

Delta R.T. -0.01 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion:244 Resp: 52790

Ion Ratio Lower Upper

244 100

212 10.5 7.1 10.7




122 20.5 5.9 8.9#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#22

Benzo(a)anthracene

Concen: 7.38 ng

RT: 21.17 min Scan# 1872

Delta R.T. 0.00 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion:228 Resp: 116625

Ion Ratio Lower Upper

228 100

226 23.2 19.0 28.6



229 21.3 16.8 25.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#23

Chrysene

Concen: 6.75 ng

RT: 21.22 min Scan# 1877

Delta R.T. 0.00 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion:228 Resp: 112394

Ion Ratio Lower Upper

228 100

226 25.3 20.6 31.0




229 21.2 17.2 25.8



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#24

Indeno(1,2,3-cd)pyrene

Concen: 7.67 ng

RT: 25.52 min Scan# 2289

Delta R.T. 0.00 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion:276 Resp: 130388

Ion Ratio Lower Upper

276 100

138 31.8 26.3 39.5



277 24.5 19.5 29.3




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#25
Pervlene-d12
Concen: 5.00 ng
RT: 23.35 min Scan# 2081
Delta R.T. -0.01 min
Lab File: BM001108.D
Acq: 22 Apr 2015 15:47




Instrument :
BNA_M
ClientSampleId :
PB82802BS


Tot Ion:264 Resp: 55695
Ion Ratio Lower Upper
264 100
260 26.9 19.0 28.6
265 21.1 17.9 26.9



#26
Benzo(b)fluoranthene
Concen: 6.73 ng
RT: 22.70 min Scan# 2019
Delta R.T. 0.00 min
Lab File: BM001108.D
Acq: 22 Apr 2015 15:47

Tgt Ion:252 Resp: 120547
Ion Ratio Lower Upper
252 100
253 21.9 18.2 27.2
125 11.9 10.9 16.3





#27

Benzo(k)fluoranthene

Concen: 6.39 ng

RT: 22.74 min Scan# 2023

Delta R.T. 0.00 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion:252 Resp: 109792

Ion Ratio Lower Upper




252 100

253 21.7 18.3 27.5

125 12.2 10.6 15.8



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#28

Benzo(a)pyrene

Concen: 7.36 ng

RT: 23.25 min Scan# 2072

Delta R.T. 0.00 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion:252 Resp: 112685

Ion Ratio Lower Upper



252 100


253 21.9 18.6 27.8

125 13.4 11.9 17.9



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#29

Dibenzo(a,h)anthracene

Concen: 6.91 ng

RT: 25.53 min Scan# 2290

Delta R.T. 0.00 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47

Instrument :

BNA_M

ClientSampleId :

PB82802BS


Tot Ion:278 Resp: 104582

Ion Ratio Lower Upper

278 100

139 20.3 18.2 27.2




279 23.2 19.1 28.7



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#30

Benzo(g,h,i)perylene

Concen: 6.73 ng

RT: 26.17 min Scan# 2352

Delta R.T. 0.00 min

Lab File: BM001108.D

Acq: 22 Apr 2015 15:47


Tgt Ion:276 Resp: 110378

Ion Ratio Lower Upper

276 100

277 23.2 19.0 28.4

138 27.5 23.4 35.2



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

