

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042224\
 Data File : BM045494.D
 Acq On : 23 Apr 2024 00:15
 Operator : MA/JU
 Sample : P2037-16MSD
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E1CY1MSD

Quant Time: Apr 23 08:31:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM042024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 23 02:57:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.597	152	690	0.400	ng/ul	#-0.02
4) Naphthalene-d8	10.375	136	2163	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.228	164	1195	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.001	188	2153	0.400	ng/ul	0.00
17) Chrysene-d12	21.205	240	1362	0.400	ng/ul	-0.01
23) Perylene-d12	23.397	264	1844	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.175	96	407	0.440	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.002	152	23	0.008	ng/ul	0.02
18) Fluoranthene-d10	19.038	212	16	0.004	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.201	88	955	1.073	ng/ul#	89
10) Acenaphthylene	13.964	152	123	0.021	ng/ul#	1
12) Fluorene	15.320	166	109	0.023	ng/ul#	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042224\
Data File : BM045494.D
Acq On : 23 Apr 2024 00:15
Operator : MA/JU
Sample : P2037-16MSD
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E1CY1MSD

Quant Time: Apr 23 08:31:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM042024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 23 02:57:45 2024
Response via : Initial Calibration

