

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042417\
 Data File : BM009698.D
 Acq On : 24 Apr 2017 19:53
 Operator : SJ/MA
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.421

Quant Time: Apr 25 04:36:22 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM042417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 24 15:43:15 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
3	Naphthalene	1.066	1.050	1.5	95	-0.01
4 SURR	2-Methylnaphthalene-d10	0.606	0.628	-3.6	101	0.00
5	2-Methylnaphthalene	0.780	0.805	-3.2	101	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
7	Acenaphthylene	1.948	2.121	-8.9	109	0.00
8 C	Acenaphthene	1.550	1.589	-2.5	90	0.00
9	Fluorene	1.843	1.831	0.7	96	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
11	Pentachlorophenol	0.155	0.144	7.1	96	0.00
12	Phenanthrene	1.195	1.245	-4.2	97	0.00
13	Anthracene	1.184	1.214	-2.5	94	0.00
14 SURR	Fluoranthene-d10	1.183	1.160	1.9	91	0.00
15 C	Fluoranthene	1.575	1.607	-2.0	95	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
17	Pyrene	1.331	1.444	-8.5	91	0.00
18	Benzo(a)anthracene	1.336	1.363	-2.0	87	0.00
19	Chrysene	1.257	1.292	-2.8	87	0.00
20 I	Perylene-d12	1.000	1.000	0.0	78	0.00
21	Benzo(b)fluoranthene	1.382	1.516	-9.7	87	0.00
22	Benzo(k)fluoranthene	1.325	1.295	2.3	78	0.00
23 C	Benzo(a)pyrene	1.329	1.355	-2.0	82	0.00
24	Indeno(1,2,3-cd)pyrene	1.528	1.511	1.1	77	0.00
25	Dibenzo(a,h)anthracene	1.255	1.197	4.6	75	0.00
26	Benzo(g,h,i)perylene	1.305	1.250	4.2	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0