

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042417\
 Data File : BM009687.D
 Acq On : 24 Apr 2017 12:00
 Operator : SJ/MA
 Sample : SST0.114
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST0.114

Manual Integrations
 APPROVED

mohammad
 4/25/2017 11:11:36 AM

Quant Time: Apr 24 14:27:27 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM042417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 24 14:08:21 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	131	0.40	ng/ul	0.00
2) Naphthalene-d8	10.63	136	414	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	249	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.23	188	658	0.40	ng/ul	0.00
16) Chrysene-d12	21.41	240	792	0.40	ng/ul	0.00
20) Perylene-d12	23.72	264	860	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.23	152	59	0.09	ng/ul	0.01
14) Fluoranthene-d10	19.26	212	197	0.10	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.68	128	109	0.09	ng/ul#	13
5) 2-Methylnaphthalene	12.29	142	74	0.09	ng/ul	89
7) Acenaphthylene	14.19	152	129	0.10	ng/ul#	34
8) Acenaphthene	14.54	153	94	0.09	ng/ul#	85
9) Fluorene	15.53	166	116	0.10	ng/ul#	81
12) Phenanthrene	17.26	178	202	0.10	ng/ul#	60
13) Anthracene	17.35	178	186	0.09	ng/ul#	41
15) Fluoranthene	19.28	202	260m	0.10	ng/ul	
17) Pyrene	19.65	202	268	0.10	ng/ul#	67
18) Benzo(a)anthracene	21.39	228	269	0.10	ng/ul#	63
19) Chrysene	21.44	228	261	0.10	ng/ul#	61
21) Benzo(b)fluoranthene	23.02	252	307	0.09	ng/ul#	50
22) Benzo(k)fluoranthene	23.06	252	305	0.10	ng/ul#	44
23) Benzo(a)pyrene	23.61	252	304	0.10	ng/ul#	53
24) Indeno(1,2,3-cd)pyrene	26.07	276	313	0.08	ng/ul#	90
25) Dibenzo(a,h)anthracene	26.08	278	265	0.08	ng/ul#	34
26) Benzo(g,h,i)perylene	26.80	276	287	0.09	ng/ul#	47

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042417\
Data File : BM009687.D
Acq On : 24 Apr 2017 12:00
Operator : SJ/MA
Sample : SST0.114
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleID :
SSTD0.114

Manual Integrations
APPROVED

mohammad
4/25/2017 11:11:36 AM

Quant Time: Apr 24 14:27:27 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM042417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 24 14:08:21 2017
Response via : Initial Calibration

