

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042516\
 Data File : BM005066.D
 Acq On : 26 Apr 2016 03:44
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02042

Quant Time: Apr 26 04:59:56 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM042516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 26 00:45:11 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.730	0.727	0.4	101	0.00
3 S	1,4-Dioxane-d8	0.394	0.394	0.0	101	0.00
4	Benzaldehyde	0.821	1.066	-29.8#	106	0.00
5 S	Phenol-d5	1.654	1.725	-4.3	105	0.00
6	Phenol	1.725	1.789	-3.7	104	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.941	0.967	-2.8	102	0.00
8	Bis(2-Chloroethyl)ether	1.322	1.346	-1.8	101	0.00
9 S	2-Chlorophenol-d4	1.309	1.362	-4.0	105	0.00
10	2-Chlorophenol	1.345	1.392	-3.5	105	0.00
11	2-Methylphenol	1.346	1.417	-5.3	107	0.00
12	2,2'-oxybis(1-Chloropropane	1.710	1.765	-3.2	102	0.00
13 S	4-Methylphenol-d8	1.403	1.489	-6.1	108	0.00
14	Acetophenone	2.111	2.301	-9.0	105	0.00
15 P	N-Nitroso-di-n-propylamine	1.057	1.151	-8.9	106	0.00
16	4-Methylphenol	1.498	1.591	-6.2	106	0.00
17	Hexachloroethane	0.536	0.538	-0.4	102	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
19 S	Nitrobenzene-d5	0.139	0.146	-5.0	106	0.00
20	Nitrobenzene	0.341	0.353	-3.5	105	0.00
21	Isophorone	0.647	0.691	-6.8	107	0.00
22 S	2-Nitrophenol-d4	0.155	0.172	-11.0	112	0.00
23 C	2-Nitrophenol	0.168	0.181	-7.7	106	0.00
24	2,4-Dimethylphenol	0.362	0.379	-4.7	106	0.00
25	Bis(2-Chloroethoxy)methane	0.402	0.417	-3.7	106	0.00
26 S	2,4-Dichlorophenol-d3	0.294	0.315	-7.1	107	0.00
27 C	2,4-Dichlorophenol	0.302	0.321	-6.3	106	0.00
28	Naphthalene	1.009	1.023	-1.4	104	0.00
29 S	4-Chloroaniline-d4	0.344	0.417	-21.2#	105	0.00
30	4-Chloroaniline	0.350	0.429	-22.6#	106	0.00
31 C	Hexachlorobutadiene	0.199	0.201	-1.0	105	0.00
32	Caprolactam	0.104	0.111	-6.7	111	0.00
33 C	4-Chloro-3-methylphenol	0.345	0.382	-10.7	110	0.00
34	2-Methylnaphthalene	0.747	0.776	-3.9	105	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
36	1,2,4,5-Tetrachlorobenzene	0.625	0.631	-1.0	104	0.00
37	Hexachlorocyclopentadiene	0.371	0.365	1.6	108	0.00
38 C	2,4,6-Trichlorophenol	0.397	0.425	-7.1	109	0.00
39	2,4,5-Trichlorophenol	0.445	0.480	-7.9	109	0.00
40	1,1'-Biphenyl	1.582	1.618	-2.3	104	0.00
41	2-Chloronaphthalene	1.203	1.229	-2.2	105	0.00
42	2-Nitroaniline	0.326	0.367	-12.6	110	0.00
43 S	Dimethylphthalate-d6	1.580	1.635	-3.5	105	0.00
44	Dimethylphthalate	1.584	1.629	-2.8	104	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042516\
 Data File : BM005066.D
 Acq On : 26 Apr 2016 03:44
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02042

Quant Time: Apr 26 04:59:56 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM042516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 26 00:45:11 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.306	0.337	-10.1	110	0.00
46 S	Acenaphthylene-d8	1.881	1.945	-3.4	104	0.00
47	Acenaphthylene	1.990	2.071	-4.1	105	0.00
48	3-Nitroaniline	0.311	0.360	-15.8	110	0.00
49 C	Acenaphthene	1.309	1.350	-3.1	104	0.00
50	2,4-Dinitrophenol	0.179	0.193	-7.8	132	0.00
51 S	4-Nitrophenol-d4	0.276	0.292	-5.8	109	0.00
52	4-Nitrophenol	0.247	0.259	-4.9	110	0.00
53	Dibenzofuran	1.931	1.986	-2.8	104	0.00
54	2,4-Dinitrotoluene	0.465	0.503	-8.2	108	0.00
55	2,3,4,6-Tetrachlorophenol	0.387	0.422	-9.0	110	0.00
56	Diethylphthalate	1.599	1.667	-4.3	106	0.00
57 S	Fluorene-d10	1.388	1.439	-3.7	105	0.00
58	Fluorene	1.499	1.579	-5.3	105	0.00
59	4-Chlorophenyl-phenylether	0.756	0.790	-4.5	105	0.00
60	4-Nitroaniline	0.336	0.379	-12.8	112	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.114	0.122	-7.0	113	0.00
63	4,6-Dinitro-2-methylphenol	0.120	0.132	-10.0	113	0.00
64	N-Nitrosodiphenylamine	0.550	0.576	-4.7	106	0.00
65	4-Bromophenyl-phenylether	0.195	0.204	-4.6	107	0.00
66	Hexachlorobenzene	0.222	0.229	-3.2	106	0.00
67	Atrazine	0.203	0.221	-8.9	108	0.00
68 C	Pentachlorophenol	0.128	0.137	-7.0	115	0.00
69	Phenanthrene	1.064	1.096	-3.0	105	0.00
70 S	Anthracene-d10	0.897	0.932	-3.9	105	0.00
71	Anthracene	1.073	1.108	-3.3	104	0.00
72	Carbazole	0.920	1.016	-10.4	106	0.00
73	Di-n-butylphthalate	1.077	1.183	-9.8	109	0.00
74 C	Fluoranthene	1.171	1.312	-12.0	105	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
76 S	Pyrene-d10	0.910	0.945	-3.8	105	0.00
77	Pyrene	1.155	1.191	-3.1	104	0.00
78	Butylbenzylphthalate	0.444	0.500	-12.6	111	0.00
79	3,3'-Dichlorobenzidine	0.358	0.406	-13.4	103	0.00
80	Benzo(a)anthracene	1.145	1.193	-4.2	104	0.00
81	Bis(2-ethylhexyl)phthalate	0.624	0.710	-13.8	110	0.00
82	Chrysene	1.095	1.121	-2.4	103	0.00
83 I	Perylene-d12	1.000	1.000	0.0	100	0.00
84	Di-n-octyl phthalate	1.243	1.374	-10.5	109	0.00
85	Benzo(b)fluoranthene	1.205	1.281	-6.3	105	0.00
86	Benzo(k)fluoranthene	1.171	1.172	-0.1	97	0.00
87 S	Benzo(a)pyrene-d12	0.896	0.928	-3.6	100	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042516\
Data File : BM005066.D
Acq On : 26 Apr 2016 03:44
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02042

Quant Time: Apr 26 04:59:56 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM042516.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 26 00:45:11 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.125	1.162	-3.3	100	-0.01
89	Indeno(1,2,3-cd)pyrene	1.073	1.119	-4.3	98	0.00
90	Dibenzo(a,h)anthracene	0.897	0.948	-5.7	99	-0.01
91	Benzo(a,h,i)perylene	0.878	0.892	-1.6	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0