

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042516\
 Data File : BM005078.D
 Acq On : 26 Apr 2016 12:58
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02043

Quant Time: Apr 26 13:32:52 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM042516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 26 00:45:11 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	183#	0.00
2	1,4-Dioxane	0.730	0.781	-7.0	191#	0.00
3 S	1,4-Dioxane-d8	0.394	0.425	-7.9	191#	0.00
4	Benzaldehyde	0.821	1.072	-30.6#	188#	0.00
5 S	Phenol-d5	1.654	1.761	-6.5	188#	0.00
6	Phenol	1.725	1.831	-6.1	187#	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.941	1.017	-8.1	188#	0.00
8	Bis(2-Chloroethyl)ether	1.322	1.440	-8.9	191#	0.00
9 S	2-Chlorophenol-d4	1.309	1.374	-5.0	185#	0.00
10	2-Chlorophenol	1.345	1.423	-5.8	188#	0.00
11	2-Methylphenol	1.346	1.437	-6.8	190#	0.00
12	2,2'-oxybis(1-Chloropropane	1.710	1.823	-6.6	186#	0.00
13 S	4-Methylphenol-d8	1.403	1.488	-6.1	190#	0.00
14	Acetophenone	2.111	2.235	-5.9	179#	0.00
15 P	N-Nitroso-di-n-propylamine	1.057	1.146	-8.4	186#	0.00
16	4-Methylphenol	1.498	1.614	-7.7	190#	0.00
17	Hexachloroethane	0.536	0.546	-1.9	182#	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	184#	0.00
19 S	Nitrobenzene-d5	0.139	0.145	-4.3	187#	0.00
20	Nitrobenzene	0.341	0.356	-4.4	187#	0.00
21	Isophorone	0.647	0.686	-6.0	188#	0.00
22 S	2-Nitrophenol-d4	0.155	0.163	-5.2	188#	0.00
23 C	2-Nitrophenol	0.168	0.179	-6.5	187#	0.00
24	2,4-Dimethylphenol	0.362	0.368	-1.7	183#	0.00
25	Bis(2-Chloroethoxy)methane	0.402	0.428	-6.5	193#	0.00
26 S	2,4-Dichlorophenol-d3	0.294	0.308	-4.8	186#	0.00
27 C	2,4-Dichlorophenol	0.302	0.312	-3.3	183#	0.00
28	Naphthalene	1.009	1.030	-2.1	185#	0.00
29 S	4-Chloroaniline-d4	0.344	0.414	-20.3#	186#	0.00
30	4-Chloroaniline	0.350	0.429	-22.6#	189#	0.00
31 C	Hexachlorobutadiene	0.199	0.185	7.0	171#	0.00
32	Caprolactam	0.104	0.106	-1.9	188#	0.00
33 C	4-Chloro-3-methylphenol	0.345	0.366	-6.1	187#	0.00
34	2-Methylnaphthalene	0.747	0.760	-1.7	182#	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	178#	0.00
36	1,2,4,5-Tetrachlorobenzene	0.625	0.617	1.3	172#	0.00
37	Hexachlorocyclopentadiene	0.371	0.337	9.2	170#	0.00
38 C	2,4,6-Trichlorophenol	0.397	0.413	-4.0	180#	0.00
39	2,4,5-Trichlorophenol	0.445	0.465	-4.5	179#	0.00
40	1,1'-Biphenyl	1.582	1.645	-4.0	180#	0.00
41	2-Chloronaphthalene	1.203	1.222	-1.6	177#	0.00
42	2-Nitroaniline	0.326	0.363	-11.3	184#	0.00
43 S	Dimethylphthalate-d6	1.580	1.590	-0.6	173#	0.00
44	Dimethylphthalate	1.584	1.590	-0.4	173#	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042516\
 Data File : BM005078.D
 Acq On : 26 Apr 2016 12:58
 Operator : UM/SJ
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02043

Quant Time: Apr 26 13:32:52 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM042516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 26 00:45:11 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.306	0.317	-3.6	175#	0.00
46 S	Acenaphthylene-d8	1.881	1.940	-3.1	177#	0.00
47	Acenaphthylene	1.990	2.096	-5.3	180#	0.00
48	3-Nitroaniline	0.311	0.351	-12.9	183#	0.00
49 C	Acenaphthene	1.309	1.356	-3.6	178#	0.00
50	2,4-Dinitrophenol	0.179	0.168	6.1	196#	0.00
51 S	4-Nitrophenol-d4	0.276	0.286	-3.6	181#	0.00
52	4-Nitrophenol	0.247	0.234	5.3	169#	0.00
53	Dibenzofuran	1.931	1.980	-2.5	176#	0.00
54	2,4-Dinitrotoluene	0.465	0.483	-3.9	175#	0.00
55	2,3,4,6-Tetrachlorophenol	0.387	0.396	-2.3	174#	0.00
56	Diethylphthalate	1.599	1.613	-0.9	174#	0.00
57 S	Fluorene-d10	1.388	1.401	-0.9	174#	0.00
58	Fluorene	1.499	1.568	-4.6	177#	0.00
59	4-Chlorophenyl-phenylether	0.756	0.757	-0.1	170#	0.00
60	4-Nitroaniline	0.336	0.366	-8.9	184#	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	169#	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.114	0.117	-2.6	174#	0.00
63	4,6-Dinitro-2-methylphenol	0.120	0.124	-3.3	172#	0.00
64	N-Nitrosodiphenylamine	0.550	0.598	-8.7	177#	0.00
65	4-Bromophenyl-phenylether	0.195	0.199	-2.1	168#	0.00
66	Hexachlorobenzene	0.222	0.220	0.9	163#	0.00
67	Atrazine	0.203	0.215	-5.9	169#	0.00
68 C	Pentachlorophenol	0.128	0.126	1.6	171#	0.00
69	Phenanthrene	1.064	1.112	-4.5	172#	0.00
70 S	Anthracene-d10	0.897	0.932	-3.9	169#	0.00
71	Anthracene	1.073	1.137	-6.0	172#	0.00
72	Carbazole	0.920	1.041	-13.2	174#	0.00
73	Di-n-butylphthalate	1.077	1.170	-8.6	173#	0.00
74 C	Fluoranthene	1.171	1.289	-10.1	166#	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	159#	0.00
76 S	Pyrene-d10	0.910	0.954	-4.8	165#	0.00
77	Pyrene	1.155	1.218	-5.5	165#	0.00
78	Butylbenzylphthalate	0.444	0.492	-10.8	169#	0.00
79	3,3'-Dichlorobenzidine	0.358	0.392	-9.5	155#	0.00
80	Benzo(a)anthracene	1.145	1.168	-2.0	157#	0.00
81	Bis(2-ethylhexyl)phthalate	0.624	0.711	-13.9	172#	0.00
82	Chrysene	1.095	1.122	-2.5	160#	0.00
83 I	Perylene-d12	1.000	1.000	0.0	145	0.00
84	Di-n-octyl phthalate	1.243	1.449	-16.6	168#	0.00
85	Benzo(b)fluoranthene	1.205	1.275	-5.8	151#	0.00
86	Benzo(k)fluoranthene	1.171	1.195	-2.0	144	0.00
87 S	Benzo(a)pyrene-d12	0.896	0.933	-4.1	147	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM042516\
Data File : BM005078.D
Acq On : 26 Apr 2016 12:58
Operator : UM/SJ
Sample : SSTDCCC020EC
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02043

Quant Time: Apr 26 13:32:52 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM042516.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 26 00:45:11 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.125	1.172	-4.2	146	0.00
89	Indeno(1,2,3-cd)pyrene	1.073	1.082	-0.8	137	0.00
90	Dibenzo(a,h)anthracene	0.897	0.915	-2.0	138	0.00
91	Benzo(a,h,i)perylene	0.878	0.867	1.3	135	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0