

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042517\
 Data File : BM009710.D
 Acq On : 25 Apr 2017 16:47
 Operator : SJ/MA
 Sample : PB98377BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK77

Quant Time: Apr 26 03:25:49 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM042517.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 25 13:32:43 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	113641	20.00	ng/ul	0.00
18) Naphthalene-d8	10.62	136	495451	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.47	164	323832	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.22	188	807763	20.00	ng/ul	0.00
75) Chrysene-d12	21.40	240	1110170	20.00	ng/ul	0.00
83) Perylene-d12	23.71	264	1042001	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.31	96	12296	4.98	ng/uL	0.00
5) Phenol-d5	6.99	99	279156	22.61	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.16	67	192607	22.88	ng/ul	0.00
9) 2-Chlorophenol-d4	7.36	132	186500	23.74	ng/ul	0.00
13) 4-Methylphenol-d8	8.52	113	226602	23.55	ng/ul	0.00
19) Nitrobenzene-d5	8.99	128	92327	22.89	ng/ul	0.00
22) 2-Nitrophenol-d4	9.71	143	104217	24.56	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.24	165	204235	24.07	ng/ul	0.00
29) 4-Chloroaniline-d4	10.77	131	105148	10.28	ng/ul	0.00
43) Dimethylphthalate-d6	13.88	166	696910	24.74	ng/ul	0.00
46) Acenaphthylene-d8	14.16	160	811161	23.88	ng/ul	0.00
51) 4-Nitrophenol-d4	14.67	143	122084	22.46	ng/ul	0.00
57) Fluorene-d10	15.47	176	612232	24.08	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.59	200	109649	19.24	ng/ul	0.00
70) Anthracene-d10	17.32	188	951201	22.93	ng/ul	0.00
76) Pyrene-d10	19.62	212	1123354	22.93	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.57	264	1122864	22.16	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042517\
Data File : BM009710.D
Acq On : 25 Apr 2017 16:47
Operator : SJ/MA
Sample : PB98377BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK77

Quant Time: Apr 26 03:25:49 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM042517.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 25 13:32:43 2017
Response via : Initial Calibration

