




Data Path : Z:\HPCHEM1\BNA M\DATA\BM042518\
Data File : BM014845.D
Acq On : 25 Apr 2018 21:59
Operator : SJ/JU
Sample : J2431-12DL2 20X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Quant Time: Apr 26 02:04:01 2018
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM041918.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Apr 26 01:59:14 2018
Response via : Initial Calibration



TIC: BM014845.D

(86) Benzo(k)fluoranthene

23.680min (-0.006) 2.94ng/ul m

response 257328

Ion	Exp%	Act%
252.00	100	100
253.00	21.40	22.68
125.00	4.30	11.41#
0.00	0.00	0.00