

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042523\
 Data File : BM039650.D
 Acq On : 25 Apr 2023 21:29
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020125

Quant Time: Apr 25 22:42:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 25 22:38:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.807	152	226692	20.000	ng/u1	0.00
20) Naphthalene-d8	10.613	136	1031727	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.466	164	632842	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.218	188	1307379	20.000	ng/u1	0.00
79) Chrysene-d12	21.418	240	1176129	20.000	ng/u1	0.00
88) Perylene-d12	23.777	264	1093323	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.202	96	45465	7.698	ng/uL	0.00
4) Pyridine-d5	3.625	84	364264	21.951	ng/u1	0.00
7) Phenol-d5	6.990	99	425627	20.514	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.143	67	304900	22.233	ng/u1	0.00
11) 2-Chlorophenol-d4	7.337	132	340846	20.942	ng/u1	0.00
15) 4-Methylphenol-d8	8.537	113	364205	21.792	ng/u1	0.00
21) Nitrobenzene-d5	8.978	128	184025	22.170	ng/u1	0.00
24) 2-Nitrophenol-d4	9.701	143	203539	21.838	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.248	165	351092	21.892	ng/u1	0.00
31) 4-Chloroaniline-d4	10.766	131	515026	21.502	ng/u1	0.00
46) Dimethylphthalate-d6	13.883	166	1091751	21.391	ng/u1	0.00
49) Acenaphthylene-d8	14.160	160	1275506	22.246	ng/u1	0.00
54) 4-Nitrophenol-d4	14.713	143	131038	16.965	ng/u1	0.00
60) Fluorene-d10	15.460	176	909418	21.749	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.601	200	177305	20.408	ng/u1	0.00
73) Anthracene-d10	17.318	188	1415512	22.513	ng/u1	0.00
81) Pyrene-d10	19.618	212	1615213	20.667	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.624	264	1296888	21.971	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.237	88	51256	7.931	ng/uL	94
5) Pyridine	3.649	79	385954	22.058	ng/u1	100
6) Benzaldehyde	6.954	77	136900	15.444	ng/u1	94
8) Phenol	7.019	94	452839	20.963	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.237	93	395477	21.958	ng/u1	97
12) 2-Chlorophenol	7.372	128	357995	21.859	ng/u1	97
13) 2-Methylphenol	8.266	108	366294	21.831	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.342	45	619396	24.760	ng/u1	98
16) Acetophenone	8.642	105	611710	22.283	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.625	70	352374	23.192	ng/u1	93
18) 4-Methylphenol	8.601	108	398415	21.994	ng/u1	99
19) Hexachloroethane	8.878	117	163300	21.954	ng/u1	97
22) Nitrobenzene	9.019	77	484887	23.659	ng/u1	94
23) Isophorone	9.548	82	978196	22.383	ng/u1	98
25) 2-Nitrophenol	9.731	139	213601	21.964	ng/u1	92
26) 2,4-Dimethylphenol	9.801	107	451037	22.696	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.031	93	570405	22.196	ng/u1	99
29) 2,4-Dichlorophenol	10.272	162	347765	22.014	ng/u1	96
30) Naphthalene	10.660	128	1272075	22.284	ng/u1	99
32) 4-Chloroaniline	10.789	127	509693	21.596	ng/u1	100
33) Hexachlorobutadiene	10.936	225	221902	21.317	ng/u1	97
34) Caprolactam	11.578	113	117065	19.899	ng/u1	94
35) 4-Chloro-3-methylphenol	11.930	107	396982	21.522	ng/u1	98
36) 2-Methylnaphthalene	12.277	142	831513	21.729	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042523\
 Data File : BM039650.D
 Acq On : 25 Apr 2023 21:29
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020125

Quant Time: Apr 25 22:42:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM041823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 25 22:38:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.501	142	831082	21.611	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.648	216	424460	22.345	ng/ul	99
40) Hexachlorocyclopentadiene	12.619	237	161553	14.274	ng/ul	100
41) 2,4,6-Trichlorophenol	12.901	196	275647	21.721	ng/ul	99
42) 2,4,5-Trichlorophenol	12.983	196	282948	21.453	ng/ul	98
43) 1,1'-Biphenyl	13.295	154	1085734	22.200	ng/ul	99
44) 2-Chloronaphthalene	13.342	162	859497	22.403	ng/ul	100
45) 2-Nitroaniline	13.560	65	269581	24.350	ng/ul	90
47) Dimethylphthalate	13.930	163	1091399	21.598	ng/ul	99
48) 2,6-Dinitrotoluene	14.060	165	215407	21.077	ng/ul	97
50) Acenaphthylene	14.189	152	1385664	22.593	ng/ul	100
51) 3-Nitroaniline	14.395	138	163244	17.042	ng/ul	93
52) Acenaphthene	14.530	153	945353	22.387	ng/ul	99
53) 2,4-Dinitrophenol	14.618	184	80892	14.870	ng/ul	94
55) 4-Nitrophenol	14.724	109	103591	17.656	ng/ul	90
56) Dibenzofuran	14.866	168	1289240	22.462	ng/ul	99
57) 2,4-Dinitrotoluene	14.854	165	314987	22.181	ng/ul#	81
58) 2,3,4,6-Tetrachlorophenol	15.101	232	252962	21.427	ng/ul	98
59) Diethylphthalate	15.289	149	1118011	21.343	ng/ul	99
61) Fluorene	15.518	166	1080950	22.767	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.513	204	520927	22.136	ng/ul	98
63) 4-Nitroaniline	15.560	138	108244	14.736	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.618	198	168818	20.211	ng/ul	93
67) N-Nitrosodiphenylamine	15.730	169	868322	22.292	ng/ul	99
68) 4-Bromophenyl-phenylether	16.407	248	298129	21.113	ng/ul	98
69) Hexachlorobenzene	16.518	284	351681	21.679	ng/ul	95
70) Atrazine	16.689	200	308223	21.325	ng/ul	98
71) Pentachlorophenol	16.877	266	186366	19.653	ng/ul	98
72) Phenanthrene	17.260	178	1621014	22.333	ng/ul	100
74) Anthracene	17.354	178	1675881	23.016	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.260	216	436726	22.430	ng/ul	99
76) Pentachlorobenzene	14.783	250	432260	22.555	ng/ul	98
77) Carbazole	17.636	167	1451877	22.439	ng/ul	100
78) Di-n-butylphthalate	18.189	149	1910061	20.871	ng/ul	99
80) Fluoranthene	19.283	202	1882067	20.546	ng/ul	99
82) Pyrene	19.648	202	1976837	20.846	ng/ul	99
83) Butylbenzylphthalate	20.548	149	852569	19.046	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.342	252	488319	18.723	ng/ul	99
85) Benzo(a)anthracene	21.400	228	1833469	21.553	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.324	149	1213127	18.224	ng/ul	100
87) Chrysene	21.453	228	1739615	21.439	ng/ul	100
89) Di-n-octyl phthalate	22.236	149	1991715	18.131	ng/ul	100
90) Benzo(b)fluoranthene	23.059	252	1671151	21.723	ng/ul	99
91) Benzo(k)fluoranthene	23.106	252	1650746	21.850	ng/ul	100
93) Benzo(a)pyrene	23.671	252	1412181	21.778	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.218	276	1569663	22.007	ng/ul	99
95) Dibenzo(a,h)anthracene	26.230	278	1317058	22.271	ng/ul	99
96) Benzo(g,h,i)perylene	26.971	276	1239664	21.818	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.931 ng/uL
RT: 3.237 min Scan# 34
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion: 88 Resp: 51256
Ion Ratio Lower Upper
88 100
43 32.3 24.6 36.8
58 77.4 56.7 85.1



#5
Pyridine
Concen: 22.058 ng/uL
RT: 3.649 min Scan# 104
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion: 79 Resp: 385954
Ion Ratio Lower Upper
79 100
52 77.3 61.9 92.9
51 37.7 29.7 44.5









#6
Benzaldehyde
Concen: 15.444 ng/ul
RT: 6.954 min Scan# 666
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion: 77 Resp: 136900
Ion Ratio Lower Upper
77 100
105 89.5 76.3 114.5
106 89.3 75.4 113.0



#8
Phenol
Concen: 20.963 ng/ul
RT: 7.019 min Scan# 677
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion: 94 Resp: 452839
Ion Ratio Lower Upper
94 100
65 29.4 22.2 33.4
66 37.4 29.6 44.4









#10
Bis(2-Chloroethyl)ether
Concen: 21.958 ng/ul
RT: 7.237 min Scan# 714
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion: 93 Resp: 395477
Ion Ratio Lower Upper
93 100
63 80.5 61.9 92.9
95 31.8 25.8 38.8



#12
2-Chlorophenol
Concen: 21.859 ng/ul
RT: 7.372 min Scan# 737
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:128 Resp: 357995
Ion Ratio Lower Upper
128 100
64 51.8 38.8 58.2
130 31.6 25.4 38.0





#13

2-Methylphenol

Concen: 21.831 ng/ul

RT: 8.266 min Scan# 890

Delta R.T. -0.006 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :





SSTD020125

Tgt Ion:108 Resp: 366294

Ion Ratio Lower Upper

108 100

107 88.8 71.7 107.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 24.760 ng/ul

RT: 8.342 min Scan# 902

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29




Tgt Ion: 45 Resp: 619396


Ion Ratio Lower Upper

45 100

77 14.5 12.3 18.5

79 11.4 9.8 14.6









#16
Acetophenone
Concen: 22.283 ng/ul
RT: 8.642 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:105 Resp: 611710
Ion Ratio Lower Upper
105 100
77 84.4 64.1 96.1
51 30.5 22.2 33.4



#17
N-Nitroso-di-n-propylamine
Concen: 23.192 ng/ul
RT: 8.625 min Scan# 950
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion: 70 Resp: 352374
Ion Ratio Lower Upper
70 100
42 60.8 43.0 64.4
101 9.1 7.4 11.2
130 18.8 16.2 24.2









#18
4-Methylphenol
Concen: 21.994 ng/ul
RT: 8.601 min Scan# 946
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:108 Resp: 398415
Ion Ratio Lower Upper
108 100
107 112.1 88.6 133.0



#19
Hexachloroethane
Concen: 21.954 ng/ul
RT: 8.878 min Scan# 993
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:117 Resp: 163300
Ion Ratio Lower Upper
117 100
201 92.9 78.3 117.5
199 58.6 47.4 71.0









#22
Nitrobenzene
Concen: 23.659 ng/ul
RT: 9.019 min Scan# 1017
Delta R.T. -0.006 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion: 77 Resp: 484887
Ion Ratio Lower Upper
77 100
123 40.6 36.1 54.1
65 13.4 10.2 15.2



#23
Isophorone
Concen: 22.383 ng/ul
RT: 9.548 min Scan# 1107
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion: 82 Resp: 978196
Ion Ratio Lower Upper
82 100
95 6.5 5.4 8.2
138 15.7 13.2 19.8









#25
2-Nitrophenol
Concen: 21.964 ng/ul
RT: 9.731 min Scan# 1138
Delta R.T. -0.006 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:139 Resp: 213601
Ion Ratio Lower Upper
139 100
65 58.8 40.7 61.1
109 30.7 23.9 35.9



#26
2,4-Dimethylphenol
Concen: 22.696 ng/ul
RT: 9.801 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:107 Resp: 451037
Ion Ratio Lower Upper
107 100
121 49.3 39.1 58.7
122 81.4 68.1 102.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 22.196 ng/ul
 RT: 10.031 min Scan# 1190
 Delta R.T. -0.006 min
 Lab File: BM039650.D
 Acq: 25 Apr 2023 21:29




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020125


Tgt Ion: 93 Resp: 570405
 Ion Ratio Lower Upper
 93 100
 95 31.9 25.8 38.6
 123 10.9 8.8 13.2



#29
 2,4-Dichlorophenol
 Concen: 22.014 ng/ul
 RT: 10.272 min Scan# 1230
 Delta R.T. -0.006 min
 Lab File: BM039650.D
 Acq: 25 Apr 2023 21:29

Tgt Ion: 162 Resp: 347765
 Ion Ratio Lower Upper
 162 100
 164 67.7 52.4 78.6
 98 40.1 29.4 44.0









#30
Naphthalene
Concen: 22.284 ng/ul
RT: 10.660 min Scan# 1297
Delta R.T. -0.006 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:128 Resp: 1272075
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 12.9 10.5 15.7



#32
4-Chloroaniline
Concen: 21.596 ng/ul
RT: 10.789 min Scan# 1318
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:127 Resp: 509693
Ion Ratio Lower Upper
127 100
129 32.2 25.8 38.8









#33
Hexachlorobutadiene
Concen: 21.317 ng/ul
RT: 10.936 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:225 Resp: 221902
Ion Ratio Lower Upper
225 100
223 64.4 50.1 75.1
227 65.2 50.4 75.6



#34
Caprolactam
Concen: 19.899 ng/ul
RT: 11.578 min Scan# 1452
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:113 Resp: 117065
Ion Ratio Lower Upper
113 100
55 188.6 145.5 218.3
56 140.0 103.9 155.9









#35
 4-Chloro-3-methylphenol
 Concen: 21.522 ng/ul
 RT: 11.930 min Scan# 1512
 Delta R.T. 0.000 min
 Lab File: BM039650.D
 Acq: 25 Apr 2023 21:29




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020125


Tgt Ion:107 Resp: 396982
 Ion Ratio Lower Upper
 107 100
 144 25.2 21.0 31.4
 142 77.3 62.9 94.3



#36
 2-Methylnaphthalene
 Concen: 21.729 ng/ul
 RT: 12.277 min Scan# 1571
 Delta R.T. -0.006 min
 Lab File: BM039650.D
 Acq: 25 Apr 2023 21:29

Tgt Ion:142 Resp: 831513
 Ion Ratio Lower Upper
 142 100
 141 87.2 71.1 106.7





#37

1-Methylnaphthalene

Concen: 21.611 ng/ul

RT: 12.501 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:142 Resp: 831082

Ion Ratio Lower Upper

142 100

141 90.7 72.0 108.0

116 3.8 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 22.345 ng/ul

RT: 12.648 min Scan# 1634

Delta R.T. -0.006 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Tgt Ion:216 Resp: 424460




Ion Ratio Lower Upper


216 100

214 78.1 62.2 93.2

179 20.7 16.5 24.7

108 19.9 14.0 21.0









#40
Hexachlorocyclopentadiene
Concen: 14.274 ng/ul
RT: 12.619 min Scan# 1629
Delta R.T. -0.006 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:237 Resp: 161553
Ion Ratio Lower Upper
237 100
235 63.6 50.9 76.3
272 12.7 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 21.721 ng/ul
RT: 12.901 min Scan# 1677
Delta R.T. -0.006 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:196 Resp: 275647
Ion Ratio Lower Upper
196 100
198 96.4 76.7 115.1
200 30.4 23.8 35.6





#42

2,4,5-Trichlorophenol

Concen: 21.453 ng/ul

RT: 12.983 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:196 Resp: 282948

Ion Ratio Lower Upper

196 100

198 92.2 75.7 113.5

200 29.6 24.6 36.8



#43

1,1'-Biphenyl

Concen: 22.200 ng/ul

RT: 13.295 min Scan# 1744

Delta R.T. -0.006 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29




Tgt Ion:154 Resp: 1085734


Ion Ratio Lower Upper

154 100

153 40.0 31.9 47.9

76 15.5 11.5 17.3









#44
2-Chloronaphthalene
Concen: 22.403 ng/ul
RT: 13.342 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:162 Resp: 859497
Ion Ratio Lower Upper
162 100
164 32.5 26.1 39.1
127 39.4 31.4 47.2



#45
2-Nitroaniline
Concen: 24.350 ng/ul
RT: 13.560 min Scan# 1789
Delta R.T. -0.006 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion: 65 Resp: 269581
Ion Ratio Lower Upper
65 100
92 65.2 53.4 80.2
138 88.2 83.2 124.8





#47

Dimethylphthalate

Concen: 21.598 ng/ul

RT: 13.930 min Scan# 1852

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:163 Resp: 1091399

Ion Ratio Lower Upper

163 100

194 4.2 3.1 4.7

164 9.9 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 21.077 ng/ul

RT: 14.060 min Scan# 1874

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29




Tgt Ion:165 Resp: 215407


Ion Ratio Lower Upper

165 100

89 66.2 50.6 76.0

121 22.0 17.3 25.9









#50
Acenaphthylene
Concen: 22.593 ng/ul
RT: 14.189 min Scan# 1896
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:152 Resp: 1385664
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.8
153 12.8 10.4 15.6



#51
3-Nitroaniline
Concen: 17.042 ng/ul
RT: 14.395 min Scan# 1931
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:138 Resp: 163244
Ion Ratio Lower Upper
138 100
108 10.5 8.8 13.2
92 128.0 95.5 143.3





#52

Acenaphthene

Concen: 22.387 ng/ul

RT: 14.530 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:153 Resp: 945353

Ion Ratio Lower Upper

153 100

152 48.0 38.2 57.4

154 88.1 71.0 106.4



#53

2,4-Dinitrophenol

Concen: 14.870 ng/ul

RT: 14.618 min Scan# 1969

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29




Tgt Ion:184 Resp: 80892


Ion Ratio Lower Upper

184 100

63 81.7 60.0 90.0

154 62.4 51.8 77.6









#55
4-Nitrophenol
Concen: 17.656 ng/ul
RT: 14.724 min Scan# 1987
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:109 Resp: 103591
Ion Ratio Lower Upper
109 100
139 124.6 106.4 159.6
65 145.6 104.8 157.2



#56
Dibenzofuran
Concen: 22.462 ng/ul
RT: 14.866 min Scan# 2011
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:168 Resp: 1289240
Ion Ratio Lower Upper
168 100
139 38.8 30.7 46.1









#57
2,4-Dinitrotoluene
Concen: 22.181 ng/ul
RT: 14.854 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:165 Resp: 314987
Ion Ratio Lower Upper
165 100
63 62.7 39.4 59.2#
182 2.7 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.427 ng/ul
RT: 15.101 min Scan# 2051
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:232 Resp: 252962
Ion Ratio Lower Upper
232 100
131 49.6 38.0 57.0
130 2.8 2.0 3.0
166 31.5 24.8 37.2





#59

Diethylphthalate

Concen: 21.343 ng/ul

RT: 15.289 min Scan# 2083

Delta R.T. -0.006 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:149 Resp: 1118011

Ion Ratio Lower Upper

149 100

177 20.9 17.3 25.9

150 12.2 9.8 14.6



#61

Fluorene

Concen: 22.767 ng/ul

RT: 15.518 min Scan# 2122

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29




Tgt Ion:166 Resp: 1080950


Ion Ratio Lower Upper

166 100

165 97.5 78.2 117.2

167 13.3 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 22.136 ng/ul
RT: 15.513 min Scan# 2121
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:204 Resp: 520927
Ion Ratio Lower Upper
204 100
206 32.2 26.8 40.2
141 58.6 48.5 72.7



#63
4-Nitroaniline
Concen: 14.736 ng/ul
RT: 15.560 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:138 Resp: 108244
Ion Ratio Lower Upper
138 100
92 57.2 44.1 66.1
108 88.3 64.6 97.0









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.211 ng/ul
 RT: 15.618 min Scan# 2139
 Delta R.T. 0.000 min
 Lab File: BM039650.D
 Acq: 25 Apr 2023 21:29




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020125


Tgt Ion:198 Resp: 168818
 Ion Ratio Lower Upper
 198 100
 182 4.4 3.1 4.7
 77 29.8 20.6 31.0



#67
 N-Nitrosodiphenylamine
 Concen: 22.292 ng/ul
 RT: 15.730 min Scan# 2158
 Delta R.T. 0.000 min
 Lab File: BM039650.D
 Acq: 25 Apr 2023 21:29

Tgt Ion:169 Resp: 868322
 Ion Ratio Lower Upper
 169 100
 168 66.1 53.2 79.8
 167 36.9 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 21.113 ng/ul
RT: 16.407 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:248 Resp: 298129
Ion Ratio Lower Upper
248 100
250 97.6 77.7 116.5
141 80.2 61.0 91.4



#69
Hexachlorobenzene
Concen: 21.679 ng/ul
RT: 16.518 min Scan# 2292
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:284 Resp: 351681
Ion Ratio Lower Upper
284 100
142 39.7 28.9 43.3
249 31.5 23.9 35.9





#70

Atrazine

Concen: 21.325 ng/ul

RT: 16.689 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:200 Resp: 308223

Ion Ratio Lower Upper

200 100

173 26.2 22.2 33.2

215 47.8 38.9 58.3



#71

Pentachlorophenol

Concen: 19.653 ng/ul

RT: 16.877 min Scan# 2353

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29




Tgt Ion:266 Resp: 186366


Ion Ratio Lower Upper

266 100

264 62.2 50.2 75.4

268 62.6 51.9 77.9





#72

Phenanthrene

Concen: 22.333 ng/ul

RT: 17.260 min Scan# 2419

Delta R.T. -0.006 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:178 Resp: 1621014

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.4

176 19.6 15.8 23.8



#74

Anthracene

Concen: 23.016 ng/ul

RT: 17.354 min Scan# 2434

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29




Tgt Ion:178 Resp: 1675881


Ion Ratio Lower Upper

178 100

179 15.4 12.1 18.1

176 19.2 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.430 ng/uL

RT: 13.260 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:216 Resp: 436726

Ion Ratio Lower Upper

216 100

214 78.5 63.0 94.4

218 48.3 38.1 57.1



#76

Pentachlorobenzene

Concen: 22.555 ng/uL

RT: 14.783 min Scan# 1997

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29




Tgt Ion:250 Resp: 432260


Ion Ratio Lower Upper

250 100

248 61.1 50.4 75.6

252 63.8 50.2 75.2









#77
 Carbazole
 Concen: 22.439 ng/ul
 RT: 17.636 min Scan# 2482
 Delta R.T. 0.000 min
 Lab File: BM039650.D
 Acq: 25 Apr 2023 21:29

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020125

Tgt Ion:167 Resp: 1451877




Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.0	25.4
139	13.2	10.6	15.8




#78
 Di-n-butylphthalate
 Concen: 20.871 ng/ul
 RT: 18.189 min Scan# 2576
 Delta R.T. 0.000 min
 Lab File: BM039650.D
 Acq: 25 Apr 2023 21:29

Tgt Ion:149 Resp: 1910061

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.0	10.6
104	5.5	4.1	6.1





#80

Fluoranthene

Concen: 20.546 ng/ul

RT: 19.283 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:202 Resp: 1882067

Ion Ratio Lower Upper

202 100

101 12.6 10.4 15.6

100 9.7 7.7 11.5



#82

Pyrene

Concen: 20.846 ng/ul

RT: 19.648 min Scan# 2824

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29




Tgt Ion:202 Resp: 1976837


Ion Ratio Lower Upper

202 100

101 15.3 12.1 18.1

100 12.3 9.6 14.4





#83

Butylbenzylphthalate

Concen: 19.046 ng/ul

RT: 20.548 min Scan# 2977

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:149 Resp: 852569

Ion Ratio Lower Upper

149 100

91 79.5 59.4 89.0

206 18.6 15.8 23.8



#84

3,3'-Dichlorobenzidine

Concen: 18.723 ng/ul

RT: 21.342 min Scan# 3112

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29




Tgt Ion:252 Resp: 488319


Ion Ratio Lower Upper

252 100

254 63.5 50.9 76.3

126 11.8 10.4 15.6





#85

Benzo(a)anthracene

Concen: 21.553 ng/ul

RT: 21.400 min Scan# 3122

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:228 Resp: 1833469

Ion Ratio Lower Upper

228 100

229 19.7 15.4 23.2

226 27.2 21.6 32.4



#86

Bis(2-ethylhexyl)phthalate

Concen: 18.224 ng/ul

RT: 21.324 min Scan# 3109

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29




Tgt Ion:149 Resp: 1213127


Ion Ratio Lower Upper

149 100

167 27.0 21.8 32.6

279 4.0 3.1 4.7





#87

Chrysene

Concen: 21.439 ng/ul

RT: 21.453 min Scan# 3131

Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:228 Resp: 1739615

Ion Ratio Lower Upper

228 100

226 30.0 23.9 35.9

229 20.0 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 18.131 ng/ul




RT: 22.236 min Scan# 3264


Delta R.T. 0.000 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Tgt Ion:149 Resp: 1991715









#90
Benzo(b)fluoranthene
Concen: 21.723 ng/ul
RT: 23.059 min Scan# 3404
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:252 Resp: 1671151
Ion Ratio Lower Upper
252 100
253 22.3 17.4 26.2
125 11.6 8.7 13.1



#91
Benzo(k)fluoranthene
Concen: 21.850 ng/ul
RT: 23.106 min Scan# 3412
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:252 Resp: 1650746
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 11.0 8.6 12.8









#93
Benzo(a)pyrene
Concen: 21.778 ng/ul
RT: 23.671 min Scan# 3508
Delta R.T. 0.000 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29




Instrument :
BNA_M
ClientSampleId :
SSTD020125


Tgt Ion:252 Resp: 1412181
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 12.2 9.4 14.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 22.007 ng/ul
RT: 26.218 min Scan# 3941
Delta R.T. -0.006 min
Lab File: BM039650.D
Acq: 25 Apr 2023 21:29

Tgt Ion:276 Resp: 1569663
Ion Ratio Lower Upper
276 100
138 26.9 21.2 31.8
277 25.3 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 22.271 ng/ul

RT: 26.230 min Scan# 3944

Delta R.T. -0.006 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Instrument :

BNA_M

ClientSampleId :

SSTD020125





Tgt Ion:278 Resp: 1317058

Ion Ratio Lower Upper

278 100

139 17.8 13.5 20.3

279 23.4 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 21.818 ng/ul

RT: 26.971 min Scan# 4069

Delta R.T. -0.006 min

Lab File: BM039650.D

Acq: 25 Apr 2023 21:29

Tgt Ion:276 Resp: 1239664

Ion Ratio Lower Upper

276 100

138 24.5 18.7 28.1

277 24.1 18.7 28.1

