


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042618\
Data File : BM014851.D
Acq On : 26 Apr 2018 08:53
Operator : SJ/JU
Sample : DFTPP30
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP30

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM041918.M
Title : SVOA CALIBRATION
Last Update : Fri Apr 27 10:59:40 2018



AutoFind: Scans 2535, 2536, 2537; Background Corrected with Scan 2530

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.0	161703	PASS
68	69	0.00	2	1.7	2605	PASS
69	198	0.00	100	46.2	149312	PASS
70	69	0.00	2	0.7	982	PASS
127	198	10	80	53.9	174314	PASS
197	198	0.00	2	0.6	2027	PASS
198	198	100	100	100.0	323520	PASS
199	198	5	9	6.6	21472	PASS
275	198	10	60	29.9	96586	PASS
365	198	1	100	4.2	13734	PASS
441	443	0.01	100	76.2	45802	PASS
442	198	50	100	93.5	302528	PASS
443	442	15	24	19.9	60117	PASS

m/z	Abundance
37.00	561.0
38.10	2483.0
39.10	12633.0
40.00	2966.0
41.10	728.0
43.00	526.0
44.00	4379.0
45.00	795.0
49.00	1632.0
50.10	38024.0
51.10	145600.0
52.10	7018.0
53.10	422.0
55.00	1090.0
56.00	4038.0
57.00	9951.0
57.90	475.0
61.00	2146.0
62.00	1762.0
63.00	5065.0
64.00	663.0
65.00	2557.0
66.90	320.0
68.10	2229.0
69.00	132096.0
70.00	1021.0
73.00	2013.0
74.00	13067.0
75.00	21192.0
76.00	6657.0
77.10	144512.0
78.00	10048.0
79.00	10072.0
80.00	7900.0
81.00	9899.0
82.00	2392.0
83.00	2457.0
83.90	820.0
85.10	1525.0
86.00	3227.0
87.10	1496.0
87.90	508.0
89.10	300.0
91.00	2365.0
92.00	2827.0
93.00	15918.0
94.00	1012.0
95.00	406.0
95.90	1460.0
97.10	377.0
98.00	12195.0
99.00	11859.0
99.90	977.0
101.00	5248.0
102.00	439.0
103.00	2047.0
104.00	3346.0
105.00	3065.0
106.00	1303.0
107.00	41328.0
108.00	6745.0
109.00	1280.0
110.00	72656.0
111.00	12333.0
112.00	1440.0
113.00	588.0
114.90	443.0
116.00	2310.0
117.00	33984.0
118.00	2936.0
119.00	661.0
120.00	564.0
122.00	2875.0
123.00	4454.0
123.90	2149.0
125.00	2085.0
127.00	144768.0
128.00	11781.0
129.00	55552.0
130.00	4803.0
131.00	1204.0
132.00	433.0
133.00	801.0
133.90	1575.0
135.00	4498.0
135.90	1687.0
137.00	3112.0
138.00	594.0
140.00	782.0
141.00	7110.0
142.00	2396.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

142.90	1828.0
143.90	497.0
144.90	400.0
145.90	1680.0
147.00	3965.0
148.00	8253.0
148.90	2707.0
150.00	695.0
151.10	760.0
151.70	702.0
153.00	2368.0
153.90	1379.0
155.00	4264.0
156.00	6009.0
157.10	1315.0
158.00	1215.0
159.00	1105.0
160.00	2171.0
161.00	3656.0
162.00	901.0
162.80	529.0
163.90	526.0
164.90	2930.0
166.00	1911.0
167.00	14334.0
168.00	12478.0
169.00	1658.0
170.00	468.0
170.90	611.0
172.00	1252.0
172.90	1631.0
174.00	2991.0
175.00	4948.0
176.00	1748.0
176.90	2906.0
178.00	1002.0
179.00	10255.0
180.00	7960.0
181.00	3101.0
181.90	458.0
182.90	335.0
184.00	750.0
185.00	4803.0
186.00	37224.0
187.00	11085.0
188.00	1327.0
189.00	2697.0
190.00	421.0
190.90	1773.0
192.00	3324.0
193.00	4149.0
193.90	972.0
195.00	397.0
196.00	6826.0
197.10	1484.0
197.90	245632.0
198.90	17968.0
200.00	1055.0
201.40	1150.0
202.90	1862.0
204.00	9153.0
205.00	15213.0
206.00	66464.0
207.00	14341.0
208.00	3247.0
209.00	1471.0
210.00	1005.0
210.90	2398.0
211.80	850.0
214.90	644.0
216.00	1357.0
217.00	17248.0
218.00	2194.0
221.00	10974.0
221.80	2178.0
222.90	3578.0
224.00	35784.0
225.00	9856.0
226.10	1414.0
227.00	14332.0
228.00	2371.0
229.00	3388.0
229.90	437.0
230.90	1152.0
234.00	856.0
234.90	1074.0
236.00	521.0
237.00	2370.0
237.90	425.0
239.00	494.0
239.90	544.0
241.00	946.0
242.00	1933.0
243.00	2397.0
244.00	27240.0

Instrument :

BNA_M

ClientSampleId :

DFTPP30

245.00	3698.0
246.00	5728.0
246.90	997.0
249.00	1072.0
250.90	365.0
253.00	688.0
254.20	1279.0
255.00	149568.0
256.00	22808.0
256.90	2144.0
258.00	8529.0
259.00	1717.0
260.90	323.0
265.00	3496.0
266.00	649.0
271.00	446.0
271.90	425.0
273.00	4373.0
274.00	10664.0
275.00	67504.0
276.00	9835.0
277.00	5513.0
278.00	1433.0
281.00	1484.0
282.10	570.0
283.00	1054.0
284.10	670.0
285.00	862.0
293.00	1157.0
293.90	611.0
295.00	458.0
296.00	19648.0
297.00	3110.0
300.80	372.0
301.90	310.0
303.00	2232.0
304.00	674.0
309.90	421.0
314.00	1089.0
314.90	2148.0
316.00	1248.0
317.00	311.0
320.90	527.0
323.00	5704.0
324.00	965.0
326.90	1636.0
327.90	931.0
331.90	395.0
333.00	611.0
334.00	3918.0
335.00	1172.0
341.00	780.0
346.00	1253.0
351.90	1728.0
353.00	1344.0
354.00	1840.0
355.00	669.0
365.00	8437.0
366.00	1364.0
371.00	495.0
372.00	3110.0
373.00	879.0
382.90	860.0
389.90	505.0
390.90	346.0
392.00	377.0
401.90	1434.0
403.00	1796.0
404.00	789.0
421.00	1406.0
422.00	1789.0
423.00	10448.0
424.00	4251.0
424.90	589.0
441.00	27192.0
442.00	174144.0
443.00	35952.0
444.00	3322.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

m/z	Abundance
36.00	320.0
37.00	889.0
38.10	2597.0
39.00	15499.0
40.00	3226.0
41.10	801.0
43.00	581.0
44.00	4842.0
45.00	676.0
47.00	301.0
49.10	2235.0
50.00	50240.0
51.10	190208.0
52.00	9449.0
53.00	456.0
55.00	1064.0
56.00	6147.0
57.00	13166.0
58.00	654.0
59.90	308.0
61.00	2761.0
62.00	2640.0
63.00	7128.0
64.00	917.0
65.10	4038.0
67.10	336.0
68.10	3278.0
69.00	178304.0
70.00	1109.0
73.00	2622.0
74.00	18208.0
75.00	28760.0
76.10	9448.0
77.10	195520.0
78.10	13742.0
79.00	13213.0
80.00	10276.0
81.00	13393.0
82.00	3288.0
83.00	3763.0
84.00	837.0
85.00	2149.0
86.00	4765.0
87.00	2047.0
88.00	862.0
89.10	380.0
91.00	3451.0
92.00	3580.0
93.00	21832.0
93.90	1268.0
95.10	386.0
96.00	1324.0
97.10	604.0
98.00	16616.0
99.00	15892.0
100.00	1441.0
101.00	8079.0
102.10	362.0
103.00	2779.0
104.00	5188.0
105.00	4376.0
106.00	1505.0
107.00	57712.0
108.00	9129.0
109.10	2014.0
110.00	103248.0
111.00	16872.0
112.00	1744.0
113.00	772.0
115.10	393.0
116.00	3173.0
117.00	49824.0
118.00	4291.0
119.00	593.0
120.00	830.0
122.00	4290.0
123.00	6373.0
124.00	2600.0
125.00	2747.0
126.10	1349.0
127.00	208256.0
128.00	17360.0
129.00	81448.0
130.00	7393.0
131.00	1436.0
131.90	745.0
133.00	777.0
134.00	2206.0
135.00	6139.0
136.00	3207.0
137.00	4016.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

137.90	992.0
140.00	1191.0
141.00	11587.0
142.00	3502.0
143.00	2038.0
144.00	681.0
144.80	571.0
145.90	2158.0
147.00	5824.0
148.00	13018.0
149.00	3058.0
150.00	693.0
151.00	1395.0
151.90	1034.0
153.00	2932.0
154.00	2444.0
155.00	5998.0
156.00	7895.0
157.00	1835.0
158.00	1771.0
159.00	1583.0
160.00	3255.0
161.00	4928.0
162.00	1395.0
162.80	644.0
163.90	605.0
165.00	4276.0
166.00	3045.0
167.00	20104.0
168.00	14599.0
169.00	2090.0
170.00	808.0
171.00	901.0
171.90	1867.0
173.00	2247.0
174.00	4029.0
175.00	8611.0
176.00	2341.0
177.00	4074.0
177.90	1450.0
179.00	14850.0
180.00	12074.0
181.00	5480.0
182.10	934.0
183.10	328.0
184.00	1312.0
185.00	7584.0
186.00	55448.0
187.00	15794.0
188.00	1980.0
188.90	3406.0
189.90	671.0
190.90	2661.0
192.00	5072.0
193.00	6187.0
193.90	1239.0
194.90	573.0
196.00	11451.0
197.10	2294.0
197.90	391232.0
199.00	25016.0
200.00	2166.0
201.40	1679.0
203.00	2431.0
204.00	14720.0
205.00	25560.0
206.00	106552.0
207.00	18568.0
208.00	4803.0
209.00	2196.0
210.20	1779.0
211.00	4430.0
212.90	300.0
215.00	1243.0
216.00	2367.0
216.90	27192.0
217.90	3466.0
221.00	16960.0
221.80	3622.0
223.00	5866.0
224.00	58360.0
225.00	15821.0
226.00	1991.0
227.00	23344.0
228.00	3545.0
229.00	5180.0
230.00	621.0
231.00	2018.0
232.10	346.0
233.00	480.0
233.90	1803.0
234.90	1651.0
235.90	1039.0
237.00	3522.0
237.80	347.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

239.00	931.0
240.00	1163.0
241.00	1265.0
242.00	3026.0
243.00	3968.0
244.00	44040.0
245.00	6864.0
246.00	8448.0
246.90	1981.0
247.80	570.0
248.90	1690.0
250.10	339.0
251.00	370.0
252.00	506.0
252.90	989.0
254.10	1780.0
255.00	247552.0
256.00	36912.0
257.00	3278.0
258.00	13886.0
259.00	2004.0
260.00	414.0
261.00	318.0
263.90	368.0
265.00	5495.0
265.90	784.0
267.00	534.0
271.00	568.0
272.00	793.0
273.00	7956.0
274.00	18568.0
275.00	114712.0
276.00	16448.0
276.90	9415.0
277.90	2065.0
279.00	388.0
280.90	1638.0
282.10	642.0
283.00	1111.0
284.10	889.0
285.00	2055.0
288.80	421.0
289.90	407.0
291.90	419.0
293.00	2386.0
293.90	592.0
295.00	811.0
296.00	32720.0
297.00	4813.0
298.00	418.0
300.80	498.0
301.90	537.0
302.90	3729.0
303.90	928.0
308.00	432.0
308.90	341.0
309.90	394.0
314.00	1813.0
315.00	3511.0
316.00	1733.0
317.00	425.0
321.00	1041.0
321.90	804.0
323.00	10614.0
324.00	1897.0
326.90	2040.0
327.90	1264.0
332.00	871.0
332.90	931.0
334.00	6987.0
335.00	1824.0
341.00	1294.0
342.80	312.0
345.90	2786.0
347.00	630.0
352.00	3545.0
353.00	2243.0
354.00	3216.0
355.00	738.0
365.00	15582.0
365.90	2477.0
371.00	971.0
372.00	5802.0
373.00	1571.0
383.00	1703.0
384.00	544.0
389.90	583.0
390.90	602.0
392.00	374.0
402.00	2423.0
403.00	3260.0
404.00	1483.0
404.90	341.0
421.00	3016.0
422.00	2880.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

423.00	19664.0
424.00	6890.0
424.90	1016.0
441.00	52464.0
442.00	339456.0
443.00	69360.0
444.00	6024.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

m/z	Abundance
37.00	901.0
38.00	1941.0
39.10	12982.0
40.00	2900.0
41.00	855.0
42.90	396.0
44.00	5187.0
45.00	549.0
48.00	317.0
49.00	1808.0
50.00	38464.0
51.10	150400.0
52.10	8524.0
53.00	392.0
55.10	1049.0
56.00	4477.0
57.00	10159.0
58.00	414.0
61.00	2129.0
62.00	2125.0
63.00	6044.0
64.00	830.0
65.10	2674.0
68.00	2308.0
69.00	137536.0
69.90	816.0
73.10	2639.0
74.00	14172.0
75.00	22280.0
76.10	6902.0
77.10	150144.0
78.10	11144.0
79.00	11487.0
80.00	7565.0
81.00	11284.0
82.00	2381.0
83.00	2396.0
84.00	770.0
85.00	1842.0
86.00	3490.0
87.00	1349.0
87.90	694.0
88.90	408.0
91.00	2573.0
92.10	2839.0
93.00	17872.0
94.00	1266.0
95.00	353.0
96.00	1291.0
96.90	507.0
98.00	13914.0
99.00	11866.0
100.00	907.0
101.00	5868.0
101.90	332.0
103.00	2326.0
104.00	3761.0
105.00	3586.0
106.00	1059.0
107.00	46888.0
108.00	6757.0
108.90	1407.0
110.00	82000.0
111.00	13239.0
112.00	1544.0
113.00	691.0
114.90	370.0
116.00	2350.0
117.00	39904.0
118.00	3278.0
119.10	828.0
120.00	455.0
122.00	3037.0
123.00	4855.0
123.90	2205.0
125.00	2147.0
126.10	751.0
127.00	169920.0
128.00	12688.0
129.00	62992.0
130.00	5068.0
131.00	1093.0
132.00	679.0
132.90	897.0
134.00	1616.0
135.00	5654.0
136.00	1747.0
137.00	3369.0
137.80	702.0
140.00	964.0
141.00	8112.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

142.00	2775.0
142.90	2532.0
143.90	536.0
145.00	556.0
145.90	1453.0
146.90	4819.0
148.00	10377.0
149.00	2590.0
149.90	691.0
151.10	1228.0
151.90	665.0
152.90	3147.0
154.00	2250.0
155.00	4938.0
156.00	6886.0
157.00	1438.0
158.00	1807.0
159.00	1152.0
159.90	2867.0
161.00	4584.0
162.00	1237.0
162.90	531.0
164.10	390.0
164.90	3191.0
166.00	2672.0
167.00	17432.0
168.00	11422.0
169.00	1612.0
169.90	481.0
171.00	709.0
171.90	1474.0
173.00	1924.0
174.00	3586.0
175.00	7310.0
176.10	1893.0
176.90	3289.0
177.90	1052.0
178.90	13163.0
180.00	9900.0
181.00	4168.0
182.00	608.0
184.00	1174.0
185.00	6573.0
186.00	49104.0
187.00	13796.0
188.00	1306.0
189.00	3553.0
190.00	661.0
191.00	2261.0
192.00	4224.0
193.00	5338.0
194.10	1421.0
196.00	10577.0
197.10	2305.0
197.90	333696.0
198.90	21432.0
200.00	2122.0
201.40	1720.0
202.90	2656.0
204.00	13514.0
205.00	21464.0
206.00	93616.0
207.00	17896.0
208.00	4509.0
209.00	1763.0
210.30	1909.0
211.00	3694.0
213.10	449.0
214.90	1163.0
215.90	2010.0
217.00	23480.0
218.00	3265.0
221.00	14737.0
221.80	3676.0
223.00	5375.0
224.00	50632.0
225.00	12787.0
226.00	1453.0
227.00	21096.0
227.90	3038.0
229.00	4562.0
229.90	551.0
231.00	1759.0
231.90	373.0
233.90	1456.0
235.00	2063.0
236.00	1016.0
237.00	2438.0
237.90	348.0
239.00	835.0
239.90	946.0
241.00	1270.0
242.00	3073.0
243.00	3079.0
244.00	41808.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

245.00	5591.0
246.00	8891.0
247.00	1514.0
248.10	352.0
248.90	1944.0
249.90	318.0
250.90	560.0
251.90	474.0
252.90	1307.0
254.10	1982.0
255.00	229184.0
256.00	31784.0
256.90	2706.0
258.00	13331.0
259.00	2156.0
260.00	354.0
261.00	375.0
264.00	395.0
265.00	6057.0
265.90	750.0
267.00	322.0
270.90	611.0
271.90	668.0
273.00	7252.0
274.00	18152.0
275.00	107544.0
276.00	13831.0
277.00	9832.0
277.90	1634.0
278.80	346.0
281.00	1585.0
282.00	671.0
283.00	1338.0
284.00	676.0
285.00	1666.0
286.00	339.0
288.80	324.0
289.90	454.0
291.90	570.0
293.00	2280.0
293.80	576.0
295.10	814.0
296.00	32216.0
297.00	4904.0
298.10	353.0
301.00	457.0
301.90	686.0
303.00	4214.0
304.00	1052.0
308.00	523.0
310.00	342.0
314.00	1579.0
315.00	3494.0
316.10	1949.0
317.00	347.0
321.00	1138.0
322.00	515.0
323.00	10449.0
324.00	1799.0
326.90	2400.0
327.90	1160.0
332.00	1037.0
333.00	1333.0
334.00	7219.0
335.00	1704.0
341.00	1440.0
342.00	368.0
345.90	2309.0
346.80	419.0
352.00	3772.0
353.00	2023.0
354.00	3520.0
355.00	893.0
364.90	17184.0
366.00	2265.0
369.90	340.0
371.00	1134.0
372.00	6539.0
373.00	1614.0
382.90	1741.0
384.00	475.0
389.90	837.0
391.00	751.0
392.00	353.0
400.90	456.0
402.00	2397.0
403.00	3841.0
404.00	1336.0
405.00	405.0
414.90	394.0
421.00	3373.0
422.00	2774.0
423.00	21832.0
424.00	6561.0
424.90	884.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

441.00	57752.0
442.00	393984.0
443.00	75040.0
444.10	6725.0
445.00	419.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30

m/z	Abundance
39.00	495.0
40.00	2571.0
41.00	435.0
43.10	454.0
44.00	4905.0
44.90	338.0
49.00	821.0
50.90	366.0
55.10	321.0
73.00	1307.0
77.10	300.0
83.90	489.0
85.90	346.0
96.00	453.0
132.90	689.0
135.00	305.0
147.00	565.0
176.90	329.0
191.00	786.0
193.00	537.0
207.00	5836.0
208.10	1308.0
209.00	648.0
281.00	1475.0
281.90	440.0
282.90	417.0

Instrument :
BNA_M
ClientSampleId :
DFTPP30