

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042715\
 Data File : BM001115.D
 Acq On : 27 Apr 2015 12:28
 Operator : TP/IZ
 Sample : SSTD0.846
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.846

Manual Integrations
 APPROVED

apatel
 4/28/2015 2:18:31 PM

Quant Time: Apr 27 14:17:46 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM042715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 27 13:50:45 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.54	152	1050	0.40	ng/ul	0.00
2) Naphthalene-d8	10.32	136	3619	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.20	164	1869	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.95	188	4162	0.40	ng/ul	0.00
16) Chrysene-d12	21.16	240	3882	0.40	ng/ul	0.00
20) Perylene-d12	23.33	264	3680	0.40	ng/ul	0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.93	152	3682	0.78	ng/ul	0.00
14) Fluoranthene-d10	19.00	212	7429	0.81	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.37	128	7487	0.80	ng/ul	97
5) 2-Methylnaphthalene	12.00	142	4925	0.79	ng/ul	100
7) Acenaphthylene	13.91	152	6796	0.81	ng/ul	97
8) Acenaphthene	14.27	153	5203	0.79	ng/ul	97
9) Fluorene	15.26	166	6069	0.80	ng/ul	99
11) Pentachlorophenol	16.63	266	583	0.63	ng/ul	98
12) Phenanthrene	17.00	178	9471	0.78	ng/ul	98
13) Anthracene	17.09	178	8814	0.81	ng/ul	97
15) Fluoranthene	19.02	202	11111	0.82	ng/ul	98
17) Pyrene	19.40	202	11273	0.77	ng/ul#	91
18) Benzo(a)anthracene	21.15	228	10050m	0.82	ng/ul	
19) Chrysene	21.19	228	10723	0.79	ng/ul	97
21) Benzo(b)fluoranthene	22.67	252	10500	0.77	ng/ul	96
22) Benzo(k)fluoranthene	22.71	252	11160	0.83	ng/ul	94
23) Benzo(a)pyrene	23.22	252	10396	0.81	ng/ul	95
24) Indeno(1,2,3-cd)pyrene	25.46	276	11976	0.84	ng/ul	99
25) Dibenzo(a,h)anthracene	25.49	278	9635	0.83	ng/ul	96
26) Benzo(g,h,i)perylene	26.12	276	10402	0.82	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042715\
Data File : BM001115.D
Acq On : 27 Apr 2015 12:28
Operator : TP/IZ
Sample : SST0.846
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleID :
SSTD0.846

Manual Integrations
APPROVED

apatel
4/28/2015 2:18:31 PM

Quant Time: Apr 27 14:17:46 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-SIM-BM042715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 27 13:50:45 2015
Response via : Initial Calibration

