


Data Path : Z:\HPCHEM1\BNA M\Data\BM042716\
 Data File : BM005095.D
 Acq On : 27 Apr 2016 11:40
 Operator : UM/SJ
 Sample : DFTPP58
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP58

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM042516.M
 Title : SVOA CALIBRATION
 Last Update : Thu Apr 28 04:13:24 2016



AutoFind: Scans 2538, 2539, 2540; Background Corrected with Scan 2533

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.3	42640	PASS
68	69	0.00	2	1.7	753	PASS
69	198	0.00	100	32.0	43557	PASS
70	69	0.00	2	0.2	104	PASS
127	198	10	80	43.5	59205	PASS
197	198	0.00	2	0.6	876	PASS
198	198	100	100	100.0	136077	PASS
199	198	5	9	6.8	9187	PASS
275	198	10	60	25.6	34858	PASS
365	198	1	100	3.1	4237	PASS
441	443	0.01	100	79.8	16798	PASS
442	198	50	100	79.3	107925	PASS
443	442	15	24	19.5	21054	PASS

Scan 2533 (17.586 min): BM005095.D
DFTPP58

m/z	Abundance
44.00	533.0
207.00	386.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

m/z	Abundance
38.00	465.0
39.10	2975.0
40.00	366.0
44.00	473.0
49.10	348.0
50.10	11328.0
51.10	45824.0
52.00	2285.0
55.00	329.0
56.00	1333.0
57.00	3210.0
61.00	617.0
62.00	682.0
63.00	1809.0
65.10	937.0
68.00	846.0
69.00	45184.0
70.00	312.0
73.10	472.0
74.10	4678.0
75.00	7757.0
76.10	2633.0
77.10	57120.0
78.10	4133.0
79.00	3541.0
80.00	2657.0
81.00	3743.0
82.00	937.0
83.00	931.0
85.00	786.0
86.00	1136.0
87.00	528.0
91.00	835.0
92.00	1010.0
93.00	6304.0
94.00	408.0
96.10	374.0
98.00	5154.0
99.00	4347.0
100.00	384.0
101.00	2316.0
102.90	795.0
104.00	1504.0
105.00	1221.0
106.00	536.0
107.00	18128.0
108.00	2939.0
109.00	527.0
110.00	32336.0
111.00	4872.0
112.00	570.0
116.00	893.0
117.00	14546.0
118.00	1206.0
122.00	1158.0
123.00	1715.0
124.00	795.0
125.00	679.0
126.20	334.0
127.00	59464.0
128.00	4877.0
129.00	22768.0
130.00	1909.0
131.00	452.0
134.00	754.0
135.00	2095.0
136.00	709.0
137.10	1193.0
141.00	3161.0
142.00	1080.0
143.00	708.0
146.00	569.0
147.00	1750.0
148.00	3647.0
149.00	947.0
151.10	344.0
153.00	1124.0
154.00	753.0
155.00	1737.0
156.00	2601.0
157.10	497.0
158.00	539.0
159.00	409.0
160.00	951.0
161.00	1412.0
162.00	353.0
165.00	1120.0
166.10	881.0
167.00	6193.0
168.00	4225.0
169.00	636.0
172.00	458.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

173.00	772.0
174.00	1335.0
175.10	2497.0
176.00	664.0
177.00	1162.0
179.00	4559.0
180.00	3572.0
181.00	1589.0
184.00	320.0
185.00	2183.0
186.10	18120.0
187.10	5493.0
188.00	740.0
189.00	986.0
191.00	522.0
192.00	1528.0
193.10	1833.0
194.00	320.0
196.10	3510.0
197.10	805.0
198.00	128560.0
199.00	8926.0
200.00	652.0
201.50	539.0
203.00	865.0
204.10	4500.0
205.10	7933.0
206.10	33392.0
207.10	4834.0
208.00	1219.0
209.00	372.0
210.00	486.0
211.00	1107.0
215.00	368.0
216.00	756.0
217.00	8803.0
218.00	1193.0
221.10	4998.0
221.90	1148.0
223.10	1802.0
224.10	18528.0
225.10	4775.0
226.00	507.0
227.00	7610.0
228.00	1119.0
229.00	1708.0
231.10	788.0
234.00	532.0
235.00	476.0
236.00	343.0
237.00	870.0
241.00	446.0
242.10	931.0
243.10	943.0
244.10	13893.0
245.10	1970.0
246.00	2611.0
247.00	446.0
249.00	525.0
253.00	344.0
254.10	631.0
255.10	69672.0
256.10	10042.0
257.00	932.0
258.00	3896.0
259.00	682.0
265.00	1799.0
273.00	1871.0
274.00	5445.0
275.10	30328.0
276.10	4192.0
277.00	2520.0
278.00	491.0
285.00	419.0
293.00	533.0
296.10	7963.0
297.00	1309.0
303.00	1046.0
314.10	361.0
315.00	818.0
316.00	524.0
323.10	2779.0
324.00	426.0
326.90	467.0
334.10	1524.0
335.00	342.0
346.00	545.0
352.00	675.0
353.10	495.0
354.10	774.0
365.00	3262.0
366.00	446.0
372.10	1187.0
373.10	311.0
383.00	330.0
402.10	496.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

403.00	604.0
421.00	558.0
422.00	606.0
423.00	4361.0
424.10	1431.0
441.10	12162.0
442.10	77304.0
443.10	15466.0
444.10	1543.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

m/z	Abundance
38.00	462.0
39.10	3171.0
40.00	320.0
44.00	477.0
49.10	340.0
50.10	11856.0
51.10	49336.0
52.10	2600.0
55.00	349.0
56.00	1305.0
57.00	3346.0
61.00	726.0
62.00	675.0
63.00	1901.0
64.00	308.0
65.10	869.0
68.00	825.0
69.00	50984.0
73.10	485.0
74.00	5363.0
75.00	8443.0
76.10	3042.0
77.10	64096.0
78.10	4664.0
79.00	3921.0
80.00	3108.0
81.00	4355.0
82.00	1180.0
83.00	1067.0
85.10	942.0
86.00	1166.0
87.00	686.0
91.00	918.0
92.10	1057.0
93.00	6950.0
94.00	449.0
96.00	364.0
98.00	5352.0
99.00	4734.0
99.90	449.0
101.00	2386.0
103.00	813.0
104.00	1616.0
105.00	1507.0
106.00	429.0
107.00	20432.0
108.00	3056.0
109.10	566.0
110.00	36344.0
111.00	5791.0
112.00	775.0
116.00	1093.0
117.00	16760.0
118.00	1307.0
122.00	1411.0
123.00	2047.0
124.00	938.0
125.00	914.0
126.20	387.0
127.10	69672.0
128.10	5141.0
129.00	26944.0
130.00	2288.0
131.00	483.0
134.00	779.0
135.00	2187.0
136.00	900.0
137.00	1265.0
140.10	387.0
141.00	3443.0
142.00	1219.0
143.00	846.0
146.00	676.0
147.00	1921.0
148.00	4481.0
149.00	803.0
151.00	514.0
153.00	1089.0
154.00	948.0
155.00	2087.0
156.10	3324.0
157.10	589.0
158.00	692.0
159.00	518.0
160.00	1011.0
161.00	1783.0
162.10	408.0
165.00	1273.0
166.00	1177.0
167.00	8027.0
168.00	4501.0
169.00	723.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

171.00	373.0
172.00	664.0
173.10	925.0
174.00	1516.0
175.10	3133.0
176.00	763.0
177.00	1325.0
178.00	403.0
179.00	5875.0
180.00	4236.0
181.00	1914.0
184.00	524.0
185.10	2859.0
186.10	21840.0
187.10	6380.0
188.00	835.0
189.00	1515.0
191.00	651.0
192.00	1928.0
193.10	2177.0
194.00	453.0
196.00	4546.0
197.10	1103.0
198.00	160000.0
199.00	10721.0
200.00	779.0
201.40	705.0
203.00	1094.0
204.10	5520.0
205.10	9951.0
206.10	43200.0
207.10	5953.0
208.00	1467.0
209.10	477.0
210.30	588.0
211.00	1600.0
215.00	453.0
216.00	903.0
217.00	11257.0
218.00	1433.0
221.10	6291.0
221.90	1577.0
223.00	2348.0
224.10	23584.0
225.10	6288.0
226.10	689.0
227.10	9536.0
228.00	1448.0
229.00	2004.0
231.10	862.0
234.00	716.0
235.00	644.0
236.00	582.0
237.10	904.0
239.00	412.0
240.00	304.0
241.00	537.0
242.00	1218.0
243.00	1295.0
244.10	18296.0
245.10	2466.0
246.10	3596.0
247.10	742.0
249.00	653.0
253.00	484.0
254.10	971.0
255.10	92040.0
256.10	13888.0
257.10	1165.0
258.00	5135.0
259.00	1000.0
265.00	2111.0
273.00	2869.0
274.10	7478.0
275.10	40824.0
276.10	5118.0
277.10	3632.0
278.00	678.0
283.00	433.0
285.00	477.0
293.00	795.0
295.00	333.0
296.10	11147.0
297.00	1564.0
303.10	1348.0
304.10	419.0
314.00	533.0
315.00	1297.0
316.00	774.0
321.00	348.0
323.10	3422.0
324.10	631.0
327.00	637.0
328.00	354.0
334.10	2379.0
335.00	519.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

341.00	451.0
346.00	815.0
352.10	1083.0
353.10	787.0
354.00	1111.0
365.00	5081.0
366.00	793.0
372.10	1849.0
373.00	442.0
383.00	536.0
402.00	794.0
403.00	1098.0
404.00	464.0
421.10	930.0
422.10	868.0
423.10	6654.0
424.10	1894.0
441.10	20000.0
442.10	127160.0
443.10	24896.0
444.10	2077.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

m/z	Abundance
38.00	356.0
39.00	2134.0
40.00	350.0
44.00	474.0
50.10	7985.0
51.10	32760.0
52.10	1611.0
56.00	940.0
57.00	2180.0
61.00	458.0
62.00	461.0
63.00	1353.0
65.00	731.0
68.00	588.0
69.00	34504.0
73.00	365.0
74.10	3470.0
75.00	5765.0
76.10	1899.0
77.10	42872.0
78.10	2956.0
79.00	2812.0
80.00	1939.0
81.10	3045.0
82.00	680.0
83.00	729.0
84.90	707.0
86.00	879.0
87.00	492.0
91.00	697.0
92.00	651.0
93.00	4652.0
93.90	336.0
98.00	3925.0
99.00	2974.0
99.90	321.0
101.00	1663.0
103.00	570.0
104.00	971.0
105.00	1077.0
106.10	369.0
107.00	13978.0
108.00	2115.0
109.10	417.0
110.00	25520.0
111.00	3724.0
112.00	423.0
116.00	747.0
117.00	11676.0
118.00	785.0
122.00	840.0
123.00	1375.0
124.00	603.0
125.00	686.0
127.00	48480.0
128.10	3680.0
129.00	18224.0
130.00	1667.0
131.00	317.0
134.00	508.0
135.00	1709.0
136.00	611.0
137.00	898.0
141.00	2664.0
142.00	1061.0
143.00	515.0
146.10	433.0
147.00	1426.0
148.00	2998.0
149.00	703.0
151.00	341.0
153.00	826.0
154.00	545.0
155.00	1212.0
156.00	2421.0
157.00	434.0
158.00	498.0
158.90	359.0
160.00	871.0
161.00	1268.0
162.00	359.0
165.00	975.0
166.00	751.0
167.00	5537.0
168.10	3467.0
169.00	375.0
172.00	447.0
173.10	603.0
174.00	1286.0
175.10	2337.0
176.10	653.0
177.00	1036.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

178.00	366.0
179.00	4127.0
180.00	3097.0
181.00	1507.0
184.00	358.0
185.00	2114.0
186.10	16640.0
187.10	4367.0
188.10	499.0
189.00	1127.0
191.00	597.0
192.00	1324.0
193.00	1592.0
194.00	377.0
196.10	2982.0
197.10	720.0
198.00	119672.0
199.00	7916.0
200.00	642.0
201.50	474.0
203.00	906.0
204.10	4368.0
205.10	7668.0
206.10	32208.0
207.10	4547.0
208.00	905.0
209.10	315.0
211.00	1176.0
215.00	331.0
216.00	765.0
217.00	8382.0
218.00	1060.0
221.10	4801.0
221.90	1189.0
223.00	1766.0
224.10	18736.0
225.10	4673.0
226.10	440.0
227.10	7293.0
228.00	1031.0
229.00	1535.0
231.10	628.0
233.90	400.0
235.00	477.0
236.00	364.0
237.00	815.0
241.00	360.0
242.10	857.0
243.10	998.0
244.10	14219.0
245.10	1777.0
246.00	2886.0
247.00	472.0
249.00	532.0
253.00	347.0
254.10	571.0
255.10	73192.0
256.10	10534.0
257.10	818.0
258.00	3926.0
259.00	701.0
265.00	1600.0
273.00	2160.0
274.10	6085.0
275.10	33424.0
276.10	4327.0
277.00	2972.0
278.00	444.0
285.10	473.0
293.00	583.0
296.00	9476.0
297.00	1374.0
303.10	1151.0
304.10	313.0
314.00	467.0
315.00	1182.0
316.10	605.0
323.10	3159.0
324.10	515.0
327.00	575.0
328.00	342.0
334.00	1771.0
335.00	482.0
341.00	399.0
346.00	853.0
352.10	1003.0
353.00	634.0
354.10	877.0
365.00	4368.0
366.00	596.0
372.10	1605.0
373.00	373.0
383.10	500.0
402.00	739.0
403.00	922.0
404.00	327.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58

421.00	989.0
422.00	859.0
423.10	6354.0
424.10	1422.0
441.10	18232.0
442.10	119312.0
443.10	22800.0
444.10	2093.0

Instrument :
BNA_M
ClientSampleId :
DFTPP58