

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042717\
 Data File : BM009754.D
 Acq On : 27 Apr 2017 23:29
 Operator : SJ/MA
 Sample : I2809-21
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 JHTE1

Quant Time: Apr 28 01:11:25 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM042517.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 28 01:10:39 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	127305	20.00	ng/ul	0.00
18) Naphthalene-d8	10.62	136	549262	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.47	164	348226	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.22	188	771100	20.00	ng/ul	0.00
75) Chrysene-d12	21.40	240	814552	20.00	ng/ul	0.00
83) Perylene-d12	23.71	264	690182	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.31	96	14614	5.29	ng/uL	0.00
5) Phenol-d5	6.99	99	78068	5.64	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.16	67	239756	25.42	ng/ul	0.00
9) 2-Chlorophenol-d4	7.35	132	187596	21.32	ng/ul	0.00
13) 4-Methylphenol-d8	8.52	113	141292	13.11	ng/ul	0.00
19) Nitrobenzene-d5	8.99	128	125019	27.96	ng/ul	0.00
22) 2-Nitrophenol-d4	9.71	143	137823	29.30	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.24	165	244020	25.94	ng/ul	0.00
29) 4-Chloroaniline-d4	10.77	131	8735	0.77	ng/ul	0.00
43) Dimethylphthalate-d6	13.87	166	874807	28.88	ng/ul	0.00
46) Acenaphthylene-d8	14.16	160	1077901	29.51	ng/ul	0.00
51) 4-Nitrophenol-d4	14.67	143	22306	3.82	ng/ul	0.00
57) Fluorene-d10	15.46	176	784386	28.69	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.59	200	133268	24.50	ng/ul	0.00
70) Anthracene-d10	17.32	188	1252372	31.63	ng/ul	0.00
76) Pyrene-d10	19.61	212	1337891	37.22	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.56	264	1077731	32.11	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042717\
Data File : BM009754.D
Acq On : 27 Apr 2017 23:29
Operator : SJ/MA
Sample : I2809-21
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
JHTE1

Quant Time: Apr 28 01:11:25 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM042517.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 28 01:10:39 2017
Response via : Initial Calibration

