

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042721\
 Data File : BM029672.D
 Acq On : 27 Apr 2021 14:03
 Operator : CG/JU
 Sample : PB136013BL
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB136013BL

Quant Time: Apr 27 14:45:34 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 26 15:01:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.610	152	65460	20.00	ng	-0.01
21) Naphthalene-d8	10.392	136	233119	20.00	ng	0.00
39) Acenaphthene-d10	14.257	164	153941	20.00	ng	-0.01
64) Phenanthrene-d10	17.004	188	327953	20.00	ng	0.00
76) Chrysene-d12	21.192	240	316280	20.00	ng	0.00
86) Perylene-d12	23.374	264	325229	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.210	112	490561	150.24	ng	0.00
7) Phenol-d6	6.793	99	598315	122.68	ng	0.00
23) Nitrobenzene-d5	8.769	82	391714	86.54	ng	0.00
42) 2,4,6-Tribromophenol	15.757	330	304011	142.80	ng	0.00
45) 2-Fluorobiphenyl	12.875	172	1107359	97.63	ng	-0.01
79) Terphenyl-d14	19.639	244	1633020	97.89	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042721\
Data File : BM029672.D
Acq On : 27 Apr 2021 14:03
Operator : CG/JU
Sample : PB136013BL
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB136013BL

Quant Time: Apr 27 14:45:34 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM042121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 26 15:01:01 2021
Response via : Initial Calibration

