

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042723\
 Data File : BM039705.D
 Acq On : 27 Apr 2023 13:00
 Operator : CG/JU
 Sample : SSTD3.251
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD3.2027

Quant Time: Apr 27 22:37:11 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 27 16:04:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.808	152	3772	0.400	ng/ul	-0.02
4) Naphthalene-d8	10.599	136	14010	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.449	164	8190	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.207	188	16283	0.400	ng/ul	0.00
17) Chrysene-d12	21.400	240	15362	0.400	ng/ul	0.00
23) Perylene-d12	23.759	264	12536	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.189	96	15319	2.756	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.226	88	16379	2.844	ng/ul#	38
14) Pentachlorophenol	16.873	266	12158	3.426	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042723\
Data File : BM039705.D
Acq On : 27 Apr 2023 13:00
Operator : CG/JU
Sample : SST03.251
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SST03.2027

Quant Time: Apr 27 22:37:11 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 27 16:04:50 2023
Response via : Initial Calibration

