

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042723\
 Data File : BM039751.D
 Acq On : 29 Apr 2023 00:06
 Operator : CG/JU
 Sample : 02417-09
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EW8X5

Quant Time: Apr 29 02:26:51 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 28 02:30:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.796	152	4637	0.400	ng/u1	-0.03
4) Naphthalene-d8	10.595	136	17865	0.400	ng/u1	-0.02
9) Acenaphthene-d10	14.446	164	9922	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.200	188	18220	0.400	ng/u1	-0.02
17) Chrysene-d12	21.399	240	10786	0.400	ng/u1	0.00
23) Perylene-d12	23.741	264	11463	0.400	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	12212	1.787	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.201	152	3352	0.146	ng/u1	-0.02
18) Fluoranthene-d10	19.234	212	5350	0.184	ng/u1	0.00
Target Compounds						
19) Fluoranthene	19.266	202	803	0.021	ng/u1#	Qvalue 56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM042723\
Data File : BM039751.D
Acq On : 29 Apr 2023 00:06
Operator : CG/JU
Sample : 02417-09
Misc :
ALS Vial : 54 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EW8X5

Quant Time: Apr 29 02:26:51 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM042723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 28 02:30:02 2023
Response via : Initial Calibration

