

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042915\
 Data File : BM001135.D
 Acq On : 28 Apr 2015 19:21
 Operator : TP/IZ
 Sample : PB82969BL
 Misc : SBLK54
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK54

Quant Time: Apr 29 05:31:28 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM042815.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 29 05:05:22 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.54	152	165219	20.00	ng/ul	0.00
18) Naphthalene-d8	10.32	136	680860	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.20	164	392588	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.96	188	868347	20.00	ng/ul	0.00
75) Chrysene-d12	21.16	240	963679	20.00	ng/ul	0.00
83) Perylene-d12	23.32	264	963511	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	2.99	96	23581	6.47	ng/uL	0.00
5) Phenol-d5	6.72	99	469877	34.46	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.87	67	314627	36.36	ng/ul	0.00
9) 2-Chlorophenol-d4	7.07	132	384892	36.33	ng/ul	0.00
13) 4-Methylphenol-d8	8.26	113	378438	34.17	ng/ul	0.00
19) Nitrobenzene-d5	8.69	128	178689	38.85	ng/ul	0.00
22) 2-Nitrophenol-d4	9.42	143	201385	40.29	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.95	165	344480	34.93	ng/ul	0.00
29) 4-Chloroaniline-d4	10.47	131	534057	45.85	ng/ul	0.00
43) Dimethylphthalate-d6	13.62	166	1117549	41.17	ng/ul	0.00
46) Acenaphthylene-d8	13.89	160	1429360	37.25	ng/ul	0.00
51) 4-Nitrophenol-d4	14.42	143	168351	29.49	ng/ul	0.00
57) Fluorene-d10	15.20	176	1013329	37.77	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.33	200	164569	31.64	ng/ul	0.00
70) Anthracene-d10	17.06	188	1564321	38.88	ng/ul	0.00
76) Pyrene-d10	19.36	212	1791513	41.94	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.19	264	1742989	39.90	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042915\
Data File : BM001135.D
Acq On : 28 Apr 2015 19:21
Operator : TP/IZ
Sample : PB82969BL
Misc : SBLK54
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK54

Quant Time: Apr 29 05:31:28 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM042815.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 29 05:05:22 2015
Response via : Initial Calibration

