


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM042915\
Data File : BM001131.D
Acq On : 28 Apr 2015 16:22
Operator : TP/IZ
Sample : DFTPP53
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP53

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM042815.M
Title : SVOA CALIBRATION
Last Update : Tue Apr 28 06:36:34 2015



AutoFind: Scans 2503, 2504, 2505; Background Corrected with Scan 2497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	55.2	144138	PASS
68	69	0.00	2	1.6	1942	PASS
69	198	0.00	100	45.5	118818	PASS
70	69	0.00	2	0.6	658	PASS
127	198	10	80	53.4	139373	PASS
197	198	0.00	2	0.3	669	PASS
198	198	100	100	100.0	261013	PASS
199	198	5	9	6.5	16901	PASS
275	198	10	60	24.2	63184	PASS
365	198	1	100	2.8	7203	PASS
441	443	0.01	100	81.0	34693	PASS
442	198	50	100	85.1	222122	PASS
443	442	15	24	19.3	42832	PASS

m/z	Abundance
37.10	774.0
38.00	1486.0
39.10	10779.0
40.00	1673.0
44.00	3995.0
45.00	682.0
49.20	749.0
50.10	40536.0
51.10	158016.0
52.00	8155.0
53.00	409.0
55.10	675.0
56.00	4010.0
57.00	10118.0
58.00	352.0
61.00	1707.0
62.10	1498.0
63.00	5223.0
64.20	707.0
65.00	2913.0
68.10	2162.0
69.00	126120.0
70.10	617.0
73.10	1415.0
74.10	12974.0
75.00	22272.0
76.10	7662.0
77.10	149184.0
78.00	10835.0
79.00	8555.0
80.00	7830.0
81.00	11812.0
82.00	3099.0
83.00	2187.0
84.00	404.0
84.90	1977.0
86.00	2573.0
86.90	881.0
87.90	529.0
89.20	454.0
90.90	2186.0
92.00	2976.0
93.00	16832.0
94.00	1346.0
94.90	336.0
96.00	1008.0
96.90	338.0
98.00	13689.0
99.00	11758.0
100.10	880.0
101.00	6063.0
101.90	400.0
103.00	1994.0
104.00	3840.0
105.00	3627.0
106.00	1026.0
107.00	46816.0
108.00	6458.0
109.00	1037.0
110.00	74224.0
111.00	12516.0
111.90	1601.0
113.00	318.0
116.00	2147.0
117.00	38264.0
118.00	3607.0
118.90	318.0
119.90	400.0
122.00	2561.0
122.90	3944.0
124.00	2275.0
125.00	1882.0
126.10	479.0
127.00	142848.0
128.00	9693.0
129.00	56008.0
129.90	4599.0
131.10	865.0
132.00	689.0
133.90	1386.0
135.00	4768.0
136.00	1567.0
137.00	2551.0
138.00	640.0
140.10	495.0
141.00	7499.0
141.90	2013.0
143.00	1690.0
144.10	503.0
145.90	1214.0
147.10	4051.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

148.00	9924.0
149.00	2647.0
150.00	820.0
150.90	866.0
151.80	439.0
153.00	1780.0
154.00	1303.0
155.00	3963.0
156.00	6158.0
157.10	1207.0
158.00	1278.0
159.00	1230.0
160.00	1894.0
160.90	3728.0
162.10	961.0
163.00	332.0
165.00	2211.0
166.10	2731.0
167.00	13529.0
168.00	9319.0
169.10	1039.0
169.90	438.0
170.90	782.0
171.90	1264.0
173.00	1509.0
174.00	3626.0
175.10	4961.0
176.10	1310.0
177.00	2715.0
178.00	1258.0
179.00	10341.0
180.00	7640.0
181.00	3312.0
182.00	419.0
184.10	692.0
185.00	4969.0
186.00	35832.0
187.00	9671.0
188.00	1192.0
189.00	2638.0
191.00	1558.0
192.00	3552.0
193.00	3301.0
194.00	706.0
195.10	586.0
196.00	6499.0
198.00	246720.0
199.00	16752.0
200.00	1655.0
201.40	1291.0
202.30	419.0
203.00	1955.0
204.00	8699.0
205.00	16327.0
206.10	63608.0
207.00	9323.0
208.00	2472.0
209.10	947.0
210.20	1042.0
211.00	2852.0
214.90	663.0
216.00	1495.0
217.00	18224.0
217.90	2529.0
218.90	461.0
221.00	10976.0
223.00	3698.0
224.10	34464.0
225.10	9426.0
226.10	1633.0
227.00	16354.0
228.00	2081.0
229.00	4009.0
230.00	380.0
230.90	700.0
233.90	809.0
235.00	1337.0
236.10	618.0
237.00	2195.0
238.10	470.0
239.00	424.0
239.70	352.0
240.90	890.0
241.90	1665.0
243.00	1949.0
244.00	25752.0
245.10	3226.0
246.00	5028.0
247.00	1420.0
248.00	336.0
249.00	796.0
250.00	359.0
252.90	487.0
254.00	957.0
255.00	130880.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

256.00	19624.0
257.10	1962.0
258.00	7446.0
258.90	1484.0
265.00	2720.0
265.90	440.0
270.90	352.0
273.00	4124.0
274.00	8663.0
275.00	55368.0
276.00	8060.0
277.10	4815.0
278.00	915.0
282.90	633.0
284.10	504.0
285.00	682.0
288.90	346.0
293.00	1172.0
293.90	428.0
296.00	14822.0
296.90	2048.0
303.00	1746.0
304.10	377.0
314.10	564.0
315.00	1835.0
316.00	975.0
316.90	315.0
321.00	519.0
323.00	4497.0
324.00	883.0
327.00	1130.0
327.90	353.0
331.90	328.0
334.00	2499.0
335.10	538.0
341.00	619.0
345.90	937.0
352.00	1665.0
353.00	913.0
354.10	1344.0
355.20	503.0
365.00	6125.0
366.00	724.0
371.00	396.0
371.90	2468.0
373.00	534.0
382.90	441.0
402.00	1120.0
403.10	1097.0
404.00	823.0
420.80	1049.0
421.90	848.0
423.10	7608.0
424.00	3761.0
425.00	654.0
441.00	25224.0
442.00	160640.0
443.00	32712.0
444.00	2735.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

m/z	Abundance
36.00	404.0
37.00	708.0
38.00	2168.0
39.00	12690.0
40.00	1306.0
41.00	756.0
44.10	4859.0
49.00	1086.0
50.10	43576.0
51.10	173184.0
52.00	8787.0
53.10	350.0
55.10	888.0
56.00	5230.0
57.00	12119.0
58.00	332.0
61.00	1871.0
62.10	2419.0
63.00	6235.0
64.00	977.0
65.00	2663.0
66.90	316.0
67.90	2277.0
69.00	143616.0
70.00	899.0
73.00	1452.0
74.00	14420.0
75.00	24288.0
76.10	7731.0
77.10	171392.0
78.00	12375.0
79.00	10101.0
80.00	7949.0
81.00	11024.0
82.00	2068.0
83.00	2901.0
84.10	427.0
85.00	1856.0
86.00	2928.0
87.00	1305.0
88.10	691.0
90.90	2393.0
92.10	3350.0
93.00	19808.0
94.00	1351.0
95.10	455.0
96.20	878.0
96.80	439.0
98.00	15418.0
99.00	12738.0
99.90	1100.0
100.90	6650.0
102.10	379.0
103.00	1732.0
103.90	3538.0
105.00	4182.0
106.00	1078.0
107.00	51792.0
108.00	7987.0
109.00	1149.0
110.00	84936.0
111.10	12369.0
112.00	1990.0
112.80	423.0
115.90	2336.0
117.00	41976.0
118.00	3659.0
118.90	374.0
120.10	737.0
122.00	3739.0
123.00	4704.0
123.90	1842.0
124.90	2370.0
125.80	341.0
126.20	418.0
127.00	170560.0
128.00	13418.0
129.00	68712.0
130.00	4757.0
131.00	1200.0
131.90	619.0
134.10	1778.0
135.10	5509.0
136.00	1648.0
137.10	2901.0
138.10	1090.0
139.00	415.0
140.10	849.0
141.00	7935.0
142.20	2021.0
143.00	1845.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

144.00	514.0
145.10	466.0
145.90	1264.0
147.00	4003.0
148.00	9948.0
149.00	2415.0
150.00	443.0
151.20	1124.0
152.00	636.0
153.00	2180.0
154.00	1974.0
155.00	4384.0
156.10	8033.0
157.00	1495.0
158.00	1458.0
159.00	1417.0
160.00	2881.0
161.00	3697.0
162.10	1142.0
162.90	557.0
163.90	369.0
165.00	2963.0
166.00	2664.0
167.00	17736.0
168.00	10235.0
169.10	1899.0
170.10	478.0
170.80	727.0
172.00	1495.0
172.90	1934.0
174.00	3700.0
175.00	5489.0
176.00	2201.0
176.90	3589.0
177.90	957.0
179.00	12563.0
180.00	9281.0
181.00	3602.0
182.00	855.0
182.70	325.0
184.00	1281.0
185.10	5154.0
186.00	46176.0
187.10	13157.0
188.00	1322.0
188.90	2804.0
190.80	1409.0
192.00	3924.0
193.00	4887.0
194.00	909.0
195.00	630.0
196.00	9130.0
196.80	2009.0
198.00	323264.0
199.00	21320.0
200.00	2057.0
200.90	329.0
201.70	938.0
203.10	2260.0
204.00	13013.0
205.10	19424.0
206.00	87928.0
207.10	11826.0
208.00	3484.0
208.90	1139.0
210.20	1175.0
211.00	4078.0
212.10	551.0
213.00	479.0
215.10	769.0
216.10	2384.0
217.00	21648.0
217.90	3212.0
221.10	15387.0
222.00	2815.0
223.00	5122.0
224.00	50232.0
225.10	11248.0
226.00	1463.0
227.00	21272.0
228.00	2924.0
229.00	3945.0
230.00	854.0
231.00	1970.0
232.10	348.0
233.00	635.0
233.90	1530.0
234.90	1454.0
235.90	1229.0
237.00	2091.0
239.10	930.0
239.90	754.0
241.00	1549.0
242.00	2218.0
243.10	1920.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

244.10	35224.0
245.10	4458.0
246.00	7095.0
246.90	1538.0
249.10	1265.0
250.80	412.0
251.90	429.0
252.90	936.0
254.10	1739.0
255.00	187008.0
256.00	27408.0
257.00	1827.0
258.00	10645.0
259.10	1431.0
259.90	351.0
261.10	495.0
265.00	3829.0
265.90	406.0
271.20	459.0
272.00	485.0
273.00	5156.0
274.00	14076.0
275.00	77216.0
276.00	9739.0
277.00	6389.0
278.00	1196.0
283.00	614.0
285.10	1035.0
292.00	368.0
293.00	1672.0
294.20	374.0
295.10	669.0
296.00	20248.0
296.90	3624.0
300.90	343.0
301.90	345.0
303.10	2651.0
304.20	622.0
310.00	468.0
314.00	576.0
314.90	2478.0
316.00	1363.0
321.00	478.0
321.90	439.0
323.10	6415.0
324.00	1024.0
327.00	1212.0
327.90	812.0
329.10	376.0
332.00	410.0
333.00	885.0
334.10	4927.0
335.00	1272.0
340.80	483.0
346.10	1465.0
352.10	2415.0
353.00	1519.0
353.90	1819.0
354.90	478.0
365.00	8937.0
366.00	1190.0
371.00	788.0
372.10	3278.0
372.90	1134.0
382.90	927.0
383.90	380.0
390.00	602.0
391.90	376.0
402.00	1136.0
403.00	2243.0
403.80	500.0
421.00	1929.0
422.00	1992.0
423.00	13438.0
424.00	4283.0
425.10	306.0
441.00	40864.0
442.00	265728.0
443.00	49280.0
444.10	4887.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

m/z	Abundance
36.00	385.0
37.10	493.0
38.00	1245.0
39.10	7767.0
40.00	1610.0
41.10	575.0
44.00	4302.0
45.00	356.0
49.00	805.0
50.00	25352.0
51.10	101216.0
52.10	4533.0
54.90	405.0
56.00	2703.0
57.00	5953.0
58.00	339.0
61.00	1556.0
62.10	1640.0
63.00	3405.0
63.90	446.0
65.00	1753.0
68.00	1388.0
69.00	86720.0
70.00	458.0
73.00	1515.0
74.00	9439.0
75.00	13551.0
76.10	4169.0
77.10	96416.0
78.10	6862.0
79.00	6377.0
80.00	4642.0
81.00	8024.0
82.00	2298.0
83.10	1677.0
84.10	340.0
85.00	1446.0
86.00	2104.0
86.80	869.0
87.90	385.0
91.00	2033.0
92.00	2025.0
93.00	10641.0
94.10	825.0
96.10	424.0
98.00	8899.0
99.00	6911.0
99.90	498.0
101.00	4284.0
103.00	1362.0
104.00	2752.0
105.10	2818.0
105.90	816.0
107.00	31320.0
108.10	3838.0
108.90	680.0
110.00	55672.0
111.10	8410.0
112.00	967.0
116.00	1630.0
117.00	27288.0
118.00	2036.0
119.00	416.0
122.00	2005.0
123.00	2616.0
123.90	1646.0
125.10	1347.0
127.00	104712.0
128.00	6872.0
129.00	42224.0
130.00	3261.0
130.90	575.0
134.10	1113.0
135.00	3322.0
136.10	1213.0
137.10	2033.0
138.20	463.0
140.00	559.0
141.00	5396.0
142.00	1696.0
143.00	1028.0
144.00	515.0
146.00	1169.0
147.00	3349.0
148.00	7365.0
148.90	2022.0
150.00	481.0
151.50	817.0
153.00	1901.0
154.00	1527.0
155.00	3186.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

156.00	4626.0
157.10	959.0
157.90	1165.0
159.00	901.0
160.00	1741.0
161.00	2627.0
162.00	453.0
165.00	1888.0
166.00	1383.0
167.00	11851.0
168.00	6316.0
168.90	872.0
170.00	328.0
170.80	725.0
172.00	768.0
173.00	1255.0
174.00	2096.0
175.00	4281.0
175.90	1593.0
177.10	2247.0
178.10	798.0
179.00	8115.0
180.00	6325.0
181.00	2971.0
182.00	465.0
184.00	772.0
185.00	4217.0
186.00	30448.0
187.00	8048.0
188.00	1106.0
189.10	1736.0
190.90	1284.0
192.00	2580.0
193.00	3168.0
194.00	459.0
194.80	364.0
196.00	5999.0
197.10	969.0
198.00	213056.0
199.00	12632.0
200.00	1645.0
201.50	1248.0
202.90	1479.0
203.90	8361.0
205.00	14093.0
206.00	56088.0
207.00	7786.0
207.90	2105.0
209.10	841.0
210.10	1207.0
210.90	2414.0
212.00	384.0
212.80	309.0
215.10	938.0
216.00	1261.0
217.00	17048.0
217.90	1836.0
220.20	505.0
221.10	7914.0
223.00	3677.0
224.00	32752.0
225.00	7825.0
226.00	1065.0
227.00	12791.0
228.00	2424.0
229.00	2676.0
231.00	1234.0
234.00	853.0
235.00	1097.0
236.00	729.0
236.90	1387.0
238.90	484.0
240.10	453.0
241.00	867.0
242.00	1796.0
243.00	2064.0
244.10	26728.0
245.10	3061.0
246.00	4155.0
246.90	671.0
249.00	1021.0
252.90	864.0
254.00	817.0
255.00	130016.0
256.00	17208.0
256.90	1429.0
258.00	8242.0
259.00	947.0
264.00	332.0
265.00	3171.0
266.00	506.0
271.00	448.0
272.90	4124.0
274.00	10336.0
275.00	56968.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

276.00	7547.0
277.00	4795.0
278.00	682.0
283.00	621.0
284.10	335.0
285.10	679.0
292.90	874.0
296.00	15779.0
297.10	2575.0
303.00	2135.0
304.10	555.0
308.90	408.0
310.10	374.0
314.00	971.0
315.00	1930.0
315.90	979.0
321.00	611.0
321.80	390.0
323.00	5844.0
324.10	853.0
327.00	1092.0
328.00	410.0
332.00	570.0
332.90	309.0
334.00	3265.0
335.10	1013.0
341.00	568.0
346.00	1283.0
352.10	1374.0
352.90	924.0
354.10	2216.0
355.10	394.0
365.00	6547.0
365.80	1158.0
370.80	376.0
372.00	2599.0
373.10	911.0
382.90	967.0
383.90	491.0
389.80	464.0
402.00	1475.0
403.10	1905.0
404.00	723.0
421.00	1514.0
422.10	2163.0
423.00	11616.0
424.10	3092.0
425.20	777.0
441.00	37992.0
442.00	240000.0
443.10	46504.0
444.00	4710.0
445.00	318.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53

Scan 2497 (17.374 min): BM001131.D
DFTPP53

m/z	Abundance
40.00	1120.0
44.00	3481.0
207.10	766.0

Instrument :
BNA_M
ClientSampleId :
DFTPP53