

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM043016\
 Data File : BM005184.D
 Acq On : 30 Apr 2016 21:24
 Operator : UM/SJ
 Sample : H2772-05
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BD3J5

Quant Time: May 01 04:16:51 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM042516.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun May 01 04:14:40 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	37863	20.00	ng/ul	0.00
18) Naphthalene-d8	10.57	136	180967	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.43	164	115455	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.17	188	254143	20.00	ng/ul	0.00
75) Chrysene-d12	21.36	240	229403	20.00	ng/ul	0.00
83) Perylene-d12	23.64	264	204239	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.29	96	988	1.33	ng/uL	0.00
5) Phenol-d5	6.96	99	22058	7.04	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.12	67	54186	30.41	ng/ul	0.00
9) 2-Chlorophenol-d4	7.32	132	58739	23.70	ng/ul	0.00
13) 4-Methylphenol-d8	8.49	113	41295	15.55	ng/ul	0.00
19) Nitrobenzene-d5	8.94	128	35460	28.29	ng/ul	0.00
22) 2-Nitrophenol-d4	9.66	143	40200	28.75	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.20	165	67505	25.41	ng/ul	0.00
29) 4-Chloroaniline-d4	0.00	131	0	0.00	ng/ul	0.00
43) Dimethylphthalate-d6	13.83	166	268914	29.49	ng/ul	0.00
46) Acenaphthylene-d8	14.12	160	314528	28.97	ng/ul	0.00
51) 4-Nitrophenol-d4	14.63	143	5512	3.46	ng/ul	0.00
57) Fluorene-d10	15.42	176	237461	29.64	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.54	200	31495	21.79	ng/ul	0.00
70) Anthracene-d10	17.27	188	365458	32.07	ng/ul	0.00
76) Pyrene-d10	19.57	212	372566	35.67	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.50	264	292588	31.97	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM043016\
Data File : BM005184.D
Acq On : 30 Apr 2016 21:24
Operator : UM/SJ
Sample : H2772-05
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BD3J5

Quant Time: May 01 04:16:51 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM042516.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun May 01 04:14:40 2016
Response via : Initial Calibration

